

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101024\
 Data File : BF139741.D
 Acq On : 10 Oct 2024 12:42
 Operator : RC/JU
 Sample : P4340-08
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 1B-3

Quant Time: Oct 10 13:06:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF100124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 02 04:58:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.869	152	81182	20.000	ng	0.00
21) Naphthalene-d8	8.145	136	304829	20.000	ng	0.00
39) Acenaphthene-d10	9.904	164	162499	20.000	ng	0.00
64) Phenanthrene-d10	11.386	188	287458	20.000	ng	0.00
76) Chrysene-d12	14.027	240	168792	20.000	ng	0.00
86) Perylene-d12	15.504	264	168324	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	587214	125.371	ng	0.01
7) Phenol-d6	6.504	99	713343	113.252	ng	0.00
23) Nitrobenzene-d5	7.428	82	556743	92.472	ng	0.00
42) 2,4,6-Tribromophenol	10.692	330	197801	126.095	ng	0.00
45) 2-Fluorobiphenyl	9.222	172	1000744	94.544	ng	0.00
79) Terphenyl-d14	12.974	244	840723	81.727	ng	0.00
Target Compounds						
77) Benzidine	12.727	184	179791	60.387	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101024\
Data File : BF139741.D
Acq On : 10 Oct 2024 12:42
Operator : RC/JU
Sample : P4340-08
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
1B-3

Quant Time: Oct 10 13:06:50 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF100124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 02 04:58:47 2024
Response via : Initial Calibration

