

Data Path : Z:\HPCHEM1\BNA F\DATA\BF101116\
 Data File : BF090496.D
 Acq On : 11 Oct 2016 11:52
 Operator : UM/SJ
 Sample : SSTDICC025
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC025

Quant Time: Oct 11 13:17:51 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF101116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 11 13:13:45 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	210407	20.00	ng	0.00
21) Naphthalene-d8	8.23	136	809068	20.00	ng	-0.01
38) Acenaphthene-d10	9.99	164	414679	20.00	ng	-0.01
63) Phenanthrene-d10	11.48	188	699001	20.00	ng	0.00
75) Chrysene-d12	14.13	240	737481	20.00	ng	0.00
86) Perylene-d12	15.61	264	571607	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.55	112	690219	48.96	ng	-0.01
7) Phenol-d6	6.58	99	852399	46.16	ng	-0.01
23) Nitrobenzene-d5	7.51	82	818449	49.21	ng	0.00
41) 2,4,6-Tribromophenol	10.78	330	195880	58.09	ng	-0.01
44) 2-Fluorobiphenyl	9.32	172	1488183	59.79	ng	0.00
78) Terphenyl-d14	13.07	244	1362679	41.54	ng	-0.01

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.61	88	168602	24.26	ng	# 84
3) Pyridine	3.37	79	438557	22.63	ng	# 77
4) n-Nitrosodimethylamine	3.31	42	155709	19.47	ng	# 44
6) Aniline	6.61	93	580813	22.77	ng	94
8) 2-Chlorophenol	6.74	128	398473	25.68	ng	90
9) Benzaldehyde	6.50	77	297062	31.83	ng	96
10) Phenol	6.59	94	461191	21.48	ng	96
11) bis(2-Chloroethyl)ether	6.68	93	352721	21.46	ng	94
12) 1,3-Dichlorobenzene	6.90	146	387302	25.04	ng	97
13) 1,4-Dichlorobenzene	6.97	146	379345	24.15	ng	98
14) 1,2-Dichlorobenzene	7.13	146	400013	26.43	ng	96
15) Benzyl Alcohol	7.09	79	368541	26.26	ng	97
16) 2,2'-oxybis(1-Chloropropan	7.23	45	517450	18.56	ng	90
17) 2-Methylphenol	7.21	107	321925	25.23	ng	97
18) Hexachloroethane	7.47	117	152103	24.56	ng	100
19) n-Nitroso-di-n-propylamine	7.37	70	288601	21.32	ng	# 75
20) 3+4-Methylphenols	7.35	107	381330	23.09	ng	# 83
22) Acetophenone	7.35	105	485248	25.32	ng	# 95
24) Nitrobenzene	7.54	77	412720	24.99	ng	# 81
25) Isophorone	7.78	82	732742	24.10	ng	95
26) 2-Nitrophenol	7.85	139	186438	26.23	ng	# 71
27) 2,4-Dimethylphenol	7.89	122	359727	26.03	ng	98
28) bis(2-Chloroethoxy)methane	7.98	93	415876	23.12	ng	97
29) 2,4-Dichlorophenol	8.10	162	291969	27.78	ng	98
30) 1,2,4-Trichlorobenzene	8.18	180	300519	27.75	ng	100
31) Naphthalene	8.26	128	1022311	26.19	ng	98
32) Benzoic acid	7.99	122	262707	27.34	ng	98
33) 4-Chloroaniline	8.30	127	425490	26.37	ng	98
34) Hexachlorobutadiene	8.38	225	183081	30.26	ng	100
35) Caprolactam	8.66	113	89356	25.29	ng	# 85
36) 4-Chloro-3-methylphenol	8.78	107	299256	22.74	ng	94
37) 2-Methylnaphthalene	8.95	142	724843	30.68	ng	98
39) 1,2,4,5-Tetrachlorobenzene	9.13	216	293080	26.81	ng	98
40) Hexachlorocyclopentadiene	9.11	237	146291	24.41	ng	97

Data Path : Z:\HPCHEM1\BNA F\DATA\BF101116\
 Data File : BF090496.D
 Acq On : 11 Oct 2016 11:52
 Operator : UM/SJ
 Sample : SSTDICC025
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC025

Quant Time: Oct 11 13:17:51 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF101116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 11 13:13:45 2016
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	9.23	196	205274	27.21	nq	97
43) 2,4,5-Trichlorophenol	9.27	196	198701	27.85	nq	# 84
45) 1,1'-Biphenyl	9.42	154	846398	27.54	nq	93
46) 2-Chloronaphthalene	9.45	162	677391	29.84	nq	94
47) 2-Nitroaniline	9.54	65	242268	27.04	nq	# 77
48) Acenaphthylene	9.86	152	985963	26.32	nq	98
49) Dimethylphthalate	9.72	163	766066	28.85	nq	97
50) 2,6-Dinitrotoluene	9.78	165	180744	30.65	nq	# 77
51) Acenaphthene	10.03	154	625771	26.79	nq	96
52) 3-Nitroaniline	9.95	138	213307	30.97	nq	# 79
53) 2,4-Dinitrophenol	10.05	184	79202	30.13	nq	# 76
54) Dibenzofuran	10.20	168	851449	28.16	nq	94
55) 4-Nitrophenol	10.10	139	147580	24.49	nq	99
56) 2,4-Dinitrotoluene	10.18	165	232114	29.58	nq	# 56
57) Fluorene	10.54	166	717292	28.23	nq	100
58) 2,3,4,6-Tetrachlorophenol	10.28	232	174065	32.67	nq	97
59) Diethylphthalate	10.42	149	781178	28.43	nq	# 93
60) 4-Chlorophenyl-phenylether	10.54	204	340877	29.51	nq	# 84
61) 4-Nitroaniline	10.55	138	199354	28.73	nq	95
62) Azobenzene	10.69	77	831634	26.20	nq	87
64) 4,6-Dinitro-2-methylphenol	10.59	198	125038	29.48	nq	# 44
65) n-Nitrosodiphenylamine	10.66	169	650810	27.78	nq	99
66) 4-Bromophenyl-phenylether	11.03	248	192436	28.09	nq	# 79
67) Hexachlorobenzene	11.09	284	204277	26.77	nq	# 76
68) Atrazine	11.18	200	209979	36.36	nq	96
69) Pentachlorophenol	11.29	266	133252	29.34	nq	99
70) Phenanthrene	11.50	178	1003398	28.03	nq	96
71) Anthracene	11.56	178	1087426	29.75	nq	98
72) Carbazole	11.71	167	915200	26.95	nq	97
73) Di-n-butylphthalate	12.05	149	1227128	29.69	nq	# 95
74) Fluoranthene	12.70	202	1108254	31.57	nq	89
76) Benzidine	12.82	184	603764	31.14	nq	99
77) Pyrene	12.93	202	1179518	23.99	nq	97
79) Butylbenzylphthalate	13.55	149	561722	22.26	nq	96
80) Benzo(a)anthracene	14.12	228	1119933	27.07	nq	98
81) 3,3'-Dichlorobenzidine	14.09	252	391247	26.08	nq	# 99
82) Chrysene	14.16	228	916952	24.08	nq	99
83) Bis(2-ethylhexyl)phthalate	14.11	149	785485	21.90	nq	# 98
84) Di-n-octyl phthalate	14.73	149	1247859	23.47	nq	# 89
85) Indeno(1,2,3-cd)pyrene	17.08	276	837419	30.84	nq	96
87) Benzo(b)fluoranthene	15.17	252	872942	23.98	nq	# 96
88) Benzo(k)fluoranthene	15.17	252	872942	25.52	nq	99
89) Benzo(a)pyrene	15.54	252	861678	27.46	nq	# 96
90) Dibenzo(a,h)anthracene	17.10	278	736917	27.60	nq	97
91) Benzo(g,h,i)perylene	17.53	276	689142	26.90	nq	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF101116\
 Data File : BF090496.D
 Acq On : 11 Oct 2016 11:52
 Operator : UM/SJ
 Sample : SSTDIC025
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDIC025

Quant Time: Oct 11 13:17:51 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF101116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 11 13:13:45 2016
 Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.95 min Scan# 404
Delta R.T. 0.00 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC025

Tot Ion:152 Resp: 210407
Ion Ratio Lower Upper
152 100
150 152.5 124.6 186.8
115 58.0 46.2 69.4




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 24.26 ng
RT: 2.61 min Scan# 24
Delta R.T. 0.00 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Tgt Ion: 88 Resp: 168602
Ion Ratio Lower Upper
88 100
58 76.6 71.2 106.8
43 24.5 29.8 44.6#



Abundance Ion 88.00 (87.70 to 88.70): BF0
Ion 58.00 (57.70 to 58.70): BF0
Ion 43.00 (42.70 to 43.70): BF0





#3

Pvridine

Concen: 22.63 ng

RT: 3.37 min Scan# 90

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC025





Tot Ion: 79 Resp: 438557

Ion Ratio Lower Upper

79 100

52 66.3 71.3 106.9#

51 31.0 34.5 51.7#



#4

n-Nitrosodimethylamine

Concen: 19.47 ng

RT: 3.31 min Scan# 85

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52




Tgt Ion: 42 Resp: 155709


Ion Ratio Lower Upper

42 100

74 175.1 88.9 133.3#

44 7.3 6.5 9.7





#5

2-Fluorophenol

Concen: 48.96 ng

RT: 5.55 min Scan# 281

Delta R.T. -0.01 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDIC025


Tot Ion:112 Resp: 690219

Ion Ratio Lower Upper

112 100

64 71.2 57.9 86.9




63 34.2 29.6 44.4



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#6

Aniline

Concen: 22.77 ng

RT: 6.61 min Scan# 374

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52


Tgt Ion: 93 Resp: 580813

Ion Ratio Lower Upper

93 100

66 39.8 34.5 51.7



65 19.5 18.2 27.2






Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#7

Phenol-d6

Concen: 46.16 ng

RT: 6.58 min Scan# 371

Delta R.T. -0.01 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDIC025





Tot Ion: 99 Resp: 852399

Ion Ratio Lower Upper

99 100

42 15.7 20.0 30.0#

71 32.1 27.9 41.9



#8

2-Chlorophenol

Concen: 25.68 ng

RT: 6.74 min Scan# 385

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52


Tgt Ion: 128 Resp: 398473


Ion Ratio Lower Upper

128 100

130 32.6 13.3 53.3

64 45.3 36.0 76.0









#9
Benzaldehyde
Concen: 31.83 ng
RT: 6.50 min Scan# 364
Delta R.T. 0.00 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDIC025


Tot Ion: 77 Resp: 297062
Ion Ratio Lower Upper
77 100
105 98.1 73.1 113.1
106 93.1 70.2 110.2



#10
Phenol
Concen: 21.48 ng
RT: 6.59 min Scan# 372
Delta R.T. -0.01 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Tgt Ion: 94 Resp: 461191
Ion Ratio Lower Upper
94 100
65 30.2 13.3 53.3
66 38.9 20.4 60.4









#11
bis(2-Chloroethyl)ether
Concen: 21.46 ng
RT: 6.68 min Scan# 380
Delta R.T. -0.01 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDIC025


Tot Ion: 93 Resp: 352721
Ion Ratio Lower Upper
93 100
63 85.5 59.7 99.7
95 31.9 14.0 54.0



#12
1,3-Dichlorobenzene
Concen: 25.04 ng
RT: 6.90 min Scan# 399
Delta R.T. 0.00 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Tgt Ion: 146 Resp: 387302
Ion Ratio Lower Upper
146 100
148 64.7 52.2 78.4
75 23.0 21.4 32.2









#13
 1,4-Dichlorobenzene
 Concen: 24.15 ng
 RT: 6.97 min Scan# 405
 Delta R.T. 0.00 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC025

Tot Ion:146 Resp: 379345
 Ion Ratio Lower Upper
 146 100
 148 62.8 51.1 76.7
 111 48.6 37.8 56.6




Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E





#14
 1,2-Dichlorobenzene
 Concen: 26.43 ng
 RT: 7.13 min Scan# 419
 Delta R.T. 0.00 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52

Tgt Ion:146 Resp: 400013
 Ion Ratio Lower Upper
 146 100
 148 65.2 52.3 78.5
 111 42.3 38.2 57.4



Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 26.26 ng

RT: 7.09 min Scan# 416

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDIC025


Tot Ion: 79 Resp: 368541

Ion Ratio Lower Upper

79 100

108 80.0 60.9 91.3




77 63.7 51.5 77.3



Abundance Ion 79.00 (78.70 to 79.70): BF0

Ion 108.00 (107.70 to 108.70): BF0

Ion 77.00 (76.70 to 77.70): BF0



#16

2,2'-oxybis(1-Chloropropane)

Concen: 18.56 ng

RT: 7.23 min Scan# 428

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52


Tgt Ion: 45 Resp: 517450

Ion Ratio Lower Upper

45 100

77 14.8 0.0 30.8



79 11.7 0.0 28.2




Abundance Ion 45.00 (44.70 to 45.70): BF0

Ion 77.00 (76.70 to 77.70): BF0

Ion 79.00 (78.70 to 79.70): BF0





#17

2-Methylphenol

Concen: 25.23 ng

RT: 7.21 min Scan# 426

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDIC025

Tot Ion:107 Resp: 321925


Ion Ratio Lower Upper

107 100

108 112.1 90.9 136.3

77 44.4 38.2 57.4

79 42.2 37.0 55.4




Abundance Ion 107.00 (106.70 to 107.70): E


Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BF0

Ion 79.00 (78.70 to 79.70): BF0



Time-->



#18

Hexachloroethane

Concen: 24.56 ng

RT: 7.47 min Scan# 449

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52


Tgt Ion:117 Resp: 152103

Ion Ratio Lower Upper

117 100

119 95.7 75.9 113.9


201 83.1 66.5 99.7






Abundance Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E







Time-->



#19
n-Nitroso-di-n-propylamine
Concen: 21.32 ng
RT: 7.37 min Scan# 440
Delta R.T. 0.00 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52


Instrument :
BNA_F
ClientSampleId :
SSTDIC025


Tot Ion	Ratio	Lower	Upper
70	100		
42	41.9	55.4	83.2#
101	10.1	7.6	11.4
130	23.3	16.5	24.7



#20
3+4-Methylphenols
Concen: 23.09 ng
RT: 7.35 min Scan# 439
Delta R.T. -0.01 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Tgt Ion	Ratio	Lower	Upper
107	100		
108	87.3	66.7	106.7
77	108.5	53.7	93.7#
79	26.6	8.4	48.4









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.23 min Scan# 516
Delta R.T. -0.01 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDIC025


Tot Ion:136	Resp:	809068
Ion Ratio	Lower	Upper
136	100	
137	11.4	9.0 13.6
54	11.0	6.6 10.0#
68	9.3	5.0 7.6#



#22
Acetophenone
Concen: 25.32 ng
RT: 7.35 min Scan# 439
Delta R.T. -0.01 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Tgt Ion:105	Resp:	485248
Ion Ratio	Lower	Upper
105	100	
71	8.3	3.8 5.6#
51	38.2	28.2 42.4
120	20.2	17.4 26.2





#23

Nitrobenzene-d5

Concen: 49.21 ng

RT: 7.51 min Scan# 453

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC025


Tot Ion: 82 Resp: 818449

Ion Ratio Lower Upper

82 100

128 44.3 40.0 60.0




54 45.2 42.6 64.0



Abundance Ion 82.00 (81.70 to 82.70): BF0

Ion 128.00 (127.70 to 128.70): BF0

Ion 54.00 (53.70 to 54.70): BF0



#24

Nitrobenzene

Concen: 24.99 ng

RT: 7.54 min Scan# 455

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52


Tgt Ion: 77 Resp: 412720

Ion Ratio Lower Upper

77 100

123 55.9 33.0 49.4#



65 11.7 12.0 18.0#




Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 123.00 (122.70 to 123.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#25

Isophorone

Concen: 24.10 ng

RT: 7.78 min Scan# 476

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDIC025





Tot Ion: 82 Resp: 732742

Ion Ratio Lower Upper

82 100

95 6.4 5.2 7.8

138 19.2 13.0 19.4



#26

2-Nitrophenol

Concen: 26.23 ng

RT: 7.85 min Scan# 482

Delta R.T. -0.01 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52




Tgt Ion: 139 Resp: 186438


Ion Ratio Lower Upper

139 100

109 26.7 18.9 28.3

65 65.3 31.6 47.4#





#27

2,4-Dimethylphenol

Concen: 26.03 ng

RT: 7.89 min Scan# 486

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDIC025


Tot Ion:122 Resp: 359727

Ion Ratio Lower Upper

122 100

107 108.8 88.6 132.8




121 57.5 47.5 71.3



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E



#28

bis(2-Chloroethoxy)methane

Concen: 23.12 ng

RT: 7.98 min Scan# 494

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52


Tgt Ion: 93 Resp: 415876

Ion Ratio Lower Upper

93 100

95 32.7 27.0 40.4



123 11.2 11.1 16.7




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 95.00 (94.70 to 95.70): BF0

Ion 123.00 (122.70 to 123.70): E









#29
 2,4-Dichlorophenol
 Concen: 27.78 ng
 RT: 8.10 min Scan# 504
 Delta R.T. 0.00 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC025

Tot Ion	Ratio	Lower	Upper
162	100		
164	64.8	45.7	85.7
98	36.1	13.3	53.3




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): BF0






#30
 1,2,4-Trichlorobenzene
 Concen: 27.75 ng
 RT: 8.18 min Scan# 511
 Delta R.T. 0.00 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52

Tgt Ion	Ratio	Lower	Upper
180	100		
182	94.1	75.3	112.9
145	35.1	27.7	41.5



Abundance Ion 180.00 (179.70 to 180.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 26.19 ng

RT: 8.26 min Scan# 518

Delta R.T. -0.01 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDIC025


Tot Ion:128 Resp: 1022311

Ion Ratio Lower Upper

128 100

129 11.5 9.7 14.5




127 13.4 11.7 17.5



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 27.34 ng

RT: 7.99 min Scan# 495

Delta R.T. -0.02 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52


Tgt Ion:122 Resp: 262707

Ion Ratio Lower Upper

122 100

105 123.0 103.9 143.9



77 98.1 74.6 114.6




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF0









#33
4-Chloroaniline
Concen: 26.37 ng
RT: 8.30 min Scan# 522
Delta R.T. -0.01 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDIC025


Tot Ion:	127	Resp:	425490
Ion	Ratio	Lower	Upper
127	100		
129	32.3	26.2	39.4
65	38.9	32.5	48.7
92	22.5	17.2	25.8



#34
Hexachlorobutadiene
Concen: 30.26 ng
RT: 8.38 min Scan# 529
Delta R.T. 0.00 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Tgt Ion:	225	Resp:	183081
Ion	Ratio	Lower	Upper
225	100		
223	63.3	51.0	76.4
227	63.2	50.8	76.2









#35
Caprolactam
Concen: 25.29 ng
RT: 8.66 min Scan# 553
Delta R.T. -0.02 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC025


Tot Ion:113 Resp: 89356
Ion Ratio Lower Upper
113 100
55 196.0 203.3 243.3#
56 145.0 140.5 180.5



#36
4-Chloro-3-methylphenol
Concen: 22.74 ng
RT: 8.78 min Scan# 564
Delta R.T. -0.01 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Tgt Ion:107 Resp: 299256
Ion Ratio Lower Upper
107 100
144 22.5 19.8 29.8
142 70.1 60.7 91.1









#37
 2-Methylnaphthalene
 Concen: 30.68 ng
 RT: 8.95 min Scan# 579
 Delta R.T. 0.00 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC025

Tot Ion	Ratio	Lower	Upper
142	100		
141	89.9	71.8	107.8
115	32.3	28.8	43.2




Abundance Ion 142.00 (141.70 to 142.70): E
 Ion 141.00 (140.70 to 141.70): E
 Ion 115.00 (114.70 to 115.70): E






#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 9.99 min Scan# 670
 Delta R.T. -0.01 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52

Tgt Ion	Ratio	Lower	Upper
164	100		
162	103.9	80.1	120.1
160	46.6	36.3	54.5



Abundance Ion 164.00 (163.70 to 164.70): E
 Ion 162.00 (161.70 to 162.70): E
 Ion 160.00 (159.70 to 160.70): E









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 26.81 ng
 RT: 9.13 min Scan# 594
 Delta R.T. 0.00 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC025

Tot Ion:216	Resp:	293080
Ion Ratio	Lower	Upper
216	100	
214	77.9	64.2 96.2
179	21.8	18.0 27.0
108	19.8	15.2 22.8




Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 108.00 (107.70 to 108.70): E






#40
 Hexachlorocyclopentadiene
 Concen: 24.41 ng
 RT: 9.11 min Scan# 593
 Delta R.T. 0.00 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52

Tgt Ion:237	Resp:	146291
Ion Ratio	Lower	Upper
237	100	
235	61.5	43.6 83.6
272	13.8	0.0 32.7



Abundance Ion 236.80 (236.50 to 237.50): E
 Ion 234.90 (234.60 to 235.60): E
 Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 58.09 ng
 RT: 10.78 min Scan# 739
 Delta R.T. -0.01 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC025


Tot Ion:330 Resp: 195880
 Ion Ratio Lower Upper
 330 100
 332 98.3 75.4 113.0
 141 42.8 26.9 40.3#



#42
 2,4,6-Trichlorophenol
 Concen: 27.21 ng
 RT: 9.23 min Scan# 603
 Delta R.T. 0.00 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52

Tgt Ion:196 Resp: 205274
 Ion Ratio Lower Upper
 196 100
 198 96.2 78.9 118.3
 200 30.1 25.4 38.2









#43
 2,4,5-Trichlorophenol
 Concen: 27.85 ng
 RT: 9.27 min Scan# 607
 Delta R.T. 0.00 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC025

Tot Ion	Ratio	Lower	Upper
196	100		
198	97.2	78.1	117.1
97	31.2	48.1	72.1#
132	22.7	27.4	41.2#




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 59.79 ng
 RT: 9.32 min Scan# 611
 Delta R.T. 0.00 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52

Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.8	31.8	47.8
170	24.6	21.7	32.5



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#45

1,1'-Biphenyl

Concen: 27.54 ng

RT: 9.42 min Scan# 620

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC025


Tot Ion:154 Resp: 846398

Ion Ratio Lower Upper

154 100

153 40.7 23.8 63.8




76 9.9 0.0 35.8



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BF0



#46

2-Chloronaphthalene

Concen: 29.84 ng

RT: 9.45 min Scan# 622

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52


Tgt Ion:162 Resp: 677391

Ion Ratio Lower Upper

162 100

127 37.6 35.6 53.4



164 33.8 27.5 41.3




Abundance Ion 162.00 (161.70 to 162.70): E

Ion 127.00 (126.70 to 127.70): E

Ion 164.00 (163.70 to 164.70): E









#47
 2-Nitroaniline
 Concen: 27.04 ng
 RT: 9.54 min Scan# 630
 Delta R.T. 0.00 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC025


Tot Ion: 65 Resp: 242268
 Ion Ratio Lower Upper
 65 100
 92 75.6 51.7 77.5
 138 127.4 79.0 118.4#



#48
 Acenaphthylene
 Concen: 26.32 ng
 RT: 9.86 min Scan# 658
 Delta R.T. 0.00 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52

Tgt Ion: 152 Resp: 985963
 Ion Ratio Lower Upper
 152 100
 151 20.5 17.4 26.0
 153 13.7 11.6 17.4









#49
Dimethylphthalate
Concen: 28.85 ng
RT: 9.72 min Scan# 646
Delta R.T. 0.00 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC025

Tot Ion:163 Resp: 766066
Ion Ratio Lower Upper
163 100
194 3.7 2.8 4.2
164 9.9 9.0 13.4




Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E






#50
2,6-Dinitrotoluene
Concen: 30.65 ng
RT: 9.78 min Scan# 651
Delta R.T. 0.00 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Tgt Ion:165 Resp: 180744
Ion Ratio Lower Upper
165 100
63 48.2 58.0 87.0#
89 51.7 51.2 76.8



Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 89.00 (88.70 to 89.70): BF0





#51

Acenaphthene

Concen: 26.79 ng

RT: 10.03 min Scan# 673

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC025


Tot Ion:154 Resp: 625771

Ion Ratio Lower Upper

154 100

153 115.4 88.6 133.0




152 54.7 42.6 64.0



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 30.97 ng

RT: 9.95 min Scan# 666

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52


Tgt Ion:138 Resp: 213307

Ion Ratio Lower Upper

138 100

108 8.7 9.7 14.5#



92 96.7 97.1 145.7#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0





#53

2,4-Dinitrophenol

Concen: 30.13 ng

RT: 10.05 min Scan# 675

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDIC025


Tot Ion:184 Resp: 79202

Ion Ratio Lower Upper

184 100

63 31.1 41.9 62.9#




154 47.5 48.9 73.3#



Abundance Ion 184.00 (183.70 to 184.70): E

Ion 63.00 (62.70 to 63.70): BF0

Ion 154.00 (153.70 to 154.70): E



#54

Dibenzofuran

Concen: 28.16 ng

RT: 10.20 min Scan# 688

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52


Tgt Ion:168 Resp: 851449

Ion Ratio Lower Upper

168 100

139 41.7 29.6 44.4



169 13.9 11.3 16.9




Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E





#55

4-Nitrophenol

Concen: 24.49 ng

RT: 10.10 min Scan# 679

Delta R.T. -0.01 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC025


Tot Ion:139 Resp: 147580

Ion Ratio Lower Upper

139 100

109 62.4 43.6 83.6




65 107.6 87.5 127.5



Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF0



#56

2,4-Dinitrotoluene

Concen: 29.58 ng

RT: 10.18 min Scan# 686

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Tgt Ion:165 Resp: 232114


Ion Ratio Lower Upper

165 100

63 41.5 17.8 26.6#

89 72.0 33.2 49.8#

182 2.4 2.1 3.1






Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF0

Ion 89.00 (88.70 to 89.70): BF0

Ion 182.00 (181.70 to 182.70): E





#57

Fluorene

Concen: 28.23 ng

RT: 10.54 min Scan# 718

Delta R.T. -0.01 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDIC025


Tot Ion:166 Resp: 717292

Ion Ratio Lower Upper

166 100

165 98.9 79.2 118.8




167 13.8 11.1 16.7



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#58

2,3,4,6-Tetrachlorophenol

Concen: 32.67 ng

RT: 10.28 min Scan# 695

Delta R.T. -0.05 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Tgt Ion:232 Resp: 174065


Ion Ratio Lower Upper

232 100

131 43.0 34.5 51.7

130 2.1 1.6 2.4

166 26.2 23.8 35.8






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E









#59
Diethylphthalate
Concen: 28.43 ng
RT: 10.42 min Scan# 707
Delta R.T. 0.00 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC025

Tot Ion:149 Resp: 781178
Ion Ratio Lower Upper
149 100
177 21.1 21.1 31.7#
150 12.8 10.2 15.2




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E






#60
4-Chlorophenyl-phenylether
Concen: 29.51 ng
RT: 10.54 min Scan# 718
Delta R.T. 0.00 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Tgt Ion:204 Resp: 340877
Ion Ratio Lower Upper
204 100
206 32.8 26.7 40.1
141 45.6 51.7 77.5#



Abundance Ion 204.00 (203.70 to 204.70): E
Ion 206.00 (205.70 to 206.70): E
Ion 141.00 (140.70 to 141.70): E





#61

4-Nitroaniline

Concen: 28.73 ng

RT: 10.55 min Scan# 719

Delta R.T. -0.01 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDIC025


Tot Ion:138 Resp: 199354

Ion Ratio Lower Upper

138 100

92 53.1 33.2 73.2




108 61.3 48.5 88.5



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BF0

Ion 108.00 (107.70 to 108.70): E



#62

Azobenzene

Concen: 26.20 ng

RT: 10.69 min Scan# 731

Delta R.T. -0.01 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Tgt Ion: 77 Resp: 831634


Ion Ratio Lower Upper

77 100

182 12.9 3.2 43.2

105 20.9 2.9 42.9

51 26.1 14.2 54.2






Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 182.00 (181.70 to 182.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 51.00 (50.70 to 51.70): BF0









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.48 min Scan# 800
Delta R.T. 0.00 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDIC025


Tot Ion:188 Resp: 699001
Ion Ratio Lower Upper
188 100
94 15.1 8.2 12.2#
80 17.9 8.8 13.2#



#64
4,6-Dinitro-2-methylphenol
Concen: 29.48 ng
RT: 10.59 min Scan# 722
Delta R.T. 0.00 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Tgt Ion:198 Resp: 125038
Ion Ratio Lower Upper
198 100
51 22.8 67.3 107.3#
105 33.1 42.6 82.6#





#65

n-Nitrosodiphenylamine

Concen: 27.78 ng

RT: 10.66 min Scan# 728

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :


BNA_F

ClientSampleId :

SSTDIC025

Tot Ion:169 Resp: 650810




Ion	Ratio	Lower	Upper
169	100		
168	68.4	53.8	80.6
167	37.4	29.8	44.8



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 28.09 ng

RT: 11.03 min Scan# 761


Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Tgt Ion:248 Resp: 192436



Ion	Ratio	Lower	Upper
248	100		
250	101.2	76.5	114.7
141	40.8	59.8	89.6




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E









#67
Hexachlorobenzene
Concen: 26.77 ng
RT: 11.09 min Scan# 766
Delta R.T. -0.01 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC025

Tot Ion:284 Resp: 204277
Ion Ratio Lower Upper
284 100
142 53.6 23.1 34.7#
249 35.6 26.3 39.5




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#68
Atrazine
Concen: 36.36 ng
RT: 11.18 min Scan# 774
Delta R.T. 0.00 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Tgt Ion:200 Resp: 209979
Ion Ratio Lower Upper
200 100
173 28.3 9.0 49.0
215 48.7 24.8 64.8



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 29.34 ng
 RT: 11.29 min Scan# 783
 Delta R.T. 0.00 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC025

Tot Ion:266 Resp: 133252
 Ion Ratio Lower Upper
 266 100
 268 64.6 50.9 76.3
 264 64.4 51.5 77.3




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E






#70
 Phenanthrene
 Concen: 28.03 ng
 RT: 11.50 min Scan# 802
 Delta R.T. -0.01 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52

Tgt Ion:178 Resp: 1003398
 Ion Ratio Lower Upper
 178 100
 176 19.6 17.3 25.9
 179 15.9 13.8 20.6



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E









#71
 Anthracene
 Concen: 29.75 ng
 RT: 11.56 min Scan# 807
 Delta R.T. 0.00 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC025


Tot Ion:178 Resp: 1087426
 Ion Ratio Lower Upper
 178 100
 176 19.5 16.8 25.2
 179 16.2 13.4 20.2



#72
 Carbazole
 Concen: 26.95 ng
 RT: 11.71 min Scan# 820
 Delta R.T. -0.01 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52

Tgt Ion:167 Resp: 915200
 Ion Ratio Lower Upper
 167 100
 166 22.0 19.1 28.7
 139 14.9 12.1 18.1





#73

Di-n-butylphthalate

Concen: 29.69 ng

RT: 12.05 min Scan# 850

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDIC025


Tot Ion:149 Resp: 1227128

Ion Ratio Lower Upper

149 100

150 9.1 8.6 12.8




104 3.6 4.7 7.1#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 31.57 ng

RT: 12.70 min Scan# 907

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52


Tgt Ion:202 Resp: 1108254

Ion Ratio Lower Upper

202 100

101 8.5 0.0 37.2



203 17.6 0.0 39.3




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 14.13 min Scan# 1032

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC025


Tot Ion:240 Resp: 737481

Ion Ratio Lower Upper

240 100

120 12.0 8.6 13.0




236 27.2 21.6 32.4



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 31.14 ng

RT: 12.82 min Scan# 917

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52


Tgt Ion:184 Resp: 603764

Ion Ratio Lower Upper

184 100

185 15.1 11.8 17.8



183 11.4 9.6 14.4




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E









#77
Pvrene
Concen: 23.99 ng
RT: 12.93 min Scan# 927
Delta R.T. 0.00 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC025

Tot Ion:202 Resp: 1179518
Ion Ratio Lower Upper
202 100
200 21.7 18.5 27.7
203 17.8 15.0 22.6




Abundance Ion 202.00 (201.70 to 202.70): E
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E






#78
Terphenyl-d14
Concen: 41.54 ng
RT: 13.07 min Scan# 939
Delta R.T. -0.01 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Tgt Ion:244 Resp: 1362679
Ion Ratio Lower Upper
244 100
212 7.5 6.8 10.2
122 16.7 13.3 19.9



Abundance Ion 244.00 (243.70 to 244.70): E
Ion 212.00 (211.70 to 212.70): E
Ion 122.00 (121.70 to 122.70): E





#79

Butylbenzylphthalate

Concen: 22.26 ng

RT: 13.55 min Scan# 981

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC025


Tot Ion:149 Resp: 561722

Ion Ratio Lower Upper

149 100

91 67.6 51.8 77.8




206 19.2 18.0 27.0



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BF0

Ion 206.00 (205.70 to 206.70): E



#80

Benzo(a)anthracene

Concen: 27.07 ng

RT: 14.12 min Scan# 1031

Delta R.T. 0.00 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52


Tgt Ion:228 Resp: 1119933

Ion Ratio Lower Upper

228 100

226 28.1 23.6 35.4



229 20.3 16.6 25.0




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#81
 3,3'-Dichlorobenzidine
 Concen: 26.08 ng
 RT: 14.09 min Scan# 1028
 Delta R.T. 0.00 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC025

Tot Ion:252 Resp: 391247
 Ion Ratio Lower Upper
 252 100
 254 64.8 52.1 78.1
 126 6.4 4.2 6.4#




Abundance Ion 252.00 (251.70 to 252.70): E
 Ion 254.00 (253.70 to 254.70): E
 Ion 126.00 (125.70 to 126.70): E






#82
 Chrysene
 Concen: 24.08 ng
 RT: 14.16 min Scan# 1034
 Delta R.T. 0.00 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52

Tgt Ion:228 Resp: 916952
 Ion Ratio Lower Upper
 228 100
 226 31.6 25.8 38.6
 229 20.5 16.9 25.3



Abundance Ion 228.00 (227.70 to 228.70): E
 Ion 226.00 (225.70 to 226.70): E
 Ion 229.00 (228.70 to 229.70): E









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 21.90 ng
 RT: 14.11 min Scan# 1030
 Delta R.T. 0.00 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC025


Tot Ion:149 Resp: 785485
 Ion Ratio Lower Upper
 149 100
 167 25.9 21.5 32.3
 279 2.5 2.6 3.8#



#84
 Di-n-octyl phthalate
 Concen: 23.47 ng
 RT: 14.73 min Scan# 1084
 Delta R.T. 0.00 min
 Lab File: BF090496.D
 Acq: 11 Oct 2016 11:52

Tgt Ion:149 Resp: 1247859
 Ion Ratio Lower Upper
 149 100
 167 1.5 1.3 1.9
 43 9.8 11.8 17.8#









#85
Indeno(1,2,3-cd)pyrene
Concen: 30.84 ng
RT: 17.08 min Scan# 1290
Delta R.T. -0.01 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDIC025


Tot Ion:276 Resp: 837419
Ion Ratio Lower Upper
276 100
138 33.0 23.8 35.8
277 25.1 20.4 30.6



#86
Perylene-d12
Concen: 20.00 ng
RT: 15.61 min Scan# 1161
Delta R.T. 0.00 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Tgt Ion:264 Resp: 571607
Ion Ratio Lower Upper
264 100
260 24.9 19.3 28.9
265 22.1 17.4 26.0









#87
Benzo(b)fluoranthene
Concen: 23.98 ng
RT: 15.17 min Scan# 1123
Delta R.T. 0.00 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC025

Tot Ion:252 Resp: 872942
Ion Ratio Lower Upper
252 100
253 22.7 18.2 27.2
125 14.4 7.0 10.4#




Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E






#88
Benzo(k)fluoranthene
Concen: 25.52 ng
RT: 15.17 min Scan# 1123
Delta R.T. -0.03 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Tgt Ion:252 Resp: 872942
Ion Ratio Lower Upper
252 100
253 22.7 18.6 27.8
125 14.4 11.7 17.5



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E









#89
Benzo(a)pyrene
Concen: 27.46 ng
RT: 15.54 min Scan# 1155
Delta R.T. -0.01 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDIC025


Tot Ion:252 Resp: 861678
Ion Ratio Lower Upper
252 100
253 22.8 18.2 27.4
125 16.6 9.8 14.8#



#90
Dibenzo(a,h)anthracene
Concen: 27.60 ng
RT: 17.10 min Scan# 1292
Delta R.T. 0.00 min
Lab File: BF090496.D
Acq: 11 Oct 2016 11:52

Tgt Ion:278 Resp: 736917
Ion Ratio Lower Upper
278 100
139 17.8 12.8 19.2
279 23.7 19.6 29.4





#91

Benzo(a,h,i)perylene

Concen: 26.90 ng

RT: 17.53 min Scan# 1329

Delta R.T. -0.01 min

Lab File: BF090496.D

Acq: 11 Oct 2016 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDIC025


Tot Ion: 276 Resp: 689142

Ion Ratio Lower Upper

276 100

277 23.4 19.4 29.0

138 27.6 20.8 31.2



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

