

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF101116\
 Data File : BF090517.D
 Acq On : 11 Oct 2016 22:13
 Operator : UM/SJ
 Sample : H5146-05
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 S8-0620

Quant Time: Oct 12 01:00:13 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF101116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 11 16:14:43 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	215726	20.00	ng	0.00
21) Naphthalene-d8	8.25	136	864471	20.00	ng	0.00
38) Acenaphthene-d10	10.01	164	475867	20.00	ng	0.00
63) Phenanthrene-d10	11.50	188	733719	20.00	ng	0.02
75) Chrysene-d12	14.13	240	486646	20.00	ng	0.00
86) Perylene-d12	15.62	264	407125	20.00	ng	0.01

System Monitoring Compounds

5) 2-Fluorophenol	5.56	112	685539	49.44	ng	0.00
7) Phenol-d6	6.59	99	745685	42.61	ng	0.00
23) Nitrobenzene-d5	7.51	82	1181910	74.15	ng	0.00
41) 2,4,6-Tribromophenol	10.81	330	588110	124.67	ng	0.01
44) 2-Fluorobiphenyl	9.32	172	2297162	81.12	ng	0.00
78) Terphenyl-d14	13.08	244	1594966	80.10	ng	0.00

Target Compounds


						Qvalue
9) Benzaldehyde	6.47	77	60727	5.51	ng	# 1
14) 1,2-Dichlorobenzene	7.13	146	49653	3.36	ng	97
15) Benzyl Alcohol	7.11	79	31730	2.36	ng	# 46
18) Hexachloroethane	7.49	117	24069	4.02	ng	# 1
19) n-Nitroso-di-n-propylamine	7.35	70	33309	2.98	ng	# 55
22) Acetophenone	7.35	105	70211	3.69	ng	# 67
27) 2,4-Dimethylphenol	7.93	122	101758	7.67	ng	# 92
31) Naphthalene	8.27	128	2345661	54.46	ng	99
32) Benzoic acid	8.05	122	102197	9.77	ng	# 24
33) 4-Chloroaniline	8.27	127	339965	18.65	ng	# 30
35) Caprolactam	8.62	113	136232	37.89	ng	# 18
36) 4-Chloro-3-methylphenol	8.83	107	47216	3.65	ng	# 33
37) 2-Methylnaphthalene	8.95	142	191565	6.63	ng	# 95
49) Dimethylphthalate	9.71	163	77130	2.41	ng	# 1
50) 2,6-Dinitrotoluene	9.78	165	43282	6.12	ng	# 76
55) 4-Nitrophenol	10.10	139	17744	2.60	ng	# 1
56) 2,4-Dinitrotoluene	10.21	165	41941	4.49	ng	# 58
59) Diethylphthalate	10.42	149	212618	6.53	ng	# 42


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF101116\
Data File : BF090517.D
Acq On : 11 Oct 2016 22:13
Operator : UM/SJ
Sample : H5146-05
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
S8-0620

Quant Time: Oct 12 01:00:13 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF101116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 11 16:14:43 2016
Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.95 min Scan# 404
Delta R.T. 0.00 min
Lab File: BF090517.D
Acq: 11 Oct 2016 22:13

Instrument :
BNA_F
ClientSampleId :
S8-0620




Tgt Ion:152 Resp: 215726
Ion Ratio Lower Upper
152 100
150 154.6 124.6 186.8
115 60.8 46.2 69.4




#5

2-Fluorophenol
Concen: 49.44 ng
RT: 5.56 min Scan# 282
Delta R.T. 0.00 min
Lab File: BF090517.D
Acq: 11 Oct 2016 22:13

Tgt Ion:112 Resp: 685539
Ion Ratio Lower Upper
112 100
64 65.8 57.9 86.9
63 36.1 29.6 44.4





#7

Phenol-d6

Concen: 42.61 ng

RT: 6.59 min Scan# 372

Delta R.T. 0.00 min

Lab File: BF090517.D

Acq: 11 Oct 2016 22:13

Instrument :

BNA_F

ClientSampleId :

S8-0620





Tgt Ion: 99 Resp: 745685

Ion Ratio Lower Upper

99 100

42 15.9 20.0 30.0#

71 59.4 27.9 41.9#



#9

Benzaldehyde

Concen: 5.51 ng

RT: 6.47 min Scan# 362

Delta R.T. -0.02 min

Lab File: BF090517.D

Acq: 11 Oct 2016 22:13




Tgt Ion: 77 Resp: 60727


Ion Ratio Lower Upper

77 100

105 755.5 73.1 113.1#

106 67.8 70.2 110.2#









#14
1,2-Dichlorobenzene
Concen: 3.36 ng
RT: 7.13 min Scan# 419
Delta R.T. 0.00 min
Lab File: BF090517.D
Acq: 11 Oct 2016 22:13




Instrument :
BNA_F
ClientSampleId :
S8-0620


Tgt Ion:146 Resp: 49653
Ion Ratio Lower Upper
146 100
148 63.3 52.3 78.5
111 45.2 38.2 57.4



#15
Benzyl Alcohol
Concen: 2.36 ng
RT: 7.11 min Scan# 418
Delta R.T. 0.02 min
Lab File: BF090517.D
Acq: 11 Oct 2016 22:13

Tgt Ion: 79 Resp: 31730
Ion Ratio Lower Upper
79 100
108 29.4 60.9 91.3#
77 106.0 51.5 77.3#





#18

Hexachloroethane

Concen: 4.02 ng

RT: 7.49 min Scan# 451

Delta R.T. 0.02 min

Lab File: BF090517.D

Acq: 11 Oct 2016 22:13

Instrument :

BNA_F

ClientSampleId :

S8-0620


Tgt Ion:117 Resp: 24069

Ion Ratio Lower Upper

117 100

119 443.1 75.9 113.9#



201 0.0 66.5 99.7#



Abundance Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E



#19

n-Nitroso-di-n-propylamine

Concen: 2.98 ng

RT: 7.35 min Scan# 439

Delta R.T. -0.01 min

Lab File: BF090517.D

Acq: 11 Oct 2016 22:13

Tgt Ion: 70 Resp: 33309



Ion Ratio Lower Upper

70 100

42 104.2 55.4 83.2#

101 33.7 7.6 11.4#

130 0.0 16.5 24.7#






Abundance Ion 70.00 (69.70 to 70.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 101.00 (100.70 to 101.70): E

Ion 130.00 (129.70 to 130.70): E









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.25 min Scan# 517
Delta R.T. 0.00 min
Lab File: BF090517.D
Acq: 11 Oct 2016 22:13




Instrument :
BNA_F
ClientSampleId :
S8-0620


Tgt Ion:	136	Resp:	864471
Ion	Ratio	Lower	Upper
136	100		
137	11.4	9.0	13.6
54	10.5	6.6	10.0#
68	10.6	5.0	7.6#



#22
Acetophenone
Concen: 3.69 ng
RT: 7.35 min Scan# 439
Delta R.T. -0.01 min
Lab File: BF090517.D
Acq: 11 Oct 2016 22:13

Tgt Ion:	105	Resp:	70211
Ion	Ratio	Lower	Upper
105	100		
71	22.1	3.8	5.6#
51	16.3	28.2	42.4#
120	7.7	17.4	26.2#









#23
 Nitrobenzene-d5
 Concen: 74.15 ng
 RT: 7.51 min Scan# 453
 Delta R.T. 0.00 min
 Lab File: BF090517.D
 Acq: 11 Oct 2016 22:13




Instrument :
 BNA_F
 ClientSampleId :
 S8-0620


Tgt Ion: 82 Resp: 1181910
 Ion Ratio Lower Upper
 82 100
 128 43.4 40.0 60.0
 54 45.4 42.6 64.0



#27
 2,4-Dimethylphenol
 Concen: 7.67 ng
 RT: 7.93 min Scan# 489
 Delta R.T. 0.03 min
 Lab File: BF090517.D
 Acq: 11 Oct 2016 22:13

Tgt Ion:122 Resp: 101758
 Ion Ratio Lower Upper
 122 100
 107 108.9 88.6 132.8
 121 44.4 47.5 71.3#









#31
Naphthalene
Concen: 54.46 ng
RT: 8.27 min Scan# 519
Delta R.T. 0.00 min
Lab File: BF090517.D
Acq: 11 Oct 2016 22:13




Instrument :
BNA_F
ClientSampleId :
S8-0620


Tgt Ion:128 Resp: 2345661
Ion Ratio Lower Upper
128 100
129 12.3 9.7 14.5
127 14.3 11.7 17.5



#32
Benzoic acid
Concen: 9.77 ng
RT: 8.05 min Scan# 500
Delta R.T. 0.03 min
Lab File: BF090517.D
Acq: 11 Oct 2016 22:13

Tgt Ion:122 Resp: 102197
Ion Ratio Lower Upper
122 100
105 22.2 103.9 143.9#
77 39.4 74.6 114.6#









#33
4-Chloroaniline
Concen: 18.65 ng
RT: 8.27 min Scan# 519
Delta R.T. -0.05 min
Lab File: BF090517.D
Acq: 11 Oct 2016 22:13




Instrument :
BNA_F
ClientSampleId :
S8-0620


Tgt Ion:	127	Resp:	339965
Ion	Ratio	Lower	Upper
127	100		
129	85.9	26.2	39.4#
65	0.0	32.5	48.7#
92	0.9	17.2	25.8#



#35
Caprolactam
Concen: 37.89 ng
RT: 8.62 min Scan# 550
Delta R.T. -0.06 min
Lab File: BF090517.D
Acq: 11 Oct 2016 22:13

Tgt Ion:	113	Resp:	136232
Ion	Ratio	Lower	Upper
113	100		
55	444.5	203.3	243.3#
56	167.7	140.5	180.5









#36
 4-Chloro-3-methylphenol
 Concen: 3.65 ng
 RT: 8.83 min Scan# 568
 Delta R.T. 0.03 min
 Lab File: BF090517.D
 Acq: 11 Oct 2016 22:13




Instrument :
 BNA_F
 ClientSampled :
 S8-0620


Tgt Ion:107 Resp: 47216
 Ion Ratio Lower Upper
 107 100
 144 38.3 19.8 29.8#
 142 7.5 60.7 91.1#



#37
 2-Methylnaphthalene
 Concen: 6.63 ng
 RT: 8.95 min Scan# 579
 Delta R.T. 0.00 min
 Lab File: BF090517.D
 Acq: 11 Oct 2016 22:13

Tgt Ion:142 Resp: 191565
 Ion Ratio Lower Upper
 142 100
 141 91.1 71.8 107.8
 115 44.9 28.8 43.2#





#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 10.01 min Scan# 671

Delta R.T. 0.00 min

Lab File: BF090517.D

Acq: 11 Oct 2016 22:13

Instrument :

BNA_F

ClientSampleId :

S8-0620


Tgt Ion:164 Resp: 475867

Ion Ratio Lower Upper

164 100

162 99.8 80.1 120.1




160 51.7 36.3 54.5



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#41

2,4,6-Tribromophenol

Concen: 124.67 ng

RT: 10.81 min Scan# 741

Delta R.T. 0.01 min

Lab File: BF090517.D

Acq: 11 Oct 2016 22:13


Tgt Ion:330 Resp: 588110

Ion Ratio Lower Upper

330 100

332 96.1 75.4 113.0



141 28.1 26.9 40.3




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E





#44

2-Fluorobiphenyl

Concen: 81.12 ng

RT: 9.32 min Scan# 611

Delta R.T. 0.00 min

Lab File: BF090517.D

Acq: 11 Oct 2016 22:13

Instrument :

BNA_F

ClientSampleId :

S8-0620


Tgt Ion:172 Resp: 2297162

Ion Ratio Lower Upper

172 100

171 36.6 31.8 47.8




170 24.5 21.7 32.5



Abundance Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E



#49

Dimethylphthalate

Concen: 2.41 ng

RT: 9.71 min Scan# 645

Delta R.T. -0.01 min

Lab File: BF090517.D

Acq: 11 Oct 2016 22:13


Tgt Ion:163 Resp: 77130

Ion Ratio Lower Upper

163 100

194 3.1 2.8 4.2



164 133.7 9.0 13.4#




Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E









#50
 2,6-Dinitrotoluene
 Concen: 6.12 ng
 RT: 9.78 min Scan# 651
 Delta R.T. 0.00 min
 Lab File: BF090517.D
 Acq: 11 Oct 2016 22:13




Instrument :
 BNA_F
 ClientSampleId :
 S8-0620


Tgt Ion:165 Resp: 43282
 Ion Ratio Lower Upper
 165 100
 63 61.1 58.0 87.0
 89 36.0 51.2 76.8#



#55
 4-Nitrophenol
 Concen: 2.60 ng
 RT: 10.10 min Scan# 679
 Delta R.T. -0.01 min
 Lab File: BF090517.D
 Acq: 11 Oct 2016 22:13

Tgt Ion:139 Resp: 17744
 Ion Ratio Lower Upper
 139 100
 109 268.5 43.6 83.6#
 65 89.4 87.5 127.5









#56
2,4-Dinitrotoluene
Concen: 4.49 ng
RT: 10.21 min Scan# 689
Delta R.T. 0.03 min
Lab File: BF090517.D
Acq: 11 Oct 2016 22:13




Instrument :
BNA_F
ClientSampleId :
S8-0620


Tgt Ion:165 Resp: 41941
Ion Ratio Lower Upper
165 100
63 68.6 17.8 26.6#
89 46.1 33.2 49.8
182 45.2 2.1 3.1#



#59
Diethylphthalate
Concen: 6.53 ng
RT: 10.42 min Scan# 707
Delta R.T. 0.00 min
Lab File: BF090517.D
Acq: 11 Oct 2016 22:13

Tgt Ion:149 Resp: 212618
Ion Ratio Lower Upper
149 100
177 24.6 21.1 31.7
150 80.9 10.2 15.2#





#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 11.50 min Scan# 802

Delta R.T. 0.02 min

Lab File: BF090517.D

Acq: 11 Oct 2016 22:13

Instrument :

BNA_F

ClientSampleId :

S8-0620


Tgt Ion:188 Resp: 733719

Ion Ratio Lower Upper

188 100

94 11.3 8.2 12.2




80 10.2 8.8 13.2



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF0

Ion 80.00 (79.70 to 80.70): BF0



#75

Chrysene-d12

Concen: 20.00 ng

RT: 14.13 min Scan# 1032

Delta R.T. 0.00 min

Lab File: BF090517.D

Acq: 11 Oct 2016 22:13


Tgt Ion:240 Resp: 486646

Ion Ratio Lower Upper

240 100

120 14.9 8.6 13.0#



236 26.9 21.6 32.4




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#78

Terphenyl-d14

Concen: 80.10 ng

RT: 13.08 min Scan# 940

Delta R.T. 0.00 min

Lab File: BF090517.D

Acq: 11 Oct 2016 22:13

Instrument :

BNA_F

ClientSampleId :

S8-0620


Tgt Ion:244 Resp: 1594966

Ion Ratio Lower Upper

244 100

212 7.3 6.8 10.2




122 14.2 13.3 19.9



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#86

Perylene-d12

Concen: 20.00 ng

RT: 15.62 min Scan# 1162

Delta R.T. 0.01 min

Lab File: BF090517.D

Acq: 11 Oct 2016 22:13


Tgt Ion:264 Resp: 407125

Ion Ratio Lower Upper

264 100

260 24.0 19.3 28.9

265 21.4 17.4 26.0



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

