

Data Path : Z:\HPCHEM1\BNA F\DATA\BF101117\
 Data File : BF099339.D
 Acq On : 11 Oct 2017 11:59
 Operator : SJ/JU
 Sample : I5745-09
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 HR-06-100617-C

Quant Time: Oct 11 16:26:25 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 11 16:25:21 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	101262	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	417611	20.00	ng	0.00
38) Acenaphthene-d10	9.89	164	193368	20.00	ng	0.00
63) Phenanthrene-d10	11.37	188	339478	20.00	ng	0.00
75) Chrysene-d12	14.01	240	192637	20.00	ng	0.00
86) Perylene-d12	15.47	264	173074	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	761727	119.75	ng	0.02
7) Phenol-d6	6.49	99	810266	103.26	ng	0.00
23) Nitrobenzene-d5	7.42	82	605651	93.01	ng	0.00
41) 2,4,6-Tribromophenol	10.68	330	225711	142.65	ng	0.00
44) 2-Fluorobiphenyl	9.21	172	1147533	99.07	ng	0.00
78) Terphenyl-d14	12.96	244	931131	109.93	ng	0.00
Target Compounds						Qvalue
15) Benzyl Alcohol	6.99	79	58348	10.136	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF101117\
Data File : BF099339.D
Acq On : 11 Oct 2017 11:59
Operator : SJ/JU
Sample : I5745-09
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
HR-06-100617-C

Quant Time: Oct 11 16:26:25 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 11 16:25:21 2017
Response via : Initial Calibration

