


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101118\
 Data File : BF109810.D
 Acq On : 12 Oct 2018 2:42
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF100818.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Oct 08 16:02:11 2018



AutoFind: Scans 1594, 1595, 1596; Background Corrected with Scan 1589

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.8	11680	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	35.8	12370	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	47.3	16316	PASS
197	198	0.00	2	0.3	120	PASS
198	198	100	100	100.0	34525	PASS
199	198	5	9	6.4	2224	PASS
275	198	10	60	26.8	9240	PASS
365	198	1	100	3.2	1121	PASS
441	198	0.01	100	13.5	4658	PASS
442	198	50	100	80.2	27690	PASS
443	442	15	24	18.6	5142	PASS