

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101124\
 Data File : BF139763.D
 Acq On : 11 Oct 2024 09:29
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 14 02:30:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF100124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 02 04:58:47 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	102	0.00
2	1,4-Dioxane	40.000	40.439	-1.1	106	-0.02
3	Pyridine	40.000	41.612	-4.0	110	-0.02
4	n-Nitrosodimethylamine	40.000	37.726	5.7	97	0.00
5 S	2-Fluorophenol	80.000	80.745	-0.9	106	0.00
6	Aniline	40.000	37.329	6.7	102	0.00
7 S	Phenol-d6	80.000	78.752	1.6	104	0.00
8	2-Chlorophenol	40.000	39.558	1.1	104	0.00
9	Benzaldehyde	40.000	38.663	3.3	103	0.00
10 C	Phenol	40.000	39.737	0.7	103	0.00
11	bis(2-Chloroethyl)ether	40.000	36.640	8.4	98	0.00
12	1,3-Dichlorobenzene	40.000	39.482	1.3	104	0.00
13 C	1,4-Dichlorobenzene	40.000	39.385	1.5	105	0.00
14	1,2-Dichlorobenzene	40.000	38.640	3.4	101	0.00
15	Benzyl Alcohol	40.000	37.643	5.9	97	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.656	8.4	96	0.00
17	2-Methylphenol	40.000	38.862	2.8	101	0.00
18	Hexachloroethane	40.000	38.477	3.8	101	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	35.887	10.3	95	0.00
20	3+4-Methylphenols	40.000	38.048	4.9	100	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	98	0.00
22	Acetophenone	40.000	38.073	4.8	98	0.00
23 S	Nitrobenzene-d5	80.000	76.980	3.8	98	0.00
24	Nitrobenzene	40.000	38.740	3.1	98	0.00
25	Isophorone	40.000	37.778	5.6	96	0.00
26 C	2-Nitrophenol	40.000	40.518	-1.3	98	0.00
27	2,4-Dimethylphenol	40.000	38.980	2.6	100	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.221	4.4	97	0.00
29 C	2,4-Dichlorophenol	40.000	39.488	1.3	99	0.00
30	1,2,4-Trichlorobenzene	40.000	39.634	0.9	100	0.00
31	Naphthalene	40.000	39.043	2.4	100	0.00
32	Benzoic acid	40.000	39.921	0.2	94	0.00
33	4-Chloroaniline	40.000	36.744	8.1	92	0.00
34 C	Hexachlorobutadiene	40.000	38.378	4.1	98	0.00
35	Caprolactam	40.000	38.265	4.3	98	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.324	6.7	96	0.00
37	2-Methylnaphthalene	40.000	37.536	6.2	96	0.00
38	1-Methylnaphthalene	40.000	37.826	5.4	97	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	96	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.828	2.9	96	0.00
41 P	Hexachlorocyclopentadiene	40.000	50.341	-25.9#	112	0.00
42 S	2,4,6-Tribromophenol	80.000	70.943	11.3	88	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.431	6.4	92	0.00
44	2,4,5-Trichlorophenol	40.000	37.855	5.4	94	0.00
45 S	2-Fluorobiphenyl	80.000	74.493	6.9	95	0.00
46	1,1'-Biphenyl	40.000	37.670	5.8	95	0.00
47	2-Chloronaphthalene	40.000	39.042	2.4	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101124\
 Data File : BF139763.D
 Acq On : 11 Oct 2024 09:29
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 14 02:30:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF100124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 02 04:58:47 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.213	7.0	91	0.00
49	Acenaphthylene	40.000	37.790	5.5	94	0.00
50	Dimethylphthalate	40.000	37.266	6.8	94	0.00
51	2,6-Dinitrotoluene	40.000	37.637	5.9	93	0.00
52 C	Acenaphthene	40.000	37.464	6.3	93	0.00
53	3-Nitroaniline	40.000	36.292	9.3	90	0.00
54 P	2,4-Dinitrophenol	40.000	32.506	18.7	76	0.00
55	Dibenzofuran	40.000	37.072	7.3	93	0.00
56 P	4-Nitrophenol	40.000	33.283	16.8	81	0.00
57	2,4-Dinitrotoluene	40.000	36.923	7.7	90	0.00
58	Fluorene	40.000	37.020	7.4	94	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	36.507	8.7	91	0.00
60	Diethylphthalate	40.000	35.893	10.3	90	0.00
61	4-Chlorophenyl-phenylether	40.000	36.755	8.1	94	0.00
62	4-Nitroaniline	40.000	35.756	10.6	89	0.00
63	Azobenzene	40.000	36.173	9.6	90	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	91	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.590	3.5	83	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.624	0.9	93	0.00
67	4-Bromophenyl-phenylether	40.000	38.768	3.1	91	0.00
68	Hexachlorobenzene	40.000	39.252	1.9	92	0.00
69	Atrazine	40.000	35.495	11.3	96	0.00
70 C	Pentachlorophenol	40.000	34.024	14.9	75	0.00
71	Phenanthrene	40.000	38.149	4.6	90	0.00
72	Anthracene	40.000	38.515	3.7	91	0.00
73	Carbazole	40.000	37.686	5.8	88	0.00
74	Di-n-butylphthalate	40.000	36.222	9.4	84	0.00
75 C	Fluoranthene	40.000	36.017	10.0	85	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	83	0.00
77	Benzydine	40.000	37.841	5.4	80	0.00
78	Pyrene	40.000	41.111	-2.8	84	0.00
79 S	Terphenyl-d14	80.000	79.468	0.7	83	0.00
80	Butylbenzylphthalate	40.000	38.750	3.1	79	0.00
81	Benzo(a)anthracene	40.000	38.584	3.5	84	0.00
82	3,3'-Dichlorobenzidine	40.000	39.588	1.0	84	0.00
83	Chrysene	40.000	39.068	2.3	82	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.572	-1.4	83	0.00
85 c	Di-n-octyl phthalate	40.000	41.327	-3.3	85	0.00
86 I	Perylene-d12	20.000	20.000	0.0	92	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	45.278	-13.2	102	0.00
88	Benzo(b)fluoranthene	40.000	35.003	12.5	91	0.00
89	Benzo(k)fluoranthene	40.000	39.461	1.3	89	0.00
90 C	Benzo(a)pyrene	40.000	39.320	1.7	93	0.00
91	Dibenzo(a,h)anthracene	40.000	45.346	-13.4	101	0.00
92	Benzo(g,h,i)perylene	40.000	46.580	-16.4	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101124\
Data File : BF139763.D
Acq On : 11 Oct 2024 09:29
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 14 02:30:57 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF100124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 02 04:58:47 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0