

Data Path : Z:\HPCHEM1\BNA F\Data\BF101215\
 Data File : BF082217.D
 Acq On : 12 Oct 2015 12:13
 Operator : UM/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 13 02:05:15 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF101015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 12 01:46:21 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
2	1,4-Dioxane	0.498	0.443	11.0	81	-0.05
3	Pyridine	1.158	1.064	8.1	80	-0.05
4	n-Nitrosodimethylamine	0.491	0.484	1.4	81	-0.06
5 S	2-Fluorophenol	0.991	0.949	4.2	78	-0.01
6	Aniline	1.663	1.638	1.5	81	-0.01
7 S	Phenol-d6	1.290	1.215	5.8	78	-0.01
8	2-Chlorophenol	1.210	1.167	3.6	84	-0.01
9	Benzaldehyde	0.659	0.676	-2.6	81	-0.01
10 C	Phenol	1.487	1.324	11.0	71	-0.01
11	bis(2-Chloroethyl)ether	1.257	1.126	10.4	75	-0.01
12	1,3-Dichlorobenzene	1.409	1.271	9.8	78	-0.01
13 C	1,4-Dichlorobenzene	1.471	1.291	12.2	74	-0.01
14	1,2-Dichlorobenzene	1.397	1.318	5.7	88	0.00
15	Benzyl Alcohol	0.890	0.956	-7.4	88	0.00
16	2,2'-oxybis(1-Chloropropane	1.751	1.666	4.9	87	0.00
17	2-Methylphenol	0.957	0.942	1.6	85	0.00
18	Hexachloroethane	0.504	0.447	11.3	78	0.00
19 P	n-Nitroso-di-n-propylamine	0.906	0.888	2.0	86	-0.01
20	3+4-Methylphenols	1.140	1.123	1.5	86	-0.01
21 I	Naphthalene-d8	1.000	1.000	0.0	83	0.00
22	Acetophenone	0.396	0.389	1.8	84	-0.01
23 S	Nitrobenzene-d5	0.298	0.287	3.7	79	-0.01
24	Nitrobenzene	0.322	0.313	2.8	90	0.00
25	Isophorone	0.601	0.583	3.0	84	0.00
26 C	2-Nitrophenol	0.161	0.159	1.2	74	-0.01
27	2,4-Dimethylphenol	0.259	0.278	-7.3	93	0.00
28	bis(2-Chloroethoxy)methane	0.350	0.335	4.3	75	-0.01
29 C	2,4-Dichlorophenol	0.259	0.250	3.5	75	0.00
30	1,2,4-Trichlorobenzene	0.299	0.281	6.0	79	0.00
31	Naphthalene	0.867	0.850	2.0	85	0.00
32	Benzoic acid	0.174	0.127	27.0#	62	-0.02
33	4-Chloroaniline	0.360	0.351	2.5	77	-0.01
34 C	Hexachlorobutadiene	0.186	0.169	9.1	77	0.00
35	Caprolactam	0.075	0.078	-4.0	81	-0.01
36 C	4-Chloro-3-methylphenol	0.254	0.278	-9.4	92	0.00
37	2-Methylnaphthalene	0.601	0.576	4.2	79	-0.01
38 I	Acenaphthene-d10	1.000	1.000	0.0	81	-0.01
39	1,2,4,5-Tetrachlorobenzene	0.595	0.593	0.3	86	0.00
40 P	Hexachlorocyclopentadiene	0.245	0.205	16.3	64	-0.01
41 S	2,4,6-Tribromophenol	0.188	0.199	-5.9	91	0.00
42 C	2,4,6-Trichlorophenol	0.395	0.400	-1.3	88	0.00
43	2,4,5-Trichlorophenol	0.428	0.415	3.0	80	-0.01
44 S	2-Fluorobiphenyl	1.206	1.143	5.2	73	-0.01

Data Path : Z:\HPCHEM1\BNA F\Data\BF101215\
 Data File : BF082217.D
 Acq On : 12 Oct 2015 12:13
 Operator : UM/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 13 02:05:15 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF101015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 12 01:46:21 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.525	1.532	-0.5	89	0.00
46	2-Chloronaphthalene	1.201	1.150	4.2	80	-0.01
47	2-Nitroaniline	0.330	0.361	-9.4	92	0.00
48	Acenaphthylene	1.870	1.783	4.7	79	-0.01
49	Dimethylphthalate	1.408	1.417	-0.6	92	0.00
50	2,6-Dinitrotoluene	0.297	0.290	2.4	77	-0.01
51 C	Acenaphthene	1.123	1.029	8.4	74	-0.01
52	3-Nitroaniline	0.336	0.337	-0.3	80	-0.01
53 P	2,4-Dinitrophenol	0.143	0.111	22.4	59	0.00
54	Dibenzofuran	1.541	1.483	3.8	78	-0.01
55 P	4-Nitrophenol	0.192	0.207	-7.8	82	0.00
56	2,4-Dinitrotoluene	0.371	0.396	-6.7	85	0.00
57	Fluorene	1.266	1.234	2.5	81	-0.01
58	2,3,4,6-Tetrachlorophenol	0.350	0.361	-3.1	91	0.00
59	Diethylphthalate	1.313	1.371	-4.4	88	0.00
60	4-Chlorophenyl-phenylether	0.580	0.622	-7.2	93	0.00
61	4-Nitroaniline	0.326	0.342	-4.9	82	-0.01
62	Azobenzene	1.285	1.315	-2.3	89	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	86	-0.01
64	4,6-Dinitro-2-methylphenol	0.112	0.096	14.3	67	-0.01
65 c	n-Nitrosodiphenylamine	0.599	0.573	4.3	84	-0.01
66	4-Bromophenyl-phenylether	0.200	0.197	1.5	87	0.00
67	Hexachlorobenzene	0.216	0.198	8.3	80	-0.01
68	Atrazine	0.190	0.211	-11.1	106	0.00
69 C	Pentachlorophenol	0.111	0.106	4.5	74	-0.01
70	Phenanthrene	0.980	1.006	-2.7	94	0.00
71	Anthracene	1.010	0.972	3.8	80	-0.01
72	Carbazole	0.922	0.916	0.7	88	0.00
73	Di-n-butylphthalate	1.137	1.139	-0.2	88	0.00
74 C	Fluoranthene	1.053	1.066	-1.2	86	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	90	0.00
76	Benzidine	0.580	0.478	17.6	73	0.00
77	Pyrene	1.187	1.149	3.2	91	0.00
78 S	Terphenyl-d14	0.755	0.635	15.9	78	-0.01
79	Butylbenzylphthalate	0.542	0.520	4.1	92	0.00
80	Benzo(a)anthracene	1.041	1.003	3.7	91	0.00
81	3,3'-Dichlorobenzidine	0.395	0.395	0.0	93	0.00
82	Chrysene	1.096	1.005	8.3	95	0.00
83	Bis(2-ethylhexyl)phthalate	0.773	0.663	14.2	78	-0.01
84 c	Di-n-octyl phthalate	1.327	1.106	16.7	80	0.00
85	Indeno(1,2,3-cd)pyrene	0.872	0.817	6.3	95	0.00
86 I	Perylene-d12	1.000	1.000	0.0	93	0.00
87	Benzo(b)fluoranthene	1.217	1.066	12.4	74	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF101215\
Data File : BF082217.D
Acq On : 12 Oct 2015 12:13
Operator : UM/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 13 02:05:15 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF101015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 12 01:46:21 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.023	1.132	-10.7	123	0.00
89 C	Benzo(a)pyrene	1.046	0.995	4.9	92	0.00
90	Dibenzo(a,h)anthracene	0.857	0.829	3.3	95	0.00
91	Benzo(a,h,i)perylene	0.832	0.803	3.5	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0