

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF101217\
 Data File : BF099419.D
 Acq On : 13 Oct 2017 2:34
 Operator : SJ/JU
 Sample : I5565-01
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RH-1

Manual Integrations
 APPROVED

Sohil
 10/13/2017 3:06:44 PM

Quant Time: Oct 13 05:30:25 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 11 16:25:21 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	105443	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	393433	20.00	ng	0.00
38) Acenaphthene-d10	9.89	164	148311	20.00	ng	0.00
63) Phenanthrene-d10	11.37	188	236881	20.00	ng	0.00
75) Chrysene-d12	14.02	240	203060	20.00	ng	0.00
86) Perylene-d12	15.50	264	143078	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.48	112	702222	106.02	ng	0.01
7) Phenol-d6	6.49	99	851852	104.25	ng	0.00
23) Nitrobenzene-d5	7.42	82	524831	85.55	ng	0.00
41) 2,4,6-Tribromophenol	10.68	330	124952	102.96	ng	0.00
44) 2-Fluorobiphenyl	9.21	172	846216	95.25	ng	0.00
78) Terphenyl-d14	12.96	244	624785	69.97	ng	0.00
Target Compounds						
10) Phenol	6.50	94	20388	2.231	ng	98
31) Naphthalene	8.16	128	126807	6.863	ng	98
37) 2-Methylnaphthalene	8.85	142	122630	10.209	ng	99
49) Dimethylphthalate	9.60	163	145417	12.991	ng	100
54) Dibenzofuran	10.09	168	35957	2.910	ng	# 90
70) Phenanthrene	11.40	178	119189	9.400	ng	98
74) Fluoranthene	12.59	202	76017	5.921	ng	98
77) Pyrene	12.82	202	79011	5.103	ng	99
80) Benzo(a)anthracene	14.01	228	50695	4.123	ng	# 90
82) Chrysene	14.04	228	73889	6.312	ng	# 93
87) Benzo(b)fluoranthene	15.06	252	60290m	6.853	ng	
89) Benzo(a)pyrene	15.43	252	29922	3.705	ng	# 80
91) Benzo(a,h,i)perylene	17.43	276	17404	2.724	ng	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF101217\
Data File : BF099419.D
Acq On : 13 Oct 2017 2:34
Operator : SJ/JU
Sample : I5565-01
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RH-1

Manual Integrations
APPROVED

Sohil
10/13/2017 3:06:44 PM

Quant Time: Oct 13 05:30:25 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 11 16:25:21 2017
Response via : Initial Calibration

