

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF101217\
 Data File : BF099421.D
 Acq On : 13 Oct 2017 3:29
 Operator : SJ/JU
 Sample : I5780-04
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PUMPING-PLANT-1-5-COMP

Quant Time: Oct 13 06:48:21 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 11 16:25:21 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	107552	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	389782	20.00	ng	0.00
38) Acenaphthene-d10	9.89	164	140572	20.00	ng	0.00
63) Phenanthrene-d10	11.37	188	224904	20.00	ng	0.00
75) Chrysene-d12	14.02	240	214164	20.00	ng	0.00
86) Perylene-d12	15.49	264	156594	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	760800	112.61	ng	0.02
7) Phenol-d6	6.49	99	785692	94.27	ng	0.00
23) Nitrobenzene-d5	7.42	82	549245	90.37	ng	0.00
41) 2,4,6-Tribromophenol	10.68	330	146213	127.11	ng	0.00
44) 2-Fluorobiphenyl	9.21	172	926989	110.09	ng	0.00
78) Terphenyl-d14	12.96	244	759762	80.68	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF101217\
Data File : BF099421.D
Acq On : 13 Oct 2017 3:29
Operator : SJ/JU
Sample : I5780-04
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PUMPING-PLANT-1-5-COMP

Quant Time: Oct 13 06:48:21 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 11 16:25:21 2017
Response via : Initial Calibration

