

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF101217\
 Data File : BF099427.D
 Acq On : 13 Oct 2017 7:53
 Operator : SJ/JU
 Sample : I5777-01
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 100917-TIDO-F-D-B

Manual Integrations
 APPROVED

Sohil
 10/13/2017 3:06:49 PM

Quant Time: Oct 13 08:26:26 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 11 16:25:21 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	102024	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	376390	20.00	ng	0.00
38) Acenaphthene-d10	9.89	164	140576	20.00	ng	0.00
63) Phenanthrene-d10	11.37	188	221779	20.00	ng	0.00
75) Chrysene-d12	14.02	240	202674	20.00	ng	0.00
86) Perylene-d12	15.49	264	153260	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.47	112	309225	48.25	ng	0.00
7) Phenol-d6	6.48	99	223361	28.25	ng	0.00
23) Nitrobenzene-d5	7.42	82	612309	104.33	ng	0.00
41) 2,4,6-Tribromophenol	10.68	330	102412	89.03	ng	0.00
44) 2-Fluorobiphenyl	9.21	172	1010399	119.99	ng	0.00
78) Terphenyl-d14	12.96	244	785471	88.14	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.60	163	25542	2.407	ng	Qvalue # 99
70) Phenanthrene	11.40	178	35324	2.976	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF101217\
Data File : BF099427.D
Acq On : 13 Oct 2017 7:53
Operator : SJ/JU
Sample : I5777-01
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
100917-TIDO-F-D-B

Manual Integrations
APPROVED

Sohil
10/13/2017 3:06:49 PM

Quant Time: Oct 13 08:26:26 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 11 16:25:21 2017
Response via : Initial Calibration

