

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF101217\
 Data File : BF099437.D
 Acq On : 13 Oct 2017 12:30
 Operator : SJ/JU
 Sample : I5729-06DL 10X
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 3W-(1-2)DL

Manual Integrations
 APPROVED

Sohil
 10/15/2017 12:26:15 PM

Quant Time: Oct 13 15:07:35 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 11 16:25:21 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	98597	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	366104	20.00	ng	0.00
38) Acenaphthene-d10	9.89	164	138267	20.00	ng	0.00
63) Phenanthrene-d10	11.37	188	211117	20.00	ng	0.00
75) Chrysene-d12	14.02	240	184674	20.00	ng	0.00
86) Perylene-d12	15.49	264	157728	20.00	ng	0.01

System Monitoring Compounds						
5) 2-Fluorophenol	5.47	112	73215	11.82	ng	0.00
7) Phenol-d6	6.48	99	84736	11.09	ng	0.00
23) Nitrobenzene-d5	7.42	82	46425	8.13	ng	0.00
41) 2,4,6-Tribromophenol	10.68	330	10729	9.48	ng	0.00
44) 2-Fluorobiphenyl	9.20	172	87413	10.55	ng	0.00
78) Terphenyl-d14	12.95	244	62086	7.65	ng	0.00

Target Compounds				Qvalue		
51) Acenaphthene	9.92	154	20923	2.418	ng	98
54) Dibenzofuran	10.09	168	28495	2.474	ng	96
57) Fluorene	10.43	166	23612	2.550	ng	96
70) Phenanthrene	11.40	178	325339	28.790	ng	99
71) Anthracene	11.45	178	79730	6.896	ng	98
72) Carbazole	11.61	167	37218	3.287	ng	99
74) Fluoranthene	12.59	202	635032	55.495	ng	99
77) Pyrene	12.82	202	537046	38.137	ng	99
80) Benzo(a)anthracene	14.01	228	233124	20.847	ng	98
82) Chrysene	14.04	228	214525	20.150	ng	97
85) Indeno(1,2,3-cd)pyrene	16.97	276	74429	8.740	ng	98
87) Benzo(b)fluoranthene	15.06	252	218003m	22.480	ng	
88) Benzo(k)fluoranthene	15.09	252	85715m	8.949	ng	
89) Benzo(a)pyrene	15.43	252	135957	15.273	ng	# 93
90) Dibenzo(a,h)anthracene	16.97	278	19230	2.699	ng	# 72
91) Benzo(g,h,i)perylene	17.42	276	70031	9.945	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF101217\
Data File : BF099437.D
Acq On : 13 Oct 2017 12:30
Operator : SJ/JU
Sample : I5729-06DL 10X
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
3W-(1-2)DL

Manual Integrations
APPROVED

Sohil
10/15/2017 12:26:15 PM

Quant Time: Oct 13 15:07:35 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 11 16:25:21 2017
Response via : Initial Calibration

