

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF101218\
 Data File : BF109855.D
 Acq On : 13 Oct 2018 1:29
 Operator : JU/SJ
 Sample : J5455-04
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MS-CON

Quant Time: Oct 13 03:50:25 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF100818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 08 16:02:11 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.69	152	85324	20.00	ng	0.00
21) Naphthalene-d8	7.97	136	343471	20.00	ng	0.00
38) Acenaphthene-d10	9.72	164	142204	20.00	ng	0.00
63) Phenanthrene-d10	11.20	188	224216	20.00	ng	0.00
75) Chrysene-d12	13.83	240	196359	20.00	ng	0.00
86) Perylene-d12	15.23	264	150153	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.32	112	391633	81.47	ng	0.00
7) Phenol-d6	6.34	99	552346	86.02	ng	-0.02
23) Nitrobenzene-d5	7.26	82	356619	57.23	ng	-0.01
41) 2,4,6-Tribromophenol	10.51	330	92405	59.64	ng	0.00
44) 2-Fluorobiphenyl	9.05	172	660841	64.00	ng	-0.01
78) Terphenyl-d14	12.77	244	474760	46.80	ng	-0.01
Target Compounds						Qvalue
49) Dimethylphthalate	9.45	163	37239	3.571	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF101218\
Data File : BF109855.D
Acq On : 13 Oct 2018 1:29
Operator : JU/SJ
Sample : J5455-04
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MS-CON

Quant Time: Oct 13 03:50:25 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF100818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 08 16:02:11 2018
Response via : Initial Calibration






#1



1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.69 min Scan# 783
Delta R.T. -0.01 min
Lab File: BF109855.D
Acq: 13 Oct 2018 1:29

Instrument :
BNA_F
ClientSampleId :
MS-CON

Tot Ion:152 Resp: 85324
Ion Ratio Lower Upper
152 100
150 163.1 122.7 184.1
115 65.4 49.1 73.7




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 81.473 ng
RT: 5.32 min Scan# 549
Delta R.T. -0.00 min
Lab File: BF109855.D
Acq: 13 Oct 2018 1:29

Tgt Ion:112 Resp: 391633
Ion Ratio Lower Upper
112 100
64 55.4 44.1 66.1
63 30.5 23.7 35.5



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 86.015 ng

RT: 6.34 min Scan# 723

Delta R.T. -0.02 min

Lab File: BF109855.D

Acq: 13 Oct 2018 1:29

Instrument :

BNA_F

ClientSampled :

MS-CON



Tot Ion: 99 Resp: 552346

Ion Ratio Lower Upper

99 100

42 20.5 15.8 23.6



71 37.9 28.0 42.0



Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 7.97 min Scan# 1001

Delta R.T. -0.01 min

Lab File: BF109855.D

Acq: 13 Oct 2018 1:29

Tgt Ion: 136 Resp: 343471



Ion Ratio Lower Upper

136 100

137 11.1 8.8 13.2

54 6.7 5.4 8.2

68 5.1 4.2 6.2





Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1









#23
 Nitrobenzene-d5
 Concen: 57.234 ng
 RT: 7.26 min Scan# 879
 Delta R.T. -0.01 min
 Lab File: BF109855.D
 Acq: 13 Oct 2018 1:29




Instrument :
 BNA_F
 ClientSampleId :
 MS-CON


Tot Ion	82	Resp	356619
Ion Ratio	Lower	Upper	
82	100		
128	41.7	34.2	51.2
54	47.2	38.6	58.0



#38
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 9.72 min Scan# 1298
 Delta R.T. -0.01 min
 Lab File: BF109855.D
 Acq: 13 Oct 2018 1:29

Tgt Ion	164	Resp	142204
Ion Ratio	Lower	Upper	
164	100		
162	103.9	83.0	124.6
160	47.4	37.0	55.6






#41
 2,4,6-Tribromophenol
 Concen: 59.637 ng
 RT: 10.51 min Scan# 1432
 Delta R.T. -0.01 min
 Lab File: BF109855.D
 Acq: 13 Oct 2018 1:29

Instrument :
 BNA_F
 ClientSampleId :
 MS-CON

Tot Ion	Ratio	Lower	Upper
330	100		
332	96.1	77.1	115.7
141	34.1	27.0	40.4






Abundance

Ion 329.80 (329.50 to 330.50): E


Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E



#44
 2-Fluorobiphenyl
 Concen: 64.003 ng
 RT: 9.05 min Scan# 1183
 Delta R.T. -0.01 min
 Lab File: BF109855.D
 Acq: 13 Oct 2018 1:29

Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.5	28.6	43.0
170	24.1	19.7	29.5






Abundance

Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E





#49

Dimethylphthalate

Concen: 3.571 ng

RT: 9.45 min Scan# 1251

Delta R.T. -0.02 min

Lab File: BF109855.D

Acq: 13 Oct 2018 1:29

Instrument :

BNA_F

ClientSampleId :

MS-CON


Tot Ion:163 Resp: 37239

Ion Ratio Lower Upper

163 100

194 4.6 3.2 4.8




164 10.5 8.1 12.1



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#63

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.20 min Scan# 1549

Delta R.T. -0.01 min

Lab File: BF109855.D

Acq: 13 Oct 2018 1:29


Tgt Ion:188 Resp: 224216

Ion Ratio Lower Upper

188 100

94 9.1 7.2 10.8



80 10.0 7.9 11.9




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1





#75

Chrysene-d12

Concen: 20.000 ng

RT: 13.83 min Scan# 1996

Delta R.T. -0.01 min

Lab File: BF109855.D

Acq: 13 Oct 2018 1:29

Instrument :

BNA_F

ClientSampleId :

MS-CON


Tot Ion:240 Resp: 196359

Ion Ratio Lower Upper

240 100

120 10.4 8.3 12.5




236 25.2 21.2 31.8



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#78

Terphenyl-d14

Concen: 46.805 ng

RT: 12.77 min Scan# 1817

Delta R.T. -0.01 min

Lab File: BF109855.D

Acq: 13 Oct 2018 1:29


Tgt Ion:244 Resp: 474760

Ion Ratio Lower Upper

244 100

212 7.2 5.8 8.8



122 11.7 8.7 13.1




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#86
Pervlene-d12
Concen: 20.000 ng
RT: 15.23 min Scan# 2234
Delta R.T. 0.00 min
Lab File: BF109855.D
Acq: 13 Oct 2018 1:29

Instrument :
BNA_F
ClientSampleId :
MS-CON

Tot Ion	Ratio	Lower	Upper
264	100		
260	23.7	18.7	28.1
265	21.1	16.7	25.1

