


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF101218\
 Data File : BF109851.D
 Acq On : 12 Oct 2018 23:40
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF100818.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Oct 08 16:02:11 2018



AutoFind: Scans 1593, 1594, 1595; Background Corrected with Scan 1589

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.0	15711	PASS
68	69	0.00	2	1.3	220	PASS
69	198	0.00	100	38.8	16495	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.8	21136	PASS
197	198	0.00	2	0.7	306	PASS
198	198	100	100	100.0	42477	PASS
199	198	5	9	6.9	2943	PASS
275	198	10	60	26.3	11158	PASS
365	198	1	100	3.3	1413	PASS
441	198	0.01	100	10.6	4512	PASS
442	198	50	100	68.2	28976	PASS
443	442	15	24	18.9	5470	PASS