



Benzaldehyde

Response Ratio




$R = -1.526e-001 A^2 + 1.232e+000 A + 6.575e-002$
Coef of Det (r^2) = 0.997023 Curve Fit: Quadratic
Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101220.M
Calibration Table Last Updated: Mon Oct 12 16:22:33 2020

Benzoic acid



$\text{Response} = 2.227\text{e-}001 * \text{Amt} - 8.556\text{e-}002$
 Coef of Det (r^2) = 0.990634 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101220.M
 Calibration Table Last Updated: Mon Oct 12 16:22:33 2020

Hexachlorocyclopentadiene




$$R = 3.469e-002 A^2 + 8.271e-002 A - 4.161e-002$$

Coef of Det (r^2) = 0.998362 Curve Fit: Quadratic

Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101220.M


Calibration Table Last Updated: Mon Oct 12 16:22:33 2020

2,4-Dinitrophenol




Response = $1.837 \times 10^{-1} \times \text{Amt} - 7.977 \times 10^{-2}$
 Coef of Det (r^2) = 0.990820 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101220.M
 Calibration Table Last Updated: Mon Oct 12 16:22:33 2020

4,6-Dinitro-2-methylphenol



Response = 1.476e-001 * Amt - 3.273e-002
 Coef of Det (r^2) = 0.995742 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101220.M
 Calibration Table Last Updated: Mon Oct 12 16:22:33 2020

Pentachlorophenol



Response = 1.142e-001 * Amt - 4.535e-002
 Coef of Det (r^2) = 0.991122 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101220.M
 Calibration Table Last Updated: Mon Oct 12 16:22:33 2020