

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101222\
 Data File : BF130629.D
 Acq On : 12 Oct 2022 14:42
 Operator : CG\JU
 Sample : N5067-14
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 C24(23-24)

Quant Time: Oct 13 05:44:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 12 06:52:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.587	152	87257	20.000	ng	0.00
21) Naphthalene-d8	7.869	136	341794	20.000	ng	0.00
39) Acenaphthene-d10	9.616	164	185372	20.000	ng	0.00
64) Phenanthrene-d10	11.092	188	310037	20.000	ng	0.00
76) Chrysene-d12	13.722	240	157741	20.000	ng	0.00
86) Perylene-d12	15.098	264	156100	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.210	112	676496	139.750	ng	0.02
7) Phenol-d6	6.240	99	734645	120.707	ng	0.00
23) Nitrobenzene-d5	7.157	82	591276	94.437	ng	0.00
42) 2,4,6-Tribromophenol	10.410	330	300834	165.853	ng	0.00
45) 2-Fluorobiphenyl	8.945	172	1129981	92.335	ng	0.00
79) Terphenyl-d14	12.680	244	1167826	122.877	ng	0.00
Target Compounds						
77) Benzidine	12.439	184	15163	3.803	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101222\
Data File : BF130629.D
Acq On : 12 Oct 2022 14:42
Operator : CG\JU
Sample : N5067-14
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
C24(23-24)

Quant Time: Oct 13 05:44:30 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 12 06:52:52 2022
Response via : Initial Calibration

