


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101222\
 Data File : BF130621.D
 Acq On : 12 Oct 2022 09:21
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101122.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Oct 12 06:52:52 2022






AutoFind: Scans 1576, 1577, 1578; Background Corrected with Scan 1571

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.1	53352	PASS
68	69	0.00	2	2.0	1099	PASS
69	198	0.00	100	35.2	55109	PASS
70	69	0.00	2	0.5	294	PASS
127	198	10	80	52.2	81645	PASS
197	198	0.00	2	0.4	637	PASS
198	198	100	100	100.0	156515	PASS
199	198	5	9	6.8	10652	PASS
275	198	10	60	29.5	46216	PASS
365	198	1	100	3.9	6167	PASS
441	198	0.01	100	15.3	23980	PASS
442	442	50	100	100.0	151077	PASS
443	442	15	24	19.1	28805	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101222\
 Data File : BF130621.D
 Acq On : 12 Oct 2022 09:21
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Oct 13 07:28:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 12 06:52:52 2022
 Response via : Initial Calibration



TIC: BF130621.D\data.ms

(70) Pentachlorophenol (C)

10.910min (-0.000) 172831.30 ng




response 123894

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.50	62.66
264.00	59.40	60.11
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101222\
Data File : BF130621.D
Acq On : 12 Oct 2022 09:21
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 13 07:28:57 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 12 06:52:52 2022
Response via : Initial Calibration



TIC: BF130621.D\data.ms

(77) Benzidine

12.445min (+ 0.006) 110415.79 ng

response 789826

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.00	14.11
183.00	11.20	11.74
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/12/2022	BNA_F	<u>BF130621.D</u>
Compound Name	Response	Retention Time
DDT	382423	13.251
DDD	3089	12.98
DDE	128	12.651
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
3217	385640	0.83

Instrument :
BNA_F
ClientSampleId :
DFTPP