

Data Path : Z:\HPCHEM1\BNA F\DATA\BF101316\
 Data File : BF090552.D
 Acq On : 13 Oct 2016 20:13
 Operator : UM/SJ
 Sample : SSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Manual Integrations
 APPROVED

sohil
 10/14/2016 6:38:47 PM

Quant Time: Oct 14 01:30:55 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF101316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 14 01:18:46 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.92	152	206955	20.00	ng	0.00
21) Naphthalene-d8	8.21	136	873169	20.00	ng	0.00
38) Acenaphthene-d10	9.97	164	469995	20.00	ng	0.00
63) Phenanthrene-d10	11.46	188	762377	20.00	ng	0.00
75) Chrysene-d12	14.10	240	633298	20.00	ng	0.00
86) Perylene-d12	15.56	264	372646	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.53	112	1199917	91.91	ng	0.00
7) Phenol-d6	6.57	99	1695817	100.61	ng	0.01
23) Nitrobenzene-d5	7.49	82	1438269	88.26	ng	0.00
41) 2,4,6-Tribromophenol	10.76	330	443447	97.65	ng	0.00
44) 2-Fluorobiphenyl	9.29	172	2509671	90.28	ng	0.00
78) Terphenyl-d14	13.05	244	2740643	105.49	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.57	88	318992	48.84	ng	95
3) Pyridine	3.31	79	733837	41.74	ng	93
4) n-Nitrosodimethylamine	3.27	42	241491	41.75	ng	78
6) Aniline	6.59	93	1010732	47.41	ng	99
8) 2-Chlorophenol	6.70	128	729532	51.15	ng	98
9) Benzaldehyde	6.47	77	470758	43.94	ng	96
10) Phenol	6.58	94	1006198	53.22	ng	94
11) bis(2-Chloroethyl)ether	6.66	93	802326	55.34	ng	99
12) 1,3-Dichlorobenzene	6.86	146	736759	48.78	ng	98
13) 1,4-Dichlorobenzene	6.94	146	764761	48.56	ng	99
14) 1,2-Dichlorobenzene	7.09	146	716845	50.40	ng	99
15) Benzyl Alcohol	7.07	79	541172	41.84	ng	96
16) 2,2'-oxybis(1-Chloropropan	7.19	45	905697	42.89	ng	91
17) 2-Methylphenol	7.18	107	570531	48.80	ng	98
18) Hexachloroethane	7.43	117	288928	49.19	ng	94
19) n-Nitroso-di-n-propylamine	7.34	70	531869	47.95	ng	# 86
20) 3+4-Methylphenols	7.34	107	653922	44.89	ng	# 81
22) Acetophenone	7.33	105	914908	47.41	ng	# 94
24) Nitrobenzene	7.51	77	738454	43.28	ng	# 80
25) Isophorone	7.74	82	1336532	45.54	ng	99
26) 2-Nitrophenol	7.82	139	383491	49.36	ng	91
27) 2,4-Dimethylphenol	7.87	122	595259	45.08	ng	98
28) bis(2-Chloroethoxy)methane	7.96	93	832490	45.44	ng	100
29) 2,4-Dichlorophenol	8.07	162	530689	48.12	ng	97
30) 1,2,4-Trichlorobenzene	8.15	180	627998	49.04	ng	99
31) Naphthalene	8.23	128	2117595	48.31	ng	99
32) Benzoic acid	8.01	122	390913	38.91	ng	93
33) 4-Chloroaniline	8.28	127	845640	46.11	ng	97
34) Hexachlorobutadiene	8.35	225	357311	50.14	ng	99
35) Caprolactam	8.66	113	148144	41.47	ng	# 82
36) 4-Chloro-3-methylphenol	8.77	107	595148	46.06	ng	# 75
37) 2-Methylnaphthalene	8.92	142	1247786	42.84	ng	99
39) 1,2,4,5-Tetrachlorobenzene	9.09	216	651806	50.38	ng	98
40) Hexachlorocyclopentadiene	9.08	237	335423	52.72	ng	97

Data Path : Z:\HPCHEM1\BNA F\DATA\BF101316\
 Data File : BF090552.D
 Acq On : 13 Oct 2016 20:13
 Operator : UM/SJ
 Sample : SSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Manual Integrations
 APPROVED

sohil
 10/14/2016 6:38:47 PM

Quant Time: Oct 14 01:30:55 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF101316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 14 01:18:46 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	9.21	196	379826	43.64	nq	99
43) 2,4,5-Trichlorophenol	9.24	196	418569	48.49	nq	97
45) 1,1'-Biphenyl	9.39	154	1645385	46.28	nq	97
46) 2-Chloronaphthalene	9.41	162	1203135	45.42	nq	98
47) 2-Nitroaniline	9.51	65	362939	37.95	nq	# 76
48) Acenaphthylene	9.83	152	2076607	48.22	nq	99
49) Dimethylphthalate	9.70	163	1503425	48.14	nq	# 98
50) 2,6-Dinitrotoluene	9.75	165	348609	49.82	nq	95
51) Acenaphthene	10.01	154	1425614	49.63	nq	99
52) 3-Nitroaniline	9.93	138	424157	48.98	nq	96
53) 2,4-Dinitrophenol	10.03	184	176499	48.13	nq	90
54) Dibenzofuran	10.18	168	1810380	47.25	nq	99
55) 4-Nitrophenol	10.10	139	265650	41.78	nq	# 83
56) 2,4-Dinitrotoluene	10.15	165	449491	48.31	nq	92
57) Fluorene	10.52	166	1521291	48.78	nq	99
58) 2,3,4,6-Tetrachlorophenol	10.29	232	346154	47.54	nq	94
59) Diethylphthalate	10.39	149	1532114	47.86	nq	# 94
60) 4-Chlorophenyl-phenylether	10.51	204	723157	50.51	nq	97
61) 4-Nitroaniline	10.55	138	431585	49.84	nq	90
62) Azobenzene	10.67	77	1849127	52.81	nq	98
64) 4,6-Dinitro-2-methylphenol	10.57	198	251583	49.54	nq	87
65) n-Nitrosodiphenylamine	10.63	169	1252025	50.34	nq	100
66) 4-Bromophenyl-phenylether	11.00	248	427135	51.05	nq	94
67) Hexachlorobenzene	11.07	284	462196	50.19	nq	# 66
68) Atrazine	11.16	200	369270	48.49	nq	93
69) Pentachlorophenol	11.26	266	225493	41.58	nq	98
70) Phenanthrene	11.48	178	2012309	48.46	nq	98
71) Anthracene	11.54	178	2120092	49.42	nq	99
72) Carbazole	11.69	167	1900711	47.79	nq	98
73) Di-n-butylphthalate	12.02	149	2259797	47.15	nq	99
74) Fluoranthene	12.67	202	2057996	48.57	nq	98
76) Benzidine	12.80	184	864333	43.73	nq	98
77) Pyrene	12.90	202	2057903	51.33	nq	99
79) Butylbenzylphthalate	13.52	149	937348	50.12	nq	98
80) Benzo(a)anthracene	14.09	228	1742277	45.83	nq	99
81) 3,3'-Dichlorobenzidine	14.05	252	636576	47.65	nq	99
82) Chrysene	14.12	228	1562666	44.66	nq	99
83) Bis(2-ethylhexyl)phthalate	14.08	149	1213549	44.95	nq	100
84) Di-n-octyl phthalate	14.69	149	1639871	40.32	nq	99
85) Indeno(1,2,3-cd)pyrene	17.01	276	1066652	38.04	nq	100
87) Benzo(b)fluoranthene	15.14	252	1279772	51.83	nq	# 95
88) Benzo(k)fluoranthene	15.16	252	1164801m	52.00	nq	
89) Benzo(a)pyrene	15.49	252	1104445	51.12	nq	100
90) Dibenzo(a,h)anthracene	17.04	278	914053	50.47	nq	96
91) Benzo(g,h,i)perylene	17.45	276	890800	52.76	nq	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\HPCHEM1\BNA F\DATA\BF101316\  
Data File : BF090552.D  
Acq On    : 13 Oct 2016   20:13  
Operator  : UM/SJ  
Sample    : SSTDICC050  
Misc      :  
ALS Vial  : 7      Sample Multiplier: 1
```


Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED

sohil
10/14/2016 6:38:47 PM

Quant Time: Oct 14 01:30:55 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF101316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 14 01:18:46 2016
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.92 min Scan# 401
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tot Ion	Ratio	Lower	Upper
152	100		
150	159.5	127.2	190.8
115	68.3	58.6	88.0

Manual Integrations
APPROVED




sohil
10/14/2016 6:38:47 PM




#2

1,4-Dioxane
Concen: 48.84 ng
RT: 2.57 min Scan# 20
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Lower	Upper
88	100		
58	80.4	67.2	100.8
43	26.6	24.1	36.1





#3

Pvridine

Concen: 41.74 ng

RT: 3.31 min Scan# 85

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Instrument :

BNA_F





ClientSampleId :

SSTDIC050

Tot Ion:	79	Resp:	733837
Ion Ratio	Lower	Upper	
79	100		
52	71.2	61.8	92.8
51	33.0	29.2	43.8

Manual Integrations
APPROVED

sohil
10/14/2016 6:38:47 PM



#4

n-Nitrosodimethylamine

Concen: 41.75 ng




RT: 3.27 min Scan# 82


Delta R.T. 0.01 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Tgt Ion:	42	Resp:	241491
Ion Ratio	Lower	Upper	
42	100		
74	169.5	113.3	169.9
44	8.0	6.0	9.0





#5

2-Fluorophenol

Concen: 91.91 ng

RT: 5.53 min Scan# 279

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Instrument :

BNA_F





ClientSampleId :

SSTDIC050

Tot Ion:	112	Resp:	1199917
Ion	Ratio	Lower	Upper
112	100		
64	64.3	52.6	78.8
63	31.6	26.0	39.0

Manual Integrations
APPROVED

sohil
10/14/2016 6:38:47 PM



#6

Aniline

Concen: 47.41 ng




RT: 6.59 min Scan# 372


Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Tgt Ion:	93	Resp:	1010732
Ion	Ratio	Lower	Upper
93	100		
66	51.4	42.2	63.2
65	28.2	22.9	34.3





#7

Phenol-d6

Concen: 100.61 ng

RT: 6.57 min Scan# 370

Delta R.T. 0.01 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Instrument :

BNA_F

ClientSampleId :

SSTDIC050

Tot Ion: 99 Resp: 1695817

Ion Ratio Lower Upper

99 100

42 14.3 14.7 22.1#





71 30.7 27.6 41.4

Manual Integrations

APPROVED

sohil

10/14/2016 6:38:47 PM



#8

2-Chlorophenol

Concen: 51.15 ng

RT: 6.70 min Scan# 382

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13




Tgt Ion: 128 Resp: 729532


Ion Ratio Lower Upper

128 100

130 31.7 12.6 52.6

64 60.7 42.6 82.6





#9

Benzaldehyde

Concen: 43.94 ng

RT: 6.47 min Scan# 362

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Instrument :

BNA_F

ClientSampleId :

SSTDIC050

Tot Ion: 77 Resp: 470758

Ion Ratio Lower Upper

77 100

105 97.2 74.3 114.3


106 91.9 67.2 107.2

Manual Integrations

APPROVED

sohil




10/14/2016 6:38:47 PM



Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 105.00 (104.70 to 105.70): BF0

Ion 106.00 (105.70 to 106.70): BF0



#10

Phenol

Concen: 53.22 ng

RT: 6.58 min Scan# 371

Delta R.T. 0.01 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13


Tgt Ion: 94 Resp: 1006198

Ion Ratio Lower Upper

94 100

65 34.4 12.5 52.5



66 50.6 25.5 65.5




Abundance Ion 94.00 (93.70 to 94.70): BF0

Ion 65.00 (64.70 to 65.70): BF0

Ion 66.00 (65.70 to 66.70): BF0









#11
bis(2-Chloroethyl)ether
Concen: 55.34 ng
RT: 6.66 min Scan# 378
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tot Ion:	93	Resp:	802326
Ion	Ratio	Lower	Upper
93	100		
63	80.9	61.6	101.6
95	31.9	11.8	51.8




Manual Integrations
APPROVED


sohil
10/14/2016 6:38:47 PM



#12
1,3-Dichlorobenzene
Concen: 48.78 ng
RT: 6.86 min Scan# 396
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tgt Ion:	146	Resp:	736759
Ion	Ratio	Lower	Upper
146	100		
148	64.9	51.4	77.2
75	31.4	27.0	40.6









#13
 1,4-Dichlorobenzene
 Concen: 48.56 ng
 RT: 6.94 min Scan# 403
 Delta R.T. 0.00 min
 Lab File: BF090552.D
 Acq: 13 Oct 2016 20:13

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Tot Ion	Ratio	Lower	Upper
146	100		
148	64.5	51.9	77.9
111	40.2	33.7	50.5




Manual Integrations
 APPROVED




sohil
 10/14/2016 6:38:47 PM



#14
 1,2-Dichlorobenzene
 Concen: 50.40 ng
 RT: 7.09 min Scan# 416
 Delta R.T. 0.00 min
 Lab File: BF090552.D
 Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Lower	Upper
146	100		
148	64.0	51.2	76.8
111	46.1	38.2	57.2









#15
Benzyl Alcohol
Concen: 41.84 ng
RT: 7.07 min Scan# 414
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tot Ion	Ratio	Lower	Upper
79	100		
108	80.0	60.2	90.4
77	63.9	52.0	78.0

Instrument :
BNA_F
ClientSampled :
SSTDIC050


Manual Integrations
APPROVED


sohil
10/14/2016 6:38:47 PM



#16
2,2'-oxybis(1-Chloropropane)
Concen: 42.89 ng
RT: 7.19 min Scan# 425
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Lower	Upper
45	100		
77	25.2	0.9	40.9
79	21.1	0.0	37.3









#17
2-Methylphenol
Concen: 48.80 ng
RT: 7.18 min Scan# 424
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tot Ion	Ratio	Lower	Upper
107	100		
108	115.0	92.6	138.8
77	49.8	42.2	63.2
79	48.1	40.7	61.1




Manual Integrations
APPROVED







sohil
10/14/2016 6:38:47 PM



#18
Hexachloroethane
Concen: 49.19 ng
RT: 7.43 min Scan# 446
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Lower	Upper
117	100		
119	94.3	75.6	113.4
201	92.5	65.4	98.0






#19
n-Nitroso-di-n-propylamine
Concen: 47.95 ng
RT: 7.34 min Scan# 438
Delta R.T. 0.01 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tot Ion	Ratio	Resp Lower	Upper
70	100		
42	43.3	43.4	65.0#
101	9.6	6.4	9.6#
130	22.1	12.1	18.1#


Manual Integrations
APPROVED


sohil
10/14/2016 6:38:47 PM



#20
3+4-Methylphenols
Concen: 44.89 ng
RT: 7.34 min Scan# 438
Delta R.T. 0.01 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Resp Lower	Upper
107	100		
108	94.6	67.7	107.7
77	107.4	123.1	163.1#
79	28.1	9.6	49.6





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 8.21 min Scan# 514

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Instrument :

BNA_F


ClientSampleId :

SSTDICC050




Tot Ion	136	Resp	873169
Ion Ratio	Lower	Upper	
136	100		
137	11.3	8.8	13.2
54	6.9	6.8	10.2
68	6.6	6.0	9.0

Manual Integrations
APPROVED

sohil
10/14/2016 6:38:47 PM



Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF0
Ion 68.00 (67.70 to 68.70): BF0



#22

Acetophenone

Concen: 47.41 ng

RT: 7.33 min Scan# 437




Delta R.T. 0.00 min


Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Tgt Ion	105	Resp	914908
Ion Ratio	Lower	Upper	
105	100		
71	7.5	8.5	12.7#
51	33.7	31.0	46.4
120	20.6	16.2	24.4

Abundance Ion 105.00 (104.70 to 105.70): E
Ion 71.00 (70.70 to 71.70): BF0
Ion 51.00 (50.70 to 51.70): BF0
Ion 120.00 (119.70 to 120.70): E









#23
Nitrobenzene-d5
Concen: 88.26 ng
RT: 7.49 min Scan# 451
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tot Ion	Ratio	Lower	Upper
82	100		
128	45.1	34.3	51.5
54	44.1	37.0	55.6




Manual Integrations
APPROVED


sohil
10/14/2016 6:38:47 PM



#24
Nitrobenzene
Concen: 43.28 ng
RT: 7.51 min Scan# 453
Delta R.T. 0.01 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Lower	Upper
77	100		
123	53.0	29.8	44.6#
65	12.0	10.9	16.3





#25

Isophorone

Concen: 45.54 ng

RT: 7.74 min Scan# 473

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Instrument :

BNA_F

ClientSampleId :

SSTDIC050

Tot Ion: 82 Resp: 1336532

Ion Ratio Lower Upper


82 100

95 6.4 5.0 7.6




138 15.1 11.5 17.3

Manual Integrations
APPROVED

sohil
10/14/2016 6:38:47 PM



Abundance Ion 82.00 (81.70 to 82.70): BF0
Ion 95.00 (94.70 to 95.70): BF0
Ion 138.00 (137.70 to 138.70): BF0



#26

2-Nitrophenol

Concen: 49.36 ng

RT: 7.82 min Scan# 480

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13


Tgt Ion: 139 Resp: 383491

Ion Ratio Lower Upper



139 100


109 24.7 21.8 32.6

65 51.5 47.7 71.5



Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BF0





#27

2,4-Dimethylphenol

Concen: 45.08 ng

RT: 7.87 min Scan# 484

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Instrument :

BNA_F


ClientSampleId :

SSTDICC050




Tot Ion:	122	Resp:	595259
Ion Ratio	Lower	Upper	
122	100		
107	109.4	89.8	134.6
121	58.5	46.1	69.1

Manual Integrations
APPROVED

sohil
10/14/2016 6:38:47 PM



Abundance Ion 122.00 (121.70 to 122.70): E
Ion 107.00 (106.70 to 107.70): E
Ion 121.00 (120.70 to 121.70): E



#28

bis(2-Chloroethoxy)methane

Concen: 45.44 ng


RT: 7.96 min Scan# 492

Delta R.T. 0.00 min



Lab File: BF090552.D


Acq: 13 Oct 2016 20:13

Tgt Ion:	93	Resp:	832490
Ion Ratio	Lower	Upper	
93	100		
95	32.6	26.2	39.2
123	11.3	8.6	13.0



Abundance Ion 93.00 (92.70 to 93.70): BF0
Ion 95.00 (94.70 to 95.70): BF0
Ion 123.00 (122.70 to 123.70): E









#29
2,4-Dichlorophenol
Concen: 48.12 ng
RT: 8.07 min Scan# 502
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tot Ion	Ratio	Resp	Lower	Upper
162	100	530689		
164	65.4		44.0	84.0
98	35.8		18.0	58.0




Manual Integrations
APPROVED


sohil
10/14/2016 6:38:47 PM



#30
1,2,4-Trichlorobenzene
Concen: 49.04 ng
RT: 8.15 min Scan# 509
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Resp	Lower	Upper
180	100	627998		
182	95.9		77.2	115.8
145	28.4		24.4	36.6









#31
Naphthalene
Concen: 48.31 ng
RT: 8.23 min Scan# 516
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tot Ion	Ratio	Lower	Upper
128	100		
129	11.7	9.0	13.6
127	13.5	10.8	16.2




Manual Integrations
APPROVED


sohil
10/14/2016 6:38:47 PM



#32
Benzoic acid
Concen: 38.91 ng
RT: 8.01 min Scan# 496
Delta R.T. 0.01 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Lower	Upper
122	100		
105	117.6	101.1	141.1
77	93.7	84.7	124.7






#33
4-Chloroaniline
Concen: 46.11 ng
RT: 8.28 min Scan# 520
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Instrument :
BNA_F
ClientSampleId :
SSTDICC050




Tot Ion	Ion	Ratio	Resp	Lower	Upper
127	100				
129	33.4	26.2	39.2		
65	33.9	30.1	45.1		
92	20.5	17.2	25.8		

Manual Integrations
APPROVED

sohil
10/14/2016 6:38:47 PM






Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BF0
Ion 92.00 (91.70 to 92.70): BF0




#34
Hexachlorobutadiene
Concen: 50.14 ng
RT: 8.35 min Scan# 526
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
225	100				
223	62.8	50.6	76.0		
227	63.4	51.4	77.2		

Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E






#35
Caprolactam
Concen: 41.47 ng
RT: 8.66 min Scan# 553
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tot Ion	Ratio	Lower	Upper
113	100		
55	160.5	170.8	210.8#
56	124.2	122.4	162.4

Manual Integrations
APPROVED




sohil
10/14/2016 6:38:47 PM



Abundance Ion 113.00 (112.70 to 113.70): E


Ion 55.00 (54.70 to 55.70): BF0

Ion 56.00 (55.70 to 56.70): BF0



#36
4-Chloro-3-methylphenol
Concen: 46.06 ng
RT: 8.77 min Scan# 563
Delta R.T. 0.01 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13



Tgt Ion	Ratio	Lower	Upper
107	100		
144	29.0	16.9	25.3#
142	87.2	52.5	78.7#




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E









#37
2-Methylnaphthalene
Concen: 42.84 ng
RT: 8.92 min Scan# 576
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tot Ion	Ratio	Lower	Upper
142	100		
141	89.2	71.3	106.9
115	34.3	28.8	43.2




Manual Integrations
APPROVED


sohil
10/14/2016 6:38:47 PM



#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.97 min Scan# 668
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Lower	Upper
164	100		
162	102.7	82.2	123.4
160	45.7	36.6	55.0





#39

1,2,4,5-Tetrachlorobenzene

Concen: 50.38 ng

RT: 9.09 min Scan# 591

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Instrument :

BNA_F





ClientSampled :

SSTDIC050

Tot Ion:	216	Resp:	651806
Ion Ratio		Lower	Upper
216	100		
214	79.2	63.5	95.3
179	22.1	18.9	28.3
108	18.2	17.8	26.6

Manual Integrations
APPROVED

sohil
10/14/2016 6:38:47 PM



#40

Hexachlorocyclopentadiene

Concen: 52.72 ng




RT: 9.08 min Scan# 590


Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Tgt Ion:	237	Resp:	335423
Ion Ratio		Lower	Upper
237	100		
235	61.1	43.8	83.8
272	13.7	0.0	33.7









#41
 2,4,6-Tribromophenol
 Concen: 97.65 ng
 RT: 10.76 min Scan# 737
 Delta R.T. 0.00 min
 Lab File: BF090552.D
 Acq: 13 Oct 2016 20:13

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Tot Ion	Ratio	Lower	Upper
330	100		
332	96.6	76.2	114.2
141	37.5	31.4	47.2




Manual Integrations
 APPROVED


sohil
 10/14/2016 6:38:47 PM



#42
 2,4,6-Trichlorophenol
 Concen: 43.64 ng
 RT: 9.21 min Scan# 601
 Delta R.T. 0.00 min
 Lab File: BF090552.D
 Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Lower	Upper
196	100		
198	96.2	77.9	116.9
200	31.7	25.1	37.7









#43
 2,4,5-Trichlorophenol
 Concen: 48.49 ng
 RT: 9.24 min Scan# 604
 Delta R.T. 0.00 min
 Lab File: BF090552.D
 Acq: 13 Oct 2016 20:13

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Tot Ion	Ion	Ratio	Resp	Lower	Upper
196	100		418569		
198	95.6	74.0	111.0		
97	64.7	54.7	82.1		
132	37.4	30.5	45.7		




Manual Integrations
 APPROVED


sohil
 10/14/2016 6:38:47 PM



#44
 2-Fluorobiphenyl
 Concen: 90.28 ng
 RT: 9.29 min Scan# 608
 Delta R.T. 0.00 min
 Lab File: BF090552.D
 Acq: 13 Oct 2016 20:13

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
172	100		2509671		
171	37.4	30.1	45.1		
170	25.4	20.6	30.8		





#45

1,1'-Biphenyl

Concen: 46.28 ng

RT: 9.39 min Scan# 617

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Instrument :

BNA_F

ClientSampleId :

SSTDIC050

Tot Ion:154 Resp: 1645385

Ion Ratio Lower Upper

154 100

153 42.0 21.5 61.5


76 13.0 0.0 36.0

Manual Integrations

APPROVED

sohil




10/14/2016 6:38:47 PM



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BF0



#46

2-Chloronaphthalene

Concen: 45.42 ng

RT: 9.41 min Scan# 619

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13


Tgt Ion:162 Resp: 1203135

Ion Ratio Lower Upper

162 100

127 43.4 35.9 53.9



164 34.1 26.7 40.1




Abundance Ion 162.00 (161.70 to 162.70): E

Ion 127.00 (126.70 to 127.70): E

Ion 164.00 (163.70 to 164.70): E









#47
2-Nitroaniline
Concen: 37.95 ng
RT: 9.51 min Scan# 628
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tot Ion	Ratio	Lower	Upper
65	100		
92	79.4	54.3	81.5
138	133.5	81.7	122.5




Manual Integrations
APPROVED


sohil
10/14/2016 6:38:47 PM



#48
Acenaphthylene
Concen: 48.22 ng
RT: 9.83 min Scan# 656
Delta R.T. 0.01 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.9	16.6	25.0
153	13.5	11.3	16.9









#49
Dimethylphthalate
Concen: 48.14 ng
RT: 9.70 min Scan# 644
Delta R.T. 0.01 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tot Ion	Ratio	Lower	Upper
163	100		
194	3.8	2.3	3.5#
164	10.1	8.6	13.0




Manual Integrations
APPROVED


sohil
10/14/2016 6:38:47 PM



#50
2,6-Dinitrotoluene
Concen: 49.82 ng
RT: 9.75 min Scan# 649
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Lower	Upper
165	100		
63	53.5	40.1	60.1
89	53.0	39.5	59.3





#51

Acenaphthene

Concen: 49.63 ng

RT: 10.01 min Scan# 671

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Instrument :

BNA_F

ClientSampleId :

SSTDIC050

Tot Ion:154 Resp: 1425614

Ion Ratio Lower Upper


154 100

153 113.0 91.2 136.8

152 54.5 44.0 66.0

Manual Integrations
APPROVED




sohil
10/14/2016 6:38:47 PM



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 48.98 ng

RT: 9.93 min Scan# 664

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13


Tgt Ion:138 Resp: 424157

Ion Ratio Lower Upper

138 100

108 10.7 8.6 12.8



92 127.3 97.8 146.8




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0






#53
2,4-Dinitrophenol
Concen: 48.13 ng
RT: 10.03 min Scan# 673
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Instrument :
BNA_F
ClientSampled :
SSTDIC050




Tot Ion	Ratio	Resp	Lower	Upper
184	100	176499		
63	70.3	60.2	90.2	
154	93.8	86.1	129.1	

Manual Integrations
APPROVED

sohil
10/14/2016 6:38:47 PM




Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 154.00 (153.70 to 154.70): E








#54
Dibenzofuran
Concen: 47.25 ng
RT: 10.18 min Scan# 686
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Resp	Lower	Upper
168	100	1810380		
139	38.5	30.5	45.7	
169	14.0	10.9	16.3	



Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E





#55

4-Nitrophenol

Concen: 41.78 ng

RT: 10.10 min Scan# 679

Delta R.T. 0.01 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Tot Ion	Ratio	Resp	Lower	Upper
139	100	265650		
109	59.2	45.9	85.9	
65	91.6	95.8	135.8	

Instrument :





BNA_F

ClientSampleId :

SSTDIC050

Manual Integrations
APPROVED

sohil
10/14/2016 6:38:47 PM



#56

2,4-Dinitrotoluene

Concen: 48.31 ng


RT: 10.15 min Scan# 684


Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Resp	Lower	Upper
165	100	449491		
63	50.1	35.4	53.0	
89	80.4	59.0	88.4	
182	2.2	1.8	2.6	





#57
Fluorene
 Concen: 48.78 ng
 RT: 10.52 min Scan# 716
 Delta R.T. 0.00 min
 Lab File: BF090552.D
 Acq: 13 Oct 2016 20:13





Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Tot Ion:166 Resp: 1521291

Ion	Ratio	Lower	Upper
166	100		
165	98.8	79.9	119.9
167	13.6	10.9	16.3

Manual Integrations
APPROVED




sohil
 10/14/2016 6:38:47 PM




#58
2,3,4,6-Tetrachlorophenol
 Concen: 47.54 ng
 RT: 10.29 min Scan# 696
 Delta R.T. 0.00 min
 Lab File: BF090552.D
 Acq: 13 Oct 2016 20:13

Tgt Ion:232 Resp: 346154

Ion	Ratio	Lower	Upper
232	100		
131	58.1	42.6	64.0
130	2.7	1.8	2.8
166	37.5	27.5	41.3









#59
Diethylphthalate
Concen: 47.86 ng
RT: 10.39 min Scan# 705
Delta R.T. 0.01 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Instrument :
BNA_F
ClientSampled :
SSTDIC050

Tot Ion:149 Resp: 1532114
Ion Ratio Lower Upper
149 100
177 23.5 15.3 22.9#
150 12.5 10.3 15.5




Manual Integrations
APPROVED


sohil
10/14/2016 6:38:47 PM



#60
4-Chlorophenyl-phenylether
Concen: 50.51 ng
RT: 10.51 min Scan# 715
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tgt Ion:204 Resp: 723157
Ion Ratio Lower Upper
204 100
206 33.8 26.5 39.7
141 62.3 47.2 70.8









#61
4-Nitroaniline
Concen: 49.84 ng
RT: 10.55 min Scan# 719
Delta R.T. 0.01 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Instrument :
BNA_F
ClientSampled :
SSTDIC050

Tot Ion	Ratio	Lower	Upper
138	100		
92	48.2	35.7	75.7
108	59.8	46.9	86.9




Manual Integrations
APPROVED


sohil
10/14/2016 6:38:47 PM



#62
Azobenzene
Concen: 52.81 ng
RT: 10.67 min Scan# 729
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Lower	Upper
77	100		
182	21.4	2.2	42.2
105	23.9	3.5	43.5
51	24.9	3.7	43.7





#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 11.46 min Scan# 798

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Instrument :

BNA_F

ClientSampleId :

SSTDIC050

Tot Ion:188 Resp: 762377

Ion Ratio Lower Upper


188 100

94 11.6 7.7 11.5#

80 12.9 8.6 12.8#

Manual Integrations
APPROVED




sohil
10/14/2016 6:38:47 PM



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF0

Ion 80.00 (79.70 to 80.70): BF0



#64

4,6-Dinitro-2-methylphenol

Concen: 49.54 ng

RT: 10.57 min Scan# 720

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13


Tgt Ion:198 Resp: 251583

Ion Ratio Lower Upper

198 100

51 47.7 19.7 59.7



105 50.9 22.6 62.6




Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BF0

Ion 105.00 (104.70 to 105.70): E





#65

n-Nitrosodiphenylamine

Concen: 50.34 ng

RT: 10.63 min Scan# 726

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Instrument :

BNA_F

ClientSampleId :

SSTDICC050

Tot Ion:169 Resp: 1252025

Ion Ratio Lower Upper

169 100

168 68.6 54.6 81.8


167 37.6 29.8 44.8

Manual Integrations

APPROVED

sohil




10/14/2016 6:38:47 PM



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 51.05 ng

RT: 11.00 min Scan# 758

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13


Tgt Ion:248 Resp: 427135

Ion Ratio Lower Upper

248 100

250 96.9 78.4 117.6



141 71.3 48.4 72.6




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E





#67

Hexachlorobenzene

Concen: 50.19 ng

RT: 11.07 min Scan# 764

Delta R.T. 0.01 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Instrument :

BNA_F





ClientSampleId :

SSTDIC050

Tot Ion	Ratio	Resp	Lower	Upper
284	100	462196		
142	20.6	44.5	66.7#	
249	28.8	29.6	44.4#	

Manual Integrations
APPROVED

sohil
10/14/2016 6:38:47 PM



#68

Atrazine

Concen: 48.49 ng




RT: 11.16 min Scan# 772


Delta R.T. 0.01 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Resp	Lower	Upper
200	100	369270		
173	27.4	12.0	52.0	
215	48.8	25.2	65.2	









#69
 Pentachlorophenol
 Concen: 41.58 ng
 RT: 11.26 min Scan# 781
 Delta R.T. 0.00 min
 Lab File: BF090552.D
 Acq: 13 Oct 2016 20:13

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Tot Ion	Ratio	Resp	Lower	Upper
266	100	225493		
268	64.6	51.5	77.3	
264	62.5	47.9	71.9	




Manual Integrations
 APPROVED


sohil
 10/14/2016 6:38:47 PM



#70
 Phenanthrene
 Concen: 48.46 ng
 RT: 11.48 min Scan# 800
 Delta R.T. 0.00 min
 Lab File: BF090552.D
 Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Resp	Lower	Upper
178	100	2012309		
176	21.0	16.1	24.1	
179	16.5	12.6	19.0	









#71
 Anthracene
 Concen: 49.42 ng
 RT: 11.54 min Scan# 805
 Delta R.T. 0.01 min
 Lab File: BF090552.D
 Acq: 13 Oct 2016 20:13

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Tot Ion:178 Resp: 2120092
 Ion Ratio Lower Upper
 178 100
 176 19.9 15.5 23.3
 179 15.9 13.0 19.4




Manual Integrations
 APPROVED


sohil
 10/14/2016 6:38:47 PM



#72
 Carbazole
 Concen: 47.79 ng
 RT: 11.69 min Scan# 818
 Delta R.T. 0.00 min
 Lab File: BF090552.D
 Acq: 13 Oct 2016 20:13

Tgt Ion:167 Resp: 1900711
 Ion Ratio Lower Upper
 167 100
 166 22.8 17.8 26.8
 139 15.0 11.3 16.9









#73
Di-n-butylphthalate
Concen: 47.15 ng
RT: 12.02 min Scan# 847
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tot Ion	Ion	Ratio	Lower	Upper
149	100			
150	10.0	7.8	11.6	
104	4.5	3.3	4.9	




Manual Integrations
APPROVED


sohil
10/14/2016 6:38:47 PM



#74
Fluoranthene
Concen: 48.57 ng
RT: 12.67 min Scan# 904
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tgt Ion	Ion	Ratio	Lower	Upper
202	100			
101	12.4	0.0	31.1	
203	18.5	0.0	38.0	





#75

Chrysene-d12

Concen: 20.00 ng

RT: 14.10 min Scan# 1029

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Instrument :

BNA_F

ClientSampleId :

SSTDIC050

Tot Ion:240 Resp: 633298

Ion Ratio Lower Upper


240 100

120 13.2 9.8 14.8

236 26.8 21.7 32.5

Manual Integrations
APPROVED

sohil
10/14/2016 6:38:47 PM






Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E

14.10



#76

Benzidine

Concen: 43.73 ng

RT: 12.80 min Scan# 915

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13


Tgt Ion:184 Resp: 864333

Ion Ratio Lower Upper

184 100

185 17.8 13.4 20.0

183 11.9 9.4 14.0






Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E

12.80









#77
 Pvrene
 Concen: 51.33 ng
 RT: 12.90 min Scan# 924
 Delta R.T. 0.00 min
 Lab File: BF090552.D
 Acq: 13 Oct 2016 20:13

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Tot Ion	Ratio	Lower	Upper
202	100		
200	22.5	17.8	26.6
203	18.8	14.6	21.8




Manual Integrations
 APPROVED


sohil
 10/14/2016 6:38:47 PM



#78
 Terphenyl-d14
 Concen: 105.49 ng
 RT: 13.05 min Scan# 937
 Delta R.T. 0.00 min
 Lab File: BF090552.D
 Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.0	7.4
122	8.8	5.6	8.4









#79
Butylbenzylphthalate
Concen: 50.12 ng
RT: 13.52 min Scan# 978
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tot Ion	Ratio	Lower	Upper
149	100		
91	69.8	54.3	81.5
206	19.1	15.8	23.6




Manual Integrations
APPROVED


sohil
10/14/2016 6:38:47 PM



#80
Benzo(a)anthracene
Concen: 45.83 ng
RT: 14.09 min Scan# 1028
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.0	22.8	34.2
229	20.8	16.6	25.0









#81
 3,3'-Dichlorobenzidine
 Concen: 47.65 ng
 RT: 14.05 min Scan# 1025
 Delta R.T. 0.00 min
 Lab File: BF090552.D
 Acq: 13 Oct 2016 20:13

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Tot Ion	Ratio	Lower	Upper
252	100		
254	64.7	52.1	78.1
126	12.2	9.3	13.9




Manual Integrations
 APPROVED


sohil
 10/14/2016 6:38:47 PM



#82
 Chrysene
 Concen: 44.66 ng
 RT: 14.12 min Scan# 1031
 Delta R.T. 0.00 min
 Lab File: BF090552.D
 Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.3	24.8	37.2
229	20.5	16.2	24.4









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 44.95 ng
 RT: 14.08 min Scan# 1027
 Delta R.T. 0.00 min
 Lab File: BF090552.D
 Acq: 13 Oct 2016 20:13

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Tot Ion:149 Resp: 1213549
 Ion Ratio Lower Upper
 149 100
 167 27.7 22.1 33.1
 279 3.3 2.6 4.0




Manual Integrations
 APPROVED


sohil
 10/14/2016 6:38:47 PM



#84
 Di-n-octyl phthalate
 Concen: 40.32 ng
 RT: 14.69 min Scan# 1081
 Delta R.T. 0.01 min
 Lab File: BF090552.D
 Acq: 13 Oct 2016 20:13

Tgt Ion:149 Resp: 1639871
 Ion Ratio Lower Upper
 149 100
 167 1.5 1.2 1.8
 43 8.2 6.7 10.1






#85
Indeno(1,2,3-cd)pyrene
Concen: 38.04 ng
RT: 17.01 min Scan# 1284
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Instrument :
BNA_F
ClientSampleId :
SSTDICC050




Tot Ion	Ratio	Lower	Upper
276	100		
138	32.1	25.8	38.6
277	24.8	20.1	30.1

Manual Integrations
APPROVED

sohil
10/14/2016 6:38:47 PM






Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E




#86
Perylene-d12
Concen: 20.00 ng
RT: 15.56 min Scan# 1157
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.6	19.3	28.9
265	21.5	16.7	25.1

Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E





#87

Benzo(b)fluoranthene

Concen: 51.83 ng

RT: 15.14 min Scan# 1120

Delta R.T. 0.01 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Instrument :

BNA_F

ClientSampleId :

SSTDIC050

Tot Ion:252 Resp: 1279772

Ion Ratio Lower Upper


252 100

253 22.4 17.8 26.8

125 10.0 12.0 18.0#

Manual Integrations
APPROVED




sohil
10/14/2016 6:38:47 PM



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 52.00 ng m

RT: 15.16 min Scan# 1122

Delta R.T. 0.00 min

Lab File: BF090552.D

Acq: 13 Oct 2016 20:13

Tgt Ion:252 Resp: 1164801

Ion Ratio Lower Upper

252 100




253 22.8 18.0 27.0


125 13.6 10.0 15.0

Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#89
Benzo(a)pyrene
Concen: 51.12 ng
RT: 15.49 min Scan# 1151
Delta R.T. 0.00 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Instrument :
BNA_F
ClientSampleId :
SSTDIC050

Tot Ion:252 Resp: 1104445
Ion Ratio Lower Upper
252 100
253 22.4 17.9 26.9
125 15.8 12.3 18.5




Manual Integrations
APPROVED


sohil
10/14/2016 6:38:47 PM



#90
Dibenzo(a,h)anthracene
Concen: 50.47 ng
RT: 17.04 min Scan# 1286
Delta R.T. 0.01 min
Lab File: BF090552.D
Acq: 13 Oct 2016 20:13

Tgt Ion:278 Resp: 914053
Ion Ratio Lower Upper
278 100
139 17.3 16.6 25.0
279 24.3 19.1 28.7





#91
 Benzo(a,h,i)perylene
 Concen: 52.76 ng
 RT: 17.45 min Scan# 1322
 Delta R.T. 0.00 min
 Lab File: BF090552.D
 Acq: 13 Oct 2016 20:13

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Tot Ion:	276	Resp:	890800
Ion Ratio	Lower	Upper	
276	100		
277	23.5	18.8	28.2
138	28.8	21.8	32.8

Manual Integrations
 APPROVED

sohil
 10/14/2016 6:38:47 PM

