




Data Path : Z:\HPCHEM1\BNA_F\DATA\BF101416\
Data File : BF090556.D
Acq On : 14 Oct 2016 8:19
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 17 02:07:35 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF101316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 14 12:44:45 2016
Response via : Initial Calibration



TIC: BF090556.D

(69) Pentachlorophenol (C)




11.252min (-0.011) 0.00ng

response 95684

Ion	Exp%	Act%
265.70	100	100
268.00	64.40	65.23
264.00	59.90	64.62
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF101416\
 Data File : BF090556.D
 Acq On : 14 Oct 2016 8:19
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 17 02:07:35 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF101316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 14 12:44:45 2016
 Response via : Initial Calibration



TIC: BF090556.D

(76) Benzidine

12.795min (-0.000) 0.00ng

response 590785

Ion	Exp%	Act%
184.00	100	100
185.00	16.70	15.20
183.00	11.70	11.59
0.00	0.00	0.00