


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101421\
 Data File : BF125838.D
 Acq On : 14 Oct 2021 10:06
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF100721.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Oct 15 15:00:46 2021




AutoFind: Scans 1618, 1619, 1620; Background Corrected with Scan 1613

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.6	57069	PASS
68	69	0.00	2	1.8	1046	PASS
69	198	0.00	100	31.5	58792	PASS
70	69	0.00	2	0.4	253	PASS
127	198	10	80	47.8	89240	PASS
197	198	0.00	2	0.4	808	PASS
198	198	100	100	100.0	186752	PASS
199	198	5	9	6.6	12365	PASS
275	198	10	60	26.1	48683	PASS
365	198	1	100	3.5	6453	PASS
441	198	0.01	100	13.5	25285	PASS
442	442	50	100	100.0	160909	PASS
443	442	15	24	18.9	30371	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101421\
Data File : BF125838.D
Acq On : 14 Oct 2021 10:06
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 15 06:36:50 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF100721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 14 12:29:41 2021
Response via : Initial Calibration



(70) Pentachlorophenol (C)

11.157min (-0.012) 31449.29 ng




response 198757

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.10	65.38
264.00	69.90	63.05
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101421\
 Data File : BF125838.D
 Acq On : 14 Oct 2021 10:06
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Oct 15 06:36:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF100721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 14 12:29:41 2021
 Response via : Initial Calibration



TIC: BF125838.D\data.ms

(77) Benzidine

12.686min (-0.006) 0.00 ng

response 1133041

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.50	13.59
183.00	10.10	11.12
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/14/2021	BNA_F	<u>BF125838.D</u>
Compound Name	Response	Retention Time
DDT	455187	13.51
DDD	12188	13.21
DDE	371	12.874
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
12559	467746	2.69

Instrument :
BNA_F
ClientSampleId :
DFTPP