

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF101517\
 Data File : BF099529.D
 Acq On : 15 Oct 2017 22:33
 Operator : SJ/JU
 Sample : I5857-01
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FK-PS-01-018

Quant Time: Oct 16 03:54:34 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 11 16:25:21 2017
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	82470	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	309284	20.00	ng	0.00
38) Acenaphthene-d10	9.88	164	113492	20.00	ng	-0.01
63) Phenanthrene-d10	11.37	188	161878	20.00	ng	0.00
75) Chrysene-d12	14.01	240	169390	20.00	ng	0.00
86) Perylene-d12	15.49	264	151849	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.47	112	477716	92.21	ng	0.00
7) Phenol-d6	6.48	99	548367	85.80	ng	0.00
23) Nitrobenzene-d5	7.41	82	312604	64.82	ng	-0.01
41) 2,4,6-Tribromophenol	10.67	330	74177	79.87	ng	-0.01
44) 2-Fluorobiphenyl	9.20	172	549857	80.88	ng	-0.01
78) Terphenyl-d14	12.95	244	373385	50.13	ng	0.00
Target Compounds						
10) Phenol	6.49	94	31902	4.463	ng	99
49) Dimethylphthalate	9.59	163	87196	10.180	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF101517\
Data File : BF099529.D
Acq On : 15 Oct 2017 22:33
Operator : SJ/JU
Sample : I5857-01
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FK-PS-01-018

Quant Time: Oct 16 03:54:34 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 11 16:25:21 2017
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.85 min Scan# 809
 Delta R.T. -0.01 min
 Lab File: BF099529.D
 Acq: 15 Oct 2017 22:33

Instrument :
 BNA_F
 ClientSampleId :
 FK-PS-01-018

Tot Ion:152 Resp: 82470
 Ion Ratio Lower Upper
 152 100
 150 156.1 124.6 186.8
 115 58.1 50.1 75.1




Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 150.00 (149.70 to 150.70): E
 Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
 Concen: 92.212 ng
 RT: 5.47 min Scan# 575
 Delta R.T. 0.00 min
 Lab File: BF099529.D
 Acq: 15 Oct 2017 22:33

Tgt Ion:112 Resp: 477716
 Ion Ratio Lower Upper
 112 100
 64 51.4 47.9 71.9
 63 25.0 23.7 35.5



Abundance Ion 112.00 (111.70 to 112.70): E
 Ion 64.00 (63.70 to 64.70): BF0
 Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 85.805 ng

RT: 6.48 min Scan# 747

Delta R.T. -0.01 min

Lab File: BF099529.D

Acq: 15 Oct 2017 22:33

Instrument :

BNA_F

ClientSampleId :

FK-PS-01-018


Tot Ion: 99 Resp: 548367

Ion Ratio Lower Upper

99 100




42 14.6 14.5 21.7

71 29.7 25.4 38.0



Abundance Ion 99.00 (98.70 to 99.70): BF0
Ion 42.00 (41.70 to 42.70): BF0
Ion 71.00 (70.70 to 71.70): BF0

6.48



#10

Phenol

Concen: 4.463 ng

RT: 6.49 min Scan# 749

Delta R.T. -0.01 min

Lab File: BF099529.D

Acq: 15 Oct 2017 22:33


Tgt Ion: 94 Resp: 31902

Ion Ratio Lower Upper

94 100



65 27.2 7.9 47.9


66 35.5 16.2 56.2



Abundance Ion 94.00 (93.70 to 94.70): BF0
Ion 65.00 (64.70 to 65.70): BF0
Ion 66.00 (65.70 to 66.70): BF0

6.49









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.13 min Scan# 1027
Delta R.T. -0.01 min
Lab File: BF099529.D
Acq: 15 Oct 2017 22:33

Instrument :
BNA_F
ClientSampleId :
FK-PS-01-018

Tot Ion:	136	Resp:	309284
Ion	Ratio	Lower	Upper
136	100		
137	10.4	8.9	13.3
54	6.7	6.6	9.8
68	5.4	5.0	7.4




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF0
Ion 68.00 (67.70 to 68.70): BF0






#23
Nitrobenzene-d5
Concen: 64.819 ng
RT: 7.41 min Scan# 905
Delta R.T. -0.01 min
Lab File: BF099529.D
Acq: 15 Oct 2017 22:33

Tgt Ion:	82	Resp:	312604
Ion	Ratio	Lower	Upper
82	100		
128	49.6	36.1	54.1
54	46.4	41.4	62.2



Abundance Ion 82.00 (81.70 to 82.70): BF0
Ion 128.00 (127.70 to 128.70): E
Ion 54.00 (53.70 to 54.70): BF0





#38

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.88 min Scan# 1325

Delta R.T. -0.01 min

Lab File: BF099529.D

Acq: 15 Oct 2017 22:33

Instrument :

BNA_F

ClientSampleId :

FK-PS-01-018


Tot Ion:164 Resp: 113492

Ion Ratio Lower Upper

164 100

162 104.3 86.1 129.1




160 45.6 38.3 57.5



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#41

2,4,6-Tribromophenol

Concen: 79.874 ng

RT: 10.67 min Scan# 1459

Delta R.T. -0.01 min

Lab File: BF099529.D

Acq: 15 Oct 2017 22:33


Tgt Ion:330 Resp: 74177

Ion Ratio Lower Upper

330 100

332 94.6 77.0 115.6



141 39.8 29.9 44.9




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#44
2-Fluorobiphenyl
Concen: 80.882 ng
RT: 9.20 min Scan# 1209
Delta R.T. -0.01 min
Lab File: BF099529.D
Acq: 15 Oct 2017 22:33




Instrument :
BNA_F
ClientSampleId :
FK-PS-01-018


Tot Ion	Ratio	Lower	Upper
172	100		
171	35.9	29.7	44.5
170	23.9	19.6	29.4



#49
Dimethylphthalate
Concen: 10.180 ng
RT: 9.59 min Scan# 1276
Delta R.T. -0.02 min
Lab File: BF099529.D
Acq: 15 Oct 2017 22:33

Tgt Ion	Ratio	Lower	Upper
163	100		
194	3.7	2.7	4.1
164	10.2	8.2	12.2









#63
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.37 min Scan# 1578
Delta R.T. -0.01 min
Lab File: BF099529.D
Acq: 15 Oct 2017 22:33




Instrument :
BNA_F
ClientSampleId :
FK-PS-01-018


Tot Ion:188 Resp: 161878
Ion Ratio Lower Upper
188 100
94 10.6 8.5 12.7
80 11.6 9.2 13.8



#75
Chrysene-d12
Concen: 20.000 ng
RT: 14.01 min Scan# 2027
Delta R.T. -0.01 min
Lab File: BF099529.D
Acq: 15 Oct 2017 22:33

Tgt Ion:240 Resp: 169390
Ion Ratio Lower Upper
240 100
120 15.9 9.5 14.3#
236 25.1 20.7 31.1





#78

Terphenyl-d14

Concen: 50.130 ng

RT: 12.95 min Scan# 1847

Delta R.T. -0.01 min

Lab File: BF099529.D

Acq: 15 Oct 2017 22:33

Instrument :

BNA_F

ClientSampleId :

FK-PS-01-018


Tot Ion:244 Resp: 373385

Ion Ratio Lower Upper

244 100

212 7.0 5.4 8.0




122 13.2 11.0 16.4



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.49 min Scan# 2278

Delta R.T. 0.01 min

Lab File: BF099529.D

Acq: 15 Oct 2017 22:33


Tgt Ion:264 Resp: 151849

Ion Ratio Lower Upper

264 100

260 23.2 19.2 28.8

265 21.0 17.5 26.3



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

