

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101520\
 Data File : BF121984.D
 Acq On : 15 Oct 2020 10:50
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 15 13:04:46 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 15 12:55:05 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.596	0.576	3.4	87	0.00
3	Pyridine	1.567	1.567	0.0	88	0.00
4	n-Nitrosodimethylamine	0.757	0.734	3.0	86	0.00
5 S	2-Fluorophenol	1.355	1.347	0.6	90	0.00
6	Aniline	2.148	2.069	3.7	88	0.00
7 S	Phenol-d6	1.744	1.674	4.0	88	0.00
8	2-Chlorophenol	1.370	1.341	2.1	89	0.00
9	Benzaldehyde	1.006	0.958	4.8	89	0.00
10 C	Phenol	1.800	1.741	3.3	89	0.00
11	bis(2-Chloroethyl)ether	1.392	1.340	3.7	87	0.00
12	1,3-Dichlorobenzene	1.541	1.495	3.0	88	0.00
13 C	1,4-Dichlorobenzene	1.547	1.508	2.5	88	0.00
14	1,2-Dichlorobenzene	1.414	1.373	2.9	88	0.00
15	Benzyl Alcohol	1.128	1.093	3.1	87	0.00
16	2,2'-oxybis(1-Chloropropane	2.142	2.059	3.9	87	0.00
17	2-Methylphenol	1.171	1.121	4.3	88	0.00
18	Hexachloroethane	0.559	0.553	1.1	89	0.00
19 P	n-Nitroso-di-n-propylamine	0.993	0.949	4.4	89	0.00
20	3+4-Methylphenols	1.412	1.368	3.1	90	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	88	0.00
22	Acetophenone	0.515	0.501	2.7	90	0.00
23 S	Nitrobenzene-d5	0.390	0.385	1.3	88	0.00
24	Nitrobenzene	0.417	0.406	2.6	87	0.00
25	Isophorone	0.737	0.708	3.9	87	0.00
26 C	2-Nitrophenol	0.192	0.193	-0.5	87	0.00
27	2,4-Dimethylphenol	0.327	0.317	3.1	87	0.00
28	bis(2-Chloroethoxy)methane	0.461	0.447	3.0	87	0.00
29 C	2,4-Dichlorophenol	0.306	0.303	1.0	88	0.00
30	1,2,4-Trichlorobenzene	0.326	0.318	2.5	88	0.00
31	Naphthalene	1.071	1.048	2.1	89	0.00
32	Benzoic acid	0.165	0.157	4.8	82	0.00
33	4-Chloroaniline	0.456	0.450	1.3	88	0.00
34 C	Hexachlorobutadiene	0.193	0.191	1.0	87	0.00
35	Caprolactam	0.097	0.098	-1.0	90	0.00
36 C	4-Chloro-3-methylphenol	0.329	0.325	1.2	89	0.00
37	2-Methylnaphthalene	0.694	0.674	2.9	88	0.00
38	1-Methylnaphthalene	0.643	0.615	4.4	88	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
40	1,2,4,5-Tetrachlorobenzene	0.646	0.635	1.7	89	0.00
41 P	Hexachlorocyclopentadiene	0.127	0.100	21.3	68	0.00
42 S	2,4,6-Tribromophenol	0.247	0.248	-0.4	89	0.00
43 C	2,4,6-Trichlorophenol	0.398	0.394	1.0	88	0.00
44	2,4,5-Trichlorophenol	0.430	0.420	2.3	88	0.00
45 S	2-Fluorobiphenyl	1.359	1.278	6.0	88	0.00
46	1,1'-Biphenyl	1.626	1.570	3.4	88	0.00
47	2-Chloronaphthalene	1.265	1.230	2.8	88	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101520\
 Data File : BF121984.D
 Acq On : 15 Oct 2020 10:50
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 15 13:04:46 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 15 12:55:05 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.417	0.419	-0.5	90	0.00
49	Acenaphthylene	1.942	1.874	3.5	88	0.00
50	Dimethylphthalate	1.462	1.416	3.1	88	0.00
51	2,6-Dinitrotoluene	0.321	0.318	0.9	87	0.00
52 C	Acenaphthene	1.155	1.113	3.6	89	0.00
53	3-Nitroaniline	0.365	0.365	0.0	91	0.00
54 P	2,4-Dinitrophenol	0.129	0.125	3.1	83	0.00
55	Dibenzofuran	1.690	1.611	4.7	87	0.00
56 P	4-Nitrophenol	0.198	0.162	18.2	71	0.00
57	2,4-Dinitrotoluene	0.395	0.392	0.8	89	0.00
58	Fluorene	1.361	1.333	2.1	90	0.00
59	2,3,4,6-Tetrachlorophenol	0.333	0.338	-1.5	88	0.00
60	Diethylphthalate	1.410	1.381	2.1	89	0.00
61	4-Chlorophenyl-phenylether	0.653	0.635	2.8	90	0.00
62	4-Nitroaniline	0.365	0.363	0.5	89	0.00
63	Azobenzene	1.491	1.469	1.5	89	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
65	4,6-Dinitro-2-methylphenol	0.115	0.123	-7.0	88	0.00
66 c	n-Nitrosodiphenylamine	0.689	0.667	3.2	90	0.00
67	4-Bromophenyl-phenylether	0.231	0.226	2.2	88	0.00
68	Hexachlorobenzene	0.262	0.251	4.2	88	0.00
69	Atrazine	0.168	0.171	-1.8	91	0.00
70 C	Pentachlorophenol	0.084	0.075	10.7	81	0.00
71	Phenanthrene	1.097	1.066	2.8	90	0.00
72	Anthracene	1.137	1.110	2.4	91	0.00
73	Carbazole	1.011	1.010	0.1	92	0.00
74	Di-n-butylphthalate	1.163	1.167	-0.3	91	0.00
75 C	Fluoranthene	1.176	1.166	0.9	91	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
77	Benzydine	0.617	0.583	5.5	90	0.00
78	Pyrene	1.524	1.489	2.3	91	0.00
79 S	Terphenyl-d14	1.049	1.011	3.6	92	0.00
80	Butylbenzylphthalate	0.604	0.622	-3.0	93	0.00
81	Benzo(a)anthracene	1.311	1.294	1.3	93	0.00
82	3,3'-Dichlorobenzidine	0.486	0.493	-1.4	92	0.00
83	Chrysene	1.287	1.266	1.6	91	0.00
84	Bis(2-ethylhexyl)phthalate	0.821	0.854	-4.0	94	0.00
85 c	Di-n-octyl phthalate	1.327	1.368	-3.1	93	0.00
86 I	Perylene-d12	1.000	1.000	0.0	91	0.00
87	Indeno(1,2,3-cd)pyrene	1.299	1.288	0.8	92	0.00
88	Benzo(b)fluoranthene	1.352	1.424	-5.3	92	0.00
89	Benzo(k)fluoranthene	1.284	1.319	-2.7	91	0.00
90 C	Benzo(a)pyrene	1.168	1.200	-2.7	91	0.00
91	Dibenzo(a,h)anthracene	1.069	1.073	-0.4	92	0.00
92	Benzo(g,h,i)perylene	1.070	1.067	0.3	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101520\
Data File : BF121984.D
Acq On : 15 Oct 2020 10:50
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 15 13:04:46 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 15 12:55:05 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0