



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101520\
 Data File : BF121983.D
 Acq On : 15 Oct 2020 10:19
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Oct 15 12:55:05 2020



AutoFind: Scans 1602, 1603, 1604; Background Corrected with Scan 1597

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.5	29472	PASS
68	69	0.00	2	1.8	464	PASS
69	198	0.00	100	37.2	25776	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	47.1	32632	PASS
197	198	0.00	2	0.2	141	PASS
198	198	100	100	100.0	69341	PASS
199	198	5	9	6.8	4687	PASS
275	198	10	60	27.7	19180	PASS
365	198	1	100	3.7	2552	PASS
441	198	0.01	100	13.8	9567	PASS
442	198	50	100	88.9	61635	PASS
443	442	15	24	18.8	11569	PASS