




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101524\
Data File : BF139804.D
Acq On : 15 Oct 2024 10:51
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 15 11:29:44 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 02 04:58:47 2024
Response via : Initial Calibration



TIC: BF139804.D\data.ms

(70) Pentachlorophenol (C)




11.181min (-0.018) 40759.65 ng

response 111004

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.50	63.20
264.00	61.60	63.34
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101524\
Data File : BF139804.D
Acq On : 15 Oct 2024 10:51
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 15 11:29:44 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 02 04:58:47 2024
Response via : Initial Calibration



TIC: BF139804.D\data.ms

(77) Benzidine

12.716min (-0.012) 68500.72 ng

response 449477

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.20	13.99
183.00	11.00	10.84
0.00	0.00	0.00