

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101524\
 Data File : BF139813.D
 Acq On : 15 Oct 2024 15:39
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF101524

Quant Time: Oct 15 16:44:19 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 15 16:41:54 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	102	0.00
2	1,4-Dioxane	40.000	38.500	3.8	102	0.00
3	Pyridine	40.000	39.616	1.0	103	0.00
4	n-Nitrosodimethylamine	40.000	39.447	1.4	102	0.00
5 S	2-Fluorophenol	80.000	77.861	2.7	102	0.00
6	Aniline	40.000	40.182	-0.5	103	0.00
7 S	Phenol-d6	80.000	77.760	2.8	102	0.00
8	2-Chlorophenol	40.000	38.912	2.7	101	0.00
9	Benzaldehyde	40.000	39.588	1.0	102	0.00
10 C	Phenol	40.000	39.048	2.4	102	0.00
11	bis(2-Chloroethyl)ether	40.000	38.993	2.5	101	0.00
12	1,3-Dichlorobenzene	40.000	38.324	4.2	101	0.00
13 C	1,4-Dichlorobenzene	40.000	38.305	4.2	100	0.00
14	1,2-Dichlorobenzene	40.000	39.243	1.9	103	0.00
15	Benzyl Alcohol	40.000	38.932	2.7	100	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.534	3.7	101	0.00
17	2-Methylphenol	40.000	39.132	2.2	102	0.00
18	Hexachloroethane	40.000	39.072	2.3	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.509	3.7	101	0.00
20	3+4-Methylphenols	40.000	38.278	4.3	101	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	101	0.00
22	Acetophenone	40.000	37.953	5.1	100	0.00
23 S	Nitrobenzene-d5	80.000	77.347	3.3	101	0.00
24	Nitrobenzene	40.000	39.408	1.5	104	0.00
25	Isophorone	40.000	38.964	2.6	102	0.00
26 C	2-Nitrophenol	40.000	40.323	-0.8	102	0.00
27	2,4-Dimethylphenol	40.000	38.752	3.1	101	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.685	3.3	102	0.00
29 C	2,4-Dichlorophenol	40.000	39.075	2.3	102	0.00
30	1,2,4-Trichlorobenzene	40.000	38.760	3.1	102	0.00
31	Naphthalene	40.000	38.486	3.8	102	0.00
32	Benzoic acid	40.000	43.071	-7.7	102	0.00
33	4-Chloroaniline	40.000	39.115	2.2	104	0.00
34 C	Hexachlorobutadiene	40.000	38.469	3.8	101	0.00
35	Caprolactam	40.000	40.138	-0.3	102	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.405	1.5	102	0.00
37	2-Methylnaphthalene	40.000	38.115	4.7	100	0.00
38	1-Methylnaphthalene	40.000	38.807	3.0	102	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.321	4.2	102	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.537	3.7	100	0.00
42 S	2,4,6-Tribromophenol	80.000	79.319	0.9	102	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.713	0.7	102	0.00
44	2,4,5-Trichlorophenol	40.000	39.399	1.5	104	0.00
45 S	2-Fluorobiphenyl	80.000	75.014	6.2	101	0.00
46	1,1'-Biphenyl	40.000	38.486	3.8	102	0.00
47	2-Chloronaphthalene	40.000	38.474	3.8	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101524\
 Data File : BF139813.D
 Acq On : 15 Oct 2024 15:39
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF101524

Quant Time: Oct 15 16:44:19 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 15 16:41:54 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.615	1.0	101	0.00
49	Acenaphthylene	40.000	38.616	3.5	101	0.00
50	Dimethylphthalate	40.000	38.938	2.7	101	0.00
51	2,6-Dinitrotoluene	40.000	39.497	1.3	100	0.00
52 C	Acenaphthene	40.000	38.023	4.9	101	0.00
53	3-Nitroaniline	40.000	39.304	1.7	100	0.00
54 P	2,4-Dinitrophenol	40.000	42.373	-5.9	99	0.00
55	Dibenzofuran	40.000	38.653	3.4	102	0.00
56 P	4-Nitrophenol	40.000	41.811	-4.5	99	0.00
57	2,4-Dinitrotoluene	40.000	40.548	-1.4	102	0.00
58	Fluorene	40.000	38.055	4.9	102	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.326	1.7	101	0.00
60	Diethylphthalate	40.000	38.426	3.9	101	0.00
61	4-Chlorophenyl-phenylether	40.000	37.663	5.8	100	0.00
62	4-Nitroaniline	40.000	39.190	2.0	100	0.00
63	Azobenzene	40.000	39.424	1.4	102	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.459	-3.6	101	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.849	2.9	102	0.00
67	4-Bromophenyl-phenylether	40.000	39.007	2.5	102	0.00
68	Hexachlorobenzene	40.000	37.830	5.4	98	0.00
69	Atrazine	40.000	31.061	22.3	99	0.00
70 C	Pentachlorophenol	40.000	40.920	-2.3	97	0.00
71	Phenanthrene	40.000	38.182	4.5	101	0.00
72	Anthracene	40.000	37.668	5.8	99	0.00
73	Carbazole	40.000	37.015	7.5	98	0.00
74	Di-n-butylphthalate	40.000	38.047	4.9	99	0.00
75 C	Fluoranthene	40.000	35.576	11.1	96	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	90	0.00
77	Benzydine	40.000	46.657	-16.6	95	0.00
78	Pyrene	40.000	34.263	14.3	94	0.00
79 S	Terphenyl-d14	80.000	81.589	-2.0	93	0.00
80	Butylbenzylphthalate	40.000	38.220	4.5	87	0.00
81	Benzo(a)anthracene	40.000	38.478	3.8	92	0.00
82	3,3'-Dichlorobenzidine	40.000	41.659	-4.1	91	0.00
83	Chrysene	40.000	39.222	1.9	89	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.601	1.0	85	0.00
85 c	Di-n-octyl phthalate	40.000	41.841	-4.6	88	0.00
86 I	Perylene-d12	20.000	20.000	0.0	97	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.498	3.8	100	0.00
88	Benzo(b)fluoranthene	40.000	40.246	-0.6	98	0.00
89	Benzo(k)fluoranthene	40.000	36.722	8.2	90	0.00
90 C	Benzo(a)pyrene	40.000	38.898	2.8	96	0.00
91	Dibenzo(a,h)anthracene	40.000	38.461	3.8	100	0.00
92	Benzo(g,h,i)perylene	40.000	37.880	5.3	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101524\
Data File : BF139813.D
Acq On : 15 Oct 2024 15:39
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF101524

Quant Time: Oct 15 16:44:19 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 15 16:41:54 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0