


Data Path : Z:\HPCHEM1\BNA_F\Data\BF101617\
Data File : BF099574.D
Acq On : 17 Oct 2017 5:36
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092317.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Oct 11 16:25:21 2017



AutoFind: Scans 1618, 1619, 1620; Background Corrected with Scan 1614

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.0	17116	PASS
68	69	0.00	2	1.2	205	PASS
69	198	0.00	100	32.9	17068	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	47.6	24682	PASS
197	198	0.00	2	0.4	182	PASS
198	198	100	100	100.0	51842	PASS
199	198	5	9	6.4	3323	PASS
275	198	10	60	23.2	12002	PASS
365	198	1	100	2.6	1363	PASS
441	442	0.01	24	15.2	6507	PASS
442	198	50	100	82.5	42760	PASS
443	442	15	24	18.7	7978	PASS