

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101722\
 Data File : BF130687.D
 Acq On : 17 Oct 2022 15:23
 Operator : CG\JU
 Sample : N5133-03
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 HAS-WATER-1013

Quant Time: Oct 18 03:21:24 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 18 03:15:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.569	152	91427	20.000	ng	0.00
21) Naphthalene-d8	7.857	136	358249	20.000	ng	0.00
39) Acenaphthene-d10	9.604	164	187150	20.000	ng	0.00
64) Phenanthrene-d10	11.080	188	302490	20.000	ng	0.00
76) Chrysene-d12	13.715	240	155362	20.000	ng	0.00
86) Perylene-d12	15.086	264	167486	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.175	112	264544	52.157	ng	0.00
7) Phenol-d6	6.222	99	206337	32.356	ng	0.00
23) Nitrobenzene-d5	7.145	82	521379	79.449	ng	0.00
42) 2,4,6-Tribromophenol	10.398	330	256051	139.823	ng	0.00
45) 2-Fluorobiphenyl	8.933	172	1034932	83.765	ng	0.00
79) Terphenyl-d14	12.668	244	825318	88.169	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101722\
Data File : BF130687.D
Acq On : 17 Oct 2022 15:23
Operator : CG\JU
Sample : N5133-03
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
HAS-WATER-1013

Quant Time: Oct 18 03:21:24 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 18 03:15:54 2022
Response via : Initial Calibration

