

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF101816\
 Data File : BF090630.D
 Acq On : 18 Oct 2016 16:18
 Operator : UM/SJ
 Sample : H5289-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SP-3-1

Quant Time: Oct 19 01:06:19 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF101316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 17 19:19:58 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.90	152	226413	20.00	ng	-0.01
21) Naphthalene-d8	8.19	136	989850	20.00	ng	-0.01
38) Acenaphthene-d10	9.95	164	509611	20.00	ng	-0.01
63) Phenanthrene-d10	11.43	188	744294	20.00	ng	-0.01
75) Chrysene-d12	14.08	240	515688	20.00	ng	-0.01
86) Perylene-d12	15.54	264	488647	20.00	ng	0.00

System Monitoring Compounds						
5) 2-Fluorophenol	5.51	112	1263586	102.31	ng	0.00
7) Phenol-d6	6.53	99	1651867	90.01	ng	-0.01
23) Nitrobenzene-d5	7.46	82	1106798	62.11	ng	-0.02
41) 2,4,6-Tribromophenol	10.74	330	343346	75.57	ng	-0.01
44) 2-Fluorobiphenyl	9.26	172	1623833	52.50	ng	-0.01
78) Terphenyl-d14	13.02	244	1352561	61.07	ng	-0.01


Target Compounds						Qvalue
4) n-Nitrosodimethylamine	3.30	42	1367	7.38	ng	# 1
19) n-Nitroso-di-n-propylamine	7.46	70	132110	10.50	ng	# 81
31) Naphthalene	8.21	128	151680	3.01	ng	100
35) Caprolactam	8.61	113	7796	2.32	ng	# 34
37) 2-Methylnaphthalene	8.90	142	132042	4.46	ng	98
48) Acenaphthylene	9.80	152	116422	2.53	ng	98
49) Dimethylphthalate	9.66	163	797146	24.11	ng	98
54) Dibenzofuran	10.15	168	91724	2.28	ng	99
70) Phenanthrene	11.46	178	372137	9.37	ng	98
71) Anthracene	11.50	178	145293	3.40	ng	100
74) Fluoranthene	12.65	202	549008	13.74	ng	95
77) Pyrene	12.87	202	493499	14.45	ng	99
80) Benzo(a)anthracene	14.10	228	821877	28.00	ng	95
82) Chrysene	14.10	228	821877	29.06	ng	98
85) Indeno(1,2,3-cd)pyrene	16.99	276	259469	15.86	ng	# 91
87) Benzo(b)fluoranthene	15.12	252	1092994	34.18	ng	97
88) Benzo(k)fluoranthene	15.12	252	1092994	31.77	ng	98
89) Benzo(a)pyrene	15.48	252	364084	12.44	ng	# 96
90) Dibenzo(a,h)anthracene	16.99	278	64450	2.89	ng	92
91) Benzo(g,h,i)perylene	17.42	276	223326	10.49	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF101816\
Data File : BF090630.D
Acq On : 18 Oct 2016 16:18
Operator : UM/SJ
Sample : H5289-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SP-3-1

Quant Time: Oct 19 01:06:19 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF101316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 17 19:19:58 2016
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.90 min Scan# 399
Delta R.T. -0.01 min
Lab File: BF090630.D
Acq: 18 Oct 2016 16:18

Instrument :
BNA_F
ClientSampleId :
SP-3-1

Tgt Ion: 152 Resp: 226413
Ion Ratio Lower Upper
152 100
150 153.7 127.2 190.8
115 66.7 58.6 88.0




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#4


n-Nitrosodimethylamine
Concen: 7.38 ng
RT: 3.30 min Scan# 84
Delta R.T. 0.07 min
Lab File: BF090630.D
Acq: 18 Oct 2016 16:18

Tgt Ion: 42 Resp: 1367
Ion Ratio Lower Upper
42 100
74 0.0 113.3 169.9#
44 80.2 6.0 9.0#



Abundance Ion 42.00 (41.70 to 42.70): BF0
Ion 74.00 (73.70 to 74.70): BF0
Ion 44.00 (43.70 to 44.70): BF0





#5

2-Fluorophenol

Concen: 102.31 ng

RT: 5.51 min Scan# 278

Delta R.T. -0.00 min

Lab File: BF090630.D

Acq: 18 Oct 2016 16:18

Instrument :

BNA_F

ClientSampleId :

SP-3-1


Tgt Ion: 112 Resp: 1263586

Ion Ratio Lower Upper

112 100

64 56.5 52.6 78.8




63 27.4 26.0 39.0



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#7

Phenol-d6

Concen: 90.01 ng

RT: 6.53 min Scan# 367

Delta R.T. -0.01 min

Lab File: BF090630.D

Acq: 18 Oct 2016 16:18


Tgt Ion: 99 Resp: 1651867

Ion Ratio Lower Upper

99 100

42 17.7 14.7 22.1



71 34.2 27.6 41.4




Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0





#19

n-Nitroso-di-n-propylamine

Concen: 10.50 ng

RT: 7.46 min Scan# 448

Delta R.T. 0.13 min

Lab File: BF090630.D

Acq: 18 Oct 2016 16:18

Instrument :

BNA_F

ClientSampleID :

SP-3-1

Tgt Ion: 70 Resp: 132110





Ion Ratio Lower Upper

70 100

42 43.6 43.4 65.0

101 0.0 6.4 9.6#

130 1.8 12.1 18.1#



#21

Naphthalene-d8

Concen: 20.00 ng

RT: 8.19 min Scan# 512

Delta R.T. -0.01 min

Lab File: BF090630.D

Acq: 18 Oct 2016 16:18

Tgt Ion: 136 Resp: 989850




Ion Ratio Lower Upper


136 100

137 11.2 8.8 13.2

54 7.4 6.8 10.2

68 6.8 6.0 9.0









#23
 Nitrobenzene-d5
 Concen: 62.11 ng
 RT: 7.46 min Scan# 448
 Delta R.T. -0.02 min
 Lab File: BF090630.D
 Acq: 18 Oct 2016 16:18




Instrument :
 BNA_F
 ClientSampleId :
 SP-3-1


Tgt Ion: 82 Resp: 1106798
 Ion Ratio Lower Upper
 82 100
 128 35.0 34.3 51.5
 54 52.5 37.0 55.6



#31
 Naphthalene
 Concen: 3.01 ng
 RT: 8.21 min Scan# 514
 Delta R.T. -0.01 min
 Lab File: BF090630.D
 Acq: 18 Oct 2016 16:18

Tgt Ion: 128 Resp: 151680
 Ion Ratio Lower Upper
 128 100
 129 11.2 9.0 13.6
 127 13.4 10.8 16.2









#35
 Caprolactam
 Concen: 2.32 ng
 RT: 8.61 min Scan# 549
 Delta R.T. -0.05 min
 Lab File: BF090630.D
 Acq: 18 Oct 2016 16:18




Instrument :
 BNA_F
 ClientSampled :
 SP-3-1


Tgt Ion:113 Resp: 7796
 Ion Ratio Lower Upper
 113 100
 55 246.0 170.8 210.8#
 56 269.3 122.4 162.4#



#37
 2-Methylnaphthalene
 Concen: 4.46 ng
 RT: 8.90 min Scan# 574
 Delta R.T. -0.01 min
 Lab File: BF090630.D
 Acq: 18 Oct 2016 16:18

Tgt Ion:142 Resp: 132042
 Ion Ratio Lower Upper
 142 100
 141 90.2 71.3 106.9
 115 37.7 28.8 43.2





#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 9.95 min Scan# 666

Delta R.T. -0.01 min

Lab File: BF090630.D

Acq: 18 Oct 2016 16:18

Instrument :

BNA_F

ClientSampleId :

SP-3-1


Tgt Ion:164 Resp: 509611

Ion Ratio Lower Upper

164 100

162 102.4 82.2 123.4




160 46.2 36.6 55.0



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#41

2,4,6-Tribromophenol

Concen: 75.57 ng

RT: 10.74 min Scan# 735

Delta R.T. -0.01 min

Lab File: BF090630.D

Acq: 18 Oct 2016 16:18


Tgt Ion:330 Resp: 343346

Ion Ratio Lower Upper

330 100

332 97.1 76.2 114.2



141 45.3 31.4 47.2




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#44
2-Fluorobiphenyl
Concen: 52.50 ng
RT: 9.26 min Scan# 606
Delta R.T. -0.01 min
Lab File: BF090630.D
Acq: 18 Oct 2016 16:18

Instrument :
BNA_F
ClientSampleId :
SP-3-1

Tgt Ion:172 Resp: 1623833
Ion Ratio Lower Upper
172 100
171 36.7 30.1 45.1
170 25.0 20.6 30.8




Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#48
Acenaphthylene
Concen: 2.53 ng
RT: 9.80 min Scan# 653
Delta R.T. -0.02 min
Lab File: BF090630.D
Acq: 18 Oct 2016 16:18

Tgt Ion:152 Resp: 116422
Ion Ratio Lower Upper
152 100
151 20.8 16.6 25.0
153 15.9 11.3 16.9



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E





#49

Dimethylphthalate

Concen: 24.11 ng

RT: 9.66 min Scan# 641

Delta R.T. -0.01 min

Lab File: BF090630.D

Acq: 18 Oct 2016 16:18

Instrument :

BNA_F

ClientSampleId :

SP-3-1


Tgt Ion:163 Resp: 797146

Ion Ratio Lower Upper

163 100

194 3.5 2.3 3.5




164 10.2 8.6 13.0



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#54

Dibenzofuran

Concen: 2.28 ng

RT: 10.15 min Scan# 684

Delta R.T. -0.01 min

Lab File: BF090630.D

Acq: 18 Oct 2016 16:18


Tgt Ion:168 Resp: 91724

Ion Ratio Lower Upper

168 100

139 38.2 30.5 45.7



169 15.1 10.9 16.3




Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E





#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 11.43 min Scan# 796

Delta R.T. -0.01 min

Lab File: BF090630.D

Acq: 18 Oct 2016 16:18

Instrument :

BNA_F

ClientSampleId :

SP-3-1


Tgt Ion:188 Resp: 744294

Ion Ratio Lower Upper

188 100

94 10.1 7.7 11.5




80 11.4 8.6 12.8



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF0

Ion 80.00 (79.70 to 80.70): BF0



#70

Phenanthrene

Concen: 9.37 ng

RT: 11.46 min Scan# 798

Delta R.T. -0.01 min

Lab File: BF090630.D

Acq: 18 Oct 2016 16:18


Tgt Ion:178 Resp: 372137

Ion Ratio Lower Upper

178 100

176 19.2 16.1 24.1



179 15.2 12.6 19.0




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#71
 Anthracene
 Concen: 3.40 ng
 RT: 11.50 min Scan# 802
 Delta R.T. -0.01 min
 Lab File: BF090630.D
 Acq: 18 Oct 2016 16:18




Instrument :
 BNA_F
 ClientSampled :
 SP-3-1


Tgt Ion:178 Resp: 145293
 Ion Ratio Lower Upper
 178 100
 176 19.2 15.5 23.3
 179 16.3 13.0 19.4



#74
 Fluoranthene
 Concen: 13.74 ng
 RT: 12.65 min Scan# 902
 Delta R.T. -0.01 min
 Lab File: BF090630.D
 Acq: 18 Oct 2016 16:18

Tgt Ion:202 Resp: 549008
 Ion Ratio Lower Upper
 202 100
 101 15.4 0.0 31.1
 203 18.2 0.0 38.0





#75

Chrysene-d12

Concen: 20.00 ng

RT: 14.08 min Scan# 1027

Delta R.T. -0.01 min

Lab File: BF090630.D

Acq: 18 Oct 2016 16:18

Instrument :

BNA_F

ClientSampleId :

SP-3-1


Tgt Ion:240 Resp: 515688

Ion Ratio Lower Upper

240 100

120 15.4 9.8 14.8#




236 26.9 21.7 32.5



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#77

Pyrene

Concen: 14.45 ng

RT: 12.87 min Scan# 922

Delta R.T. -0.01 min

Lab File: BF090630.D

Acq: 18 Oct 2016 16:18


Tgt Ion:202 Resp: 493499

Ion Ratio Lower Upper

202 100

200 21.8 17.8 26.6



203 18.3 14.6 21.8




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#78

Terphenyl-d14

Concen: 61.07 ng

RT: 13.02 min Scan# 935

Delta R.T. -0.01 min

Lab File: BF090630.D

Acq: 18 Oct 2016 16:18

Instrument :

BNA_F

ClientSampleId :

SP-3-1


Tgt Ion:244 Resp: 1352561

Ion Ratio Lower Upper

244 100

212 6.7 5.0 7.4




122 10.1 5.6 8.4#



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Benzo(a)anthracene

Concen: 28.00 ng

RT: 14.10 min Scan# 1029

Delta R.T. 0.02 min

Lab File: BF090630.D

Acq: 18 Oct 2016 16:18


Tgt Ion:228 Resp: 821877

Ion Ratio Lower Upper

228 100

226 31.0 22.8 34.2



229 22.8 16.6 25.0




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#82

Chrysene

Concen: 29.06 ng

RT: 14.10 min Scan# 1029

Delta R.T. -0.01 min

Lab File: BF090630.D

Acq: 18 Oct 2016 16:18

Instrument :

BNA_F

ClientSampleId :

SP-3-1


Tgt Ion:228 Resp: 821877

Ion Ratio Lower Upper

228 100

226 31.0 24.8 37.2




229 22.8 16.2 24.4



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#85

Indeno(1,2,3-cd)pyrene

Concen: 15.86 ng

RT: 16.99 min Scan# 1282

Delta R.T. -0.00 min

Lab File: BF090630.D

Acq: 18 Oct 2016 16:18


Tgt Ion:276 Resp: 259469

Ion Ratio Lower Upper

276 100

138 25.6 25.8 38.6#



277 27.5 20.1 30.1




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#86
Perylene-d12
Concen: 20.00 ng
RT: 15.54 min Scan# 1155
Delta R.T. -0.00 min
Lab File: BF090630.D
Acq: 18 Oct 2016 16:18

Instrument :
BNA_F
ClientSampleId :
SP-3-1

Tgt Ion: 264 Resp: 488647




Ion	Ratio	Lower	Upper
264	100		
260	25.2	19.3	28.9
265	23.6	16.7	25.1




#87
Benzo(b)fluoranthene
Concen: 34.18 ng
RT: 15.12 min Scan# 1118
Delta R.T. -0.00 min
Lab File: BF090630.D
Acq: 18 Oct 2016 16:18

Tgt Ion: 252 Resp: 1092994

Ion	Ratio	Lower	Upper
252	100		
253	23.3	17.8	26.8
125	13.7	12.0	18.0









#88
Benzo(k)fluoranthene
Concen: 31.77 ng
RT: 15.12 min Scan# 1118
Delta R.T. -0.03 min
Lab File: BF090630.D
Acq: 18 Oct 2016 16:18




Instrument :
BNA_F
ClientSampleId :
SP-3-1


Tgt Ion:252 Resp: 1092994
Ion Ratio Lower Upper
252 100
253 23.3 18.0 27.0
125 13.7 10.0 15.0



#89
Benzo(a)pyrene
Concen: 12.44 ng
RT: 15.48 min Scan# 1150
Delta R.T. -0.00 min
Lab File: BF090630.D
Acq: 18 Oct 2016 16:18

Tgt Ion:252 Resp: 364084
Ion Ratio Lower Upper
252 100
253 21.8 17.9 26.9
125 12.0 12.3 18.5#





#90

Dibenzo(a,h)anthracene

Concen: 2.89 ng

RT: 16.99 min Scan# 1282

Delta R.T. -0.02 min

Lab File: BF090630.D

Acq: 18 Oct 2016 16:18

Instrument :

BNA_F

ClientSampleId :

SP-3-1


Tgt Ion:278 Resp: 64450

Ion Ratio Lower Upper

278 100

139 24.8 16.6 25.0




279 27.8 19.1 28.7



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#91

Benzo(g,h,i)perylene

Concen: 10.49 ng

RT: 17.42 min Scan# 1320

Delta R.T. -0.00 min

Lab File: BF090630.D

Acq: 18 Oct 2016 16:18


Tgt Ion:276 Resp: 223326

Ion Ratio Lower Upper

276 100

277 24.3 18.8 28.2

138 25.3 21.8 32.8



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

