

Data Path : Z:\HPCHEM1\BNA F\DATA\BF101817\
 Data File : BF099623.D
 Acq On : 18 Oct 2017 17:59
 Operator : SJ/JU
 Sample : I5834-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ENV-115-4(5-10)

Manual Integrations
 APPROVED

Sohil
 10/19/2017 5:14:52 PM

Quant Time: Oct 18 22:33:24 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 18 15:18:49 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.83	152	108916	20.00	ng	0.00
21) Naphthalene-d8	8.10	136	454984	20.00	ng	0.00
38) Acenaphthene-d10	9.86	164	204748	20.00	ng	0.00
63) Phenanthrene-d10	11.35	188	338151	20.00	ng	0.00
75) Chrysene-d12	13.99	240	193554	20.00	ng	0.00
86) Perylene-d12	15.45	264	191043	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.46	112	664644	97.14	ng	0.00
7) Phenol-d6	6.47	99	800747	94.87	ng	0.00
23) Nitrobenzene-d5	7.39	82	480027	67.66	ng	0.00
41) 2,4,6-Tribromophenol	10.66	330	171137	102.15	ng	0.00
44) 2-Fluorobiphenyl	9.18	172	802076	65.40	ng	0.00
78) Terphenyl-d14	12.93	244	572179	67.23	ng	0.00
Target Compounds						
10) Phenol	6.48	94	53180	5.634	ng	90
49) Dimethylphthalate	9.57	163	193829	12.543	ng	99
74) Fluoranthene	12.56	202	79272	4.325	ng	96
77) Pyrene	12.79	202	66787	4.525	ng	99
80) Benzo(a)anthracene	13.98	228	24233	2.068	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF101817\
Data File : BF099623.D
Acq On : 18 Oct 2017 17:59
Operator : SJ/JU
Sample : I5834-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ENV-115-4(5-10)

Manual Integrations
APPROVED

Sohil
10/19/2017 5:14:52 PM

Quant Time: Oct 18 22:33:24 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 18 15:18:49 2017
Response via : Initial Calibration

