

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF101818\
 Data File : BF109954.D
 Acq On : 18 Oct 2018 12:04
 Operator : JU/SJ
 Sample : J5543-08
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 101618-SIDI-E-U-S

Quant Time: Oct 19 02:36:30 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 16 14:58:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.97	152	279668	20.00	ng	0.00
21) Naphthalene-d8	8.26	136	1144560	20.00	ng	-0.01
39) Acenaphthene-d10	10.02	164	515955	20.00	ng	-0.01
64) Phenanthrene-d10	11.51	188	903270	20.00	ng	0.00
76) Chrysene-d12	14.16	240	496702	20.00	ng	0.00
87) Perylene-d12	15.69	264	516716	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	732517	41.52	ng	0.00
7) Phenol-d6	6.59	99	563276	25.71	ng	-0.02
23) Nitrobenzene-d5	7.54	82	1720273	101.25	ng	-0.01
42) 2,4,6-Tribromophenol	10.81	330	421599	94.38	ng	0.00
45) 2-Fluorobiphenyl	9.34	172	2789268	86.26	ng	0.00
79) Terphenyl-d14	13.10	244	1970411	83.30	ng	0.00
Target Compounds						
10) Phenol	6.60	94	59667	2.523	ng	Qvalue # 60
50) Dimethylphthalate	9.72	163	87620	2.398	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF101818\
Data File : BF109954.D
Acq On : 18 Oct 2018 12:04
Operator : JU/SJ
Sample : J5543-08
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
101618-SIDI-E-U-S

Quant Time: Oct 19 02:36:30 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 16 14:58:24 2018
Response via : Initial Calibration

