


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF101818\
 Data File : BF109950.D
 Acq On : 18 Oct 2018 10:15
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101518.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Oct 16 14:58:24 2018



AutoFind: Scans 1641, 1642, 1643; Background Corrected with Scan 1636

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.5	69949	PASS
68	69	0.00	2	1.9	1227	PASS
69	198	0.00	100	37.2	64229	PASS
70	69	0.00	2	0.4	285	PASS
127	198	10	80	49.3	85245	PASS
197	198	0.00	2	0.8	1467	PASS
198	198	100	100	100.0	172797	PASS
199	198	5	9	7.0	12028	PASS
275	198	10	60	25.0	43176	PASS
365	198	1	100	2.7	4741	PASS
441	198	0.01	100	13.7	23636	PASS
442	198	50	100	87.5	151210	PASS
443	442	15	24	18.3	27731	PASS