


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101821\
 Data File : BF125870.D
 Acq On : 18 Oct 2021 11:05
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF100721.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Oct 18 12:43:03 2021






AutoFind: Scans 1616, 1617, 1618; Background Corrected with Scan 1611

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.8	53763	PASS
68	69	0.00	2	1.7	994	PASS
69	198	0.00	100	32.2	58139	PASS
70	69	0.00	2	0.4	213	PASS
127	198	10	80	48.1	86739	PASS
197	198	0.00	2	0.6	1114	PASS
198	198	100	100	100.0	180301	PASS
199	198	5	9	6.7	12096	PASS
275	198	10	60	26.1	47069	PASS
365	198	1	100	3.1	5638	PASS
441	198	0.01	100	12.9	23320	PASS
442	442	50	100	100.0	150277	PASS
443	442	15	24	18.7	28149	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101821\
Data File : BF125870.D
Acq On : 18 Oct 2021 11:05
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 18 12:36:30 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF100721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 15 15:00:46 2021
Response via : Initial Calibration



TIC: BF125870.D\data.ms

(70) Pentachlorophenol (C)

11.145min (-0.012) 39466.11 ng




response 179256

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.10	62.96
264.00	69.90	61.37
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101821\
Data File : BF125870.D
Acq On : 18 Oct 2021 11:05
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 18 12:36:30 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF100721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 15 15:00:46 2021
Response via : Initial Calibration



TIC: BF125870.D\data.ms

(77) Benzidine

12.674min (-0.006) 0.00 ng

response 817813

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.50	13.57
183.00	10.10	11.13
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/18/2021	BNA_F	<u>BF125870.D</u>
Compound Name	Response	Retention Time
DDT	343020	13.49
DDD	7245	13.19
DDE	468	12.86
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
7713	350733	2.20

Instrument :
BNA_F
ClientSampleId :
DFTPP