

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: _____

Lab Code: CHEM Case No.: BF101924 SAS No.: BF101924 SDG No.: BF101924

Instrument ID: BNA_F Calibration Date/Time: 10/19/2024 : 13:19

Lab File ID: BF139868.D Init. Calib. Date(s): _____

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): _____

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
n-Nitrosodimethylamine	0.793	0.785		-1.009	
Pyridine	1.476	1.457		-1.287	
2-Fluorophenol	1.278	1.257		-1.643	
Benzaldehyde	0.931	0.857		-7.948	
Aniline	1.542	1.588		2.983	
Phenol-d6	1.655	1.616		-2.356	
Phenol	1.706	1.691		-0.879	20.0
bis(2-Chloroethyl) ether	1.319	1.311		-0.607	
2-Chlorophenol	1.306	1.290		-1.225	
1,2-Dichlorobenzene	1.398	1.374		-1.717	
1,3-Dichlorobenzene	1.499	1.481		-1.201	
1,4-Dichlorobenzene	1.498	1.477		-1.402	20.0
Benzyl Alcohol	1.214	1.235		1.73	
2-Methylphenol	1.114	1.119		0.449	
2,2-oxybis(1-Chloropropane)	2.248	2.235		-0.578	
Acetophenone	0.490	0.473		-3.469	
3+4-Methylphenols	1.424	1.412		-0.843	
n-Nitroso-di-n-propylamine	1.004	0.992	0.050	-1.195	
Nitrobenzene-d5	0.361	0.368		1.939	
Hexachloroethane	0.534	0.529		-0.936	
Nitrobenzene	0.396	0.395		-0.253	
Isophorone	0.685	0.695		1.46	
2-Nitrophenol	0.149	0.163		9.396	20.0
2,4-Dimethylphenol	0.249	0.247		-0.803	
bis(2-Chloroethoxy) methane	0.415	0.414		-0.241	
2,4-Dichlorophenol	0.283	0.291		2.827	20.0
1,2,4-Trichlorobenzene	0.314	0.312		-0.637	
Benzoic acid	0.217	0.240		10.599	
Naphthalene	1.031	1.003		-2.716	
4-Chloroaniline	0.352	0.357		1.42	
Hexachlorobutadiene	0.197	0.195		-1.015	20.0
Caprolactam	0.090	0.097		7.778	
4-Chloro-3-methylphenol	0.314	0.319		1.592	20.0
2-Methylnaphthalene	0.632	0.623		-1.424	
Hexachlorocyclopentadiene	0.188	0.203	0.050	7.979	
2,4,6-Trichlorophenol	0.382	0.375		-1.832	20.0
2-Fluorobiphenyl	1.210	1.118		-7.603	
2,4,5-Trichlorophenol	0.394	0.408		3.553	
1,1-Biphenyl	1.392	1.331		-4.382	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: _____

Lab Code: CHEM Case No.: BF101924 SAS No.: BF101924 SDG No.: BF101924

Instrument ID: BNA_F Calibration Date/Time: 10/19/2024 : 13:19

Lab File ID: BF139868.D Init. Calib. Date(s): _____

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): _____

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Chloronaphthalene	1.121	1.086		-3.122	
2-Nitroaniline	0.343	0.368		7.289	
Dimethylphthalate	1.246	1.253		0.562	
Acenaphthylene	1.621	1.602		-1.172	
2,6-Dinitrotoluene	0.270	0.288		6.667	
3-Nitroaniline	0.278	0.294		5.755	
Acenaphthene	1.049	1.033		-1.525	20.0
2,4-Dinitrophenol	0.093	0.111	0.050	19.355	
4-Nitrophenol	0.214	0.240	0.050	12.15	
Dibenzofuran	1.505	1.468		-2.458	
2,4-Dinitrotoluene	0.332	0.373		12.349	
Diethylphthalate	1.223	1.247		1.962	
4-Chlorophenyl-phenylether	0.579	0.566		-2.245	
Fluorene	1.146	1.104		-3.665	
4-Nitroaniline	0.263	0.287		9.125	
4,6-Dinitro-2-methylphenol	0.082	0.095		15.854	
n-Nitrosodiphenylamine	0.599	0.574		-4.174	20.0
Azobenzene	1.297	1.285		-0.925	
2,4,6-Tribromophenol	0.187	0.197		5.348	
4-Bromophenyl-phenylether	0.206	0.206		0.0	
Hexachlorobenzene	0.232	0.227		-2.155	
Atrazine	0.162	0.160		-1.235	
Pentachlorophenol	0.141	0.153		8.511	20.0
Phenanthrene	0.945	0.895		-5.291	
Anthracene	0.922	0.896		-2.82	
Carbazole	0.857	0.833		-2.8	
Di-n-butylphthalate	0.990	1.009		1.919	
Fluoranthene	0.955	0.934		-2.199	20.0
Benzidine	0.329	0.292		-11.246	
Pyrene	1.751	1.863		6.396	
Terphenyl-d14	1.227	1.294		5.46	
Butylbenzylphthalate	0.525	0.590		12.381	
3,3-Dichlorobenzidine	0.380	0.367		-3.421	
Benzo(a)anthracene	1.302	1.324		1.69	
Chrysene	1.194	1.212		1.508	
Bis(2-ethylhexyl)phthalate	0.589	0.631		7.131	
Di-n-octyl phthalate	1.071	1.129		5.415	20.0
Benzo(b)fluoranthene	1.218	1.157		-5.008	
Benzo(k)fluoranthene	1.051	1.103		4.948	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: _____

Lab Code: CHEM Case No.: BF101924 SAS No.: BF101924 SDG No.: BF101924

Instrument ID: BNA_F Calibration Date/Time: 10/19/2024 : 13:19

Lab File ID: BF139868.D Init. Calib. Date(s): _____

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): _____

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Benzo(a)pyrene	1.002	1.023		2.096	20.0
Indeno(1,2,3-cd)pyrene	1.287	1.426		10.8	
Dibenzo(a,h)anthracene	1.073	1.186		10.531	
Benzo(g,h,i)perylene	1.072	1.225		14.272	
1,2,4,5-Tetrachlorobenzene	0.540	0.528		-2.222	
1,4-Dioxane	0.596	0.589		-1.174	20.0
2,3,4,6-Tetrachlorophenol	0.309	0.321		3.883	
1-Methylnaphthalene	0.620	0.612		-1.29	

All other compounds must meet a minimum RRF of 0.010.

Report of Analysis

Client:		Date Collected:	10/17/2024 12:00:00 AM
Project:		Date Received:	10/17/2024 12:00:00 AM
Client Sample ID:	PB164216BL	SDG No.:	BF101924
Lab Sample ID:	PB164216BL	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	30.01	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139869.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	86.8	U	86.8	270	330	ug/Kg
110-86-1	Pyridine	79.9	U	79.9	130	170	ug/Kg
100-52-7	Benzaldehyde	180	U	180	270	330	ug/Kg
62-53-3	Aniline	96.9	U	96.9	130	170	ug/Kg
108-95-2	Phenol	82.9	U	82.9	130	170	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	83.7	U	83.7	130	170	ug/Kg
95-57-8	2-Chlorophenol	83.5	U	83.5	130	170	ug/Kg
95-50-1	1,2-Dichlorobenzene	84.6	U	84.6	130	170	ug/Kg
541-73-1	1,3-Dichlorobenzene	86	U	86	130	170	ug/Kg
106-46-7	1,4-Dichlorobenzene	86.5	U	86.5	130	170	ug/Kg
100-51-6	Benzyl Alcohol	83	U	83	270	330	ug/Kg
95-48-7	2-Methylphenol	80.6	U	80.6	130	170	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	90.9	U	90.9	130	170	ug/Kg
98-86-2	Acetophenone	86.9	U	86.9	130	170	ug/Kg
65794-96-9	3+4-Methylphenols	79.8	U	79.8	270	330	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	40.3	U	40.3	80	80	ug/Kg
67-72-1	Hexachloroethane	83	U	83	130	170	ug/Kg
98-95-3	Nitrobenzene	90.8	U	90.8	130	170	ug/Kg
78-59-1	Isophorone	84.6	U	84.6	130	170	ug/Kg
88-75-5	2-Nitrophenol	94.5	U	94.5	130	170	ug/Kg
105-67-9	2,4-Dimethylphenol	93.2	U	93.2	130	170	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	85.8	U	85.8	130	170	ug/Kg
120-83-2	2,4-Dichlorophenol	75.5	U	75.5	130	170	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	85.4	U	85.4	130	170	ug/Kg
65-85-0	Benzoic acid	240	U	240	270	330	ug/Kg
91-20-3	Naphthalene	82.6	U	82.6	130	170	ug/Kg
106-47-8	4-Chloroaniline	82.6	U	82.6	130	170	ug/Kg
87-68-3	Hexachlorobutadiene	83.3	U	83.3	130	170	ug/Kg

Report of Analysis

Client:		Date Collected:	10/17/2024 12:00:00 AM
Project:		Date Received:	10/17/2024 12:00:00 AM
Client Sample ID:	PB164216BL	SDG No.:	BF101924
Lab Sample ID:	PB164216BL	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	30.01 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139869.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	86.8	U	86.8	270	330	ug/Kg
59-50-7	4-Chloro-3-methylphenol	77.5	U	77.5	130	170	ug/Kg
91-57-6	2-Methylnaphthalene	82.5	U	82.5	130	170	ug/Kg
77-47-4	Hexachlorocyclopentadiene	160	U	160	270	330	ug/Kg
88-06-2	2,4,6-Trichlorophenol	71.4	U	71.4	130	170	ug/Kg
95-95-4	2,4,5-Trichlorophenol	74	U	74	130	170	ug/Kg
92-52-4	1,1-Biphenyl	87.4	U	87.4	130	170	ug/Kg
91-58-7	2-Chloronaphthalene	83.3	U	83.3	130	170	ug/Kg
88-74-4	2-Nitroaniline	95	U	95	130	170	ug/Kg
131-11-3	Dimethylphthalate	81.7	U	81.7	130	170	ug/Kg
208-96-8	Acenaphthylene	86.5	U	86.5	130	170	ug/Kg
606-20-2	2,6-Dinitrotoluene	83.2	U	83.2	130	170	ug/Kg
99-09-2	3-Nitroaniline	89.2	U	89.2	130	170	ug/Kg
83-32-9	Acenaphthene	81.1	U	81.1	130	170	ug/Kg
Total PAH	Total PAH	1325.1	U	1325.1	130	2720	ug/Kg
51-28-5	2,4-Dinitrophenol	240	U	240	270	330	ug/Kg
100-02-7	4-Nitrophenol	120	U	120	270	330	ug/Kg
132-64-9	Dibenzofuran	84.4	U	84.4	130	170	ug/Kg
121-14-2	2,4-Dinitrotoluene	86.2	U	86.2	130	170	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	169.4	U	169.4	130	340	ug/Kg
84-66-2	Diethylphthalate	80.1	U	80.1	130	170	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	85.6	U	85.6	130	170	ug/Kg
86-73-7	Fluorene	85.5	U	85.5	130	170	ug/Kg
100-01-6	4-Nitroaniline	110	U	110	130	170	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	120	U	120	270	330	ug/Kg
86-30-6	n-Nitrosodiphenylamine	81.6	U	81.6	130	170	ug/Kg
103-33-3	Azobenzene	79.6	U	79.6	130	170	ug/Kg
101-55-3	4-Bromophenyl-phenylether	78.9	U	78.9	130	170	ug/Kg
118-74-1	Hexachlorobenzene	85	U	85	130	170	ug/Kg
1912-24-9	Atrazine	91.4	U	91.4	130	170	ug/Kg

Report of Analysis

Client:		Date Collected:	10/17/2024 12:00:00 AM
Project:		Date Received:	10/17/2024 12:00:00 AM
Client Sample ID:	PB164216BL	SDG No.:	BF101924
Lab Sample ID:	PB164216BL	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	30.01 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139869.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	77.3	U	77.3	270	330	ug/Kg
85-01-8	Phenanthrene	84	U	84	130	170	ug/Kg
120-12-7	Anthracene	84.4	U	84.4	130	170	ug/Kg
86-74-8	Carbazole	80.3	U	80.3	130	170	ug/Kg
84-74-2	Di-n-butylphthalate	84.3	U	84.3	130	170	ug/Kg
206-44-0	Fluoranthene	81.7	U	81.7	130	170	ug/Kg
92-87-5	Benzidine	160	U	160	270	330	ug/Kg
129-00-0	Pyrene	83	U	83	130	170	ug/Kg
85-68-7	Butylbenzylphthalate	96.8	U	96.8	130	170	ug/Kg
91-94-1	3,3-Dichlorobenzidine	98.6	U	98.6	270	330	ug/Kg
56-55-3	Benzo(a)anthracene	80.7	U	80.7	130	170	ug/Kg
218-01-9	Chrysene	79.5	U	79.5	130	170	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	91	U	91	130	170	ug/Kg
117-84-0	Di-n-octyl phthalate	110	U	110	270	330	ug/Kg
205-99-2	Benzo(b)fluoranthene	81.1	U	81.1	130	170	ug/Kg
207-08-9	Benzo(k)fluoranthene	82.6	U	82.6	130	170	ug/Kg
50-32-8	Benzo(a)pyrene	93	U	93	130	170	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	78.1	U	78.1	130	170	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	81.2	U	81.2	130	170	ug/Kg
191-24-2	Benzo(g,h,i)perylene	80.1	U	80.1	130	170	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	86.8	U	86.8	130	170	ug/Kg
123-91-1	1,4-Dioxane	110	U	110	130	170	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	74.7	U	74.7	130	170	ug/Kg
90-12-0	1-Methylnaphthalene	83.2	U	83.2	170	170	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	127.174		18-112		85	SPK: -150
13127-88-3	Phenol-d6	120.843		15-107		81	SPK: -150
4165-60-0	Nitrobenzene-d5	90.639		18-107		91	SPK: -100
321-60-8	2-Fluorobiphenyl	84.528		20-109		85	SPK: -100

Report of Analysis

Client:		Date Collected:	10/17/2024 12:00:00 AM
Project:		Date Received:	10/17/2024 12:00:00 AM
Client Sample ID:	PB164216BL	SDG No.:	BF101924
Lab Sample ID:	PB164216BL	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	30.01 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139869.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	137.52		10-116		92	SPK: -150
1718-51-0	Terphenyl-d14	79.195		10-105		79	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	200094	6.887				
1146-65-2	Naphthalene-d8	794942	8.169				
15067-26-2	Acenaphthene-d10	453346	9.928				
1517-22-2	Phenanthrene-d10	848578	11.41				
1719-03-5	Chrysene-d12	603034	14.051				
1520-96-3	Perylene-d12	446604	15.527				
TENTATIVE IDENTIFIED COMPOUNDS							
000994-05-8	Butane, 2-methoxy-2-methyl-	160	J			2.257	
000123-42-2	2-Pentanone, 4-hydroxy-4-methyl-	240	A			5.14	
000541-02-6	Cyclopentasiloxane, decamethyl-	210	J			7.645	

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	WALL	SDG No.:	BF101924
Lab Sample ID:	P4426-37	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.05	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139870.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	52	U	52	160	200	ug/Kg
110-86-1	Pyridine	47.9	U	47.9	77.9	100	ug/Kg
100-52-7	Benzaldehyde	110	U	110	160	200	ug/Kg
62-53-3	Aniline	58.1	U	58.1	77.9	100	ug/Kg
108-95-2	Phenol	49.7	U	49.7	77.9	100	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	50.2	U	50.2	77.9	100	ug/Kg
95-57-8	2-Chlorophenol	50	U	50	77.9	100	ug/Kg
95-50-1	1,2-Dichlorobenzene	50.7	U	50.7	77.9	100	ug/Kg
541-73-1	1,3-Dichlorobenzene	51.5	U	51.5	77.9	100	ug/Kg
106-46-7	1,4-Dichlorobenzene	51.8	U	51.8	77.9	100	ug/Kg
100-51-6	Benzyl Alcohol	49.8	U	49.8	160	200	ug/Kg
95-48-7	2-Methylphenol	48.3	U	48.3	77.9	100	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	54.5	U	54.5	77.9	100	ug/Kg
98-86-2	Acetophenone	52.1	U	52.1	77.9	100	ug/Kg
65794-96-9	3+4-Methylphenols	47.8	U	47.8	160	200	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	24.2	U	24.2	48	48	ug/Kg
67-72-1	Hexachloroethane	49.8	U	49.8	77.9	100	ug/Kg
98-95-3	Nitrobenzene	54.4	U	54.4	77.9	100	ug/Kg
78-59-1	Isophorone	50.7	U	50.7	77.9	100	ug/Kg
88-75-5	2-Nitrophenol	56.6	U	56.6	77.9	100	ug/Kg
105-67-9	2,4-Dimethylphenol	55.9	U	55.9	77.9	100	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	51.4	U	51.4	77.9	100	ug/Kg
120-83-2	2,4-Dichlorophenol	45.3	U	45.3	77.9	100	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	51.2	U	51.2	77.9	100	ug/Kg
65-85-0	Benzoic acid	140	U	140	160	200	ug/Kg
91-20-3	Naphthalene	49.5	U	49.5	77.9	100	ug/Kg
106-47-8	4-Chloroaniline	49.5	U	49.5	77.9	100	ug/Kg
87-68-3	Hexachlorobutadiene	49.9	U	49.9	77.9	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	WALL	SDG No.:	BF101924
Lab Sample ID:	P4426-37	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.05	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139870.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	52	U	52	160	200	ug/Kg
59-50-7	4-Chloro-3-methylphenol	46.5	U	46.5	77.9	100	ug/Kg
91-57-6	2-Methylnaphthalene	49.5	U	49.5	77.9	100	ug/Kg
77-47-4	Hexachlorocyclopentadiene	93.5	U	93.5	160	200	ug/Kg
88-06-2	2,4,6-Trichlorophenol	42.8	U	42.8	77.9	100	ug/Kg
95-95-4	2,4,5-Trichlorophenol	44.4	U	44.4	77.9	100	ug/Kg
92-52-4	1,1-Biphenyl	52.4	U	52.4	77.9	100	ug/Kg
91-58-7	2-Chloronaphthalene	49.9	U	49.9	77.9	100	ug/Kg
88-74-4	2-Nitroaniline	56.9	U	56.9	77.9	100	ug/Kg
131-11-3	Dimethylphthalate	49	U	49	77.9	100	ug/Kg
208-96-8	Acenaphthylene	51.8	U	51.8	77.9	100	ug/Kg
606-20-2	2,6-Dinitrotoluene	49.9	U	49.9	77.9	100	ug/Kg
99-09-2	3-Nitroaniline	53.5	U	53.5	77.9	100	ug/Kg
83-32-9	Acenaphthene	48.6	U	48.6	77.9	100	ug/Kg
Total PAH	Total PAH	794.2	U	794.2	77.9	1600	ug/Kg
51-28-5	2,4-Dinitrophenol	150	U	150	160	200	ug/Kg
100-02-7	4-Nitrophenol	69.5	U	69.5	160	200	ug/Kg
132-64-9	Dibenzofuran	50.6	U	50.6	77.9	100	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	101.6	U	101.6	77.9	200	ug/Kg
121-14-2	2,4-Dinitrotoluene	51.7	U	51.7	77.9	100	ug/Kg
84-66-2	Diethylphthalate	48	U	48	77.9	100	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	51.3	U	51.3	77.9	100	ug/Kg
86-73-7	Fluorene	51.2	U	51.2	77.9	100	ug/Kg
100-01-6	4-Nitroaniline	64.1	U	64.1	77.9	100	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	70.1	U	70.1	160	200	ug/Kg
86-30-6	n-Nitrosodiphenylamine	48.9	U	48.9	77.9	100	ug/Kg
103-33-3	Azobenzene	47.7	U	47.7	77.9	100	ug/Kg
101-55-3	4-Bromophenyl-phenylether	47.3	U	47.3	77.9	100	ug/Kg
118-74-1	Hexachlorobenzene	50.9	U	50.9	77.9	100	ug/Kg
1912-24-9	Atrazine	54.8	U	54.8	77.9	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	WALL	SDG No.:	BF101924
Lab Sample ID:	P4426-37	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.05	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139870.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	46.3	U	46.3	160	200	ug/Kg
85-01-8	Phenanthrene	50.3	U	50.3	77.9	100	ug/Kg
120-12-7	Anthracene	50.6	U	50.6	77.9	100	ug/Kg
86-74-8	Carbazole	48.1	U	48.1	77.9	100	ug/Kg
84-74-2	Di-n-butylphthalate	50.5	U	50.5	77.9	100	ug/Kg
206-44-0	Fluoranthene	49	U	49	77.9	100	ug/Kg
92-87-5	Benzidine	97.7	U	97.7	160	200	ug/Kg
129-00-0	Pyrene	49.8	U	49.8	77.9	100	ug/Kg
85-68-7	Butylbenzylphthalate	58	U	58	77.9	100	ug/Kg
91-94-1	3,3-Dichlorobenzidine	59.1	U	59.1	160	200	ug/Kg
56-55-3	Benzo(a)anthracene	48.4	U	48.4	77.9	100	ug/Kg
218-01-9	Chrysene	47.7	U	47.7	77.9	100	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	54.5	U	54.5	77.9	100	ug/Kg
117-84-0	Di-n-octyl phthalate	65.9	U	65.9	160	200	ug/Kg
205-99-2	Benzo(b)fluoranthene	48.6	U	48.6	77.9	100	ug/Kg
207-08-9	Benzo(k)fluoranthene	49.5	U	49.5	77.9	100	ug/Kg
50-32-8	Benzo(a)pyrene	55.7	U	55.7	77.9	100	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	46.8	U	46.8	77.9	100	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	48.7	U	48.7	77.9	100	ug/Kg
191-24-2	Benzo(g,h,i)perylene	48	U	48	77.9	100	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	52	U	52	77.9	100	ug/Kg
123-91-1	1,4-Dioxane	65.9	U	65.9	77.9	100	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	44.8	U	44.8	77.9	100	ug/Kg
90-12-0	1-Methylnaphthalene	49.9	U	49.9	100	100	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	42.187		18-112		28	SPK: -150
13127-88-3	Phenol-d6	46.11		15-107		31	SPK: -150
4165-60-0	Nitrobenzene-d5	41.433		18-107		41	SPK: -100
321-60-8	2-Fluorobiphenyl	45.257		20-109		45	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	WALL	SDG No.:	BF101924
Lab Sample ID:	P4426-37	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.05	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139870.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	37.446		10-116		25	SPK: -150
1718-51-0	Terphenyl-d14	40.085		10-105		40	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	190163	6.887				
1146-65-2	Naphthalene-d8	745893	8.169				
15067-26-2	Acenaphthene-d10	402060	9.922				
1517-22-2	Phenanthrene-d10	639889	11.41				
1719-03-5	Chrysene-d12	370008	14.051				
1520-96-3	Perylene-d12	387887	15.527				



Client:			Date Collected:	10/16/2024 12:00:00 AM	
Project:			Date Received:	10/16/2024 12:00:00 AM	
Client Sample ID:	CURBS-AND-OTHER		SDG No.:	BF101924	
Lab Sample ID:	P4426-39		Matrix:	Solid	
Analytical Method:	8270E		% Moisture:	0	
Sample Wt/Vol:	50.02	Units: g	Final Vol:	1000	uL
Soil Aliquot Vol:	uL		Test:		
Extraction Type :	Decanted :		Level :	LOW	
Injection Volume :	GPC Factor :		GPC Cleanup :	PH :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139871.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	52.1	U	52.1	160	200	ug/Kg
110-86-1	Pyridine	47.9	U	47.9	78	100	ug/Kg
100-52-7	Benzaldehyde	110	U	110	160	200	ug/Kg
62-53-3	Aniline	58.1	U	58.1	78	100	ug/Kg
108-95-2	Phenol	49.7	U	49.7	78	100	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	50.2	U	50.2	78	100	ug/Kg
95-57-8	2-Chlorophenol	50.1	U	50.1	78	100	ug/Kg
95-50-1	1,2-Dichlorobenzene	50.7	U	50.7	78	100	ug/Kg
541-73-1	1,3-Dichlorobenzene	51.6	U	51.6	78	100	ug/Kg
106-46-7	1,4-Dichlorobenzene	51.9	U	51.9	78	100	ug/Kg
100-51-6	Benzyl Alcohol	49.8	U	49.8	160	200	ug/Kg
95-48-7	2-Methylphenol	48.3	U	48.3	78	100	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	54.5	U	54.5	78	100	ug/Kg
98-86-2	Acetophenone	52.1	U	52.1	78	100	ug/Kg
65794-96-9	3+4-Methylphenols	47.9	U	47.9	160	200	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	24.2	U	24.2	48	48	ug/Kg
67-72-1	Hexachloroethane	49.8	U	49.8	78	100	ug/Kg
98-95-3	Nitrobenzene	54.5	U	54.5	78	100	ug/Kg
78-59-1	Isophorone	50.7	U	50.7	78	100	ug/Kg
88-75-5	2-Nitrophenol	56.7	U	56.7	78	100	ug/Kg
105-67-9	2,4-Dimethylphenol	55.9	U	55.9	78	100	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	51.5	U	51.5	78	100	ug/Kg
120-83-2	2,4-Dichlorophenol	45.3	U	45.3	78	100	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	51.2	U	51.2	78	100	ug/Kg
65-85-0	Benzoic acid	140	U	140	160	200	ug/Kg
91-20-3	Naphthalene	49.5	U	49.5	78	100	ug/Kg
106-47-8	4-Chloroaniline	49.5	U	49.5	78	100	ug/Kg
87-68-3	Hexachlorobutadiene	50	U	50	78	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	CURBS-AND-OTHER	SDG No.:	BF101924
Lab Sample ID:	P4426-39	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.02	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139871.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	52.1	U	52.1	160	200	ug/Kg
59-50-7	4-Chloro-3-methylphenol	46.5	U	46.5	78	100	ug/Kg
91-57-6	2-Methylnaphthalene	49.5	U	49.5	78	100	ug/Kg
77-47-4	Hexachlorocyclopentadiene	93.6	U	93.6	160	200	ug/Kg
88-06-2	2,4,6-Trichlorophenol	42.8	U	42.8	78	100	ug/Kg
95-95-4	2,4,5-Trichlorophenol	44.4	U	44.4	78	100	ug/Kg
92-52-4	1,1-Biphenyl	52.4	U	52.4	78	100	ug/Kg
91-58-7	2-Chloronaphthalene	50	U	50	78	100	ug/Kg
88-74-4	2-Nitroaniline	57	U	57	78	100	ug/Kg
131-11-3	Dimethylphthalate	49	U	49	78	100	ug/Kg
208-96-8	Acenaphthylene	51.9	U	51.9	78	100	ug/Kg
606-20-2	2,6-Dinitrotoluene	49.9	U	49.9	78	100	ug/Kg
99-09-2	3-Nitroaniline	53.5	U	53.5	78	100	ug/Kg
83-32-9	Acenaphthene	48.6	U	48.6	78	100	ug/Kg
Total PAH	Total PAH	794.6	U	794.6	78	1600	ug/Kg
51-28-5	2,4-Dinitrophenol	150	U	150	160	200	ug/Kg
100-02-7	4-Nitrophenol	69.6	U	69.6	160	200	ug/Kg
132-64-9	Dibenzofuran	50.6	U	50.6	78	100	ug/Kg
121-14-2	2,4-Dinitrotoluene	51.7	U	51.7	78	100	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	101.6	U	101.6	78	200	ug/Kg
84-66-2	Diethylphthalate	48	U	48	78	100	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	51.3	U	51.3	78	100	ug/Kg
86-73-7	Fluorene	51.3	U	51.3	78	100	ug/Kg
100-01-6	4-Nitroaniline	64.2	U	64.2	78	100	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	70.2	U	70.2	160	200	ug/Kg
86-30-6	n-Nitrosodiphenylamine	48.9	U	48.9	78	100	ug/Kg
103-33-3	Azobenzene	47.7	U	47.7	78	100	ug/Kg
101-55-3	4-Bromophenyl-phenylether	47.3	U	47.3	78	100	ug/Kg
118-74-1	Hexachlorobenzene	51	U	51	78	100	ug/Kg
1912-24-9	Atrazine	54.8	U	54.8	78	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	CURBS-AND-OTHER	SDG No.:	BF101924
Lab Sample ID:	P4426-39	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.02	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139871.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	46.4	U	46.4	160	200	ug/Kg
85-01-8	Phenanthrene	50.4	U	50.4	78	100	ug/Kg
120-12-7	Anthracene	50.6	U	50.6	78	100	ug/Kg
86-74-8	Carbazole	48.2	U	48.2	78	100	ug/Kg
84-74-2	Di-n-butylphthalate	50.6	U	50.6	78	100	ug/Kg
206-44-0	Fluoranthene	49	U	49	78	100	ug/Kg
92-87-5	Benzidine	97.8	U	97.8	160	200	ug/Kg
129-00-0	Pyrene	49.8	U	49.8	78	100	ug/Kg
85-68-7	Butylbenzylphthalate	58.1	U	58.1	78	100	ug/Kg
91-94-1	3,3-Dichlorobenzidine	59.1	U	59.1	160	200	ug/Kg
56-55-3	Benzo(a)anthracene	48.4	U	48.4	78	100	ug/Kg
218-01-9	Chrysene	47.7	U	47.7	78	100	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	54.6	U	54.6	78	100	ug/Kg
117-84-0	Di-n-octyl phthalate	66	U	66	160	200	ug/Kg
205-99-2	Benzo(b)fluoranthene	48.6	U	48.6	78	100	ug/Kg
207-08-9	Benzo(k)fluoranthene	49.5	U	49.5	78	100	ug/Kg
50-32-8	Benzo(a)pyrene	55.8	U	55.8	78	100	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	46.8	U	46.8	78	100	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	48.7	U	48.7	78	100	ug/Kg
191-24-2	Benzo(g,h,i)perylene	48	U	48	78	100	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	52.1	U	52.1	78	100	ug/Kg
123-91-1	1,4-Dioxane	66	U	66	78	100	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	44.8	U	44.8	78	100	ug/Kg
90-12-0	1-Methylnaphthalene	49.9	U	49.9	100	100	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	9.538	*	18-112		6	SPK: -150
13127-88-3	Phenol-d6	20.073	*	15-107		13	SPK: -150
4165-60-0	Nitrobenzene-d5	33.846		18-107		34	SPK: -100
321-60-8	2-Fluorobiphenyl	39.731		20-109		40	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	CURBS-AND-OTHER	SDG No.:	BF101924
Lab Sample ID:	P4426-39	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.02 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139871.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	5.554	*	10-116		4	SPK: -150
1718-51-0	Terphenyl-d14	40.308		10-105		40	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	195900	6.887				
1146-65-2	Naphthalene-d8	753548	8.169				
15067-26-2	Acenaphthene-d10	418416	9.922				
1517-22-2	Phenanthrene-d10	677607	11.41				
1719-03-5	Chrysene-d12	365941	14.045				
1520-96-3	Perylene-d12	394368	15.527				

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TRANSFORMER-PAD	SDG No.:	BF101924
Lab Sample ID:	P4426-35	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.06	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139872.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	52	U	52	160	200	ug/Kg
110-86-1	Pyridine	47.9	U	47.9	77.9	100	ug/Kg
100-52-7	Benzaldehyde	110	U	110	160	200	ug/Kg
62-53-3	Aniline	58.1	U	58.1	77.9	100	ug/Kg
108-95-2	Phenol	49.7	U	49.7	77.9	100	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	50.2	U	50.2	77.9	100	ug/Kg
95-57-8	2-Chlorophenol	50	U	50	77.9	100	ug/Kg
95-50-1	1,2-Dichlorobenzene	50.7	U	50.7	77.9	100	ug/Kg
541-73-1	1,3-Dichlorobenzene	51.5	U	51.5	77.9	100	ug/Kg
106-46-7	1,4-Dichlorobenzene	51.8	U	51.8	77.9	100	ug/Kg
100-51-6	Benzyl Alcohol	49.7	U	49.7	160	200	ug/Kg
95-48-7	2-Methylphenol	48.3	U	48.3	77.9	100	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	54.5	U	54.5	77.9	100	ug/Kg
98-86-2	Acetophenone	52.1	U	52.1	77.9	100	ug/Kg
65794-96-9	3+4-Methylphenols	47.8	U	47.8	160	200	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	24.2	U	24.2	47.9	47.9	ug/Kg
67-72-1	Hexachloroethane	49.7	U	49.7	77.9	100	ug/Kg
98-95-3	Nitrobenzene	54.4	U	54.4	77.9	100	ug/Kg
78-59-1	Isophorone	50.7	U	50.7	77.9	100	ug/Kg
88-75-5	2-Nitrophenol	56.6	U	56.6	77.9	100	ug/Kg
105-67-9	2,4-Dimethylphenol	55.9	U	55.9	77.9	100	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	51.4	U	51.4	77.9	100	ug/Kg
120-83-2	2,4-Dichlorophenol	45.2	U	45.2	77.9	100	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	51.2	U	51.2	77.9	100	ug/Kg
65-85-0	Benzoic acid	140	U	140	160	200	ug/Kg
91-20-3	Naphthalene	49.5	U	49.5	77.9	100	ug/Kg
106-47-8	4-Chloroaniline	49.5	U	49.5	77.9	100	ug/Kg
87-68-3	Hexachlorobutadiene	49.9	U	49.9	77.9	100	ug/Kg



Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TRANSFORMER-PAD	SDG No.:	BF101924
Lab Sample ID:	P4426-35	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.06	Units:	g
Soil Aliquot Vol:			
		Final Vol:	1000 uL
		Test:	
Extraction Type :		Decanted :	
		Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139872.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	52	U	52	160	200	ug/Kg
59-50-7	4-Chloro-3-methylphenol	46.4	U	46.4	77.9	100	ug/Kg
91-57-6	2-Methylnaphthalene	49.4	U	49.4	77.9	100	ug/Kg
77-47-4	Hexachlorocyclopentadiene	93.5	U	93.5	160	200	ug/Kg
88-06-2	2,4,6-Trichlorophenol	42.8	U	42.8	77.9	100	ug/Kg
95-95-4	2,4,5-Trichlorophenol	44.3	U	44.3	77.9	100	ug/Kg
92-52-4	1,1-Biphenyl	52.4	U	52.4	77.9	100	ug/Kg
91-58-7	2-Chloronaphthalene	49.9	U	49.9	77.9	100	ug/Kg
88-74-4	2-Nitroaniline	56.9	U	56.9	77.9	100	ug/Kg
131-11-3	Dimethylphthalate	49	U	49	77.9	100	ug/Kg
208-96-8	Acenaphthylene	51.8	U	51.8	77.9	100	ug/Kg
606-20-2	2,6-Dinitrotoluene	49.9	U	49.9	77.9	100	ug/Kg
99-09-2	3-Nitroaniline	53.5	U	53.5	77.9	100	ug/Kg
83-32-9	Acenaphthene	48.6	U	48.6	77.9	100	ug/Kg
Total PAH	Total PAH	794	U	794	77.9	1600	ug/Kg
51-28-5	2,4-Dinitrophenol	150	U	150	160	200	ug/Kg
100-02-7	4-Nitrophenol	69.5	U	69.5	160	200	ug/Kg
132-64-9	Dibenzofuran	50.6	U	50.6	77.9	100	ug/Kg
121-14-2	2,4-Dinitrotoluene	51.7	U	51.7	77.9	100	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	101.6	U	101.6	77.9	200	ug/Kg
84-66-2	Diethylphthalate	48	U	48	77.9	100	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	51.3	U	51.3	77.9	100	ug/Kg
86-73-7	Fluorene	51.2	U	51.2	77.9	100	ug/Kg
100-01-6	4-Nitroaniline	64.1	U	64.1	77.9	100	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	70.1	U	70.1	160	200	ug/Kg
86-30-6	n-Nitrosodiphenylamine	48.9	U	48.9	77.9	100	ug/Kg
103-33-3	Azobenzene	47.7	U	47.7	77.9	100	ug/Kg
101-55-3	4-Bromophenyl-phenylether	47.3	U	47.3	77.9	100	ug/Kg
118-74-1	Hexachlorobenzene	50.9	U	50.9	77.9	100	ug/Kg
1912-24-9	Atrazine	54.8	U	54.8	77.9	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TRANSFORMER-PAD	SDG No.:	BF101924
Lab Sample ID:	P4426-35	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.06	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139872.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	46.3	U	46.3	160	200	ug/Kg
85-01-8	Phenanthrene	50.3	U	50.3	77.9	100	ug/Kg
120-12-7	Anthracene	50.6	U	50.6	77.9	100	ug/Kg
86-74-8	Carbazole	48.1	U	48.1	77.9	100	ug/Kg
84-74-2	Di-n-butylphthalate	50.5	U	50.5	77.9	100	ug/Kg
206-44-0	Fluoranthene	49	U	49	77.9	100	ug/Kg
92-87-5	Benzidine	97.7	U	97.7	160	200	ug/Kg
129-00-0	Pyrene	49.7	U	49.7	77.9	100	ug/Kg
85-68-7	Butylbenzylphthalate	58	U	58	77.9	100	ug/Kg
91-94-1	3,3-Dichlorobenzidine	59.1	U	59.1	160	200	ug/Kg
56-55-3	Benzo(a)anthracene	48.4	U	48.4	77.9	100	ug/Kg
218-01-9	Chrysene	47.6	U	47.6	77.9	100	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	54.5	U	54.5	77.9	100	ug/Kg
117-84-0	Di-n-octyl phthalate	65.9	U	65.9	160	200	ug/Kg
205-99-2	Benzo(b)fluoranthene	48.6	U	48.6	77.9	100	ug/Kg
207-08-9	Benzo(k)fluoranthene	49.5	U	49.5	77.9	100	ug/Kg
50-32-8	Benzo(a)pyrene	55.7	U	55.7	77.9	100	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	46.8	U	46.8	77.9	100	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	48.7	U	48.7	77.9	100	ug/Kg
191-24-2	Benzo(g,h,i)perylene	48	U	48	77.9	100	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	52	U	52	77.9	100	ug/Kg
123-91-1	1,4-Dioxane	65.9	U	65.9	77.9	100	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	44.8	U	44.8	77.9	100	ug/Kg
90-12-0	1-Methylnaphthalene	49.9	U	49.9	100	100	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	24.024	*	18-112		16	SPK: -150
13127-88-3	Phenol-d6	32.175		15-107		21	SPK: -150
4165-60-0	Nitrobenzene-d5	38.91		18-107		39	SPK: -100
321-60-8	2-Fluorobiphenyl	47.747		20-109		48	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TRANSFORMER-PAD	SDG No.:	BF101924
Lab Sample ID:	P4426-35	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.06 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139872.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	12.924	*	10-116		9	SPK: -150
1718-51-0	Terphenyl-d14	25.809		10-105		26	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	171810	6.887				
1146-65-2	Naphthalene-d8	614846	8.169				
15067-26-2	Acenaphthene-d10	262055	9.928				
1517-22-2	Phenanthrene-d10	226931	11.445				
1719-03-5	Chrysene-d12	333883	14.074				
1520-96-3	Perylene-d12	363697	15.533				

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-11	SDG No.:	BF101924
Lab Sample ID:	P4426-21	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.03	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139873.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	52	U	52	160	200	ug/Kg
110-86-1	Pyridine	47.9	U	47.9	78	100	ug/Kg
100-52-7	Benzaldehyde	110	U	110	160	200	ug/Kg
62-53-3	Aniline	58.1	U	58.1	78	100	ug/Kg
108-95-2	Phenol	49.7	U	49.7	78	100	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	50.2	U	50.2	78	100	ug/Kg
95-57-8	2-Chlorophenol	50.1	U	50.1	78	100	ug/Kg
95-50-1	1,2-Dichlorobenzene	50.7	U	50.7	78	100	ug/Kg
541-73-1	1,3-Dichlorobenzene	51.6	U	51.6	78	100	ug/Kg
106-46-7	1,4-Dichlorobenzene	51.9	U	51.9	78	100	ug/Kg
100-51-6	Benzyl Alcohol	49.8	U	49.8	160	200	ug/Kg
95-48-7	2-Methylphenol	48.3	U	48.3	78	100	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	54.5	U	54.5	78	100	ug/Kg
98-86-2	Acetophenone	52.1	U	52.1	78	100	ug/Kg
65794-96-9	3+4-Methylphenols	47.9	U	47.9	160	200	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	24.2	U	24.2	48	48	ug/Kg
67-72-1	Hexachloroethane	49.8	U	49.8	78	100	ug/Kg
98-95-3	Nitrobenzene	54.4	U	54.4	78	100	ug/Kg
78-59-1	Isophorone	50.7	U	50.7	78	100	ug/Kg
88-75-5	2-Nitrophenol	56.7	U	56.7	78	100	ug/Kg
105-67-9	2,4-Dimethylphenol	55.9	U	55.9	78	100	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	51.4	U	51.4	78	100	ug/Kg
120-83-2	2,4-Dichlorophenol	45.3	U	45.3	78	100	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	51.2	U	51.2	78	100	ug/Kg
65-85-0	Benzoic acid	140	U	140	160	200	ug/Kg
91-20-3	Naphthalene	49.5	U	49.5	78	100	ug/Kg
106-47-8	4-Chloroaniline	49.5	U	49.5	78	100	ug/Kg
87-68-3	Hexachlorobutadiene	50	U	50	78	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-11	SDG No.:	BF101924
Lab Sample ID:	P4426-21	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.03	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139873.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	52	U	52	160	200	ug/Kg
59-50-7	4-Chloro-3-methylphenol	46.5	U	46.5	78	100	ug/Kg
91-57-6	2-Methylnaphthalene	49.5	U	49.5	78	100	ug/Kg
77-47-4	Hexachlorocyclopentadiene	93.5	U	93.5	160	200	ug/Kg
88-06-2	2,4,6-Trichlorophenol	42.8	U	42.8	78	100	ug/Kg
95-95-4	2,4,5-Trichlorophenol	44.4	U	44.4	78	100	ug/Kg
92-52-4	1,1-Biphenyl	52.4	U	52.4	78	100	ug/Kg
91-58-7	2-Chloronaphthalene	50	U	50	78	100	ug/Kg
88-74-4	2-Nitroaniline	57	U	57	78	100	ug/Kg
131-11-3	Dimethylphthalate	49	U	49	78	100	ug/Kg
208-96-8	Acenaphthylene	51.9	U	51.9	78	100	ug/Kg
606-20-2	2,6-Dinitrotoluene	49.9	U	49.9	78	100	ug/Kg
99-09-2	3-Nitroaniline	53.5	U	53.5	78	100	ug/Kg
83-32-9	Acenaphthene	48.6	U	48.6	78	100	ug/Kg
Total PAH	Total PAH	794.6	U	794.6	78	1600	ug/Kg
51-28-5	2,4-Dinitrophenol	150	U	150	160	200	ug/Kg
100-02-7	4-Nitrophenol	69.6	U	69.6	160	200	ug/Kg
132-64-9	Dibenzofuran	50.6	U	50.6	78	100	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	101.6	U	101.6	78	200	ug/Kg
121-14-2	2,4-Dinitrotoluene	51.7	U	51.7	78	100	ug/Kg
84-66-2	Diethylphthalate	48	U	48	78	100	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	51.3	U	51.3	78	100	ug/Kg
86-73-7	Fluorene	51.3	U	51.3	78	100	ug/Kg
100-01-6	4-Nitroaniline	64.2	U	64.2	78	100	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	70.2	U	70.2	160	200	ug/Kg
86-30-6	n-Nitrosodiphenylamine	48.9	U	48.9	78	100	ug/Kg
103-33-3	Azobenzene	47.7	U	47.7	78	100	ug/Kg
101-55-3	4-Bromophenyl-phenylether	47.3	U	47.3	78	100	ug/Kg
118-74-1	Hexachlorobenzene	51	U	51	78	100	ug/Kg
1912-24-9	Atrazine	54.8	U	54.8	78	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-11	SDG No.:	BF101924
Lab Sample ID:	P4426-21	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.03	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139873.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	46.4	U	46.4	160	200	ug/Kg
85-01-8	Phenanthrene	50.4	U	50.4	78	100	ug/Kg
120-12-7	Anthracene	50.6	U	50.6	78	100	ug/Kg
86-74-8	Carbazole	48.2	U	48.2	78	100	ug/Kg
84-74-2	Di-n-butylphthalate	50.5	U	50.5	78	100	ug/Kg
206-44-0	Fluoranthene	49	U	49	78	100	ug/Kg
92-87-5	Benzidine	97.7	U	97.7	160	200	ug/Kg
129-00-0	Pyrene	49.8	U	49.8	78	100	ug/Kg
85-68-7	Butylbenzylphthalate	58	U	58	78	100	ug/Kg
91-94-1	3,3-Dichlorobenzidine	59.1	U	59.1	160	200	ug/Kg
56-55-3	Benzo(a)anthracene	48.4	U	48.4	78	100	ug/Kg
218-01-9	Chrysene	47.7	U	47.7	78	100	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	54.6	U	54.6	78	100	ug/Kg
117-84-0	Di-n-octyl phthalate	66	U	66	160	200	ug/Kg
205-99-2	Benzo(b)fluoranthene	48.6	U	48.6	78	100	ug/Kg
207-08-9	Benzo(k)fluoranthene	49.5	U	49.5	78	100	ug/Kg
50-32-8	Benzo(a)pyrene	55.8	U	55.8	78	100	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	46.8	U	46.8	78	100	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	48.7	U	48.7	78	100	ug/Kg
191-24-2	Benzo(g,h,i)perylene	48	U	48	78	100	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	52	U	52	78	100	ug/Kg
123-91-1	1,4-Dioxane	66	U	66	78	100	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	44.8	U	44.8	78	100	ug/Kg
90-12-0	1-Methylnaphthalene	49.9	U	49.9	100	100	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	27.137		18-112		18	SPK: -150
13127-88-3	Phenol-d6	41.017		15-107		27	SPK: -150
4165-60-0	Nitrobenzene-d5	47.193		18-107		47	SPK: -100
321-60-8	2-Fluorobiphenyl	52.411		20-109		52	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-11	SDG No.:	BF101924
Lab Sample ID:	P4426-21	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.03 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139873.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	16.923		10-116		11	SPK: -150
1718-51-0	Terphenyl-d14	46.859		10-105		47	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	196276	6.887				
1146-65-2	Naphthalene-d8	749941	8.169				
15067-26-2	Acenaphthene-d10	407911	9.922				
1517-22-2	Phenanthrene-d10	666291	11.41				
1719-03-5	Chrysene-d12	368761	14.045				
1520-96-3	Perylene-d12	386969	15.527				

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-7	SDG No.:	BF101924
Lab Sample ID:	P4426-13	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.08	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139874.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	52	U	52	160	200	ug/Kg
110-86-1	Pyridine	47.9	U	47.9	77.9	100	ug/Kg
100-52-7	Benzaldehyde	110	U	110	160	200	ug/Kg
62-53-3	Aniline	58	U	58	77.9	100	ug/Kg
108-95-2	Phenol	49.7	U	49.7	77.9	100	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	50.1	U	50.1	77.9	100	ug/Kg
95-57-8	2-Chlorophenol	50	U	50	77.9	100	ug/Kg
95-50-1	1,2-Dichlorobenzene	50.7	U	50.7	77.9	100	ug/Kg
541-73-1	1,3-Dichlorobenzene	51.5	U	51.5	77.9	100	ug/Kg
106-46-7	1,4-Dichlorobenzene	51.8	U	51.8	77.9	100	ug/Kg
100-51-6	Benzyl Alcohol	49.7	U	49.7	160	200	ug/Kg
95-48-7	2-Methylphenol	48.3	U	48.3	77.9	100	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	54.5	U	54.5	77.9	100	ug/Kg
98-86-2	Acetophenone	52.1	U	52.1	77.9	100	ug/Kg
65794-96-9	3+4-Methylphenols	47.8	U	47.8	160	200	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	24.1	U	24.1	47.9	47.9	ug/Kg
67-72-1	Hexachloroethane	49.7	U	49.7	77.9	100	ug/Kg
98-95-3	Nitrobenzene	54.4	U	54.4	77.9	100	ug/Kg
78-59-1	Isophorone	50.7	U	50.7	77.9	100	ug/Kg
88-75-5	2-Nitrophenol	56.6	U	56.6	77.9	100	ug/Kg
105-67-9	2,4-Dimethylphenol	55.8	U	55.8	77.9	100	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	51.4	U	51.4	77.9	100	ug/Kg
120-83-2	2,4-Dichlorophenol	45.2	U	45.2	77.9	100	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	51.2	U	51.2	77.9	100	ug/Kg
65-85-0	Benzoic acid	140	U	140	160	200	ug/Kg
91-20-3	Naphthalene	49.5	U	49.5	77.9	100	ug/Kg
106-47-8	4-Chloroaniline	49.5	U	49.5	77.9	100	ug/Kg
87-68-3	Hexachlorobutadiene	49.9	U	49.9	77.9	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-7	SDG No.:	BF101924
Lab Sample ID:	P4426-13	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.08	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139874.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	52	U	52	160	200	ug/Kg
59-50-7	4-Chloro-3-methylphenol	46.4	U	46.4	77.9	100	ug/Kg
91-57-6	2-Methylnaphthalene	49.4	U	49.4	77.9	100	ug/Kg
77-47-4	Hexachlorocyclopentadiene	93.5	U	93.5	160	200	ug/Kg
88-06-2	2,4,6-Trichlorophenol	42.8	U	42.8	77.9	100	ug/Kg
95-95-4	2,4,5-Trichlorophenol	44.3	U	44.3	77.9	100	ug/Kg
92-52-4	1,1-Biphenyl	52.4	U	52.4	77.9	100	ug/Kg
91-58-7	2-Chloronaphthalene	49.9	U	49.9	77.9	100	ug/Kg
88-74-4	2-Nitroaniline	56.9	U	56.9	77.9	100	ug/Kg
131-11-3	Dimethylphthalate	48.9	U	48.9	77.9	100	ug/Kg
208-96-8	Acenaphthylene	51.8	U	51.8	77.9	100	ug/Kg
606-20-2	2,6-Dinitrotoluene	49.8	U	49.8	77.9	100	ug/Kg
99-09-2	3-Nitroaniline	53.4	U	53.4	77.9	100	ug/Kg
83-32-9	Acenaphthene	48.6	U	48.6	77.9	100	ug/Kg
Total PAH	Total PAH	793.7	U	793.7	77.9	1600	ug/Kg
51-28-5	2,4-Dinitrophenol	150	U	150	160	200	ug/Kg
100-02-7	4-Nitrophenol	69.5	U	69.5	160	200	ug/Kg
132-64-9	Dibenzofuran	50.6	U	50.6	77.9	100	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	101.4	U	101.4	77.9	200	ug/Kg
121-14-2	2,4-Dinitrotoluene	51.6	U	51.6	77.9	100	ug/Kg
84-66-2	Diethylphthalate	48	U	48	77.9	100	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	51.3	U	51.3	77.9	100	ug/Kg
86-73-7	Fluorene	51.2	U	51.2	77.9	100	ug/Kg
100-01-6	4-Nitroaniline	64.1	U	64.1	77.9	100	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	70.1	U	70.1	160	200	ug/Kg
86-30-6	n-Nitrosodiphenylamine	48.9	U	48.9	77.9	100	ug/Kg
103-33-3	Azobenzene	47.7	U	47.7	77.9	100	ug/Kg
101-55-3	4-Bromophenyl-phenylether	47.3	U	47.3	77.9	100	ug/Kg
118-74-1	Hexachlorobenzene	50.9	U	50.9	77.9	100	ug/Kg
1912-24-9	Atrazine	54.8	U	54.8	77.9	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-7	SDG No.:	BF101924
Lab Sample ID:	P4426-13	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.08	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139874.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	46.3	U	46.3	160	200	ug/Kg
85-01-8	Phenanthrene	50.3	U	50.3	77.9	100	ug/Kg
120-12-7	Anthracene	50.6	U	50.6	77.9	100	ug/Kg
86-74-8	Carbazole	48.1	U	48.1	77.9	100	ug/Kg
84-74-2	Di-n-butylphthalate	50.5	U	50.5	77.9	100	ug/Kg
206-44-0	Fluoranthene	48.9	U	48.9	77.9	100	ug/Kg
92-87-5	Benzidine	97.6	U	97.6	160	200	ug/Kg
129-00-0	Pyrene	49.7	U	49.7	77.9	100	ug/Kg
85-68-7	Butylbenzylphthalate	58	U	58	77.9	100	ug/Kg
91-94-1	3,3-Dichlorobenzidine	59.1	U	59.1	160	200	ug/Kg
56-55-3	Benzo(a)anthracene	48.3	U	48.3	77.9	100	ug/Kg
218-01-9	Chrysene	47.6	U	47.6	77.9	100	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	54.5	U	54.5	77.9	100	ug/Kg
117-84-0	Di-n-octyl phthalate	65.9	U	65.9	160	200	ug/Kg
205-99-2	Benzo(b)fluoranthene	48.6	U	48.6	77.9	100	ug/Kg
207-08-9	Benzo(k)fluoranthene	49.5	U	49.5	77.9	100	ug/Kg
50-32-8	Benzo(a)pyrene	55.7	U	55.7	77.9	100	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	46.8	U	46.8	77.9	100	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	48.6	U	48.6	77.9	100	ug/Kg
191-24-2	Benzo(g,h,i)perylene	48	U	48	77.9	100	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	52	U	52	77.9	100	ug/Kg
123-91-1	1,4-Dioxane	65.9	U	65.9	77.9	100	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	44.7	U	44.7	77.9	100	ug/Kg
90-12-0	1-Methylnaphthalene	49.8	U	49.8	100	100	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	20.389	*	18-112		14	SPK: -150
13127-88-3	Phenol-d6	33.937		15-107		23	SPK: -150
4165-60-0	Nitrobenzene-d5	33.18		18-107		33	SPK: -100
321-60-8	2-Fluorobiphenyl	37.035		20-109		37	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-7	SDG No.:	BF101924
Lab Sample ID:	P4426-13	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.08	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139874.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	9.356	*	10-116		6	SPK: -150
1718-51-0	Terphenyl-d14	32.641		10-105		33	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	193840	6.887				
1146-65-2	Naphthalene-d8	750911	8.169				
15067-26-2	Acenaphthene-d10	409213	9.928				
1517-22-2	Phenanthrene-d10	649825	11.41				
1719-03-5	Chrysene-d12	355728	14.051				
1520-96-3	Perylene-d12	376625	15.527				

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-1	SDG No.:	BF101924
Lab Sample ID:	P4425-02	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139875.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	10.3	U	10.3	80	100	ug/L
110-86-1	Pyridine	15.5	U	15.5	40	50	ug/L
100-52-7	Benzaldehyde	40	U	40	80	100	ug/L
62-53-3	Aniline	16	U	16	40	50	ug/L
108-95-2	Phenol	9.3	U	9.3	40	50	ug/L
111-44-4	bis(2-Chloroethyl)ether	11.9	U	11.9	40	50	ug/L
95-57-8	2-Chlorophenol	7.1	U	7.1	40	50	ug/L
95-50-1	1,2-Dichlorobenzene	8.8	U	8.8	40	50	ug/L
541-73-1	1,3-Dichlorobenzene	8	U	8	40	50	ug/L
106-46-7	1,4-Dichlorobenzene	8.4	U	8.4	40	50	ug/L
100-51-6	Benzyl Alcohol	16	U	16	80	100	ug/L
95-48-7	2-Methylphenol	11.3	U	11.3	40	50	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	13.5	U	13.5	40	50	ug/L
98-86-2	Acetophenone	11	U	11	40	50	ug/L
65794-96-9	3+4-Methylphenols	11.5	U	11.5	80	100	ug/L
621-64-7	n-Nitroso-di-n-propylamine	14.8	U	14.8	25	25	ug/L
67-72-1	Hexachloroethane	10.1	U	10.1	40	50	ug/L
98-95-3	Nitrobenzene	12.7	U	12.7	40	50	ug/L
78-59-1	Isophorone	11.4	U	11.4	40	50	ug/L
88-75-5	2-Nitrophenol	19.6	U	19.6	40	50	ug/L
105-67-9	2,4-Dimethylphenol	15.1	U	15.1	40	50	ug/L
111-91-1	bis(2-Chloroethoxy)methane	10.2	U	10.2	40	50	ug/L
120-83-2	2,4-Dichlorophenol	8.8	U	8.8	40	50	ug/L
120-82-1	1,2,4-Trichlorobenzene	11	U	11	40	50	ug/L
65-85-0	Benzoic acid	54.7	U	54.7	80	100	ug/L
91-20-3	Naphthalene	10.2	U	10.2	40	50	ug/L
106-47-8	4-Chloroaniline	13	U	13	40	50	ug/L
87-68-3	Hexachlorobutadiene	12.7	U	12.7	40	50	ug/L

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-1	SDG No.:	BF101924
Lab Sample ID:	P4425-02	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000
		Test:	uL
Extraction Type :		Decanted :	Level :
			LOW
Injection Volume :		GPC Factor :	GPC Cleanup :
			PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139875.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	16.5	U	16.5	80	100	ug/L
59-50-7	4-Chloro-3-methylphenol	8.4	U	8.4	40	50	ug/L
91-57-6	2-Methylnaphthalene	11.3	U	11.3	40	50	ug/L
77-47-4	Hexachlorocyclopentadiene	50.2	U	50.2	80	100	ug/L
88-06-2	2,4,6-Trichlorophenol	8.9	U	8.9	40	50	ug/L
95-95-4	2,4,5-Trichlorophenol	10.1	U	10.1	40	50	ug/L
92-52-4	1,1-Biphenyl	9.1	U	9.1	40	50	ug/L
91-58-7	2-Chloronaphthalene	9.7	U	9.7	40	50	ug/L
88-74-4	2-Nitroaniline	14.1	U	14.1	40	50	ug/L
131-11-3	Dimethylphthalate	32.6	J	9.3	40	50	ug/L
208-96-8	Acenaphthylene	10.4	U	10.4	40	50	ug/L
606-20-2	2,6-Dinitrotoluene	12.4	U	12.4	40	50	ug/L
99-09-2	3-Nitroaniline	13.7	U	13.7	40	50	ug/L
83-32-9	Acenaphthene	8.1	U	8.1	40	50	ug/L
Total PAH	Total PAH	172.9	U	172.9	40	800	ug/L
51-28-5	2,4-Dinitrophenol	64.2	U	64.2	80	100	ug/L
100-02-7	4-Nitrophenol	20	U	20	80	100	ug/L
132-64-9	Dibenzofuran	9.3	U	9.3	40	50	ug/L
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	27.6	U	27.6	40	100	ug/L
121-14-2	2,4-Dinitrotoluene	15.2	U	15.2	40	50	ug/L
84-66-2	Diethylphthalate	10.4	U	10.4	40	50	ug/L
7005-72-3	4-Chlorophenyl-phenylether	9.8	U	9.8	40	50	ug/L
86-73-7	Fluorene	9.6	U	9.6	40	50	ug/L
100-01-6	4-Nitroaniline	20.4	U	20.4	40	50	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	30.7	U	30.7	80	100	ug/L
86-30-6	n-Nitrosodiphenylamine	8.9	U	8.9	40	50	ug/L
103-33-3	Azobenzene	12	U	12	40	50	ug/L
101-55-3	4-Bromophenyl-phenylether	9.5	U	9.5	40	50	ug/L
118-74-1	Hexachlorobenzene	11.4	U	11.4	40	50	ug/L
1912-24-9	Atrazine	12.6	U	12.6	40	50	ug/L

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-1	SDG No.:	BF101924
Lab Sample ID:	P4425-02	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139875.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	18.5	U	18.5	80	100	ug/L
85-01-8	Phenanthrene	8.9	U	8.9	40	50	ug/L
120-12-7	Anthracene	10.7	U	10.7	40	50	ug/L
86-74-8	Carbazole	11.5	U	11.5	40	50	ug/L
84-74-2	Di-n-butylphthalate	14.7	U	14.7	40	50	ug/L
206-44-0	Fluoranthene	12.9	U	12.9	40	50	ug/L
92-87-5	Benzidine	40.7	U	40.7	80	100	ug/L
129-00-0	Pyrene	10.6	U	10.6	40	50	ug/L
85-68-7	Butylbenzylphthalate	21	U	21	40	50	ug/L
91-94-1	3,3-Dichlorobenzidine	12.8	U	12.8	80	100	ug/L
56-55-3	Benzo(a)anthracene	9.4	U	9.4	40	50	ug/L
218-01-9	Chrysene	8.6	U	8.6	40	50	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	18.9	U	18.9	40	50	ug/L
117-84-0	Di-n-octyl phthalate	25	U	25	80	100	ug/L
205-99-2	Benzo(b)fluoranthene	11.4	U	11.4	40	50	ug/L
207-08-9	Benzo(k)fluoranthene	11.9	U	11.9	40	50	ug/L
50-32-8	Benzo(a)pyrene	16.7	U	16.7	40	50	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	10.2	U	10.2	40	50	ug/L
53-70-3	Dibenzo(a,h)anthracene	11.5	U	11.5	40	50	ug/L
191-24-2	Benzo(g,h,i)perylene	11.8	U	11.8	40	50	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	10.9	U	10.9	40	50	ug/L
123-91-1	1,4-Dioxane	12.5	U	12.5	40	50	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	7.9	U	7.9	40	50	ug/L
90-12-0	1-Methylnaphthalene	8.6	U	8.6	40	50	ug/L
SURROGATES							
367-12-4	2-Fluorophenol	107.967		10-139		72	SPK: -150
13127-88-3	Phenol-d6	97.791		10-134		65	SPK: -150
4165-60-0	Nitrobenzene-d5	89.208		49-133		89	SPK: -100
321-60-8	2-Fluorobiphenyl	83.254		52-132		83	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-1	SDG No.:	BF101924
Lab Sample ID:	P4425-02	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139875.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	132.324		44-137		88	SPK: -150
1718-51-0	Terphenyl-d14	81.619		48-125		82	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	199577	6.893				
1146-65-2	Naphthalene-d8	755660	8.169				
15067-26-2	Acenaphthene-d10	412131	9.928				
1517-22-2	Phenanthrene-d10	670097	11.41				
1719-03-5	Chrysene-d12	370038	14.045				
1520-96-3	Perylene-d12	367770	15.527				

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-2	SDG No.:	BF101924
Lab Sample ID:	P4425-04	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139876.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	10.3	U	10.3	80	100	ug/L
110-86-1	Pyridine	15.5	U	15.5	40	50	ug/L
100-52-7	Benzaldehyde	40	U	40	80	100	ug/L
62-53-3	Aniline	16	U	16	40	50	ug/L
108-95-2	Phenol	9.3	U	9.3	40	50	ug/L
111-44-4	bis(2-Chloroethyl)ether	11.9	U	11.9	40	50	ug/L
95-57-8	2-Chlorophenol	7.1	U	7.1	40	50	ug/L
95-50-1	1,2-Dichlorobenzene	8.8	U	8.8	40	50	ug/L
541-73-1	1,3-Dichlorobenzene	8	U	8	40	50	ug/L
106-46-7	1,4-Dichlorobenzene	8.4	U	8.4	40	50	ug/L
100-51-6	Benzyl Alcohol	16	U	16	80	100	ug/L
95-48-7	2-Methylphenol	11.3	U	11.3	40	50	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	13.5	U	13.5	40	50	ug/L
98-86-2	Acetophenone	11	U	11	40	50	ug/L
65794-96-9	3+4-Methylphenols	11.5	U	11.5	80	100	ug/L
621-64-7	n-Nitroso-di-n-propylamine	14.8	U	14.8	25	25	ug/L
67-72-1	Hexachloroethane	10.1	U	10.1	40	50	ug/L
98-95-3	Nitrobenzene	12.7	U	12.7	40	50	ug/L
78-59-1	Isophorone	11.4	U	11.4	40	50	ug/L
88-75-5	2-Nitrophenol	19.6	U	19.6	40	50	ug/L
105-67-9	2,4-Dimethylphenol	15.1	U	15.1	40	50	ug/L
111-91-1	bis(2-Chloroethoxy)methane	10.2	U	10.2	40	50	ug/L
120-83-2	2,4-Dichlorophenol	8.8	U	8.8	40	50	ug/L
120-82-1	1,2,4-Trichlorobenzene	11	U	11	40	50	ug/L
65-85-0	Benzoic acid	54.7	U	54.7	80	100	ug/L
91-20-3	Naphthalene	10.2	U	10.2	40	50	ug/L
106-47-8	4-Chloroaniline	13	U	13	40	50	ug/L
87-68-3	Hexachlorobutadiene	12.7	U	12.7	40	50	ug/L

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-2	SDG No.:	BF101924
Lab Sample ID:	P4425-04	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139876.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	16.5	U	16.5	80	100	ug/L
59-50-7	4-Chloro-3-methylphenol	8.4	U	8.4	40	50	ug/L
91-57-6	2-Methylnaphthalene	11.3	U	11.3	40	50	ug/L
77-47-4	Hexachlorocyclopentadiene	50.2	U	50.2	80	100	ug/L
88-06-2	2,4,6-Trichlorophenol	8.9	U	8.9	40	50	ug/L
95-95-4	2,4,5-Trichlorophenol	10.1	U	10.1	40	50	ug/L
92-52-4	1,1-Biphenyl	9.1	U	9.1	40	50	ug/L
91-58-7	2-Chloronaphthalene	9.7	U	9.7	40	50	ug/L
88-74-4	2-Nitroaniline	14.1	U	14.1	40	50	ug/L
131-11-3	Dimethylphthalate	9.3	U	9.3	40	50	ug/L
208-96-8	Acenaphthylene	10.4	U	10.4	40	50	ug/L
606-20-2	2,6-Dinitrotoluene	12.4	U	12.4	40	50	ug/L
99-09-2	3-Nitroaniline	13.7	U	13.7	40	50	ug/L
83-32-9	Acenaphthene	8.1	U	8.1	40	50	ug/L
Total PAH	Total PAH	172.9	U	172.9	40	800	ug/L
51-28-5	2,4-Dinitrophenol	64.2	U	64.2	80	100	ug/L
100-02-7	4-Nitrophenol	20	U	20	80	100	ug/L
132-64-9	Dibenzofuran	9.3	U	9.3	40	50	ug/L
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	27.6	U	27.6	40	100	ug/L
121-14-2	2,4-Dinitrotoluene	15.2	U	15.2	40	50	ug/L
84-66-2	Diethylphthalate	10.4	U	10.4	40	50	ug/L
7005-72-3	4-Chlorophenyl-phenylether	9.8	U	9.8	40	50	ug/L
86-73-7	Fluorene	9.6	U	9.6	40	50	ug/L
100-01-6	4-Nitroaniline	20.4	U	20.4	40	50	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	30.7	U	30.7	80	100	ug/L
86-30-6	n-Nitrosodiphenylamine	8.9	U	8.9	40	50	ug/L
103-33-3	Azobenzene	12	U	12	40	50	ug/L
101-55-3	4-Bromophenyl-phenylether	9.5	U	9.5	40	50	ug/L
118-74-1	Hexachlorobenzene	11.4	U	11.4	40	50	ug/L
1912-24-9	Atrazine	12.6	U	12.6	40	50	ug/L

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-2	SDG No.:	BF101924
Lab Sample ID:	P4425-04	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139876.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	18.5	U	18.5	80	100	ug/L
85-01-8	Phenanthrene	8.9	U	8.9	40	50	ug/L
120-12-7	Anthracene	10.7	U	10.7	40	50	ug/L
86-74-8	Carbazole	11.5	U	11.5	40	50	ug/L
84-74-2	Di-n-butylphthalate	14.7	U	14.7	40	50	ug/L
206-44-0	Fluoranthene	12.9	U	12.9	40	50	ug/L
92-87-5	Benzidine	40.7	U	40.7	80	100	ug/L
129-00-0	Pyrene	10.6	U	10.6	40	50	ug/L
85-68-7	Butylbenzylphthalate	21	U	21	40	50	ug/L
91-94-1	3,3-Dichlorobenzidine	12.8	U	12.8	80	100	ug/L
56-55-3	Benzo(a)anthracene	9.4	U	9.4	40	50	ug/L
218-01-9	Chrysene	8.6	U	8.6	40	50	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	18.9	U	18.9	40	50	ug/L
117-84-0	Di-n-octyl phthalate	25	U	25	80	100	ug/L
205-99-2	Benzo(b)fluoranthene	11.4	U	11.4	40	50	ug/L
207-08-9	Benzo(k)fluoranthene	11.9	U	11.9	40	50	ug/L
50-32-8	Benzo(a)pyrene	16.7	U	16.7	40	50	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	10.2	U	10.2	40	50	ug/L
53-70-3	Dibenzo(a,h)anthracene	11.5	U	11.5	40	50	ug/L
191-24-2	Benzo(g,h,i)perylene	11.8	U	11.8	40	50	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	10.9	U	10.9	40	50	ug/L
123-91-1	1,4-Dioxane	12.5	U	12.5	40	50	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	7.9	U	7.9	40	50	ug/L
90-12-0	1-Methylnaphthalene	8.6	U	8.6	40	50	ug/L
SURROGATES							
367-12-4	2-Fluorophenol	113.766		10-139		76	SPK: -150
13127-88-3	Phenol-d6	103.899		10-134		69	SPK: -150
4165-60-0	Nitrobenzene-d5	88.783		49-133		89	SPK: -100
321-60-8	2-Fluorobiphenyl	85.039		52-132		85	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-2	SDG No.:	BF101924
Lab Sample ID:	P4425-04	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139876.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	132.555		44-137		88	SPK: -150
1718-51-0	Terphenyl-d14	80.787		48-125		81	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	195862	6.893				
1146-65-2	Naphthalene-d8	748996	8.169				
15067-26-2	Acenaphthene-d10	405573	9.922				
1517-22-2	Phenanthrene-d10	673234	11.41				
1719-03-5	Chrysene-d12	377546	14.045				
1520-96-3	Perylene-d12	257410	15.527				

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-3	SDG No.:	BF101924
Lab Sample ID:	P4425-06	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139877.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	10.3	U	10.3	80	100	ug/L
110-86-1	Pyridine	15.5	U	15.5	40	50	ug/L
100-52-7	Benzaldehyde	40	U	40	80	100	ug/L
62-53-3	Aniline	16	U	16	40	50	ug/L
108-95-2	Phenol	9.3	U	9.3	40	50	ug/L
111-44-4	bis(2-Chloroethyl)ether	11.9	U	11.9	40	50	ug/L
95-57-8	2-Chlorophenol	7.1	U	7.1	40	50	ug/L
95-50-1	1,2-Dichlorobenzene	8.8	U	8.8	40	50	ug/L
541-73-1	1,3-Dichlorobenzene	8	U	8	40	50	ug/L
106-46-7	1,4-Dichlorobenzene	8.4	U	8.4	40	50	ug/L
100-51-6	Benzyl Alcohol	16	U	16	80	100	ug/L
95-48-7	2-Methylphenol	11.3	U	11.3	40	50	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	13.5	U	13.5	40	50	ug/L
98-86-2	Acetophenone	11	U	11	40	50	ug/L
65794-96-9	3+4-Methylphenols	11.5	U	11.5	80	100	ug/L
621-64-7	n-Nitroso-di-n-propylamine	14.8	U	14.8	25	25	ug/L
67-72-1	Hexachloroethane	10.1	U	10.1	40	50	ug/L
98-95-3	Nitrobenzene	12.7	U	12.7	40	50	ug/L
78-59-1	Isophorone	11.4	U	11.4	40	50	ug/L
88-75-5	2-Nitrophenol	19.6	U	19.6	40	50	ug/L
105-67-9	2,4-Dimethylphenol	15.1	U	15.1	40	50	ug/L
111-91-1	bis(2-Chloroethoxy)methane	10.2	U	10.2	40	50	ug/L
120-83-2	2,4-Dichlorophenol	8.8	U	8.8	40	50	ug/L
120-82-1	1,2,4-Trichlorobenzene	11	U	11	40	50	ug/L
65-85-0	Benzoic acid	54.7	U	54.7	80	100	ug/L
91-20-3	Naphthalene	10.2	U	10.2	40	50	ug/L
106-47-8	4-Chloroaniline	13	U	13	40	50	ug/L
87-68-3	Hexachlorobutadiene	12.7	U	12.7	40	50	ug/L

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-3	SDG No.:	BF101924
Lab Sample ID:	P4425-06	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139877.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	16.5	U	16.5	80	100	ug/L
59-50-7	4-Chloro-3-methylphenol	8.4	U	8.4	40	50	ug/L
91-57-6	2-Methylnaphthalene	11.3	U	11.3	40	50	ug/L
77-47-4	Hexachlorocyclopentadiene	50.2	U	50.2	80	100	ug/L
88-06-2	2,4,6-Trichlorophenol	8.9	U	8.9	40	50	ug/L
95-95-4	2,4,5-Trichlorophenol	10.1	U	10.1	40	50	ug/L
92-52-4	1,1-Biphenyl	9.1	U	9.1	40	50	ug/L
91-58-7	2-Chloronaphthalene	9.7	U	9.7	40	50	ug/L
88-74-4	2-Nitroaniline	14.1	U	14.1	40	50	ug/L
131-11-3	Dimethylphthalate	9.3	U	9.3	40	50	ug/L
208-96-8	Acenaphthylene	10.4	U	10.4	40	50	ug/L
606-20-2	2,6-Dinitrotoluene	12.4	U	12.4	40	50	ug/L
99-09-2	3-Nitroaniline	13.7	U	13.7	40	50	ug/L
83-32-9	Acenaphthene	8.1	U	8.1	40	50	ug/L
Total PAH	Total PAH	172.9	U	172.9	40	800	ug/L
51-28-5	2,4-Dinitrophenol	64.2	U	64.2	80	100	ug/L
100-02-7	4-Nitrophenol	20	U	20	80	100	ug/L
132-64-9	Dibenzofuran	9.3	U	9.3	40	50	ug/L
121-14-2	2,4-Dinitrotoluene	15.2	U	15.2	40	50	ug/L
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	27.6	U	27.6	40	100	ug/L
84-66-2	Diethylphthalate	10.4	U	10.4	40	50	ug/L
7005-72-3	4-Chlorophenyl-phenylether	9.8	U	9.8	40	50	ug/L
86-73-7	Fluorene	9.6	U	9.6	40	50	ug/L
100-01-6	4-Nitroaniline	20.4	U	20.4	40	50	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	30.7	U	30.7	80	100	ug/L
86-30-6	n-Nitrosodiphenylamine	8.9	U	8.9	40	50	ug/L
103-33-3	Azobenzene	12	U	12	40	50	ug/L
101-55-3	4-Bromophenyl-phenylether	9.5	U	9.5	40	50	ug/L
118-74-1	Hexachlorobenzene	11.4	U	11.4	40	50	ug/L
1912-24-9	Atrazine	12.6	U	12.6	40	50	ug/L

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-3	SDG No.:	BF101924
Lab Sample ID:	P4425-06	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139877.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	18.5	U	18.5	80	100	ug/L
85-01-8	Phenanthrene	8.9	U	8.9	40	50	ug/L
120-12-7	Anthracene	10.7	U	10.7	40	50	ug/L
86-74-8	Carbazole	11.5	U	11.5	40	50	ug/L
84-74-2	Di-n-butylphthalate	14.7	U	14.7	40	50	ug/L
206-44-0	Fluoranthene	12.9	U	12.9	40	50	ug/L
92-87-5	Benzidine	40.7	U	40.7	80	100	ug/L
129-00-0	Pyrene	10.6	U	10.6	40	50	ug/L
85-68-7	Butylbenzylphthalate	21	U	21	40	50	ug/L
91-94-1	3,3-Dichlorobenzidine	12.8	U	12.8	80	100	ug/L
56-55-3	Benzo(a)anthracene	9.4	U	9.4	40	50	ug/L
218-01-9	Chrysene	8.6	U	8.6	40	50	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	18.9	U	18.9	40	50	ug/L
117-84-0	Di-n-octyl phthalate	25	U	25	80	100	ug/L
205-99-2	Benzo(b)fluoranthene	11.4	U	11.4	40	50	ug/L
207-08-9	Benzo(k)fluoranthene	11.9	U	11.9	40	50	ug/L
50-32-8	Benzo(a)pyrene	16.7	U	16.7	40	50	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	10.2	U	10.2	40	50	ug/L
53-70-3	Dibenzo(a,h)anthracene	11.5	U	11.5	40	50	ug/L
191-24-2	Benzo(g,h,i)perylene	11.8	U	11.8	40	50	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	10.9	U	10.9	40	50	ug/L
123-91-1	1,4-Dioxane	12.5	U	12.5	40	50	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	7.9	U	7.9	40	50	ug/L
90-12-0	1-Methylnaphthalene	8.6	U	8.6	40	50	ug/L
SURROGATES							
367-12-4	2-Fluorophenol	116.899		10-139		78	SPK: -150
13127-88-3	Phenol-d6	108.523		10-134		72	SPK: -150
4165-60-0	Nitrobenzene-d5	92.275		49-133		92	SPK: -100
321-60-8	2-Fluorobiphenyl	86.92		52-132		87	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-3	SDG No.:	BF101924
Lab Sample ID:	P4425-06	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139877.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	136.113		44-137		91	SPK: -150
1718-51-0	Terphenyl-d14	84.901		48-125		85	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	200976	6.892				
1146-65-2	Naphthalene-d8	775912	8.169				
15067-26-2	Acenaphthene-d10	422411	9.922				
1517-22-2	Phenanthrene-d10	695149	11.41				
1719-03-5	Chrysene-d12	391319	14.045				
1520-96-3	Perylene-d12	342377	15.527				

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-4	SDG No.:	BF101924
Lab Sample ID:	P4425-08	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139878.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	10.3	U	10.3	80	100	ug/L
110-86-1	Pyridine	15.5	U	15.5	40	50	ug/L
100-52-7	Benzaldehyde	40	U	40	80	100	ug/L
62-53-3	Aniline	16	U	16	40	50	ug/L
108-95-2	Phenol	9.3	U	9.3	40	50	ug/L
111-44-4	bis(2-Chloroethyl)ether	11.9	U	11.9	40	50	ug/L
95-57-8	2-Chlorophenol	7.1	U	7.1	40	50	ug/L
95-50-1	1,2-Dichlorobenzene	8.8	U	8.8	40	50	ug/L
541-73-1	1,3-Dichlorobenzene	8	U	8	40	50	ug/L
106-46-7	1,4-Dichlorobenzene	8.4	U	8.4	40	50	ug/L
100-51-6	Benzyl Alcohol	16	U	16	80	100	ug/L
95-48-7	2-Methylphenol	11.3	U	11.3	40	50	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	13.5	U	13.5	40	50	ug/L
98-86-2	Acetophenone	11	U	11	40	50	ug/L
65794-96-9	3+4-Methylphenols	11.5	U	11.5	80	100	ug/L
621-64-7	n-Nitroso-di-n-propylamine	14.8	U	14.8	25	25	ug/L
67-72-1	Hexachloroethane	10.1	U	10.1	40	50	ug/L
98-95-3	Nitrobenzene	12.7	U	12.7	40	50	ug/L
78-59-1	Isophorone	11.4	U	11.4	40	50	ug/L
88-75-5	2-Nitrophenol	19.6	U	19.6	40	50	ug/L
105-67-9	2,4-Dimethylphenol	15.1	U	15.1	40	50	ug/L
111-91-1	bis(2-Chloroethoxy)methane	10.2	U	10.2	40	50	ug/L
120-83-2	2,4-Dichlorophenol	8.8	U	8.8	40	50	ug/L
120-82-1	1,2,4-Trichlorobenzene	11	U	11	40	50	ug/L
65-85-0	Benzoic acid	54.7	U	54.7	80	100	ug/L
91-20-3	Naphthalene	10.2	U	10.2	40	50	ug/L
106-47-8	4-Chloroaniline	13	U	13	40	50	ug/L
87-68-3	Hexachlorobutadiene	12.7	U	12.7	40	50	ug/L

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-4	SDG No.:	BF101924
Lab Sample ID:	P4425-08	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139878.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	16.5	U	16.5	80	100	ug/L
59-50-7	4-Chloro-3-methylphenol	8.4	U	8.4	40	50	ug/L
91-57-6	2-Methylnaphthalene	11.3	U	11.3	40	50	ug/L
77-47-4	Hexachlorocyclopentadiene	50.2	U	50.2	80	100	ug/L
88-06-2	2,4,6-Trichlorophenol	8.9	U	8.9	40	50	ug/L
95-95-4	2,4,5-Trichlorophenol	10.1	U	10.1	40	50	ug/L
92-52-4	1,1-Biphenyl	9.1	U	9.1	40	50	ug/L
91-58-7	2-Chloronaphthalene	9.7	U	9.7	40	50	ug/L
88-74-4	2-Nitroaniline	14.1	U	14.1	40	50	ug/L
131-11-3	Dimethylphthalate	9.3	U	9.3	40	50	ug/L
208-96-8	Acenaphthylene	10.4	U	10.4	40	50	ug/L
606-20-2	2,6-Dinitrotoluene	12.4	U	12.4	40	50	ug/L
99-09-2	3-Nitroaniline	13.7	U	13.7	40	50	ug/L
83-32-9	Acenaphthene	8.1	U	8.1	40	50	ug/L
Total PAH	Total PAH	172.9	U	172.9	40	800	ug/L
51-28-5	2,4-Dinitrophenol	64.2	U	64.2	80	100	ug/L
100-02-7	4-Nitrophenol	20	U	20	80	100	ug/L
132-64-9	Dibenzofuran	9.3	U	9.3	40	50	ug/L
121-14-2	2,4-Dinitrotoluene	15.2	U	15.2	40	50	ug/L
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	27.6	U	27.6	40	100	ug/L
84-66-2	Diethylphthalate	10.4	U	10.4	40	50	ug/L
7005-72-3	4-Chlorophenyl-phenylether	9.8	U	9.8	40	50	ug/L
86-73-7	Fluorene	9.6	U	9.6	40	50	ug/L
100-01-6	4-Nitroaniline	20.4	U	20.4	40	50	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	30.7	U	30.7	80	100	ug/L
86-30-6	n-Nitrosodiphenylamine	8.9	U	8.9	40	50	ug/L
103-33-3	Azobenzene	12	U	12	40	50	ug/L
101-55-3	4-Bromophenyl-phenylether	9.5	U	9.5	40	50	ug/L
118-74-1	Hexachlorobenzene	11.4	U	11.4	40	50	ug/L
1912-24-9	Atrazine	12.6	U	12.6	40	50	ug/L

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-4	SDG No.:	BF101924
Lab Sample ID:	P4425-08	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139878.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	18.5	U	18.5	80	100	ug/L
85-01-8	Phenanthrene	8.9	U	8.9	40	50	ug/L
120-12-7	Anthracene	10.7	U	10.7	40	50	ug/L
86-74-8	Carbazole	11.5	U	11.5	40	50	ug/L
84-74-2	Di-n-butylphthalate	14.7	U	14.7	40	50	ug/L
206-44-0	Fluoranthene	12.9	U	12.9	40	50	ug/L
92-87-5	Benzidine	40.7	U	40.7	80	100	ug/L
129-00-0	Pyrene	10.6	U	10.6	40	50	ug/L
85-68-7	Butylbenzylphthalate	21	U	21	40	50	ug/L
91-94-1	3,3-Dichlorobenzidine	12.8	U	12.8	80	100	ug/L
56-55-3	Benzo(a)anthracene	9.4	U	9.4	40	50	ug/L
218-01-9	Chrysene	8.6	U	8.6	40	50	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	18.9	U	18.9	40	50	ug/L
117-84-0	Di-n-octyl phthalate	25	U	25	80	100	ug/L
205-99-2	Benzo(b)fluoranthene	11.4	U	11.4	40	50	ug/L
207-08-9	Benzo(k)fluoranthene	11.9	U	11.9	40	50	ug/L
50-32-8	Benzo(a)pyrene	16.7	U	16.7	40	50	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	10.2	U	10.2	40	50	ug/L
53-70-3	Dibenzo(a,h)anthracene	11.5	U	11.5	40	50	ug/L
191-24-2	Benzo(g,h,i)perylene	11.8	U	11.8	40	50	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	10.9	U	10.9	40	50	ug/L
123-91-1	1,4-Dioxane	12.5	U	12.5	40	50	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	7.9	U	7.9	40	50	ug/L
90-12-0	1-Methylnaphthalene	8.6	U	8.6	40	50	ug/L
SURROGATES							
367-12-4	2-Fluorophenol	112.779		10-139		75	SPK: -150
13127-88-3	Phenol-d6	105.369		10-134		70	SPK: -150
4165-60-0	Nitrobenzene-d5	89.927		49-133		90	SPK: -100
321-60-8	2-Fluorobiphenyl	84.175		52-132		84	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-4	SDG No.:	BF101924
Lab Sample ID:	P4425-08	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139878.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	133.883		44-137		89	SPK: -150
1718-51-0	Terphenyl-d14	83.447		48-125		83	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	205805	6.892				
1146-65-2	Naphthalene-d8	797916	8.169				
15067-26-2	Acenaphthene-d10	434693	9.927				
1517-22-2	Phenanthrene-d10	696781	11.41				
1719-03-5	Chrysene-d12	383604	14.045				
1520-96-3	Perylene-d12	380773	15.527				

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-5	SDG No.:	BF101924
Lab Sample ID:	P4425-10	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139879.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	10.3	U	10.3	80	100	ug/L
110-86-1	Pyridine	15.5	U	15.5	40	50	ug/L
100-52-7	Benzaldehyde	40	U	40	80	100	ug/L
62-53-3	Aniline	16	U	16	40	50	ug/L
108-95-2	Phenol	9.3	U	9.3	40	50	ug/L
111-44-4	bis(2-Chloroethyl)ether	11.9	U	11.9	40	50	ug/L
95-57-8	2-Chlorophenol	7.1	U	7.1	40	50	ug/L
95-50-1	1,2-Dichlorobenzene	8.8	U	8.8	40	50	ug/L
541-73-1	1,3-Dichlorobenzene	8	U	8	40	50	ug/L
106-46-7	1,4-Dichlorobenzene	8.4	U	8.4	40	50	ug/L
100-51-6	Benzyl Alcohol	16	U	16	80	100	ug/L
95-48-7	2-Methylphenol	11.3	U	11.3	40	50	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	13.5	U	13.5	40	50	ug/L
98-86-2	Acetophenone	11	U	11	40	50	ug/L
65794-96-9	3+4-Methylphenols	11.5	U	11.5	80	100	ug/L
621-64-7	n-Nitroso-di-n-propylamine	14.8	U	14.8	25	25	ug/L
67-72-1	Hexachloroethane	10.1	U	10.1	40	50	ug/L
98-95-3	Nitrobenzene	12.7	U	12.7	40	50	ug/L
78-59-1	Isophorone	11.4	U	11.4	40	50	ug/L
88-75-5	2-Nitrophenol	19.6	U	19.6	40	50	ug/L
105-67-9	2,4-Dimethylphenol	15.1	U	15.1	40	50	ug/L
111-91-1	bis(2-Chloroethoxy)methane	10.2	U	10.2	40	50	ug/L
120-83-2	2,4-Dichlorophenol	8.8	U	8.8	40	50	ug/L
120-82-1	1,2,4-Trichlorobenzene	11	U	11	40	50	ug/L
65-85-0	Benzoic acid	54.7	U	54.7	80	100	ug/L
91-20-3	Naphthalene	10.2	U	10.2	40	50	ug/L
106-47-8	4-Chloroaniline	13	U	13	40	50	ug/L
87-68-3	Hexachlorobutadiene	12.7	U	12.7	40	50	ug/L

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-5	SDG No.:	BF101924
Lab Sample ID:	P4425-10	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139879.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	16.5	U	16.5	80	100	ug/L
59-50-7	4-Chloro-3-methylphenol	8.4	U	8.4	40	50	ug/L
91-57-6	2-Methylnaphthalene	11.3	U	11.3	40	50	ug/L
77-47-4	Hexachlorocyclopentadiene	50.2	U	50.2	80	100	ug/L
88-06-2	2,4,6-Trichlorophenol	8.9	U	8.9	40	50	ug/L
95-95-4	2,4,5-Trichlorophenol	10.1	U	10.1	40	50	ug/L
92-52-4	1,1-Biphenyl	9.1	U	9.1	40	50	ug/L
91-58-7	2-Chloronaphthalene	9.7	U	9.7	40	50	ug/L
88-74-4	2-Nitroaniline	14.1	U	14.1	40	50	ug/L
131-11-3	Dimethylphthalate	9.3	U	9.3	40	50	ug/L
208-96-8	Acenaphthylene	10.4	U	10.4	40	50	ug/L
606-20-2	2,6-Dinitrotoluene	12.4	U	12.4	40	50	ug/L
99-09-2	3-Nitroaniline	13.7	U	13.7	40	50	ug/L
83-32-9	Acenaphthene	8.1	U	8.1	40	50	ug/L
Total PAH	Total PAH	172.9	U	172.9	40	800	ug/L
51-28-5	2,4-Dinitrophenol	64.2	U	64.2	80	100	ug/L
100-02-7	4-Nitrophenol	20	U	20	80	100	ug/L
132-64-9	Dibenzofuran	9.3	U	9.3	40	50	ug/L
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	27.6	U	27.6	40	100	ug/L
121-14-2	2,4-Dinitrotoluene	15.2	U	15.2	40	50	ug/L
84-66-2	Diethylphthalate	10.4	U	10.4	40	50	ug/L
7005-72-3	4-Chlorophenyl-phenylether	9.8	U	9.8	40	50	ug/L
86-73-7	Fluorene	9.6	U	9.6	40	50	ug/L
100-01-6	4-Nitroaniline	20.4	U	20.4	40	50	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	30.7	U	30.7	80	100	ug/L
86-30-6	n-Nitrosodiphenylamine	8.9	U	8.9	40	50	ug/L
103-33-3	Azobenzene	12	U	12	40	50	ug/L
101-55-3	4-Bromophenyl-phenylether	9.5	U	9.5	40	50	ug/L
118-74-1	Hexachlorobenzene	11.4	U	11.4	40	50	ug/L
1912-24-9	Atrazine	12.6	U	12.6	40	50	ug/L

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-5	SDG No.:	BF101924
Lab Sample ID:	P4425-10	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139879.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	18.5	U	18.5	80	100	ug/L
85-01-8	Phenanthrene	8.9	U	8.9	40	50	ug/L
120-12-7	Anthracene	10.7	U	10.7	40	50	ug/L
86-74-8	Carbazole	11.5	U	11.5	40	50	ug/L
84-74-2	Di-n-butylphthalate	14.7	U	14.7	40	50	ug/L
206-44-0	Fluoranthene	12.9	U	12.9	40	50	ug/L
92-87-5	Benzidine	40.7	U	40.7	80	100	ug/L
129-00-0	Pyrene	10.6	U	10.6	40	50	ug/L
85-68-7	Butylbenzylphthalate	21	U	21	40	50	ug/L
91-94-1	3,3-Dichlorobenzidine	12.8	U	12.8	80	100	ug/L
56-55-3	Benzo(a)anthracene	9.4	U	9.4	40	50	ug/L
218-01-9	Chrysene	8.6	U	8.6	40	50	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	18.9	U	18.9	40	50	ug/L
117-84-0	Di-n-octyl phthalate	25	U	25	80	100	ug/L
205-99-2	Benzo(b)fluoranthene	11.4	U	11.4	40	50	ug/L
207-08-9	Benzo(k)fluoranthene	11.9	U	11.9	40	50	ug/L
50-32-8	Benzo(a)pyrene	16.7	U	16.7	40	50	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	10.2	U	10.2	40	50	ug/L
53-70-3	Dibenzo(a,h)anthracene	11.5	U	11.5	40	50	ug/L
191-24-2	Benzo(g,h,i)perylene	11.8	U	11.8	40	50	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	10.9	U	10.9	40	50	ug/L
123-91-1	1,4-Dioxane	12.5	U	12.5	40	50	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	7.9	U	7.9	40	50	ug/L
90-12-0	1-Methylnaphthalene	8.6	U	8.6	40	50	ug/L
SURROGATES							
367-12-4	2-Fluorophenol	114.465		10-139		76	SPK: -150
13127-88-3	Phenol-d6	105.273		10-134		70	SPK: -150
4165-60-0	Nitrobenzene-d5	90.21		49-133		90	SPK: -100
321-60-8	2-Fluorobiphenyl	85.125		52-132		85	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	TP-5	SDG No.:	BF101924
Lab Sample ID:	P4425-10	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139879.D	1	10/18/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	134.392		44-137		90	SPK: -150
1718-51-0	Terphenyl-d14	85.386		48-125		85	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	195941	6.892				
1146-65-2	Naphthalene-d8	751190	8.169				
15067-26-2	Acenaphthene-d10	414555	9.922				
1517-22-2	Phenanthrene-d10	687642	11.41				
1719-03-5	Chrysene-d12	376430	14.045				
1520-96-3	Perylene-d12	359501	15.527				

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-8	SDG No.:	BF101924
Lab Sample ID:	P4426-15	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.04	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139880.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	52	U	52	160	200	ug/Kg
110-86-1	Pyridine	47.9	U	47.9	77.9	100	ug/Kg
100-52-7	Benzaldehyde	110	U	110	160	200	ug/Kg
62-53-3	Aniline	58.1	U	58.1	77.9	100	ug/Kg
108-95-2	Phenol	49.7	U	49.7	77.9	100	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	50.2	U	50.2	77.9	100	ug/Kg
95-57-8	2-Chlorophenol	50.1	U	50.1	77.9	100	ug/Kg
95-50-1	1,2-Dichlorobenzene	50.7	U	50.7	77.9	100	ug/Kg
541-73-1	1,3-Dichlorobenzene	51.6	U	51.6	77.9	100	ug/Kg
106-46-7	1,4-Dichlorobenzene	51.9	U	51.9	77.9	100	ug/Kg
100-51-6	Benzyl Alcohol	49.8	U	49.8	160	200	ug/Kg
95-48-7	2-Methylphenol	48.3	U	48.3	77.9	100	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	54.5	U	54.5	77.9	100	ug/Kg
98-86-2	Acetophenone	52.1	U	52.1	77.9	100	ug/Kg
65794-96-9	3+4-Methylphenols	47.8	U	47.8	160	200	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	24.2	U	24.2	48	48	ug/Kg
67-72-1	Hexachloroethane	49.8	U	49.8	77.9	100	ug/Kg
98-95-3	Nitrobenzene	54.4	U	54.4	77.9	100	ug/Kg
78-59-1	Isophorone	50.7	U	50.7	77.9	100	ug/Kg
88-75-5	2-Nitrophenol	56.7	U	56.7	77.9	100	ug/Kg
105-67-9	2,4-Dimethylphenol	55.9	U	55.9	77.9	100	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	51.4	U	51.4	77.9	100	ug/Kg
120-83-2	2,4-Dichlorophenol	45.3	U	45.3	77.9	100	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	51.2	U	51.2	77.9	100	ug/Kg
65-85-0	Benzoic acid	140	U	140	160	200	ug/Kg
91-20-3	Naphthalene	49.5	U	49.5	77.9	100	ug/Kg
106-47-8	4-Chloroaniline	49.5	U	49.5	77.9	100	ug/Kg
87-68-3	Hexachlorobutadiene	49.9	U	49.9	77.9	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-8	SDG No.:	BF101924
Lab Sample ID:	P4426-15	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.04	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139880.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	52	U	52	160	200	ug/Kg
59-50-7	4-Chloro-3-methylphenol	46.5	U	46.5	77.9	100	ug/Kg
91-57-6	2-Methylnaphthalene	49.5	U	49.5	77.9	100	ug/Kg
77-47-4	Hexachlorocyclopentadiene	93.5	U	93.5	160	200	ug/Kg
88-06-2	2,4,6-Trichlorophenol	42.8	U	42.8	77.9	100	ug/Kg
95-95-4	2,4,5-Trichlorophenol	44.4	U	44.4	77.9	100	ug/Kg
92-52-4	1,1-Biphenyl	52.4	U	52.4	77.9	100	ug/Kg
91-58-7	2-Chloronaphthalene	49.9	U	49.9	77.9	100	ug/Kg
88-74-4	2-Nitroaniline	57	U	57	77.9	100	ug/Kg
131-11-3	Dimethylphthalate	49	U	49	77.9	100	ug/Kg
208-96-8	Acenaphthylene	51.9	U	51.9	77.9	100	ug/Kg
606-20-2	2,6-Dinitrotoluene	49.9	U	49.9	77.9	100	ug/Kg
99-09-2	3-Nitroaniline	53.5	U	53.5	77.9	100	ug/Kg
83-32-9	Acenaphthene	48.6	U	48.6	77.9	100	ug/Kg
Total PAH	Total PAH	794.6	U	794.6	77.9	1600	ug/Kg
51-28-5	2,4-Dinitrophenol	150	U	150	160	200	ug/Kg
100-02-7	4-Nitrophenol	69.5	U	69.5	160	200	ug/Kg
132-64-9	Dibenzofuran	50.6	U	50.6	77.9	100	ug/Kg
121-14-2	2,4-Dinitrotoluene	51.7	U	51.7	77.9	100	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	101.6	U	101.6	77.9	200	ug/Kg
84-66-2	Diethylphthalate	48	U	48	77.9	100	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	51.3	U	51.3	77.9	100	ug/Kg
86-73-7	Fluorene	51.3	U	51.3	77.9	100	ug/Kg
100-01-6	4-Nitroaniline	64.1	U	64.1	77.9	100	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	70.1	U	70.1	160	200	ug/Kg
86-30-6	n-Nitrosodiphenylamine	48.9	U	48.9	77.9	100	ug/Kg
103-33-3	Azobenzene	47.7	U	47.7	77.9	100	ug/Kg
101-55-3	4-Bromophenyl-phenylether	47.3	U	47.3	77.9	100	ug/Kg
118-74-1	Hexachlorobenzene	51	U	51	77.9	100	ug/Kg
1912-24-9	Atrazine	54.8	U	54.8	77.9	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-8	SDG No.:	BF101924
Lab Sample ID:	P4426-15	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.04	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139880.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	46.3	U	46.3	160	200	ug/Kg
85-01-8	Phenanthrene	50.4	U	50.4	77.9	100	ug/Kg
120-12-7	Anthracene	50.6	U	50.6	77.9	100	ug/Kg
86-74-8	Carbazole	48.1	U	48.1	77.9	100	ug/Kg
84-74-2	Di-n-butylphthalate	50.5	U	50.5	77.9	100	ug/Kg
206-44-0	Fluoranthene	49	U	49	77.9	100	ug/Kg
92-87-5	Benzidine	97.7	U	97.7	160	200	ug/Kg
129-00-0	Pyrene	49.8	U	49.8	77.9	100	ug/Kg
85-68-7	Butylbenzylphthalate	58	U	58	77.9	100	ug/Kg
91-94-1	3,3-Dichlorobenzidine	59.1	U	59.1	160	200	ug/Kg
56-55-3	Benzo(a)anthracene	48.4	U	48.4	77.9	100	ug/Kg
218-01-9	Chrysene	47.7	U	47.7	77.9	100	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	54.6	U	54.6	77.9	100	ug/Kg
117-84-0	Di-n-octyl phthalate	65.9	U	65.9	160	200	ug/Kg
205-99-2	Benzo(b)fluoranthene	48.6	U	48.6	77.9	100	ug/Kg
207-08-9	Benzo(k)fluoranthene	49.5	U	49.5	77.9	100	ug/Kg
50-32-8	Benzo(a)pyrene	55.8	U	55.8	77.9	100	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	46.8	U	46.8	77.9	100	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	48.7	U	48.7	77.9	100	ug/Kg
191-24-2	Benzo(g,h,i)perylene	48	U	48	77.9	100	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	52	U	52	77.9	100	ug/Kg
123-91-1	1,4-Dioxane	65.9	U	65.9	77.9	100	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	44.8	U	44.8	77.9	100	ug/Kg
90-12-0	1-Methylnaphthalene	49.9	U	49.9	100	100	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	12.991	*	18-112		9	SPK: -150
13127-88-3	Phenol-d6	28.218		15-107		19	SPK: -150
4165-60-0	Nitrobenzene-d5	41.121		18-107		41	SPK: -100
321-60-8	2-Fluorobiphenyl	45.792		20-109		46	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-8	SDG No.:	BF101924
Lab Sample ID:	P4426-15	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.04	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139880.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	5.31	*	10-116		4	SPK: -150
1718-51-0	Terphenyl-d14	42.599		10-105		43	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	200447	6.887				
1146-65-2	Naphthalene-d8	768121	8.169				
15067-26-2	Acenaphthene-d10	421896	9.922				
1517-22-2	Phenanthrene-d10	668798	11.41				
1719-03-5	Chrysene-d12	364704	14.045				
1520-96-3	Perylene-d12	378848	15.527				

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-15	SDG No.:	BF101924
Lab Sample ID:	P4426-29	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.04	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139881.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	52	U	52	160	200	ug/Kg
110-86-1	Pyridine	47.9	U	47.9	77.9	100	ug/Kg
100-52-7	Benzaldehyde	110	U	110	160	200	ug/Kg
62-53-3	Aniline	58.1	U	58.1	77.9	100	ug/Kg
108-95-2	Phenol	49.7	U	49.7	77.9	100	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	50.2	U	50.2	77.9	100	ug/Kg
95-57-8	2-Chlorophenol	50.1	U	50.1	77.9	100	ug/Kg
95-50-1	1,2-Dichlorobenzene	50.7	U	50.7	77.9	100	ug/Kg
541-73-1	1,3-Dichlorobenzene	51.6	U	51.6	77.9	100	ug/Kg
106-46-7	1,4-Dichlorobenzene	51.9	U	51.9	77.9	100	ug/Kg
100-51-6	Benzyl Alcohol	49.8	U	49.8	160	200	ug/Kg
95-48-7	2-Methylphenol	48.3	U	48.3	77.9	100	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	54.5	U	54.5	77.9	100	ug/Kg
98-86-2	Acetophenone	52.1	U	52.1	77.9	100	ug/Kg
65794-96-9	3+4-Methylphenols	47.8	U	47.8	160	200	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	24.2	U	24.2	48	48	ug/Kg
67-72-1	Hexachloroethane	49.8	U	49.8	77.9	100	ug/Kg
98-95-3	Nitrobenzene	54.4	U	54.4	77.9	100	ug/Kg
78-59-1	Isophorone	50.7	U	50.7	77.9	100	ug/Kg
88-75-5	2-Nitrophenol	56.7	U	56.7	77.9	100	ug/Kg
105-67-9	2,4-Dimethylphenol	55.9	U	55.9	77.9	100	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	51.4	U	51.4	77.9	100	ug/Kg
120-83-2	2,4-Dichlorophenol	45.3	U	45.3	77.9	100	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	51.2	U	51.2	77.9	100	ug/Kg
65-85-0	Benzoic acid	140	U	140	160	200	ug/Kg
91-20-3	Naphthalene	49.5	U	49.5	77.9	100	ug/Kg
106-47-8	4-Chloroaniline	49.5	U	49.5	77.9	100	ug/Kg
87-68-3	Hexachlorobutadiene	49.9	U	49.9	77.9	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-15	SDG No.:	BF101924
Lab Sample ID:	P4426-29	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.04	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139881.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	52	U	52	160	200	ug/Kg
59-50-7	4-Chloro-3-methylphenol	46.5	U	46.5	77.9	100	ug/Kg
91-57-6	2-Methylnaphthalene	49.5	U	49.5	77.9	100	ug/Kg
77-47-4	Hexachlorocyclopentadiene	93.5	U	93.5	160	200	ug/Kg
88-06-2	2,4,6-Trichlorophenol	42.8	U	42.8	77.9	100	ug/Kg
95-95-4	2,4,5-Trichlorophenol	44.4	U	44.4	77.9	100	ug/Kg
92-52-4	1,1-Biphenyl	52.4	U	52.4	77.9	100	ug/Kg
91-58-7	2-Chloronaphthalene	49.9	U	49.9	77.9	100	ug/Kg
88-74-4	2-Nitroaniline	57	U	57	77.9	100	ug/Kg
131-11-3	Dimethylphthalate	49	U	49	77.9	100	ug/Kg
208-96-8	Acenaphthylene	51.9	U	51.9	77.9	100	ug/Kg
606-20-2	2,6-Dinitrotoluene	49.9	U	49.9	77.9	100	ug/Kg
99-09-2	3-Nitroaniline	53.5	U	53.5	77.9	100	ug/Kg
83-32-9	Acenaphthene	48.6	U	48.6	77.9	100	ug/Kg
Total PAH	Total PAH	794.6	U	794.6	77.9	1600	ug/Kg
51-28-5	2,4-Dinitrophenol	150	U	150	160	200	ug/Kg
100-02-7	4-Nitrophenol	69.5	U	69.5	160	200	ug/Kg
132-64-9	Dibenzofuran	50.6	U	50.6	77.9	100	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	101.6	U	101.6	77.9	200	ug/Kg
121-14-2	2,4-Dinitrotoluene	51.7	U	51.7	77.9	100	ug/Kg
84-66-2	Diethylphthalate	48	U	48	77.9	100	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	51.3	U	51.3	77.9	100	ug/Kg
86-73-7	Fluorene	51.3	U	51.3	77.9	100	ug/Kg
100-01-6	4-Nitroaniline	64.1	U	64.1	77.9	100	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	70.1	U	70.1	160	200	ug/Kg
86-30-6	n-Nitrosodiphenylamine	48.9	U	48.9	77.9	100	ug/Kg
103-33-3	Azobenzene	47.7	U	47.7	77.9	100	ug/Kg
101-55-3	4-Bromophenyl-phenylether	47.3	U	47.3	77.9	100	ug/Kg
118-74-1	Hexachlorobenzene	51	U	51	77.9	100	ug/Kg
1912-24-9	Atrazine	54.8	U	54.8	77.9	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-15	SDG No.:	BF101924
Lab Sample ID:	P4426-29	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.04	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139881.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	46.3	U	46.3	160	200	ug/Kg
85-01-8	Phenanthrene	50.4	U	50.4	77.9	100	ug/Kg
120-12-7	Anthracene	50.6	U	50.6	77.9	100	ug/Kg
86-74-8	Carbazole	48.1	U	48.1	77.9	100	ug/Kg
84-74-2	Di-n-butylphthalate	50.5	U	50.5	77.9	100	ug/Kg
206-44-0	Fluoranthene	49	U	49	77.9	100	ug/Kg
92-87-5	Benzidine	97.7	U	97.7	160	200	ug/Kg
129-00-0	Pyrene	49.8	U	49.8	77.9	100	ug/Kg
85-68-7	Butylbenzylphthalate	58	U	58	77.9	100	ug/Kg
91-94-1	3,3-Dichlorobenzidine	59.1	U	59.1	160	200	ug/Kg
56-55-3	Benzo(a)anthracene	48.4	U	48.4	77.9	100	ug/Kg
218-01-9	Chrysene	47.7	U	47.7	77.9	100	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	54.6	U	54.6	77.9	100	ug/Kg
117-84-0	Di-n-octyl phthalate	65.9	U	65.9	160	200	ug/Kg
205-99-2	Benzo(b)fluoranthene	48.6	U	48.6	77.9	100	ug/Kg
207-08-9	Benzo(k)fluoranthene	49.5	U	49.5	77.9	100	ug/Kg
50-32-8	Benzo(a)pyrene	55.8	U	55.8	77.9	100	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	46.8	U	46.8	77.9	100	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	48.7	U	48.7	77.9	100	ug/Kg
191-24-2	Benzo(g,h,i)perylene	48	U	48	77.9	100	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	52	U	52	77.9	100	ug/Kg
123-91-1	1,4-Dioxane	65.9	U	65.9	77.9	100	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	44.8	U	44.8	77.9	100	ug/Kg
90-12-0	1-Methylnaphthalene	49.9	U	49.9	100	100	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	8.801	*	18-112		6	SPK: -150
13127-88-3	Phenol-d6	20.165	*	15-107		13	SPK: -150
4165-60-0	Nitrobenzene-d5	36.172		18-107		36	SPK: -100
321-60-8	2-Fluorobiphenyl	48.126		20-109		48	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-15	SDG No.:	BF101924
Lab Sample ID:	P4426-29	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.04	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139881.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	5.843	*	10-116		4	SPK: -150
1718-51-0	Terphenyl-d14	34.333		10-105		34	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	195168	6.887				
1146-65-2	Naphthalene-d8	723646	8.169				
15067-26-2	Acenaphthene-d10	347098	9.928				
1517-22-2	Phenanthrene-d10	428172	11.416				
1719-03-5	Chrysene-d12	361459	14.051				
1520-96-3	Perylene-d12	381877	15.527				

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-10	SDG No.:	BF101924
Lab Sample ID:	P4426-19	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.02	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139882.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	52.1	U	52.1	160	200	ug/Kg
110-86-1	Pyridine	47.9	U	47.9	78	100	ug/Kg
100-52-7	Benzaldehyde	110	U	110	160	200	ug/Kg
62-53-3	Aniline	58.1	U	58.1	78	100	ug/Kg
108-95-2	Phenol	49.7	U	49.7	78	100	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	50.2	U	50.2	78	100	ug/Kg
95-57-8	2-Chlorophenol	50.1	U	50.1	78	100	ug/Kg
95-50-1	1,2-Dichlorobenzene	50.7	U	50.7	78	100	ug/Kg
541-73-1	1,3-Dichlorobenzene	51.6	U	51.6	78	100	ug/Kg
106-46-7	1,4-Dichlorobenzene	51.9	U	51.9	78	100	ug/Kg
100-51-6	Benzyl Alcohol	49.8	U	49.8	160	200	ug/Kg
95-48-7	2-Methylphenol	48.3	U	48.3	78	100	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	54.5	U	54.5	78	100	ug/Kg
98-86-2	Acetophenone	52.1	U	52.1	78	100	ug/Kg
65794-96-9	3+4-Methylphenols	47.9	U	47.9	160	200	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	24.2	U	24.2	48	48	ug/Kg
67-72-1	Hexachloroethane	49.8	U	49.8	78	100	ug/Kg
98-95-3	Nitrobenzene	54.5	U	54.5	78	100	ug/Kg
78-59-1	Isophorone	50.7	U	50.7	78	100	ug/Kg
88-75-5	2-Nitrophenol	56.7	U	56.7	78	100	ug/Kg
105-67-9	2,4-Dimethylphenol	55.9	U	55.9	78	100	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	51.5	U	51.5	78	100	ug/Kg
120-83-2	2,4-Dichlorophenol	45.3	U	45.3	78	100	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	51.2	U	51.2	78	100	ug/Kg
65-85-0	Benzoic acid	140	U	140	160	200	ug/Kg
91-20-3	Naphthalene	49.5	U	49.5	78	100	ug/Kg
106-47-8	4-Chloroaniline	49.5	U	49.5	78	100	ug/Kg
87-68-3	Hexachlorobutadiene	50	U	50	78	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-10	SDG No.:	BF101924
Lab Sample ID:	P4426-19	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.02 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139882.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	52.1	U	52.1	160	200	ug/Kg
59-50-7	4-Chloro-3-methylphenol	46.5	U	46.5	78	100	ug/Kg
91-57-6	2-Methylnaphthalene	49.5	U	49.5	78	100	ug/Kg
77-47-4	Hexachlorocyclopentadiene	93.6	U	93.6	160	200	ug/Kg
88-06-2	2,4,6-Trichlorophenol	42.8	U	42.8	78	100	ug/Kg
95-95-4	2,4,5-Trichlorophenol	44.4	U	44.4	78	100	ug/Kg
92-52-4	1,1-Biphenyl	52.4	U	52.4	78	100	ug/Kg
91-58-7	2-Chloronaphthalene	50	U	50	78	100	ug/Kg
88-74-4	2-Nitroaniline	57	U	57	78	100	ug/Kg
131-11-3	Dimethylphthalate	49	U	49	78	100	ug/Kg
208-96-8	Acenaphthylene	51.9	U	51.9	78	100	ug/Kg
606-20-2	2,6-Dinitrotoluene	49.9	U	49.9	78	100	ug/Kg
99-09-2	3-Nitroaniline	53.5	U	53.5	78	100	ug/Kg
83-32-9	Acenaphthene	48.6	U	48.6	78	100	ug/Kg
Total PAH	Total PAH	794.6	U	794.6	78	1600	ug/Kg
51-28-5	2,4-Dinitrophenol	150	U	150	160	200	ug/Kg
100-02-7	4-Nitrophenol	69.6	U	69.6	160	200	ug/Kg
132-64-9	Dibenzofuran	50.6	U	50.6	78	100	ug/Kg
121-14-2	2,4-Dinitrotoluene	51.7	U	51.7	78	100	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	101.6	U	101.6	78	200	ug/Kg
84-66-2	Diethylphthalate	48	U	48	78	100	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	51.3	U	51.3	78	100	ug/Kg
86-73-7	Fluorene	51.3	U	51.3	78	100	ug/Kg
100-01-6	4-Nitroaniline	64.2	U	64.2	78	100	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	70.2	U	70.2	160	200	ug/Kg
86-30-6	n-Nitrosodiphenylamine	48.9	U	48.9	78	100	ug/Kg
103-33-3	Azobenzene	47.7	U	47.7	78	100	ug/Kg
101-55-3	4-Bromophenyl-phenylether	47.3	U	47.3	78	100	ug/Kg
118-74-1	Hexachlorobenzene	51	U	51	78	100	ug/Kg
1912-24-9	Atrazine	54.8	U	54.8	78	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-10	SDG No.:	BF101924
Lab Sample ID:	P4426-19	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.02	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139882.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	46.4	U	46.4	160	200	ug/Kg
85-01-8	Phenanthrene	50.4	U	50.4	78	100	ug/Kg
120-12-7	Anthracene	50.6	U	50.6	78	100	ug/Kg
86-74-8	Carbazole	48.2	U	48.2	78	100	ug/Kg
84-74-2	Di-n-butylphthalate	50.6	U	50.6	78	100	ug/Kg
206-44-0	Fluoranthene	49	U	49	78	100	ug/Kg
92-87-5	Benzidine	97.8	U	97.8	160	200	ug/Kg
129-00-0	Pyrene	49.8	U	49.8	78	100	ug/Kg
85-68-7	Butylbenzylphthalate	58.1	U	58.1	78	100	ug/Kg
91-94-1	3,3-Dichlorobenzidine	59.1	U	59.1	160	200	ug/Kg
56-55-3	Benzo(a)anthracene	48.4	U	48.4	78	100	ug/Kg
218-01-9	Chrysene	47.7	U	47.7	78	100	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	54.6	U	54.6	78	100	ug/Kg
117-84-0	Di-n-octyl phthalate	66	U	66	160	200	ug/Kg
205-99-2	Benzo(b)fluoranthene	48.6	U	48.6	78	100	ug/Kg
207-08-9	Benzo(k)fluoranthene	49.5	U	49.5	78	100	ug/Kg
50-32-8	Benzo(a)pyrene	55.8	U	55.8	78	100	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	46.8	U	46.8	78	100	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	48.7	U	48.7	78	100	ug/Kg
191-24-2	Benzo(g,h,i)perylene	48	U	48	78	100	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	52.1	U	52.1	78	100	ug/Kg
123-91-1	1,4-Dioxane	66	U	66	78	100	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	44.8	U	44.8	78	100	ug/Kg
90-12-0	1-Methylnaphthalene	49.9	U	49.9	100	100	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	34.222		18-112		23	SPK: -150
13127-88-3	Phenol-d6	46.063		15-107		31	SPK: -150
4165-60-0	Nitrobenzene-d5	42.68		18-107		43	SPK: -100
321-60-8	2-Fluorobiphenyl	48.266		20-109		48	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-10	SDG No.:	BF101924
Lab Sample ID:	P4426-19	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.02	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139882.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	26.103		10-116		17	SPK: -150
1718-51-0	Terphenyl-d14	39.782		10-105		40	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	197521	6.886				
1146-65-2	Naphthalene-d8	771255	8.169				
15067-26-2	Acenaphthene-d10	412785	9.922				
1517-22-2	Phenanthrene-d10	636993	11.41				
1719-03-5	Chrysene-d12	370206	14.045				
1520-96-3	Perylene-d12	387933	15.527				

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-13	SDG No.:	BF101924
Lab Sample ID:	P4426-25	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.07	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139883.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	52	U	52	160	200	ug/Kg
110-86-1	Pyridine	47.9	U	47.9	77.9	100	ug/Kg
100-52-7	Benzaldehyde	110	U	110	160	200	ug/Kg
62-53-3	Aniline	58.1	U	58.1	77.9	100	ug/Kg
108-95-2	Phenol	49.7	U	49.7	77.9	100	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	50.1	U	50.1	77.9	100	ug/Kg
95-57-8	2-Chlorophenol	50	U	50	77.9	100	ug/Kg
95-50-1	1,2-Dichlorobenzene	50.7	U	50.7	77.9	100	ug/Kg
541-73-1	1,3-Dichlorobenzene	51.5	U	51.5	77.9	100	ug/Kg
106-46-7	1,4-Dichlorobenzene	51.8	U	51.8	77.9	100	ug/Kg
100-51-6	Benzyl Alcohol	49.7	U	49.7	160	200	ug/Kg
95-48-7	2-Methylphenol	48.3	U	48.3	77.9	100	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	54.5	U	54.5	77.9	100	ug/Kg
98-86-2	Acetophenone	52.1	U	52.1	77.9	100	ug/Kg
65794-96-9	3+4-Methylphenols	47.8	U	47.8	160	200	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	24.1	U	24.1	47.9	47.9	ug/Kg
67-72-1	Hexachloroethane	49.7	U	49.7	77.9	100	ug/Kg
98-95-3	Nitrobenzene	54.4	U	54.4	77.9	100	ug/Kg
78-59-1	Isophorone	50.7	U	50.7	77.9	100	ug/Kg
88-75-5	2-Nitrophenol	56.6	U	56.6	77.9	100	ug/Kg
105-67-9	2,4-Dimethylphenol	55.8	U	55.8	77.9	100	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	51.4	U	51.4	77.9	100	ug/Kg
120-83-2	2,4-Dichlorophenol	45.2	U	45.2	77.9	100	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	51.2	U	51.2	77.9	100	ug/Kg
65-85-0	Benzoic acid	140	U	140	160	200	ug/Kg
91-20-3	Naphthalene	49.5	U	49.5	77.9	100	ug/Kg
106-47-8	4-Chloroaniline	49.5	U	49.5	77.9	100	ug/Kg
87-68-3	Hexachlorobutadiene	49.9	U	49.9	77.9	100	ug/Kg

Report of Analysis

Client:				Date Collected:	10/16/2024 12:00:00 AM		
Project:				Date Received:	10/16/2024 12:00:00 AM		
Client Sample ID:	PAD-13			SDG No.:	BF101924		
Lab Sample ID:	P4426-25			Matrix:	Solid		
Analytical Method:	8270E			% Moisture:	0		
Sample Wt/Vol:	50.07	Units:	g	Final Vol:	1000	uL	
Soil Aliquot Vol:	uL			Test:			
Extraction Type :	Decanted :			Level :	LOW		
Injection Volume :	GPC Factor :			GPC Cleanup :	PH :		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139883.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	52	U	52	160	200	ug/Kg
59-50-7	4-Chloro-3-methylphenol	46.4	U	46.4	77.9	100	ug/Kg
91-57-6	2-Methylnaphthalene	49.4	U	49.4	77.9	100	ug/Kg
77-47-4	Hexachlorocyclopentadiene	93.5	U	93.5	160	200	ug/Kg
88-06-2	2,4,6-Trichlorophenol	42.8	U	42.8	77.9	100	ug/Kg
95-95-4	2,4,5-Trichlorophenol	44.3	U	44.3	77.9	100	ug/Kg
92-52-4	1,1-Biphenyl	52.4	U	52.4	77.9	100	ug/Kg
91-58-7	2-Chloronaphthalene	49.9	U	49.9	77.9	100	ug/Kg
88-74-4	2-Nitroaniline	56.9	U	56.9	77.9	100	ug/Kg
131-11-3	Dimethylphthalate	49	U	49	77.9	100	ug/Kg
208-96-8	Acenaphthylene	51.8	U	51.8	77.9	100	ug/Kg
606-20-2	2,6-Dinitrotoluene	49.9	U	49.9	77.9	100	ug/Kg
99-09-2	3-Nitroaniline	53.4	U	53.4	77.9	100	ug/Kg
83-32-9	Acenaphthene	48.6	U	48.6	77.9	100	ug/Kg
Total PAH	Total PAH	794	U	794	77.9	1600	ug/Kg
51-28-5	2,4-Dinitrophenol	150	U	150	160	200	ug/Kg
100-02-7	4-Nitrophenol	69.5	U	69.5	160	200	ug/Kg
132-64-9	Dibenzofuran	50.6	U	50.6	77.9	100	ug/Kg
121-14-2	2,4-Dinitrotoluene	51.6	U	51.6	77.9	100	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	101.5	U	101.5	77.9	200	ug/Kg
84-66-2	Diethylphthalate	48	U	48	77.9	100	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	51.3	U	51.3	77.9	100	ug/Kg
86-73-7	Fluorene	51.2	U	51.2	77.9	100	ug/Kg
100-01-6	4-Nitroaniline	64.1	U	64.1	77.9	100	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	70.1	U	70.1	160	200	ug/Kg
86-30-6	n-Nitrosodiphenylamine	48.9	U	48.9	77.9	100	ug/Kg
103-33-3	Azobenzene	47.7	U	47.7	77.9	100	ug/Kg
101-55-3	4-Bromophenyl-phenylether	47.3	U	47.3	77.9	100	ug/Kg
118-74-1	Hexachlorobenzene	50.9	U	50.9	77.9	100	ug/Kg
1912-24-9	Atrazine	54.8	U	54.8	77.9	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-13	SDG No.:	BF101924
Lab Sample ID:	P4426-25	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.07	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139883.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	46.3	U	46.3	160	200	ug/Kg
85-01-8	Phenanthrene	50.3	U	50.3	77.9	100	ug/Kg
120-12-7	Anthracene	50.6	U	50.6	77.9	100	ug/Kg
86-74-8	Carbazole	48.1	U	48.1	77.9	100	ug/Kg
84-74-2	Di-n-butylphthalate	50.5	U	50.5	77.9	100	ug/Kg
206-44-0	Fluoranthene	49	U	49	77.9	100	ug/Kg
92-87-5	Benzidine	97.7	U	97.7	160	200	ug/Kg
129-00-0	Pyrene	49.7	U	49.7	77.9	100	ug/Kg
85-68-7	Butylbenzylphthalate	58	U	58	77.9	100	ug/Kg
91-94-1	3,3-Dichlorobenzidine	59.1	U	59.1	160	200	ug/Kg
56-55-3	Benzo(a)anthracene	48.4	U	48.4	77.9	100	ug/Kg
218-01-9	Chrysene	47.6	U	47.6	77.9	100	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	54.5	U	54.5	77.9	100	ug/Kg
117-84-0	Di-n-octyl phthalate	65.9	U	65.9	160	200	ug/Kg
205-99-2	Benzo(b)fluoranthene	48.6	U	48.6	77.9	100	ug/Kg
207-08-9	Benzo(k)fluoranthene	49.5	U	49.5	77.9	100	ug/Kg
50-32-8	Benzo(a)pyrene	55.7	U	55.7	77.9	100	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	46.8	U	46.8	77.9	100	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	48.7	U	48.7	77.9	100	ug/Kg
191-24-2	Benzo(g,h,i)perylene	48	U	48	77.9	100	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	52	U	52	77.9	100	ug/Kg
123-91-1	1,4-Dioxane	65.9	U	65.9	77.9	100	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	44.8	U	44.8	77.9	100	ug/Kg
90-12-0	1-Methylnaphthalene	49.9	U	49.9	100	100	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	40.727		18-112		27	SPK: -150
13127-88-3	Phenol-d6	61.452		15-107		41	SPK: -150
4165-60-0	Nitrobenzene-d5	71.37		18-107		71	SPK: -100
321-60-8	2-Fluorobiphenyl	74.814		20-109		75	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-13	SDG No.:	BF101924
Lab Sample ID:	P4426-25	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.07	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139883.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	33.834		10-116		23	SPK: -150
1718-51-0	Terphenyl-d14	87.335		10-105		87	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	204683	6.887				
1146-65-2	Naphthalene-d8	784106	8.169				
15067-26-2	Acenaphthene-d10	407138	9.922				
1517-22-2	Phenanthrene-d10	608764	11.41				
1719-03-5	Chrysene-d12	270871	14.045				
1520-96-3	Perylene-d12	378751	15.527				

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-17	SDG No.:	BF101924
Lab Sample ID:	P4426-33	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.08	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139884.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	52	U	52	160	200	ug/Kg
110-86-1	Pyridine	47.9	U	47.9	77.9	100	ug/Kg
100-52-7	Benzaldehyde	110	U	110	160	200	ug/Kg
62-53-3	Aniline	58	U	58	77.9	100	ug/Kg
108-95-2	Phenol	87.7	J	49.7	77.9	100	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	50.1	U	50.1	77.9	100	ug/Kg
95-57-8	2-Chlorophenol	50	U	50	77.9	100	ug/Kg
95-50-1	1,2-Dichlorobenzene	50.7	U	50.7	77.9	100	ug/Kg
541-73-1	1,3-Dichlorobenzene	51.5	U	51.5	77.9	100	ug/Kg
106-46-7	1,4-Dichlorobenzene	51.8	U	51.8	77.9	100	ug/Kg
100-51-6	Benzyl Alcohol	49.7	U	49.7	160	200	ug/Kg
95-48-7	2-Methylphenol	48.3	U	48.3	77.9	100	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	54.5	U	54.5	77.9	100	ug/Kg
98-86-2	Acetophenone	61.3	J	52.1	77.9	100	ug/Kg
65794-96-9	3+4-Methylphenols	88.7	J	47.8	160	200	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	24.1	U	24.1	47.9	47.9	ug/Kg
67-72-1	Hexachloroethane	49.7	U	49.7	77.9	100	ug/Kg
98-95-3	Nitrobenzene	54.4	U	54.4	77.9	100	ug/Kg
78-59-1	Isophorone	50.7	U	50.7	77.9	100	ug/Kg
88-75-5	2-Nitrophenol	56.6	U	56.6	77.9	100	ug/Kg
105-67-9	2,4-Dimethylphenol	55.8	U	55.8	77.9	100	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	51.4	U	51.4	77.9	100	ug/Kg
120-83-2	2,4-Dichlorophenol	45.2	U	45.2	77.9	100	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	51.2	U	51.2	77.9	100	ug/Kg
65-85-0	Benzoic acid	140	U	140	160	200	ug/Kg
91-20-3	Naphthalene	49.5	U	49.5	77.9	100	ug/Kg
106-47-8	4-Chloroaniline	49.5	U	49.5	77.9	100	ug/Kg
87-68-3	Hexachlorobutadiene	49.9	U	49.9	77.9	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-17	SDG No.:	BF101924
Lab Sample ID:	P4426-33	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.08	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139884.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	52	U	52	160	200	ug/Kg
59-50-7	4-Chloro-3-methylphenol	46.4	U	46.4	77.9	100	ug/Kg
91-57-6	2-Methylnaphthalene	49.4	U	49.4	77.9	100	ug/Kg
77-47-4	Hexachlorocyclopentadiene	93.5	U	93.5	160	200	ug/Kg
88-06-2	2,4,6-Trichlorophenol	42.8	U	42.8	77.9	100	ug/Kg
95-95-4	2,4,5-Trichlorophenol	44.3	U	44.3	77.9	100	ug/Kg
92-52-4	1,1-Biphenyl	52.4	U	52.4	77.9	100	ug/Kg
91-58-7	2-Chloronaphthalene	49.9	U	49.9	77.9	100	ug/Kg
88-74-4	2-Nitroaniline	56.9	U	56.9	77.9	100	ug/Kg
131-11-3	Dimethylphthalate	48.9	U	48.9	77.9	100	ug/Kg
208-96-8	Acenaphthylene	51.8	U	51.8	77.9	100	ug/Kg
606-20-2	2,6-Dinitrotoluene	49.8	U	49.8	77.9	100	ug/Kg
99-09-2	3-Nitroaniline	53.4	U	53.4	77.9	100	ug/Kg
83-32-9	Acenaphthene	48.6	U	48.6	77.9	100	ug/Kg
Total PAH	Total PAH	2166	U	793.7	77.9	1600	ug/Kg
51-28-5	2,4-Dinitrophenol	150	U	150	160	200	ug/Kg
100-02-7	4-Nitrophenol	69.5	U	69.5	160	200	ug/Kg
132-64-9	Dibenzofuran	50.6	U	50.6	77.9	100	ug/Kg
121-14-2	2,4-Dinitrotoluene	51.6	U	51.6	77.9	100	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	101.4	U	101.4	77.9	200	ug/Kg
84-66-2	Diethylphthalate	48	U	48	77.9	100	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	51.3	U	51.3	77.9	100	ug/Kg
86-73-7	Fluorene	51.2	U	51.2	77.9	100	ug/Kg
100-01-6	4-Nitroaniline	64.1	U	64.1	77.9	100	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	70.1	U	70.1	160	200	ug/Kg
86-30-6	n-Nitrosodiphenylamine	48.9	U	48.9	77.9	100	ug/Kg
103-33-3	Azobenzene	47.7	U	47.7	77.9	100	ug/Kg
101-55-3	4-Bromophenyl-phenylether	47.3	U	47.3	77.9	100	ug/Kg
118-74-1	Hexachlorobenzene	50.9	U	50.9	77.9	100	ug/Kg
1912-24-9	Atrazine	54.8	U	54.8	77.9	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-17	SDG No.:	BF101924
Lab Sample ID:	P4426-33	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.08	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139884.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	46.3	U	46.3	160	200	ug/Kg
85-01-8	Phenanthrene	290		50.3	77.9	100	ug/Kg
120-12-7	Anthracene	50.6	U	50.6	77.9	100	ug/Kg
86-74-8	Carbazole	48.1	U	48.1	77.9	100	ug/Kg
84-74-2	Di-n-butylphthalate	50.5	U	50.5	77.9	100	ug/Kg
206-44-0	Fluoranthene	530		48.9	77.9	100	ug/Kg
92-87-5	Benzidine	97.6	U	97.6	160	200	ug/Kg
129-00-0	Pyrene	150		49.7	77.9	100	ug/Kg
85-68-7	Butylbenzylphthalate	58	U	58	77.9	100	ug/Kg
91-94-1	3,3-Dichlorobenzidine	59.1	U	59.1	160	200	ug/Kg
56-55-3	Benzo(a)anthracene	69.8	J	48.3	77.9	100	ug/Kg
218-01-9	Chrysene	370		47.6	77.9	100	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	54.5	U	54.5	77.9	100	ug/Kg
117-84-0	Di-n-octyl phthalate	65.9	U	65.9	160	200	ug/Kg
205-99-2	Benzo(b)fluoranthene	450		48.6	77.9	100	ug/Kg
207-08-9	Benzo(k)fluoranthene	170		49.5	77.9	100	ug/Kg
50-32-8	Benzo(a)pyrene	55.7	U	55.7	77.9	100	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	68.5	J	46.8	77.9	100	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	48.6	U	48.6	77.9	100	ug/Kg
191-24-2	Benzo(g,h,i)perylene	67.7	J	48	77.9	100	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	52	U	52	77.9	100	ug/Kg
123-91-1	1,4-Dioxane	65.9	U	65.9	77.9	100	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	44.7	U	44.7	77.9	100	ug/Kg
90-12-0	1-Methylnaphthalene	49.8	U	49.8	100	100	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	29.722		18-112		20	SPK: -150
13127-88-3	Phenol-d6	37.198		15-107		25	SPK: -150
4165-60-0	Nitrobenzene-d5	38.176		18-107		38	SPK: -100
321-60-8	2-Fluorobiphenyl	45.519		20-109		46	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-17	SDG No.:	BF101924
Lab Sample ID:	P4426-33	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.08	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139884.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	24.222		10-116		16	SPK: -150
1718-51-0	Terphenyl-d14	34.738		10-105		35	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	203314	6.887				
1146-65-2	Naphthalene-d8	753115	8.169				
15067-26-2	Acenaphthene-d10	357579	9.922				
1517-22-2	Phenanthrene-d10	505963	11.422				
1719-03-5	Chrysene-d12	403624	14.057				
1520-96-3	Perylene-d12	349762	15.533				

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-16	SDG No.:	BF101924
Lab Sample ID:	P4426-31	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.03	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139885.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	52	U	52	160	200	ug/Kg
110-86-1	Pyridine	47.9	U	47.9	78	100	ug/Kg
100-52-7	Benzaldehyde	110	U	110	160	200	ug/Kg
62-53-3	Aniline	58.1	U	58.1	78	100	ug/Kg
108-95-2	Phenol	49.7	U	49.7	78	100	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	50.2	U	50.2	78	100	ug/Kg
95-57-8	2-Chlorophenol	50.1	U	50.1	78	100	ug/Kg
95-50-1	1,2-Dichlorobenzene	50.7	U	50.7	78	100	ug/Kg
541-73-1	1,3-Dichlorobenzene	51.6	U	51.6	78	100	ug/Kg
106-46-7	1,4-Dichlorobenzene	51.9	U	51.9	78	100	ug/Kg
100-51-6	Benzyl Alcohol	49.8	U	49.8	160	200	ug/Kg
95-48-7	2-Methylphenol	48.3	U	48.3	78	100	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	54.5	U	54.5	78	100	ug/Kg
98-86-2	Acetophenone	52.1	U	52.1	78	100	ug/Kg
65794-96-9	3+4-Methylphenols	47.9	U	47.9	160	200	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	24.2	U	24.2	48	48	ug/Kg
67-72-1	Hexachloroethane	49.8	U	49.8	78	100	ug/Kg
98-95-3	Nitrobenzene	54.4	U	54.4	78	100	ug/Kg
78-59-1	Isophorone	50.7	U	50.7	78	100	ug/Kg
88-75-5	2-Nitrophenol	56.7	U	56.7	78	100	ug/Kg
105-67-9	2,4-Dimethylphenol	55.9	U	55.9	78	100	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	51.4	U	51.4	78	100	ug/Kg
120-83-2	2,4-Dichlorophenol	45.3	U	45.3	78	100	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	51.2	U	51.2	78	100	ug/Kg
65-85-0	Benzoic acid	140	U	140	160	200	ug/Kg
91-20-3	Naphthalene	49.5	U	49.5	78	100	ug/Kg
106-47-8	4-Chloroaniline	49.5	U	49.5	78	100	ug/Kg
87-68-3	Hexachlorobutadiene	50	U	50	78	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-16	SDG No.:	BF101924
Lab Sample ID:	P4426-31	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.03 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139885.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	52	U	52	160	200	ug/Kg
59-50-7	4-Chloro-3-methylphenol	46.5	U	46.5	78	100	ug/Kg
91-57-6	2-Methylnaphthalene	49.5	U	49.5	78	100	ug/Kg
77-47-4	Hexachlorocyclopentadiene	93.5	U	93.5	160	200	ug/Kg
88-06-2	2,4,6-Trichlorophenol	42.8	U	42.8	78	100	ug/Kg
95-95-4	2,4,5-Trichlorophenol	44.4	U	44.4	78	100	ug/Kg
92-52-4	1,1-Biphenyl	52.4	U	52.4	78	100	ug/Kg
91-58-7	2-Chloronaphthalene	50	U	50	78	100	ug/Kg
88-74-4	2-Nitroaniline	57	U	57	78	100	ug/Kg
131-11-3	Dimethylphthalate	49	U	49	78	100	ug/Kg
208-96-8	Acenaphthylene	51.9	U	51.9	78	100	ug/Kg
606-20-2	2,6-Dinitrotoluene	49.9	U	49.9	78	100	ug/Kg
99-09-2	3-Nitroaniline	53.5	U	53.5	78	100	ug/Kg
83-32-9	Acenaphthene	48.6	U	48.6	78	100	ug/Kg
Total PAH	Total PAH	794.6	U	794.6	78	1600	ug/Kg
51-28-5	2,4-Dinitrophenol	150	U	150	160	200	ug/Kg
100-02-7	4-Nitrophenol	69.6	U	69.6	160	200	ug/Kg
132-64-9	Dibenzofuran	50.6	U	50.6	78	100	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	101.6	U	101.6	78	200	ug/Kg
121-14-2	2,4-Dinitrotoluene	51.7	U	51.7	78	100	ug/Kg
84-66-2	Diethylphthalate	48	U	48	78	100	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	51.3	U	51.3	78	100	ug/Kg
86-73-7	Fluorene	51.3	U	51.3	78	100	ug/Kg
100-01-6	4-Nitroaniline	64.2	U	64.2	78	100	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	70.2	U	70.2	160	200	ug/Kg
86-30-6	n-Nitrosodiphenylamine	48.9	U	48.9	78	100	ug/Kg
103-33-3	Azobenzene	47.7	U	47.7	78	100	ug/Kg
101-55-3	4-Bromophenyl-phenylether	47.3	U	47.3	78	100	ug/Kg
118-74-1	Hexachlorobenzene	51	U	51	78	100	ug/Kg
1912-24-9	Atrazine	54.8	U	54.8	78	100	ug/Kg

Report of Analysis

Client:			Date Collected:	10/16/2024 12:00:00 AM	
Project:			Date Received:	10/16/2024 12:00:00 AM	
Client Sample ID:	PAD-16		SDG No.:	BF101924	
Lab Sample ID:	P4426-31		Matrix:	Solid	
Analytical Method:	8270E		% Moisture:	0	
Sample Wt/Vol:	50.03	Units: g	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:		
Extraction Type :		Decanted :	Level :	LOW	
Injection Volume :		GPC Factor :	GPC Cleanup :	PH :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139885.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	46.4	U	46.4	160	200	ug/Kg
85-01-8	Phenanthrene	50.4	U	50.4	78	100	ug/Kg
120-12-7	Anthracene	50.6	U	50.6	78	100	ug/Kg
86-74-8	Carbazole	48.2	U	48.2	78	100	ug/Kg
84-74-2	Di-n-butylphthalate	55.3	J	50.5	78	100	ug/Kg
206-44-0	Fluoranthene	49	U	49	78	100	ug/Kg
92-87-5	Benzidine	97.7	U	97.7	160	200	ug/Kg
129-00-0	Pyrene	49.8	U	49.8	78	100	ug/Kg
85-68-7	Butylbenzylphthalate	58	U	58	78	100	ug/Kg
91-94-1	3,3-Dichlorobenzidine	59.1	U	59.1	160	200	ug/Kg
56-55-3	Benzo(a)anthracene	48.4	U	48.4	78	100	ug/Kg
218-01-9	Chrysene	47.7	U	47.7	78	100	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	54.6	U	54.6	78	100	ug/Kg
117-84-0	Di-n-octyl phthalate	66	U	66	160	200	ug/Kg
205-99-2	Benzo(b)fluoranthene	48.6	U	48.6	78	100	ug/Kg
207-08-9	Benzo(k)fluoranthene	49.5	U	49.5	78	100	ug/Kg
50-32-8	Benzo(a)pyrene	55.8	U	55.8	78	100	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	46.8	U	46.8	78	100	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	48.7	U	48.7	78	100	ug/Kg
191-24-2	Benzo(g,h,i)perylene	48	U	48	78	100	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	52	U	52	78	100	ug/Kg
123-91-1	1,4-Dioxane	66	U	66	78	100	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	44.8	U	44.8	78	100	ug/Kg
90-12-0	1-Methylnaphthalene	49.9	U	49.9	100	100	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	16.898	*	18-112		11	SPK: -150
13127-88-3	Phenol-d6	27.739		15-107		18	SPK: -150
4165-60-0	Nitrobenzene-d5	41.049		18-107		41	SPK: -100
321-60-8	2-Fluorobiphenyl	49.1		20-109		49	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-16	SDG No.:	BF101924
Lab Sample ID:	P4426-31	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.03	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139885.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	10.606	*	10-116		7	SPK: -150
1718-51-0	Terphenyl-d14	38.957		10-105		39	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	195373	6.887				
1146-65-2	Naphthalene-d8	714977	8.169				
15067-26-2	Acenaphthene-d10	338275	9.922				
1517-22-2	Phenanthrene-d10	475082	11.41				
1719-03-5	Chrysene-d12	372525	14.045				
1520-96-3	Perylene-d12	269044	15.527				

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-12	SDG No.:	BF101924
Lab Sample ID:	P4426-23	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.09	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139886.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	52	U	52	160	200	ug/Kg
110-86-1	Pyridine	47.9	U	47.9	77.9	100	ug/Kg
100-52-7	Benzaldehyde	110	U	110	160	200	ug/Kg
62-53-3	Aniline	58	U	58	77.9	100	ug/Kg
108-95-2	Phenol	49.7	U	49.7	77.9	100	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	50.1	U	50.1	77.9	100	ug/Kg
95-57-8	2-Chlorophenol	50	U	50	77.9	100	ug/Kg
95-50-1	1,2-Dichlorobenzene	50.7	U	50.7	77.9	100	ug/Kg
541-73-1	1,3-Dichlorobenzene	51.5	U	51.5	77.9	100	ug/Kg
106-46-7	1,4-Dichlorobenzene	51.8	U	51.8	77.9	100	ug/Kg
100-51-6	Benzyl Alcohol	49.7	U	49.7	160	200	ug/Kg
95-48-7	2-Methylphenol	48.3	U	48.3	77.9	100	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	54.4	U	54.4	77.9	100	ug/Kg
98-86-2	Acetophenone	52	U	52	77.9	100	ug/Kg
65794-96-9	3+4-Methylphenols	47.8	U	47.8	160	200	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	24.1	U	24.1	47.9	47.9	ug/Kg
67-72-1	Hexachloroethane	49.7	U	49.7	77.9	100	ug/Kg
98-95-3	Nitrobenzene	54.4	U	54.4	77.9	100	ug/Kg
78-59-1	Isophorone	50.7	U	50.7	77.9	100	ug/Kg
88-75-5	2-Nitrophenol	56.6	U	56.6	77.9	100	ug/Kg
105-67-9	2,4-Dimethylphenol	55.8	U	55.8	77.9	100	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	51.4	U	51.4	77.9	100	ug/Kg
120-83-2	2,4-Dichlorophenol	45.2	U	45.2	77.9	100	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	51.1	U	51.1	77.9	100	ug/Kg
65-85-0	Benzoic acid	140	U	140	160	200	ug/Kg
91-20-3	Naphthalene	49.5	U	49.5	77.9	100	ug/Kg
106-47-8	4-Chloroaniline	49.5	U	49.5	77.9	100	ug/Kg
87-68-3	Hexachlorobutadiene	49.9	U	49.9	77.9	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-12	SDG No.:	BF101924
Lab Sample ID:	P4426-23	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.09	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139886.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	52	U	52	160	200	ug/Kg
59-50-7	4-Chloro-3-methylphenol	46.4	U	46.4	77.9	100	ug/Kg
91-57-6	2-Methylnaphthalene	49.4	U	49.4	77.9	100	ug/Kg
77-47-4	Hexachlorocyclopentadiene	93.4	U	93.4	160	200	ug/Kg
88-06-2	2,4,6-Trichlorophenol	42.8	U	42.8	77.9	100	ug/Kg
95-95-4	2,4,5-Trichlorophenol	44.3	U	44.3	77.9	100	ug/Kg
92-52-4	1,1-Biphenyl	52.3	U	52.3	77.9	100	ug/Kg
91-58-7	2-Chloronaphthalene	49.9	U	49.9	77.9	100	ug/Kg
88-74-4	2-Nitroaniline	56.9	U	56.9	77.9	100	ug/Kg
131-11-3	Dimethylphthalate	48.9	U	48.9	77.9	100	ug/Kg
208-96-8	Acenaphthylene	51.8	U	51.8	77.9	100	ug/Kg
606-20-2	2,6-Dinitrotoluene	49.8	U	49.8	77.9	100	ug/Kg
99-09-2	3-Nitroaniline	53.4	U	53.4	77.9	100	ug/Kg
83-32-9	Acenaphthene	48.6	U	48.6	77.9	100	ug/Kg
Total PAH	Total PAH	793.6	U	793.6	77.9	1600	ug/Kg
51-28-5	2,4-Dinitrophenol	150	U	150	160	200	ug/Kg
100-02-7	4-Nitrophenol	69.5	U	69.5	160	200	ug/Kg
132-64-9	Dibenzofuran	50.5	U	50.5	77.9	100	ug/Kg
121-14-2	2,4-Dinitrotoluene	51.6	U	51.6	77.9	100	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	101.4	U	101.4	77.9	200	ug/Kg
84-66-2	Diethylphthalate	48	U	48	77.9	100	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	51.3	U	51.3	77.9	100	ug/Kg
86-73-7	Fluorene	51.2	U	51.2	77.9	100	ug/Kg
100-01-6	4-Nitroaniline	64.1	U	64.1	77.9	100	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	70.1	U	70.1	160	200	ug/Kg
86-30-6	n-Nitrosodiphenylamine	48.9	U	48.9	77.9	100	ug/Kg
103-33-3	Azobenzene	47.7	U	47.7	77.9	100	ug/Kg
101-55-3	4-Bromophenyl-phenylether	47.3	U	47.3	77.9	100	ug/Kg
118-74-1	Hexachlorobenzene	50.9	U	50.9	77.9	100	ug/Kg
1912-24-9	Atrazine	54.7	U	54.7	77.9	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-12	SDG No.:	BF101924
Lab Sample ID:	P4426-23	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.09	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139886.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	46.3	U	46.3	160	200	ug/Kg
85-01-8	Phenanthrene	50.3	U	50.3	77.9	100	ug/Kg
120-12-7	Anthracene	50.5	U	50.5	77.9	100	ug/Kg
86-74-8	Carbazole	48.1	U	48.1	77.9	100	ug/Kg
84-74-2	Di-n-butylphthalate	50.5	U	50.5	77.9	100	ug/Kg
206-44-0	Fluoranthene	48.9	U	48.9	77.9	100	ug/Kg
92-87-5	Benzidine	97.6	U	97.6	160	200	ug/Kg
129-00-0	Pyrene	49.7	U	49.7	77.9	100	ug/Kg
85-68-7	Butylbenzylphthalate	58	U	58	77.9	100	ug/Kg
91-94-1	3,3-Dichlorobenzidine	59.1	U	59.1	160	200	ug/Kg
56-55-3	Benzo(a)anthracene	48.3	U	48.3	77.9	100	ug/Kg
218-01-9	Chrysene	47.6	U	47.6	77.9	100	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	54.5	U	54.5	77.9	100	ug/Kg
117-84-0	Di-n-octyl phthalate	65.9	U	65.9	160	200	ug/Kg
205-99-2	Benzo(b)fluoranthene	48.6	U	48.6	77.9	100	ug/Kg
207-08-9	Benzo(k)fluoranthene	49.5	U	49.5	77.9	100	ug/Kg
50-32-8	Benzo(a)pyrene	55.7	U	55.7	77.9	100	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	46.8	U	46.8	77.9	100	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	48.6	U	48.6	77.9	100	ug/Kg
191-24-2	Benzo(g,h,i)perylene	48	U	48	77.9	100	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	52	U	52	77.9	100	ug/Kg
123-91-1	1,4-Dioxane	65.9	U	65.9	77.9	100	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	44.7	U	44.7	77.9	100	ug/Kg
90-12-0	1-Methylnaphthalene	49.8	U	49.8	100	100	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	20.262	*	18-112		14	SPK: -150
13127-88-3	Phenol-d6	33.682		15-107		22	SPK: -150
4165-60-0	Nitrobenzene-d5	47.71		18-107		48	SPK: -100
321-60-8	2-Fluorobiphenyl	54.703		20-109		55	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-12	SDG No.:	BF101924
Lab Sample ID:	P4426-23	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.09	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139886.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	11.998	*	10-116		8	SPK: -150
1718-51-0	Terphenyl-d14	68.596		10-105		69	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	197123	6.892				
1146-65-2	Naphthalene-d8	719768	8.169				
15067-26-2	Acenaphthene-d10	338510	9.922				
1517-22-2	Phenanthrene-d10	490725	11.41				
1719-03-5	Chrysene-d12	235740	14.051				
1520-96-3	Perylene-d12	275588	15.527				

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-3	SDG No.:	BF101924
Lab Sample ID:	P4426-05	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.01	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139887.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	52.1	U	52.1	160	200	ug/Kg
110-86-1	Pyridine	47.9	U	47.9	78	100	ug/Kg
100-52-7	Benzaldehyde	110	U	110	160	200	ug/Kg
62-53-3	Aniline	58.1	U	58.1	78	100	ug/Kg
108-95-2	Phenol	49.7	U	49.7	78	100	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	50.2	U	50.2	78	100	ug/Kg
95-57-8	2-Chlorophenol	50.1	U	50.1	78	100	ug/Kg
95-50-1	1,2-Dichlorobenzene	50.7	U	50.7	78	100	ug/Kg
541-73-1	1,3-Dichlorobenzene	51.6	U	51.6	78	100	ug/Kg
106-46-7	1,4-Dichlorobenzene	51.9	U	51.9	78	100	ug/Kg
100-51-6	Benzyl Alcohol	49.8	U	49.8	160	200	ug/Kg
95-48-7	2-Methylphenol	48.4	U	48.4	78	100	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	54.5	U	54.5	78	100	ug/Kg
98-86-2	Acetophenone	52.1	U	52.1	78	100	ug/Kg
65794-96-9	3+4-Methylphenols	47.9	U	47.9	160	200	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	24.2	U	24.2	48	48	ug/Kg
67-72-1	Hexachloroethane	49.8	U	49.8	78	100	ug/Kg
98-95-3	Nitrobenzene	54.5	U	54.5	78	100	ug/Kg
78-59-1	Isophorone	50.7	U	50.7	78	100	ug/Kg
88-75-5	2-Nitrophenol	56.7	U	56.7	78	100	ug/Kg
105-67-9	2,4-Dimethylphenol	55.9	U	55.9	78	100	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	51.5	U	51.5	78	100	ug/Kg
120-83-2	2,4-Dichlorophenol	45.3	U	45.3	78	100	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	51.2	U	51.2	78	100	ug/Kg
65-85-0	Benzoic acid	140	U	140	160	200	ug/Kg
91-20-3	Naphthalene	49.6	U	49.6	78	100	ug/Kg
106-47-8	4-Chloroaniline	49.6	U	49.6	78	100	ug/Kg
87-68-3	Hexachlorobutadiene	50	U	50	78	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-3	SDG No.:	BF101924
Lab Sample ID:	P4426-05	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.01	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139887.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	52.1	U	52.1	160	200	ug/Kg
59-50-7	4-Chloro-3-methylphenol	46.5	U	46.5	78	100	ug/Kg
91-57-6	2-Methylnaphthalene	49.5	U	49.5	78	100	ug/Kg
77-47-4	Hexachlorocyclopentadiene	93.6	U	93.6	160	200	ug/Kg
88-06-2	2,4,6-Trichlorophenol	42.8	U	42.8	78	100	ug/Kg
95-95-4	2,4,5-Trichlorophenol	44.4	U	44.4	78	100	ug/Kg
92-52-4	1,1-Biphenyl	52.4	U	52.4	78	100	ug/Kg
91-58-7	2-Chloronaphthalene	50	U	50	78	100	ug/Kg
88-74-4	2-Nitroaniline	57	U	57	78	100	ug/Kg
131-11-3	Dimethylphthalate	49	U	49	78	100	ug/Kg
208-96-8	Acenaphthylene	51.9	U	51.9	78	100	ug/Kg
606-20-2	2,6-Dinitrotoluene	49.9	U	49.9	78	100	ug/Kg
99-09-2	3-Nitroaniline	53.5	U	53.5	78	100	ug/Kg
83-32-9	Acenaphthene	48.7	U	48.7	78	100	ug/Kg
Total PAH	Total PAH	795.2	U	795.2	78	1600	ug/Kg
51-28-5	2,4-Dinitrophenol	150	U	150	160	200	ug/Kg
100-02-7	4-Nitrophenol	69.6	U	69.6	160	200	ug/Kg
132-64-9	Dibenzofuran	50.6	U	50.6	78	100	ug/Kg
121-14-2	2,4-Dinitrotoluene	51.7	U	51.7	78	100	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	101.6	U	101.6	78	200	ug/Kg
84-66-2	Diethylphthalate	48.1	U	48.1	78	100	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	51.3	U	51.3	78	100	ug/Kg
86-73-7	Fluorene	51.3	U	51.3	78	100	ug/Kg
100-01-6	4-Nitroaniline	64.2	U	64.2	78	100	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	70.2	U	70.2	160	200	ug/Kg
86-30-6	n-Nitrosodiphenylamine	49	U	49	78	100	ug/Kg
103-33-3	Azobenzene	47.8	U	47.8	78	100	ug/Kg
101-55-3	4-Bromophenyl-phenylether	47.3	U	47.3	78	100	ug/Kg
118-74-1	Hexachlorobenzene	51	U	51	78	100	ug/Kg
1912-24-9	Atrazine	54.8	U	54.8	78	100	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-3	SDG No.:	BF101924
Lab Sample ID:	P4426-05	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.01	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139887.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	46.4	U	46.4	160	200	ug/Kg
85-01-8	Phenanthrene	50.4	U	50.4	78	100	ug/Kg
120-12-7	Anthracene	50.6	U	50.6	78	100	ug/Kg
86-74-8	Carbazole	48.2	U	48.2	78	100	ug/Kg
84-74-2	Di-n-butylphthalate	50.6	U	50.6	78	100	ug/Kg
206-44-0	Fluoranthene	49	U	49	78	100	ug/Kg
92-87-5	Benzidine	97.8	U	97.8	160	200	ug/Kg
129-00-0	Pyrene	49.8	U	49.8	78	100	ug/Kg
85-68-7	Butylbenzylphthalate	58.1	U	58.1	78	100	ug/Kg
91-94-1	3,3-Dichlorobenzidine	59.1	U	59.1	160	200	ug/Kg
56-55-3	Benzo(a)anthracene	48.4	U	48.4	78	100	ug/Kg
218-01-9	Chrysene	47.7	U	47.7	78	100	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	54.6	U	54.6	78	100	ug/Kg
117-84-0	Di-n-octyl phthalate	66	U	66	160	200	ug/Kg
205-99-2	Benzo(b)fluoranthene	48.7	U	48.7	78	100	ug/Kg
207-08-9	Benzo(k)fluoranthene	49.6	U	49.6	78	100	ug/Kg
50-32-8	Benzo(a)pyrene	55.8	U	55.8	78	100	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	46.9	U	46.9	78	100	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	48.7	U	48.7	78	100	ug/Kg
191-24-2	Benzo(g,h,i)perylene	48.1	U	48.1	78	100	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	52.1	U	52.1	78	100	ug/Kg
123-91-1	1,4-Dioxane	66	U	66	78	100	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	44.8	U	44.8	78	100	ug/Kg
90-12-0	1-Methylnaphthalene	49.9	U	49.9	100	100	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	27.592		18-112		18	SPK: -150
13127-88-3	Phenol-d6	41.101		15-107		27	SPK: -150
4165-60-0	Nitrobenzene-d5	46.935		18-107		47	SPK: -100
321-60-8	2-Fluorobiphenyl	52.671		20-109		53	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-3	SDG No.:	BF101924
Lab Sample ID:	P4426-05	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.01	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139887.D	1	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	14.717		10-116		10	SPK: -150
1718-51-0	Terphenyl-d14	41.717		10-105		42	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	201874	6.893				
1146-65-2	Naphthalene-d8	725305	8.169				
15067-26-2	Acenaphthene-d10	339096	9.928				
1517-22-2	Phenanthrene-d10	498866	11.41				
1719-03-5	Chrysene-d12	384159	14.051				
1520-96-3	Perylene-d12	273829	15.527				

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-14	SDG No.:	BF101924
Lab Sample ID:	P4426-27	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.06	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139888.D	5	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	260	U	260	810	990	ug/Kg
110-86-1	Pyridine	240	U	240	390	510	ug/Kg
100-52-7	Benzaldehyde	550	U	550	810	990	ug/Kg
62-53-3	Aniline	290	U	290	390	510	ug/Kg
108-95-2	Phenol	250	U	250	390	510	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	250	U	250	390	510	ug/Kg
95-57-8	2-Chlorophenol	250	U	250	390	510	ug/Kg
95-50-1	1,2-Dichlorobenzene	250	U	250	390	510	ug/Kg
541-73-1	1,3-Dichlorobenzene	260	U	260	390	510	ug/Kg
106-46-7	1,4-Dichlorobenzene	260	U	260	390	510	ug/Kg
100-51-6	Benzyl Alcohol	250	U	250	810	990	ug/Kg
95-48-7	2-Methylphenol	240	U	240	390	510	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	270	U	270	390	510	ug/Kg
98-86-2	Acetophenone	260	U	260	390	510	ug/Kg
65794-96-9	3+4-Methylphenols	240	U	240	810	990	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	120	U	120	240	240	ug/Kg
67-72-1	Hexachloroethane	250	U	250	390	510	ug/Kg
98-95-3	Nitrobenzene	270	U	270	390	510	ug/Kg
78-59-1	Isophorone	250	U	250	390	510	ug/Kg
88-75-5	2-Nitrophenol	280	U	280	390	510	ug/Kg
105-67-9	2,4-Dimethylphenol	280	U	280	390	510	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	260	U	260	390	510	ug/Kg
120-83-2	2,4-Dichlorophenol	230	U	230	390	510	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	260	U	260	390	510	ug/Kg
65-85-0	Benzoic acid	710	U	710	810	990	ug/Kg
91-20-3	Naphthalene	250	U	250	390	510	ug/Kg
106-47-8	4-Chloroaniline	250	U	250	390	510	ug/Kg
87-68-3	Hexachlorobutadiene	250	U	250	390	510	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-14	SDG No.:	BF101924
Lab Sample ID:	P4426-27	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.06	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139888.D	5	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	260	U	260	810	990	ug/Kg
59-50-7	4-Chloro-3-methylphenol	230	U	230	390	510	ug/Kg
91-57-6	2-Methylnaphthalene	250	U	250	390	510	ug/Kg
77-47-4	Hexachlorocyclopentadiene	470	U	470	810	990	ug/Kg
88-06-2	2,4,6-Trichlorophenol	210	U	210	390	510	ug/Kg
95-95-4	2,4,5-Trichlorophenol	220	U	220	390	510	ug/Kg
92-52-4	1,1-Biphenyl	260	U	260	390	510	ug/Kg
91-58-7	2-Chloronaphthalene	250	U	250	390	510	ug/Kg
88-74-4	2-Nitroaniline	280	U	280	390	510	ug/Kg
131-11-3	Dimethylphthalate	240	U	240	390	510	ug/Kg
208-96-8	Acenaphthylene	260	U	260	390	510	ug/Kg
606-20-2	2,6-Dinitrotoluene	250	U	250	390	510	ug/Kg
99-09-2	3-Nitroaniline	270	U	270	390	510	ug/Kg
83-32-9	Acenaphthene	240	U	240	390	510	ug/Kg
Total PAH	Total PAH	3960	U	3960	390	8160	ug/Kg
51-28-5	2,4-Dinitrophenol	730	U	730	810	990	ug/Kg
100-02-7	4-Nitrophenol	350	U	350	810	990	ug/Kg
132-64-9	Dibenzofuran	250	U	250	390	510	ug/Kg
121-14-2	2,4-Dinitrotoluene	260	U	260	390	510	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	510	U	510	390	1020	ug/Kg
84-66-2	Diethylphthalate	240	U	240	390	510	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	260	U	260	390	510	ug/Kg
86-73-7	Fluorene	260	U	260	390	510	ug/Kg
100-01-6	4-Nitroaniline	320	U	320	390	510	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	350	U	350	810	990	ug/Kg
86-30-6	n-Nitrosodiphenylamine	240	U	240	390	510	ug/Kg
103-33-3	Azobenzene	240	U	240	390	510	ug/Kg
101-55-3	4-Bromophenyl-phenylether	240	U	240	390	510	ug/Kg
118-74-1	Hexachlorobenzene	250	U	250	390	510	ug/Kg
1912-24-9	Atrazine	270	U	270	390	510	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-14	SDG No.:	BF101924
Lab Sample ID:	P4426-27	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.06	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139888.D	5	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	230	U	230	810	990	ug/Kg
85-01-8	Phenanthrene	250	U	250	390	510	ug/Kg
120-12-7	Anthracene	250	U	250	390	510	ug/Kg
86-74-8	Carbazole	240	U	240	390	510	ug/Kg
84-74-2	Di-n-butylphthalate	250	U	250	390	510	ug/Kg
206-44-0	Fluoranthene	240	U	240	390	510	ug/Kg
92-87-5	Benzidine	490	U	490	810	990	ug/Kg
129-00-0	Pyrene	250	U	250	390	510	ug/Kg
85-68-7	Butylbenzylphthalate	290	U	290	390	510	ug/Kg
91-94-1	3,3-Dichlorobenzidine	300	U	300	810	990	ug/Kg
56-55-3	Benzo(a)anthracene	240	U	240	390	510	ug/Kg
218-01-9	Chrysene	240	U	240	390	510	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	270	U	270	390	510	ug/Kg
117-84-0	Di-n-octyl phthalate	330	U	330	810	990	ug/Kg
205-99-2	Benzo(b)fluoranthene	240	U	240	390	510	ug/Kg
207-08-9	Benzo(k)fluoranthene	250	U	250	390	510	ug/Kg
50-32-8	Benzo(a)pyrene	280	U	280	390	510	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	230	U	230	390	510	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	240	U	240	390	510	ug/Kg
191-24-2	Benzo(g,h,i)perylene	240	U	240	390	510	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	260	U	260	390	510	ug/Kg
123-91-1	1,4-Dioxane	330	U	330	390	510	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	220	U	220	390	510	ug/Kg
90-12-0	1-Methylnaphthalene	250	U	250	510	510	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	56.285		18-112		38	SPK: -150
13127-88-3	Phenol-d6	56.53		15-107		38	SPK: -150
4165-60-0	Nitrobenzene-d5	46.56		18-107		47	SPK: -100
321-60-8	2-Fluorobiphenyl	57.235		20-109		57	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-14	SDG No.:	BF101924
Lab Sample ID:	P4426-27	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.06	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139888.D	5	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	50.245		10-116		33	SPK: -150
1718-51-0	Terphenyl-d14	48.585		10-105		49	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	170784	6.887				
1146-65-2	Naphthalene-d8	588934	8.169				
15067-26-2	Acenaphthene-d10	263847	9.922				
1517-22-2	Phenanthrene-d10	437146	11.41				
1719-03-5	Chrysene-d12	297758	14.051				
1520-96-3	Perylene-d12	222325	15.533				

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-1	SDG No.:	BF101924
Lab Sample ID:	P4426-01	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.06	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139889.D	5	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	260	U	260	810	990	ug/Kg
110-86-1	Pyridine	240	U	240	390	510	ug/Kg
100-52-7	Benzaldehyde	550	U	550	810	990	ug/Kg
62-53-3	Aniline	290	U	290	390	510	ug/Kg
108-95-2	Phenol	250	U	250	390	510	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	250	U	250	390	510	ug/Kg
95-57-8	2-Chlorophenol	250	U	250	390	510	ug/Kg
95-50-1	1,2-Dichlorobenzene	250	U	250	390	510	ug/Kg
541-73-1	1,3-Dichlorobenzene	260	U	260	390	510	ug/Kg
106-46-7	1,4-Dichlorobenzene	260	U	260	390	510	ug/Kg
100-51-6	Benzyl Alcohol	250	U	250	810	990	ug/Kg
95-48-7	2-Methylphenol	240	U	240	390	510	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	270	U	270	390	510	ug/Kg
98-86-2	Acetophenone	260	U	260	390	510	ug/Kg
65794-96-9	3+4-Methylphenols	240	U	240	810	990	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	120	U	120	240	240	ug/Kg
67-72-1	Hexachloroethane	250	U	250	390	510	ug/Kg
98-95-3	Nitrobenzene	270	U	270	390	510	ug/Kg
78-59-1	Isophorone	250	U	250	390	510	ug/Kg
88-75-5	2-Nitrophenol	280	U	280	390	510	ug/Kg
105-67-9	2,4-Dimethylphenol	280	U	280	390	510	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	260	U	260	390	510	ug/Kg
120-83-2	2,4-Dichlorophenol	230	U	230	390	510	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	260	U	260	390	510	ug/Kg
65-85-0	Benzoic acid	710	U	710	810	990	ug/Kg
91-20-3	Naphthalene	250	U	250	390	510	ug/Kg
106-47-8	4-Chloroaniline	250	U	250	390	510	ug/Kg
87-68-3	Hexachlorobutadiene	250	U	250	390	510	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-1	SDG No.:	BF101924
Lab Sample ID:	P4426-01	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.06	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139889.D	5	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	260	U	260	810	990	ug/Kg
59-50-7	4-Chloro-3-methylphenol	230	U	230	390	510	ug/Kg
91-57-6	2-Methylnaphthalene	250	U	250	390	510	ug/Kg
77-47-4	Hexachlorocyclopentadiene	470	U	470	810	990	ug/Kg
88-06-2	2,4,6-Trichlorophenol	210	U	210	390	510	ug/Kg
95-95-4	2,4,5-Trichlorophenol	220	U	220	390	510	ug/Kg
92-52-4	1,1-Biphenyl	260	U	260	390	510	ug/Kg
91-58-7	2-Chloronaphthalene	250	U	250	390	510	ug/Kg
88-74-4	2-Nitroaniline	280	U	280	390	510	ug/Kg
131-11-3	Dimethylphthalate	240	U	240	390	510	ug/Kg
208-96-8	Acenaphthylene	260	U	260	390	510	ug/Kg
606-20-2	2,6-Dinitrotoluene	250	U	250	390	510	ug/Kg
99-09-2	3-Nitroaniline	270	U	270	390	510	ug/Kg
83-32-9	Acenaphthene	240	U	240	390	510	ug/Kg
Total PAH	Total PAH	3960	U	3960	390	8160	ug/Kg
51-28-5	2,4-Dinitrophenol	730	U	730	810	990	ug/Kg
100-02-7	4-Nitrophenol	350	U	350	810	990	ug/Kg
132-64-9	Dibenzofuran	250	U	250	390	510	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	510	U	510	390	1020	ug/Kg
121-14-2	2,4-Dinitrotoluene	260	U	260	390	510	ug/Kg
84-66-2	Diethylphthalate	240	U	240	390	510	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	260	U	260	390	510	ug/Kg
86-73-7	Fluorene	260	U	260	390	510	ug/Kg
100-01-6	4-Nitroaniline	320	U	320	390	510	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	350	U	350	810	990	ug/Kg
86-30-6	n-Nitrosodiphenylamine	240	U	240	390	510	ug/Kg
103-33-3	Azobenzene	240	U	240	390	510	ug/Kg
101-55-3	4-Bromophenyl-phenylether	240	U	240	390	510	ug/Kg
118-74-1	Hexachlorobenzene	250	U	250	390	510	ug/Kg
1912-24-9	Atrazine	270	U	270	390	510	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-1	SDG No.:	BF101924
Lab Sample ID:	P4426-01	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.06	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139889.D	5	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	230	U	230	810	990	ug/Kg
85-01-8	Phenanthrene	250	U	250	390	510	ug/Kg
120-12-7	Anthracene	250	U	250	390	510	ug/Kg
86-74-8	Carbazole	240	U	240	390	510	ug/Kg
84-74-2	Di-n-butylphthalate	250	U	250	390	510	ug/Kg
206-44-0	Fluoranthene	240	U	240	390	510	ug/Kg
92-87-5	Benzidine	490	U	490	810	990	ug/Kg
129-00-0	Pyrene	250	U	250	390	510	ug/Kg
85-68-7	Butylbenzylphthalate	290	U	290	390	510	ug/Kg
91-94-1	3,3-Dichlorobenzidine	300	U	300	810	990	ug/Kg
56-55-3	Benzo(a)anthracene	240	U	240	390	510	ug/Kg
218-01-9	Chrysene	240	U	240	390	510	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	270	U	270	390	510	ug/Kg
117-84-0	Di-n-octyl phthalate	330	U	330	810	990	ug/Kg
205-99-2	Benzo(b)fluoranthene	240	U	240	390	510	ug/Kg
207-08-9	Benzo(k)fluoranthene	250	U	250	390	510	ug/Kg
50-32-8	Benzo(a)pyrene	280	U	280	390	510	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	230	U	230	390	510	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	240	U	240	390	510	ug/Kg
191-24-2	Benzo(g,h,i)perylene	240	U	240	390	510	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	260	U	260	390	510	ug/Kg
123-91-1	1,4-Dioxane	330	U	330	390	510	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	220	U	220	390	510	ug/Kg
90-12-0	1-Methylnaphthalene	250	U	250	510	510	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	20.42	*	18-112		14	SPK: -150
13127-88-3	Phenol-d6	27.385		15-107		18	SPK: -150
4165-60-0	Nitrobenzene-d5	33.31		18-107		33	SPK: -100
321-60-8	2-Fluorobiphenyl	41.445		20-109		41	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-1	SDG No.:	BF101924
Lab Sample ID:	P4426-01	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.06	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139889.D	5	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	18.825		10-116		13	SPK: -150
1718-51-0	Terphenyl-d14	39.28		10-105		39	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	166324	6.892				
1146-65-2	Naphthalene-d8	554266	8.169				
15067-26-2	Acenaphthene-d10	249336	9.928				
1517-22-2	Phenanthrene-d10	436102	11.41				
1719-03-5	Chrysene-d12	274554	14.051				
1520-96-3	Perylene-d12	211115	15.533				

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-5	SDG No.:	BF101924
Lab Sample ID:	P4426-09	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.06	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139890.D	5	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	260	U	260	810	990	ug/Kg
110-86-1	Pyridine	240	U	240	390	510	ug/Kg
100-52-7	Benzaldehyde	550	U	550	810	990	ug/Kg
62-53-3	Aniline	290	U	290	390	510	ug/Kg
108-95-2	Phenol	250	U	250	390	510	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	250	U	250	390	510	ug/Kg
95-57-8	2-Chlorophenol	250	U	250	390	510	ug/Kg
95-50-1	1,2-Dichlorobenzene	250	U	250	390	510	ug/Kg
541-73-1	1,3-Dichlorobenzene	260	U	260	390	510	ug/Kg
106-46-7	1,4-Dichlorobenzene	260	U	260	390	510	ug/Kg
100-51-6	Benzyl Alcohol	250	U	250	810	990	ug/Kg
95-48-7	2-Methylphenol	240	U	240	390	510	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	270	U	270	390	510	ug/Kg
98-86-2	Acetophenone	260	U	260	390	510	ug/Kg
65794-96-9	3+4-Methylphenols	240	U	240	810	990	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	120	U	120	240	240	ug/Kg
67-72-1	Hexachloroethane	250	U	250	390	510	ug/Kg
98-95-3	Nitrobenzene	270	U	270	390	510	ug/Kg
78-59-1	Isophorone	250	U	250	390	510	ug/Kg
88-75-5	2-Nitrophenol	280	U	280	390	510	ug/Kg
105-67-9	2,4-Dimethylphenol	280	U	280	390	510	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	260	U	260	390	510	ug/Kg
120-83-2	2,4-Dichlorophenol	230	U	230	390	510	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	260	U	260	390	510	ug/Kg
65-85-0	Benzoic acid	710	U	710	810	990	ug/Kg
91-20-3	Naphthalene	250	U	250	390	510	ug/Kg
106-47-8	4-Chloroaniline	250	U	250	390	510	ug/Kg
87-68-3	Hexachlorobutadiene	250	U	250	390	510	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-5	SDG No.:	BF101924
Lab Sample ID:	P4426-09	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.06	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139890.D	5	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	260	U	260	810	990	ug/Kg
59-50-7	4-Chloro-3-methylphenol	230	U	230	390	510	ug/Kg
91-57-6	2-Methylnaphthalene	250	U	250	390	510	ug/Kg
77-47-4	Hexachlorocyclopentadiene	470	U	470	810	990	ug/Kg
88-06-2	2,4,6-Trichlorophenol	210	U	210	390	510	ug/Kg
95-95-4	2,4,5-Trichlorophenol	220	U	220	390	510	ug/Kg
92-52-4	1,1-Biphenyl	260	U	260	390	510	ug/Kg
91-58-7	2-Chloronaphthalene	250	U	250	390	510	ug/Kg
88-74-4	2-Nitroaniline	280	U	280	390	510	ug/Kg
131-11-3	Dimethylphthalate	240	U	240	390	510	ug/Kg
208-96-8	Acenaphthylene	260	U	260	390	510	ug/Kg
606-20-2	2,6-Dinitrotoluene	250	U	250	390	510	ug/Kg
99-09-2	3-Nitroaniline	270	U	270	390	510	ug/Kg
83-32-9	Acenaphthene	240	U	240	390	510	ug/Kg
Total PAH	Total PAH	3960	U	3960	390	8160	ug/Kg
51-28-5	2,4-Dinitrophenol	730	U	730	810	990	ug/Kg
100-02-7	4-Nitrophenol	350	U	350	810	990	ug/Kg
132-64-9	Dibenzofuran	250	U	250	390	510	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	610	U	510	390	1020	ug/Kg
121-14-2	2,4-Dinitrotoluene	260	U	260	390	510	ug/Kg
84-66-2	Diethylphthalate	240	U	240	390	510	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	260	U	260	390	510	ug/Kg
86-73-7	Fluorene	260	U	260	390	510	ug/Kg
100-01-6	4-Nitroaniline	320	U	320	390	510	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	350	U	350	810	990	ug/Kg
86-30-6	n-Nitrosodiphenylamine	240	U	240	390	510	ug/Kg
103-33-3	Azobenzene	240	U	240	390	510	ug/Kg
101-55-3	4-Bromophenyl-phenylether	240	U	240	390	510	ug/Kg
118-74-1	Hexachlorobenzene	250	U	250	390	510	ug/Kg
1912-24-9	Atrazine	270	U	270	390	510	ug/Kg

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-5	SDG No.:	BF101924
Lab Sample ID:	P4426-09	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.06	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139890.D	5	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	230	U	230	810	990	ug/Kg
85-01-8	Phenanthrene	250	U	250	390	510	ug/Kg
120-12-7	Anthracene	250	U	250	390	510	ug/Kg
86-74-8	Carbazole	240	U	240	390	510	ug/Kg
84-74-2	Di-n-butylphthalate	250	U	250	390	510	ug/Kg
206-44-0	Fluoranthene	240	U	240	390	510	ug/Kg
92-87-5	Benzidine	490	U	490	810	990	ug/Kg
129-00-0	Pyrene	250	U	250	390	510	ug/Kg
85-68-7	Butylbenzylphthalate	290	U	290	390	510	ug/Kg
91-94-1	3,3-Dichlorobenzidine	300	U	300	810	990	ug/Kg
56-55-3	Benzo(a)anthracene	240	U	240	390	510	ug/Kg
218-01-9	Chrysene	240	U	240	390	510	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	270	U	270	390	510	ug/Kg
117-84-0	Di-n-octyl phthalate	330	U	330	810	990	ug/Kg
205-99-2	Benzo(b)fluoranthene	240	U	240	390	510	ug/Kg
207-08-9	Benzo(k)fluoranthene	250	U	250	390	510	ug/Kg
50-32-8	Benzo(a)pyrene	280	U	280	390	510	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	230	U	230	390	510	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	240	U	240	390	510	ug/Kg
191-24-2	Benzo(g,h,i)perylene	240	U	240	390	510	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	260	U	260	390	510	ug/Kg
123-91-1	1,4-Dioxane	330	U	330	390	510	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	220	U	220	390	510	ug/Kg
90-12-0	1-Methylnaphthalene	250	U	250	510	510	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	26.975		18-112		18	SPK: -150
13127-88-3	Phenol-d6	33.985		15-107		23	SPK: -150
4165-60-0	Nitrobenzene-d5	37.155		18-107		37	SPK: -100
321-60-8	2-Fluorobiphenyl	47.54		20-109		48	SPK: -100

Report of Analysis

Client:		Date Collected:	10/16/2024 12:00:00 AM
Project:		Date Received:	10/16/2024 12:00:00 AM
Client Sample ID:	PAD-5	SDG No.:	BF101924
Lab Sample ID:	P4426-09	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	50.06	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139890.D	5	10/17/2024 12:00:00 AM	10/19/2024 12:00:00 AM	PB164216

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	23.21		10-116		15	SPK: -150
1718-51-0	Terphenyl-d14	45.465		10-105		45	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	170997	6.892				
1146-65-2	Naphthalene-d8	573389	8.169				
15067-26-2	Acenaphthene-d10	255342	9.928				
1517-22-2	Phenanthrene-d10	452819	11.41				
1719-03-5	Chrysene-d12	288225	14.051				
1520-96-3	Perylene-d12	224163	15.527				



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet SW-846

SDG No.: BF101924

Client:

Sample ID	Client ID		Parameter	Concentration	C	MDL	RDL	Units
Client ID : TP-1								
P4425-02	TP-1	water	Dimethylphthalate	32.600	J	9.3	50	ug/L
			Total Svoc :			32.60		
			Total Concentration:			32.60		
Client ID : PAD-16								
P4426-31	PAD-16	Solid	Di-n-butylphthalate	55.300	J	50.5	100	ug/Kg
			Total Svoc :			55.30		
			Total Concentration:			55.30		
Client ID : PAD-17								
P4426-33	PAD-17	Solid	3+4-Methylphenols	88.700	J	47.8	200	ug/Kg
P4426-33	PAD-17	Solid	Acetophenone	61.300	J	52.1	100	ug/Kg
P4426-33	PAD-17	Solid	Benzo(a)anthracene	69.800	J	48.3	100	ug/Kg
P4426-33	PAD-17	Solid	Benzo(b)fluoranthene	450.000		48.6	100	ug/Kg
P4426-33	PAD-17	Solid	Benzo(g,h,i)perylene	67.700	J	48	100	ug/Kg
P4426-33	PAD-17	Solid	Benzo(k)fluoranthene	170.000		49.5	100	ug/Kg
P4426-33	PAD-17	Solid	Chrysene	370.000		47.6	100	ug/Kg
P4426-33	PAD-17	Solid	Fluoranthene	530.000		48.9	100	ug/Kg
P4426-33	PAD-17	Solid	Indeno(1,2,3-cd)pyrene	68.500	J	46.8	100	ug/Kg
P4426-33	PAD-17	Solid	Phenanthrene	290.000		50.3	100	ug/Kg
P4426-33	PAD-17	Solid	Phenol	87.700	J	49.7	100	ug/Kg
P4426-33	PAD-17	Solid	Pyrene	150.000		49.7	100	ug/Kg
			Total Svoc :			2,403.70		
			Total Concentration:			2,403.70		



6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH

Contract:

Lab Code: CHEM Case No.: BF101824

SAS No.: BF101824

SDG No.: BF101824

Instrument ID:

Calibration Date(s): 10/18/2024

Calibration Time(s): 10:27

LAB FILE ID:		RRF2.5 = BF139844.D			RRF005 = BF139845.D		RRF010 = BF139846.D	
		RRF020 = BF139847.D			RRF040 = BF139848.D		RRF050 = BF139849.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
n-Nitrosodimethylamine		0.815	0.800	0.799	0.806	0.782	0.793	2.4
Pyridine		1.622	1.570	1.522	1.504	1.386	1.476	7.3
2-Fluorophenol		1.465	1.398	1.331	1.278	1.179	1.278	10.1
Benzaldehyde			1.137	1.042	0.940	0.873	0.931	15.2
Aniline		1.673	1.649	1.618	1.582	1.471	1.542	8.1
Phenol-d6		1.900	1.818	1.709	1.647	1.538	1.655	10.1
Phenol		1.952	1.832	1.760	1.712	1.583	1.706	9.2
bis(2-Chloroethyl)ether		1.470	1.423	1.359	1.294	1.251	1.319	7.8
2-Chlorophenol		1.503	1.422	1.358	1.310	1.212	1.306	10.2
1,2-Dichlorobenzene		1.660	1.579	1.478	1.379	1.273	1.398	13.2
1,3-Dichlorobenzene		1.718	1.656	1.544	1.499	1.392	1.499	10.2
1,4-Dichlorobenzene		1.723	1.641	1.558	1.487	1.392	1.498	10.2
Benzyl Alcohol		1.355	1.299	1.257	1.213	1.154	1.214	8.1
2-Methylphenol		1.264	1.164	1.134	1.114	1.053	1.114	7.7
2,2-oxybis(1-Chloropropane)		2.524	2.409	2.353	2.255	2.117	2.248	8.7
Acetophenone		0.578	0.533	0.516	0.481	0.452	0.490	11.4
3+4-Methylphenols		1.678	1.573	1.520	1.418	1.309	1.424	12.4
n-Nitroso-di-n-propylamine	1.105	1.141	1.066	1.025	0.974	0.921	1.004	9.6
Nitrobenzene-d5		0.365	0.365	0.372	0.371	0.354	0.361	2.9
Hexachloroethane		0.583	0.571	0.549	0.541	0.507	0.534	7.1
Nitrobenzene		0.417	0.402	0.408	0.403	0.383	0.396	4.1
Isophorone		0.755	0.709	0.702	0.679	0.652	0.685	5.9
2-Nitrophenol		0.124	0.137	0.150	0.157	0.158	0.149	9.2
2,4-Dimethylphenol		0.285	0.259	0.256	0.248	0.237	0.249	8.1
bis(2-Chloroethoxy)methane		0.468	0.440	0.432	0.412	0.394	0.415	8.2
2,4-Dichlorophenol		0.308	0.297	0.293	0.283	0.272	0.283	6.2
1,2,4-Trichlorobenzene		0.351	0.331	0.325	0.315	0.298	0.314	7.7
Benzoic acid			0.175	0.207	0.220	0.231	0.217	10.7
Naphthalene		1.209	1.121	1.091	1.023	0.958	1.031	11.2
4-Chloroaniline		0.397	0.382	0.368	0.351	0.332	0.352	9.3
Hexachlorobutadiene		0.224	0.206	0.203	0.197	0.185	0.197	8.0
Caprolactam		0.092	0.092	0.092	0.092	0.088	0.090	2.9
4-Chloro-3-methylphenol		0.341	0.328	0.324	0.315	0.299	0.314	6.0
2-Methylnaphthalene		0.740	0.689	0.666	0.621	0.587	0.632	11.1
Hexachlorocyclopentadiene		0.185	0.193	0.198	0.198	0.186	0.188	5.3
2,4,6-Trichlorophenol		0.405	0.382	0.400	0.387	0.363	0.382	4.8

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: _____

Lab Code: CHEM Case No.: BF101824 SAS No.: BF101824 SDG No.: BF101824

Instrument ID: _____ Calibration Date(s): 10/18/2024

Calibration Time(s): 10:27

LAB FILE ID:		RRF2.5 = BF139844.D		RRF005 = BF139845.D		RRF010 = BF139846.D		
		RRF020 = BF139847.D		RRF040 = BF139848.D		RRF050 = BF139849.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorobiphenyl		1.512	1.396	1.280	1.162	1.088	1.210	16.0
2,4,5-Trichlorophenol		0.426	0.425	0.398	0.401	0.381	0.394	6.6
1,1-Biphenyl		1.640	1.536	1.483	1.379	1.285	1.392	12.2
2-Chloronaphthalene		1.288	1.225	1.164	1.111	1.048	1.121	10.0
2-Nitroaniline		0.299	0.318	0.353	0.365	0.354	0.343	7.2
Dimethylphthalate		1.410	1.344	1.283	1.237	1.172	1.246	8.6
Acenaphthylene		1.854	1.756	1.717	1.615	1.507	1.621	10.1
2,6-Dinitrotoluene		0.256	0.266	0.278	0.280	0.273	0.270	3.4
3-Nitroaniline		0.270	0.275	0.296	0.292	0.276	0.278	4.8
Acenaphthene		1.200	1.136	1.089	1.044	0.977	1.049	9.5
2,4-Dinitrophenol			0.056	0.077	0.101	0.101	0.093	23.9
4-Nitrophenol		0.187	0.207	0.225	0.232	0.219	0.214	6.8
Dibenzofuran		1.767	1.659	1.579	1.486	1.389	1.505	11.5
2,4-Dinitrotoluene		0.280	0.314	0.337	0.356	0.344	0.332	7.9
Diethylphthalate		1.366	1.308	1.267	1.214	1.165	1.223	8.0
4-Chlorophenyl-phenylether		0.698	0.638	0.617	0.569	0.526	0.579	13.1
Fluorene		1.409	1.309	1.201	1.110	1.029	1.146	14.7
4-Nitroaniline		0.249	0.259	0.270	0.275	0.263	0.263	3.6
4,6-Dinitro-2-methylphenol			0.055	0.072	0.087	0.090	0.082	18.6
n-Nitrosodiphenylamine		0.672	0.641	0.621	0.595	0.568	0.599	8.1
Azobenzene		1.459	1.385	1.355	1.306	1.216	1.297	8.6
2,4,6-Tribromophenol		0.201	0.196	0.193	0.188	0.179	0.187	5.5
4-Bromophenyl-phenylether		0.230	0.217	0.211	0.202	0.197	0.206	7.0
Hexachlorobenzene		0.258	0.245	0.237	0.231	0.217	0.232	7.2
Atrazine		0.189	0.175	0.156	0.175	0.135	0.162	11.4
Pentachlorophenol		0.118	0.137	0.148	0.152	0.145	0.141	8.0
Phenanthrene		1.121	1.035	0.994	0.933	0.863	0.945	11.8
Anthracene		1.082	1.014	0.972	0.913	0.851	0.922	11.5
Carbazole		1.002	0.964	0.923	0.846	0.776	0.857	12.6
Di-n-butylphthalate		1.104	1.071	1.062	1.014	0.923	0.990	9.8
Fluoranthene		1.149	1.108	1.036	0.943	0.842	0.955	15.3
Benzidine		0.457	0.461	0.293	0.366	0.276	0.329	30.9
Pyrene		1.892	1.828	1.900	1.805	1.685	1.751	8.4
Terphenyl-d14		1.381	1.335	1.340	1.244	1.154	1.227	11.0
Butylbenzylphthalate		0.490	0.513	0.536	0.555	0.535	0.525	4.0
3,3-Dichlorobenzidine		0.368	0.372	0.390	0.387	0.374	0.380	2.2

All other compounds must meet a minimum RRF of 0.010.



6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: _____
Lab Code: CHEM Case No.: BF101824 SAS No.: BF101824 SDG No.: BF101824
Instrument ID: _____ Calibration Date(s): 10/18/2024 _____
Calibration Time(s): 10:27 _____

LAB FILE ID:		RRF2.5 = BF139844.D			RRF005 = BF139845.D		RRF010 = BF139846.D	
		RRF020 = BF139847.D			RRF040 = BF139848.D		RRF050 = BF139849.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Benzo(a)anthracene		1.425	1.355	1.329	1.331	1.267	1.302	6.5
Chrysene		1.325	1.244	1.234	1.167	1.134	1.194	6.5
Bis(2-ethylhexyl)phthalate		0.520	0.539	0.577	0.626	0.620	0.589	7.5
Di-n-octyl phthalate			0.786	0.930	1.135	1.182	1.071	16.1
Benzo(b)fluoranthene		1.317	1.240	1.319	1.196	1.105	1.218	7.1
Benzo(k)fluoranthene		1.213	1.177	0.992	1.066	1.030	1.051	10.4
Benzo(a)pyrene		1.030	1.024	1.018	1.025	0.974	1.002	3.1
Indeno(1,2,3-cd)pyrene		1.209	1.261	1.307	1.352	1.299	1.287	3.8
Dibenzo(a,h)anthracene		1.021	1.064	1.103	1.120	1.083	1.073	3.3
Benzo(g,h,i)perylene		1.030	1.046	1.090	1.128	1.081	1.072	3.4
1,2,4,5-Tetrachlorobenzene		0.609	0.579	0.562	0.537	0.512	0.540	8.7
1,4-Dioxane		0.651	0.625	0.603	0.591	0.571	0.596	5.7
2,3,4,6-Tetrachlorophenol		0.334	0.322	0.326	0.310	0.296	0.309	6.3
1-Methylnaphthalene		0.726	0.683	0.655	0.612	0.569	0.620	11.5

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: BF101924 SAS No.: BF101924 SDG NO.: BF101924

EPA Sample No.: SSTDICCC040 Date Analyzed: Oct 19 2024 12:00AM

Lab File ID: BF139848.D Time Analyzed: 13:47

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

		IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
	12 HOUR STD	181899	6.892	688880	8.18	379753	9.93
	UPPER LIMIT	363798	7.392	1377760	8.675	759506	10.427
	LOWER LIMIT	90949.5	6.392	344440	7.675	189876.5	9.427
	EPA SAMPLE NO.						
01	PB164216BL	200094	6.89	794942	8.17	453346	9.93
02	WALL	190163	6.89	745893	8.17	402060	9.92
03	CURBS-AND-OTHER	195900	6.89	753548	8.17	418416	9.92
04	TRANSFORMER-PAD	171810	6.89	614846	8.17	262055	9.93
05	PAD-11	196276	6.89	749941	8.17	407911	9.92
06	PAD-7	193840	6.89	750911	8.17	409213	9.93
07	TP-1	199577	6.89	755660	8.17	412131	9.93
08	TP-2	195862	6.89	748996	8.17	405573	9.92
09	TP-3	200976	6.89	775912	8.17	422411	9.92
10	TP-4	205805	6.89	797916	8.17	434693	9.93
11	TP-5	195941	6.89	751190	8.17	414555	9.92
12	PAD-8	200447	6.89	768121	8.17	421896	9.92
13	PAD-15	195168	6.89	723646	8.17	347098	9.93
14	PAD-10	197521	6.89	771255	8.17	412785	9.92
15	PAD-13	204683	6.89	784106	8.17	407138	9.92
16	PAD-17	203314	6.89	753115	8.17	357579	9.92
17	PAD-16	195373	6.89	714977	8.17	338275	9.92
18	PAD-12	197123	6.89	719768	8.17	338510	9.92
19	PAD-3	201874	6.89	725305	8.17	339096	9.93
20	PAD-14	170784	6.89	588934	8.17	263847	9.92
21	PAD-1	166324	6.89	554266	8.17	249336	9.93
22	PAD-5	170997	6.89	573389	8.17	255342	9.93

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: BF101924 SAS No.: BF101924 SDG NO.: BF101924

EPA Sample No.: SSTDICCC040 Date Analyzed: Oct 19 2024 12:00AM

Lab File ID: BF139848.D Time Analyzed: 23:57

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	181899	6.892	688880	8.18	379753	9.93
UPPER LIMIT	363798	7.392	1377760	8.675	759506	10.427
LOWER LIMIT	90949.5	6.392	344440	7.675	189876.5	9.427
EPA SAMPLE NO.						

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: BF101924 SAS No.: BF101924 SDG NO.: BF101924

EPA Sample No.: SSTDCCC040 Date Analyzed: Oct 19 2024 12:00AM

Lab File ID: BF139868.D Time Analyzed: 13:47

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

		IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
	12 HOUR STD	219099	6.892	841132	8.18	477909	9.93
	UPPER LIMIT	438198	7.392	1682264	8.675	955818	10.433
	LOWER LIMIT	109549.5	6.392	420566	7.675	238954.5	9.433
	EPA SAMPLE NO.						
01	PB164216BL	200094	6.89	794942	8.17	453346	9.93
02	WALL	190163	6.89	745893	8.17	402060	9.92
03	CURBS-AND-OTHER	195900	6.89	753548	8.17	418416	9.92
04	TRANSFORMER-PAD	171810	6.89	614846	8.17	262055	9.93
05	PAD-11	196276	6.89	749941	8.17	407911	9.92
06	PAD-7	193840	6.89	750911	8.17	409213	9.93
07	TP-1	199577	6.89	755660	8.17	412131	9.93
08	TP-2	195862	6.89	748996	8.17	405573	9.92
09	TP-3	200976	6.89	775912	8.17	422411	9.92
10	TP-4	205805	6.89	797916	8.17	434693	9.93
11	TP-5	195941	6.89	751190	8.17	414555	9.92
12	PAD-8	200447	6.89	768121	8.17	421896	9.92
13	PAD-15	195168	6.89	723646	8.17	347098	9.93
14	PAD-10	197521	6.89	771255	8.17	412785	9.92
15	PAD-13	204683	6.89	784106	8.17	407138	9.92
16	PAD-17	203314	6.89	753115	8.17	357579	9.92
17	PAD-16	195373	6.89	714977	8.17	338275	9.92
18	PAD-12	197123	6.89	719768	8.17	338510	9.92
19	PAD-3	201874	6.89	725305	8.17	339096	9.93
20	PAD-14	170784	6.89	588934	8.17	263847	9.92
21	PAD-1	166324	6.89	554266	8.17	249336	9.93
22	PAD-5	170997	6.89	573389	8.17	255342	9.93

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: BF101924 SAS No.: BF101924 SDG NO.: BF101924

EPA Sample No.: SSTDCCC040 Date Analyzed: Oct 19 2024 12:00AM

Lab File ID: BF139868.D Time Analyzed: 23:57

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	219099	6.892	841132	8.18	477909	9.93
UPPER LIMIT	438198	7.392	1682264	8.675	955818	10.433
LOWER LIMIT	109549.5	6.392	420566	7.675	238954.5	9.433
EPA SAMPLE NO.						

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: BF101924 SAS No.: BF101924 SDG NO.: BF101924

EPA Sample No.: SSTDICCC040 Date Analyzed: Oct 19 2024 12:00AM

Lab File ID: BF139848.D Time Analyzed: 13:47

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	663459	11.416	345800	14.051	365076	15.527
	UPPER LIMIT	1326918	11.916	691600	14.551	730152	16.027
	LOWER LIMIT	331729.5	10.916	172900	13.551	182538	15.027
	EPA SAMPLE NO.						
01	PB164216BL	848578	11.41	603034	14.05	446604	15.53
02	WALL	639889	11.41	370008	14.05	387887	15.53
03	CURBS-AND-OTHER	677607	11.41	365941	14.05	394368	15.53
04	TRANSFORMER-PAD	226931 *	11.45	333883	14.07	363697	15.53
05	PAD-11	666291	11.41	368761	14.05	386969	15.53
06	PAD-7	649825	11.41	355728	14.05	376625	15.53
07	TP-1	670097	11.41	370038	14.05	367770	15.53
08	TP-2	673234	11.41	377546	14.05	257410	15.53
09	TP-3	695149	11.41	391319	14.05	342377	15.53
10	TP-4	696781	11.41	383604	14.05	380773	15.53
11	TP-5	687642	11.41	376430	14.05	359501	15.53
12	PAD-8	668798	11.41	364704	14.05	378848	15.53
13	PAD-15	428172	11.42	361459	14.05	381877	15.53
14	PAD-10	636993	11.41	370206	14.05	387933	15.53
15	PAD-13	608764	11.41	270871	14.05	378751	15.53
16	PAD-17	505963	11.42	403624	14.06	349762	15.53
17	PAD-16	475082	11.41	372525	14.05	269044	15.53
18	PAD-12	490725	11.41	235740	14.05	275588	15.53
19	PAD-3	498866	11.41	384159	14.05	273829	15.53
20	PAD-14	437146	11.41	297758	14.05	222325	15.53
21	PAD-1	436102	11.41	274554	14.05	211115	15.53

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: BF101924 SAS No.: BF101924 SDG NO.: BF101924

EPA Sample No.: SSTDICCC040 Date Analyzed: Oct 19 2024 12:00AM

Lab File ID: BF139848.D Time Analyzed: 23:57

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	663459	11.416	345800	14.051	365076	15.527
UPPER LIMIT	1326918	11.916	691600	14.551	730152	16.027
LOWER LIMIT	331729.5	10.916	172900	13.551	182538	15.027
EPA SAMPLE NO.						
22 PAD-5	452819	11.41	288225	14.05	224163	15.53

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: BF101924 SAS No.: BF101924 SDG NO.: BF101924

EPA Sample No.: SSTDCCC040 Date Analyzed: Oct 19 2024 12:00AM

Lab File ID: BF139868.D Time Analyzed: 13:47

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	863533	11.416	433293	14.057	428998	15.527
	UPPER LIMIT	1727066	11.916	866586	14.557	857996	16.027
	LOWER LIMIT	431766.5	10.916	216646.5	13.557	214499	15.027
	EPA SAMPLE NO.						
01	PB164216BL	848578	11.41	603034	14.05	446604	15.53
02	WALL	639889	11.41	370008	14.05	387887	15.53
03	CURBS-AND-OTHER	677607	11.41	365941	14.05	394368	15.53
04	TRANSFORMER-PAD	226931 *	11.45	333883	14.07	363697	15.53
05	PAD-11	666291	11.41	368761	14.05	386969	15.53
06	PAD-7	649825	11.41	355728	14.05	376625	15.53
07	TP-1	670097	11.41	370038	14.05	367770	15.53
08	TP-2	673234	11.41	377546	14.05	257410	15.53
09	TP-3	695149	11.41	391319	14.05	342377	15.53
10	TP-4	696781	11.41	383604	14.05	380773	15.53
11	TP-5	687642	11.41	376430	14.05	359501	15.53
12	PAD-8	668798	11.41	364704	14.05	378848	15.53
13	PAD-15	428172 *	11.42	361459	14.05	381877	15.53
14	PAD-10	636993	11.41	370206	14.05	387933	15.53
15	PAD-13	608764	11.41	270871	14.05	378751	15.53
16	PAD-17	505963	11.42	403624	14.06	349762	15.53
17	PAD-16	475082	11.41	372525	14.05	269044	15.53
18	PAD-12	490725	11.41	235740	14.05	275588	15.53
19	PAD-3	498866	11.41	384159	14.05	273829	15.53
20	PAD-14	437146	11.41	297758	14.05	222325	15.53
21	PAD-1	436102	11.41	274554	14.05	211115 *	15.53

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: BF101924 SAS No.: BF101924 SDG NO.: BF101924

EPA Sample No.: SSTDCCC040 Date Analyzed: Oct 19 2024 12:00AM

Lab File ID: BF139868.D Time Analyzed: 23:57

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	863533	11.416	433293	14.057	428998	15.527
UPPER LIMIT	1727066	11.916	866586	14.557	857996	16.027
LOWER LIMIT	431766.5	10.916	216646.5	13.557	214499	15.027
EPA SAMPLE NO.						
22 PAD-5	452819	11.41	288225	14.05	224163	15.53

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB164216BL

Lab Name: CHEMTECH

Contract: _____

Lab Code: CHEM Case No.: BF101924

SAS No.: BF101924 SDG NO.: BF101924

Lab File ID: BF139869.D

Lab Sample ID: PB164216BL

Instrument ID: BNA_F

Date Extracted: 10/17/2024

Matrix: (soil/water) Solid

Date Analyzed: 10/19/2024

Level: (low/med) LOW

Time Analyzed: 13:47

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
WALL	P4426-37	BF139870.D	10/19/2024
CURBS-AND-OTHER	P4426-39	BF139871.D	10/19/2024
TRANSFORMER-PAD	P4426-35	BF139872.D	10/19/2024
PAD-11	P4426-21	BF139873.D	10/19/2024
PAD-7	P4426-13	BF139874.D	10/19/2024
PAD-8	P4426-15	BF139880.D	10/19/2024
PAD-15	P4426-29	BF139881.D	10/19/2024
PAD-10	P4426-19	BF139882.D	10/19/2024
PAD-13	P4426-25	BF139883.D	10/19/2024
PAD-17	P4426-33	BF139884.D	10/19/2024
PAD-16	P4426-31	BF139885.D	10/19/2024
PAD-12	P4426-23	BF139886.D	10/19/2024
PAD-3	P4426-05	BF139887.D	10/19/2024
PAD-14	P4426-27	BF139888.D	10/19/2024
PAD-1	P4426-01	BF139889.D	10/19/2024
PAD-5	P4426-09	BF139890.D	10/19/2024

COMMENTS: _____

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

TP-1

Lab Name: CHEMTECH

Contract: _____

Lab Code: CHEM Case No.: BF101924

SAS No.: BF101924 SDG NO.: BF101924

Lab File ID: BF139875.D

Lab Sample ID: P4425-02

Instrument ID: BNA_F

Date Extracted: 10/18/2024

Matrix: (soil/water) water

Date Analyzed: 10/19/2024

Level: (low/med) LOW

Time Analyzed: 16:53

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
TP-1	P4425-02	BF139875.D	10/19/2024
TP-2	P4425-04	BF139876.D	10/19/2024
TP-3	P4425-06	BF139877.D	10/19/2024
TP-4	P4425-08	BF139878.D	10/19/2024
TP-5	P4425-10	BF139879.D	10/19/2024

COMMENTS: _____

Surrogate Summary

SW-846

SDG No.: **BF101924**

Client: _____

Analytical Method: **8270E**

Lab Sample ID	Client ID	Datafile	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
								Low	High
P4426-01	PAD-1	BF139889.D	2-Fluorophenol	150	20.42	14	*	18	112
			Phenol-d6	150	27.39	18		15	107
			Nitrobenzene-d5	100	33.31	33		18	107
			2-Fluorobiphenyl	100	41.45	41		20	109
			2,4,6-Tribromophenol	150	18.83	13		10	116
			Terphenyl-d14	100	39.28	39		10	105
P4426-05	PAD-3	BF139887.D	2-Fluorophenol	150	27.59	18		18	112
			Phenol-d6	150	41.10	27		15	107
			Nitrobenzene-d5	100	46.94	47		18	107
			2-Fluorobiphenyl	100	52.67	53		20	109
			2,4,6-Tribromophenol	150	14.72	10		10	116
			Terphenyl-d14	100	41.72	42		10	105
P4426-09	PAD-5	BF139890.D	2-Fluorophenol	150	26.98	18		18	112
			Phenol-d6	150	33.99	23		15	107
			Nitrobenzene-d5	100	37.16	37		18	107
			2-Fluorobiphenyl	100	47.54	48		20	109
			2,4,6-Tribromophenol	150	23.21	15		10	116
			Terphenyl-d14	100	45.47	45		10	105
P4426-13	PAD-7	BF139874.D	2-Fluorophenol	150	20.39	14	*	18	112
			Phenol-d6	150	33.94	23		15	107
			Nitrobenzene-d5	100	33.18	33		18	107
			2-Fluorobiphenyl	100	37.04	37		20	109
			2,4,6-Tribromophenol	150	9.36	6		10	116
			Terphenyl-d14	100	32.64	33		10	105
P4426-15	PAD-8	BF139880.D	2-Fluorophenol	150	12.99	9	*	18	112
			Phenol-d6	150	28.22	19		15	107
			Nitrobenzene-d5	100	41.12	41		18	107
			2-Fluorobiphenyl	100	45.79	46		20	109
			2,4,6-Tribromophenol	150	5.31	4		10	116
			Terphenyl-d14	100	42.60	43		10	105
P4426-19	PAD-10	BF139882.D	2-Fluorophenol	150	34.22	23		18	112
			Phenol-d6	150	46.06	31		15	107
			Nitrobenzene-d5	100	42.68	43		18	107
			2-Fluorobiphenyl	100	48.27	48		20	109
			2,4,6-Tribromophenol	150	26.10	17		10	116
			Terphenyl-d14	100	39.78	40		10	105
P4426-21	PAD-11	BF139873.D	2-Fluorophenol	150	27.14	18		18	112
			Phenol-d6	150	41.02	27		15	107
			Nitrobenzene-d5	100	47.19	47		18	107
			2-Fluorobiphenyl	100	52.41	52		20	109
			2,4,6-Tribromophenol	150	16.92	11		10	116
			Terphenyl-d14	100	46.86	47		10	105
P4426-23	PAD-12	BF139886.D	2-Fluorophenol	150	20.26	14	*	18	112
			Phenol-d6	150	33.68	22		15	107
			Nitrobenzene-d5	100	47.71	48		18	107
			2-Fluorobiphenyl	100	54.70	55		20	109
			2,4,6-Tribromophenol	150	12.00	8		10	116
			Terphenyl-d14	100	68.60	69		10	105
P4426-25	PAD-13	BF139883.D	2-Fluorophenol	150	40.73	27		18	112

Surrogate Summary

SW-846

SDG No.: **BF101924**

Client: _____

Analytical Method: **8270E**

Lab Sample ID	Client ID	Datafile	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
								Low	High
P4426-25	PAD-13	BF139883.D	Phenol-d6	150	61.45	41		15	107
			Nitrobenzene-d5	100	71.37	71		18	107
			2-Fluorobiphenyl	100	74.81	75		20	109
			2,4,6-Tribromophenol	150	33.83	23		10	116
			Terphenyl-d14	100	87.34	87		10	105
P4426-27	PAD-14	BF139888.D	2-Fluorophenol	150	56.29	38		18	112
			Phenol-d6	150	56.53	38		15	107
			Nitrobenzene-d5	100	46.56	47		18	107
			2-Fluorobiphenyl	100	57.24	57		20	109
			2,4,6-Tribromophenol	150	50.25	33		10	116
P4426-29	PAD-15	BF139881.D	Terphenyl-d14	100	48.59	49		10	105
			2-Fluorophenol	150	8.80	6	*	18	112
			Phenol-d6	150	20.17	13	*	15	107
			Nitrobenzene-d5	100	36.17	36		18	107
			2-Fluorobiphenyl	100	48.13	48		20	109
P4426-31	PAD-16	BF139885.D	2,4,6-Tribromophenol	150	5.84	4	*	10	116
			Terphenyl-d14	100	34.33	34		10	105
			2-Fluorophenol	150	16.90	11	*	18	112
			Phenol-d6	150	27.74	18		15	107
			Nitrobenzene-d5	100	41.05	41		18	107
P4426-33	PAD-17	BF139884.D	2-Fluorobiphenyl	100	49.10	49		20	109
			2,4,6-Tribromophenol	150	10.61	7	*	10	116
			Terphenyl-d14	100	38.96	39		10	105
			2-Fluorophenol	150	29.72	20		18	112
			Phenol-d6	150	37.20	25		15	107
P4426-35	TRANSFORMER-IBF139872.D		Nitrobenzene-d5	100	38.18	38		18	107
			2-Fluorobiphenyl	100	45.52	46		20	109
			2,4,6-Tribromophenol	150	24.22	16		10	116
			Terphenyl-d14	100	34.74	35		10	105
			2-Fluorophenol	150	24.02	16	*	18	112
P4426-37	WALL	BF139870.D	Phenol-d6	150	32.18	21		15	107
			Nitrobenzene-d5	100	38.91	39		18	107
			2-Fluorobiphenyl	100	47.75	48		20	109
			2,4,6-Tribromophenol	150	12.92	9	*	10	116
			Terphenyl-d14	100	25.81	26		10	105
P4426-39	CURBS-AND-OTHB139871.D		2-Fluorophenol	150	42.19	28		18	112
			Phenol-d6	150	46.11	31		15	107
			Nitrobenzene-d5	100	41.43	41		18	107
			2-Fluorobiphenyl	100	45.26	45		20	109
			2,4,6-Tribromophenol	150	37.45	25		10	116
PB164216BL	PB164216BL	BF139869.D	Terphenyl-d14	100	40.09	40		10	105
			2-Fluorophenol	150	9.54	6	*	18	112
			Phenol-d6	150	20.07	13	*	15	107
			Nitrobenzene-d5	100	33.85	34		18	107
			2-Fluorobiphenyl	100	39.73	40		20	109
			2,4,6-Tribromophenol	150	5.55	4	*	10	116
			Terphenyl-d14	100	40.31	40		10	105
			2-Fluorophenol	150	127.17	85		18	112
			Phenol-d6	150	120.84	81		15	107

Surrogate Summary

SW-846

SDG No.: BF101924

Client: _____

Analytical Method: 8270E

Lab Sample ID	Client ID	Datafile	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
								Low	High
PB164216BL	PB164216BL	BF139869.D	Nitrobenzene-d5	100	90.64	91		18	107
			2-Fluorobiphenyl	100	84.53	85		20	109
			2,4,6-Tribromophenol	150	137.52	92		10	116
			Terphenyl-d14	100	79.20	79		10	105

Surrogate Summary

SW-846

SDG No.: **BF101924**

Client: _____

Analytical Method: **8270E**

Lab Sample ID	Client ID	Datafile	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
								Low	High
P4425-02	TP-1	BF139875.D	2-Fluorophenol	150	107.97	72		10	139
			Phenol-d6	150	97.79	65		10	134
			Nitrobenzene-d5	100	89.21	89		49	133
			2-Fluorobiphenyl	100	83.25	83		52	132
			2,4,6-Tribromophenol	150	132.32	88		44	137
			Terphenyl-d14	100	81.62	82		48	125
P4425-04	TP-2	BF139876.D	2-Fluorophenol	150	113.77	76		10	139
			Phenol-d6	150	103.90	69		10	134
			Nitrobenzene-d5	100	88.78	89		49	133
			2-Fluorobiphenyl	100	85.04	85		52	132
			2,4,6-Tribromophenol	150	132.56	88		44	137
			Terphenyl-d14	100	80.79	81		48	125
P4425-06	TP-3	BF139877.D	2-Fluorophenol	150	116.90	78		10	139
			Phenol-d6	150	108.52	72		10	134
			Nitrobenzene-d5	100	92.28	92		49	133
			2-Fluorobiphenyl	100	86.92	87		52	132
			2,4,6-Tribromophenol	150	136.11	91		44	137
			Terphenyl-d14	100	84.90	85		48	125
P4425-08	TP-4	BF139878.D	2-Fluorophenol	150	112.78	75		10	139
			Phenol-d6	150	105.37	70		10	134
			Nitrobenzene-d5	100	89.93	90		49	133
			2-Fluorobiphenyl	100	84.18	84		52	132
			2,4,6-Tribromophenol	150	133.88	89		44	137
			Terphenyl-d14	100	83.45	83		48	125
P4425-10	TP-5	BF139879.D	2-Fluorophenol	150	114.47	76		10	139
			Phenol-d6	150	105.27	70		10	134
			Nitrobenzene-d5	100	90.21	90		49	133
			2-Fluorobiphenyl	100	85.13	85		52	132
			2,4,6-Tribromophenol	150	134.39	90		44	137
			Terphenyl-d14	100	85.39	85		48	125

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: BF101924

Lab Code: CHEM

SAS No.: BF101924

SDG NO.: BF101924

Lab File ID: BF139867.D

DFTPP Injection Date: 10/19/2024

Instrument ID: BNA_F

DFTPP Injection Time: 12:51

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	41.5
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	36.3
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	45
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	27.4
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	15
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.1 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF139868.D	10/19/2024	13:19
PB164216BL	PB164216BL	BF139869.D	10/19/2024	13:47
WALL	P4426-37	BF139870.D	10/19/2024	14:31
CURBS-AND-OTHER	P4426-39	BF139871.D	10/19/2024	14:59
TRANSFORMER-PAD	P4426-35	BF139872.D	10/19/2024	15:27
PAD-11	P4426-21	BF139873.D	10/19/2024	15:56
PAD-7	P4426-13	BF139874.D	10/19/2024	16:24
TP-1	P4425-02	BF139875.D	10/19/2024	16:53
TP-2	P4425-04	BF139876.D	10/19/2024	17:21
TP-3	P4425-06	BF139877.D	10/19/2024	17:49
TP-4	P4425-08	BF139878.D	10/19/2024	18:18
TP-5	P4425-10	BF139879.D	10/19/2024	18:46
PAD-8	P4426-15	BF139880.D	10/19/2024	19:14
PAD-15	P4426-29	BF139881.D	10/19/2024	19:42
PAD-10	P4426-19	BF139882.D	10/19/2024	20:10
PAD-13	P4426-25	BF139883.D	10/19/2024	20:38
PAD-17	P4426-33	BF139884.D	10/19/2024	21:06
PAD-16	P4426-31	BF139885.D	10/19/2024	21:35

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: BF101924

Lab Code: CHEM

SAS No.: BF101924 SDG NO.: BF101924

Lab File ID: BF139867.D

DFTPP Injection Date: 10/19/2024

Instrument ID: BNA_F

DFTPP Injection Time: 12:51

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	41.5
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	36.3
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	45
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	27.4
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	15
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.1 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
PAD-12	P4426-23	BF139886.D	10/19/2024	22:03
PAD-3	P4426-05	BF139887.D	10/19/2024	22:31
PAD-14	P4426-27	BF139888.D	10/19/2024	23:00
PAD-1	P4426-01	BF139889.D	10/19/2024	23:28
PAD-5	P4426-09	BF139890.D	10/19/2024	23:57