

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101924\
 Data File : BF139868.D
 Acq On : 19 Oct 2024 13:19
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 21 01:39:32 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	120	0.00
2	1,4-Dioxane	40.000	39.572	1.1	120	-0.01
3	Pyridine	40.000	39.481	1.3	117	0.00
4	n-Nitrosodimethylamine	40.000	39.608	1.0	117	0.00
5 S	2-Fluorophenol	80.000	78.684	1.6	118	0.00
6	Aniline	40.000	41.200	-3.0	121	0.00
7 S	Phenol-d6	80.000	78.153	2.3	118	0.00
8	2-Chlorophenol	40.000	39.519	1.2	119	0.00
9	Benzaldehyde	40.000	36.838	7.9	110	0.00
10 C	Phenol	40.000	39.643	0.9	119	0.00
11	bis(2-Chloroethyl)ether	40.000	39.778	0.6	122	0.00
12	1,3-Dichlorobenzene	40.000	39.500	1.3	119	0.00
13 C	1,4-Dichlorobenzene	40.000	39.443	1.4	120	0.00
14	1,2-Dichlorobenzene	40.000	39.303	1.7	120	0.00
15	Benzyl Alcohol	40.000	40.708	-1.8	123	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.773	0.6	119	0.00
17	2-Methylphenol	40.000	40.193	-0.5	121	0.00
18	Hexachloroethane	40.000	39.592	1.0	118	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.532	1.2	123	0.00
20	3+4-Methylphenols	40.000	39.656	0.9	120	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	122	0.00
22	Acetophenone	40.000	38.609	3.5	120	0.00
23 S	Nitrobenzene-d5	80.000	81.585	-2.0	121	0.00
24	Nitrobenzene	40.000	39.912	0.2	120	0.00
25	Isophorone	40.000	40.596	-1.5	125	0.00
26 C	2-Nitrophenol	40.000	43.619	-9.0	126	0.00
27	2,4-Dimethylphenol	40.000	39.689	0.8	121	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.944	0.1	123	0.00
29 C	2,4-Dichlorophenol	40.000	40.999	-2.5	125	0.00
30	1,2,4-Trichlorobenzene	40.000	39.811	0.5	121	0.00
31	Naphthalene	40.000	38.877	2.8	120	0.00
32	Benzoic acid	40.000	44.223	-10.6	133	0.01
33	4-Chloroaniline	40.000	40.552	-1.4	124	0.00
34 C	Hexachlorobutadiene	40.000	39.531	1.2	121	0.00
35	Caprolactam	40.000	42.956	-7.4	129	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.669	-1.7	124	0.00
37	2-Methylnaphthalene	40.000	39.466	1.3	122	0.00
38	1-Methylnaphthalene	40.000	39.471	1.3	122	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	126	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.132	2.2	124	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.317	-8.3	129	0.00
42 S	2,4,6-Tribromophenol	80.000	84.192	-5.2	132	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.230	1.9	122	0.00
44	2,4,5-Trichlorophenol	40.000	41.357	-3.4	128	0.00
45 S	2-Fluorobiphenyl	80.000	73.938	7.6	121	0.00
46	1,1'-Biphenyl	40.000	38.227	4.4	121	0.00
47	2-Chloronaphthalene	40.000	38.741	3.1	123	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101924\
 Data File : BF139868.D
 Acq On : 19 Oct 2024 13:19
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 21 01:39:32 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.913	-7.3	127	0.00
49	Acenaphthylene	40.000	39.535	1.2	125	0.00
50	Dimethylphthalate	40.000	40.237	-0.6	127	0.00
51	2,6-Dinitrotoluene	40.000	42.765	-6.9	129	0.00
52 C	Acenaphthene	40.000	39.402	1.5	125	0.00
53	3-Nitroaniline	40.000	42.318	-5.8	127	0.00
54 P	2,4-Dinitrophenol	40.000	43.008	-7.5	138	0.00
55	Dibenzofuran	40.000	39.037	2.4	124	0.00
56 P	4-Nitrophenol	40.000	44.908	-12.3	131	0.00
57	2,4-Dinitrotoluene	40.000	44.979	-12.4	132	0.00
58	Fluorene	40.000	38.544	3.6	125	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.516	-3.8	130	0.00
60	Diethylphthalate	40.000	40.787	-2.0	129	0.00
61	4-Chlorophenyl-phenylether	40.000	39.118	2.2	125	0.00
62	4-Nitroaniline	40.000	43.707	-9.3	131	0.00
63	Azobenzene	40.000	39.651	0.9	124	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	130	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.619	-16.5	142	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.351	4.1	126	0.00
67	4-Bromophenyl-phenylether	40.000	39.953	0.1	132	0.00
68	Hexachlorobenzene	40.000	39.177	2.1	128	0.00
69	Atrazine	40.000	39.455	1.4	119	0.00
70 C	Pentachlorophenol	40.000	43.456	-8.6	131	0.00
71	Phenanthrene	40.000	37.903	5.2	125	0.00
72	Anthracene	40.000	38.872	2.8	128	0.00
73	Carbazole	40.000	38.890	2.8	128	0.00
74	Di-n-butylphthalate	40.000	40.768	-1.9	130	0.00
75 C	Fluoranthene	40.000	39.138	2.2	129	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	125	0.00
77	Benzydine	40.000	35.521	11.2	100	0.00
78	Pyrene	40.000	42.565	-6.4	129	0.00
79 S	Terphenyl-d14	80.000	84.333	-5.4	130	0.00
80	Butylbenzylphthalate	40.000	44.974	-12.4	133	0.00
81	Benzo(a)anthracene	40.000	40.667	-1.7	125	0.00
82	3,3'-Dichlorobenzidine	40.000	38.635	3.4	119	0.00
83	Chrysene	40.000	40.614	-1.5	130	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.876	-7.2	126	0.00
85 c	Di-n-octyl phthalate	40.000	42.178	-5.4	125	0.00
86 I	Perylene-d12	20.000	20.000	0.0	118	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	44.331	-10.8	124	0.00
88	Benzo(b)fluoranthene	40.000	37.997	5.0	114	0.00
89	Benzo(k)fluoranthene	40.000	42.002	-5.0	122	0.00
90 C	Benzo(a)pyrene	40.000	40.801	-2.0	117	0.00
91	Dibenzo(a,h)anthracene	40.000	44.191	-10.5	124	0.00
92	Benzo(g,h,i)perylene	40.000	45.696	-14.2	128	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101924\
Data File : BF139868.D
Acq On : 19 Oct 2024 13:19
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 21 01:39:32 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0