

Data Path : Z:\HPCHEM1\BNA F\DATA\BF102016\
 Data File : BF090685.D
 Acq On : 20 Oct 2016 16:36
 Operator : UM/SJ
 Sample : H5315-04 2X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FO-TANK-SIDEWALL-SOUTH

Manual Integrations
 APPROVED

sohil
 10/21/2016 6:58:33 PM

Quant Time: Oct 20 18:59:46 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF101316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 20 18:12:00 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	221078	20.00	ng	-0.01
21) Naphthalene-d8	8.17	136	1036012	20.00	ng	0.00
38) Acenaphthene-d10	9.93	164	451438	20.00	ng	0.00
63) Phenanthrene-d10	11.41	188	592126	20.00	ng	0.00
75) Chrysene-d12	14.05	240	512862	20.00	ng	0.00
86) Perylene-d12	15.49	264	455012	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.48	112	766441	63.56	ng	0.00
7) Phenol-d6	6.51	99	982732	54.84	ng	-0.01
23) Nitrobenzene-d5	7.45	82	602026	32.28	ng	-0.01
41) 2,4,6-Tribromophenol	10.71	330	130397	32.40	ng	0.00
44) 2-Fluorobiphenyl	9.24	172	864323	31.55	ng	-0.01
78) Terphenyl-d14	12.99	244	515795	23.42	ng	-0.01
Target Compounds						
49) Dimethylphthalate	9.64	163	240742	8.22	ng	Qvalue # 96
74) Fluoranthene	12.61	202	92633	2.91	ng	89
77) Pyrene	12.84	202	86026	2.53	ng	96
80) Benzo(a)anthracene	14.04	228	72549m	2.48	ng	
87) Benzo(b)fluoranthene	15.08	252	73389m	2.46	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF102016\
Data File : BF090685.D
Acq On : 20 Oct 2016 16:36
Operator : UM/SJ
Sample : H5315-04 2X
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FO-TANK-SIDEWALL-SOUTH

Manual Integrations
APPROVED

sohil
10/21/2016 6:58:33 PM

Quant Time: Oct 20 18:59:46 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF101316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 20 18:12:00 2016
Response via : Initial Calibration

