

Data Path : Z:\HPCHEM1\BNA F\DATA\BF102017\
 Data File : BF099714.D
 Acq On : 20 Oct 2017 21:25
 Operator : SJ/JU
 Sample : I5950-02
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 20170719-C-COMP2

Quant Time: Oct 21 22:02:05 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 20 13:11:43 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.81	152	91696	20.00	ng	-0.01
21) Naphthalene-d8	8.09	136	369865	20.00	ng	-0.01
38) Acenaphthene-d10	9.85	164	149261	20.00	ng	-0.01
63) Phenanthrene-d10	11.33	188	214204	20.00	ng	-0.01
75) Chrysene-d12	13.97	240	165277	20.00	ng	-0.01
86) Perylene-d12	15.44	264	166778	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.44	112	580128	100.71	ng	0.00
7) Phenol-d6	6.46	99	677258	95.31	ng	0.00
23) Nitrobenzene-d5	7.38	82	379803	65.85	ng	-0.01
41) 2,4,6-Tribromophenol	10.64	330	110612	90.56	ng	-0.01
44) 2-Fluorobiphenyl	9.17	172	644119	72.04	ng	0.00
78) Terphenyl-d14	12.92	244	365913	50.35	ng	0.00
Target Compounds						
10) Phenol	6.47	94	23335	2.936	ng	96
49) Dimethylphthalate	9.56	163	60844	5.401	ng	98
83) Bis(2-ethylhexyl)phthalate	13.94	149	17052	2.091	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF102017\
Data File : BF099714.D
Acq On : 20 Oct 2017 21:25
Operator : SJ/JU
Sample : I5950-02
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
20170719-C-COMP2

Quant Time: Oct 21 22:02:05 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 20 13:11:43 2017
Response via : Initial Calibration

