


Data Path : Z:\HPCHEM1\BNA F\Data\BF102017\
 Data File : BF099715.D
 Acq On : 20 Oct 2017 23:17
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092317.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Oct 20 13:11:43 2017



AutoFind: Scans 1612, 1613, 1614; Background Corrected with Scan 1608

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.8	15117	PASS
68	69	0.00	2	0.8	108	PASS
69	198	0.00	100	31.0	14329	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	47.5	21917	PASS
197	198	0.00	2	0.7	326	PASS
198	198	100	100	100.0	46154	PASS
199	198	5	9	6.6	3033	PASS
275	198	10	60	22.6	10416	PASS
365	198	1	100	1.8	839	PASS
441	442	0.01	24	16.2	5916	PASS
442	198	50	100	79.2	36554	PASS
443	442	15	24	18.1	6616	PASS