

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102020\
 Data File : BF122028.D
 Acq On : 20 Oct 2020 14:28
 Operator : JU/CG
 Sample : PB132494BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB132494BL

Quant Time: Oct 20 14:54:02 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 15 12:55:05 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.728	152	128212	20.00	ng	-0.02
21) Naphthalene-d8	8.010	136	525361	20.00	ng	-0.02
39) Acenaphthene-d10	9.769	164	283995	20.00	ng	-0.01
64) Phenanthrene-d10	11.251	188	537815	20.00	ng	-0.01
76) Chrysene-d12	13.898	240	428735	20.00	ng	-0.01
86) Perylene-d12	15.333	264	354177	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.357	112	970007	111.66	ng	0.00
7) Phenol-d6	6.381	99	1241102	110.99	ng	-0.01
23) Nitrobenzene-d5	7.298	82	879086	85.82	ng	-0.02
42) 2,4,6-Tribromophenol	10.563	330	411803	117.22	ng	-0.01
45) 2-Fluorobiphenyl	9.092	172	1492975	77.35	ng	-0.01
79) Terphenyl-d14	12.845	244	1690860	75.16	ng	0.00


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102020\
Data File : BF122028.D
Acq On : 20 Oct 2020 14:28
Operator : JU/CG
Sample : PB132494BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB132494BL

Quant Time: Oct 20 14:54:02 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 15 12:55:05 2020
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.728 min Scan# 789
Delta R.T. -0.017 min
Lab File: BF122028.D
Acq: 20 Oct 2020 14:28




Instrument :
BNA_F
ClientSampleId :
PB132494BL


Tgt Ion:152 Resp: 128212
Ion Ratio Lower Upper
152 100
150 157.9 125.3 187.9
115 63.4 51.9 77.9



#5
2-Fluorophenol
Concen: 111.66 ng
RT: 5.357 min Scan# 556
Delta R.T. -0.006 min
Lab File: BF122028.D
Acq: 20 Oct 2020 14:28

Tgt Ion:112 Resp: 970007
Ion Ratio Lower Upper
112 100
64 66.1 50.4 75.6
63 34.8 25.8 38.6









#7
Phenol-d6
Concen: 110.99 ng
RT: 6.381 min Scan# 730
Delta R.T. -0.012 min
Lab File: BF122028.D
Acq: 20 Oct 2020 14:28

Instrument :
BNA_F
ClientSampleId :
PB132494BL

Tgt Ion: 99 Resp: 1241102




Ion	Ratio	Lower	Upper
99	100		
42	20.4	15.9	23.9
71	35.5	27.1	40.7




#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.010 min Scan# 1007
Delta R.T. -0.018 min
Lab File: BF122028.D
Acq: 20 Oct 2020 14:28

Tgt Ion: 136 Resp: 525361

Ion	Ratio	Lower	Upper
136	100		
137	11.4	8.6	13.0
54	9.9	7.1	10.7
68	6.7	5.0	7.4





#23

Nitrobenzene-d5

Concen: 85.82 ng

RT: 7.298 min Scan# 886





Delta R.T. -0.018 min

Lab File: BF122028.D

Acq: 20 Oct 2020 14:28

Tgt Ion: 82 Resp: 879086

Ion	Ratio	Lower	Upper
82	100		
128	41.5	35.0	52.4
54	55.0	43.9	65.9



#39

Acenaphthene-d10

Concen: 20.00 ng

RT: 9.769 min Scan# 1306




Delta R.T. -0.012 min


Lab File: BF122028.D

Acq: 20 Oct 2020 14:28

Tgt Ion: 164 Resp: 283995

Ion	Ratio	Lower	Upper
164	100		
162	103.1	83.2	124.8
160	46.6	37.5	56.3









#42
2,4,6-Tribromophenol
Concen: 117.22 ng
RT: 10.563 min Scan# 1441
Delta R.T. -0.012 min
Lab File: BF122028.D
Acq: 20 Oct 2020 14:28




Instrument :
BNA_F
ClientSampleId :
PB132494BL


Tgt Ion:330 Resp: 411803
Ion Ratio Lower Upper
330 100
332 96.2 77.3 115.9
141 27.4 22.3 33.5



#45
2-Fluorobiphenyl
Concen: 77.35 ng
RT: 9.092 min Scan# 1191
Delta R.T. -0.012 min
Lab File: BF122028.D
Acq: 20 Oct 2020 14:28

Tgt Ion:172 Resp: 1492975
Ion Ratio Lower Upper
172 100
171 36.4 29.1 43.7
170 24.3 19.8 29.6









#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.251 min Scan# 1558
Delta R.T. -0.012 min
Lab File: BF122028.D
Acq: 20 Oct 2020 14:28




Instrument :
BNA_F
ClientSampleId :
PB132494BL


Tgt Ion:188	Resp: 537815
Ion Ratio	Lower Upper
188	100
94	11.2 8.3 12.5
80	12.1 9.1 13.7



#76
Chrysene-d12
Concen: 20.00 ng
RT: 13.898 min Scan# 2008
Delta R.T. -0.012 min
Lab File: BF122028.D
Acq: 20 Oct 2020 14:28


Tgt Ion:240	Resp: 428735
Ion Ratio	Lower Upper
240	100
120	10.7 9.2 13.8
236	25.7 20.9 31.3








#79
 Terphenyl-d14
 Concen: 75.16 ng
 RT: 12.845 min Scan# 1829
 Delta R.T. -0.006 min
 Lab File: BF122028.D
 Acq: 20 Oct 2020 14:28


Instrument :
 BNA_F
 ClientSampleId :
 PB132494BL



Tgt Ion:244 Resp: 1690860
 Ion Ratio Lower Upper
 244 100
 212 7.3 5.4 8.2
 122 9.9 8.1 12.1



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 15.333 min Scan# 2252
 Delta R.T. -0.006 min
 Lab File: BF122028.D
 Acq: 20 Oct 2020 14:28



Tgt Ion:264 Resp: 354177
 Ion Ratio Lower Upper
 264 100
 260 24.3 18.7 28.1
 265 21.7 17.3 25.9

