

Data Path : Z:\HPCHEM1\BNA F\DATA\BF102115\
 Data File : BF082445.D
 Acq On : 22 Oct 2015 11:57
 Operator : UM/IZ
 Sample : G4111-03
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CLEARWELL

Quant Time: Oct 22 14:52:02 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF101015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 21 13:20:24 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.77	152	102770	20.00	ng	-0.01
21) Naphthalene-d8	8.06	136	422521	20.00	ng	-0.01
38) Acenaphthene-d10	9.82	164	173928	20.00	ng	0.00
63) Phenanthrene-d10	11.30	188	252095	20.00	ng	0.00
75) Chrysene-d12	14.04	240	248263	20.00	ng	0.02
86) Perylene-d12	15.59	264	232929	20.00	ng	0.06
System Monitoring Compounds						
5) 2-Fluorophenol	5.36	112	187021	36.73	ng	0.00
7) Phenol-d6	6.41	99	145386	21.94	ng	-0.01
23) Nitrobenzene-d5	7.35	82	483627	76.87	ng	0.00
41) 2,4,6-Tribromophenol	10.61	330	86233	52.87	ng	-0.01
44) 2-Fluorobiphenyl	9.14	172	950240	90.60	ng	0.00
78) Terphenyl-d14	12.90	244	381162	40.68	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.53	163	25382	2.07	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF102115\
Data File : BF082445.D
Acq On : 22 Oct 2015 11:57
Operator : UM/IZ
Sample : G4111-03
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CLEARWELL

Quant Time: Oct 22 14:52:02 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF101015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 21 13:20:24 2015
Response via : Initial Calibration

