

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF102116\
 Data File : BF090703.D
 Acq On : 21 Oct 2016 10:31
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Oct 21 23:11:34 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 21 17:36:31 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	219755	20.00	ng	0.00
21) Naphthalene-d8	8.14	136	923498	20.00	ng	0.00
38) Acenaphthene-d10	9.90	164	467316	20.00	ng	0.00
63) Phenanthrene-d10	11.39	188	709691	20.00	ng	0.00
75) Chrysene-d12	14.02	240	499586	20.00	ng	0.00
86) Perylene-d12	15.46	264	338350	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.46	112	1086552	79.74	ng	0.00
7) Phenol-d6	6.50	99	1485690	80.50	ng	0.00
23) Nitrobenzene-d5	7.42	82	1396679	82.95	ng	0.00
41) 2,4,6-Tribromophenol	10.69	330	358867	97.02	ng	0.00
44) 2-Fluorobiphenyl	9.23	172	2302519	76.49	ng	0.00
78) Terphenyl-d14	12.97	244	1870316	85.80	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.49	88	289740	36.75	ng	89
3) Pyridine	3.21	79	735704	39.82	ng	# 79
4) n-Nitrosodimethylamine	3.16	42	263662	41.71	ng	87
6) Aniline	6.52	93	893511	42.70	ng	# 71
8) 2-Chlorophenol	6.65	128	688495	42.09	ng	91
9) Benzaldehyde	6.41	77	404420	40.08	ng	# 76
10) Phenol	6.51	94	796219	38.38	ng	# 52
11) bis(2-Chloroethyl)ether	6.60	93	736352	42.24	ng	93
12) 1,3-Dichlorobenzene	6.79	146	629389	38.31	ng	# 87
13) 1,4-Dichlorobenzene	6.87	146	687935	39.47	ng	# 89
14) 1,2-Dichlorobenzene	7.03	146	665692	39.02	ng	96
15) Benzyl Alcohol	7.00	79	535669	42.73	ng	# 72
16) 2,2'-oxybis(1-Chloropropan	7.14	45	1029379	39.43	ng	77
17) 2-Methylphenol	7.11	107	512411	41.58	ng	# 82
18) Hexachloroethane	7.37	117	272124	38.46	ng	91
19) n-Nitroso-di-n-propylamine	7.27	70	482702	35.12	ng	# 72
20) 3+4-Methylphenols	7.27	107	611853	41.88	ng	# 62
22) Acetophenone	7.27	105	850817	41.34	ng	# 80
24) Nitrobenzene	7.45	77	796961	43.66	ng	# 68
25) Isophorone	7.69	82	1277470	40.08	ng	# 87
26) 2-Nitrophenol	7.77	139	326092	43.73	ng	# 77
27) 2,4-Dimethylphenol	7.80	122	518596	39.22	ng	# 78
28) bis(2-Chloroethoxy)methane	7.89	93	714626	40.98	ng	# 92
29) 2,4-Dichlorophenol	8.01	162	485249	44.72	ng	94
30) 1,2,4-Trichlorobenzene	8.09	180	528411	41.38	ng	# 96
31) Naphthalene	8.17	128	1767196	39.12	ng	96
32) Benzoic acid	7.94	122	369764	40.98	ng	# 59
33) 4-Chloroaniline	8.22	127	726769	43.38	ng	# 93
34) Hexachlorobutadiene	8.29	225	297182	41.39	ng	99
35) Caprolactam	8.59	113	141703	45.72	ng	# 69
36) 4-Chloro-3-methylphenol	8.70	107	564102	44.69	ng	# 82
37) 2-Methylnaphthalene	8.86	142	1134555	42.97	ng	# 93
39) 1,2,4,5-Tetrachlorobenzene	9.02	216	498679	39.52	ng	97
40) Hexachlorocyclopentadiene	9.01	237	236436	39.91	ng	96

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF102116\
 Data File : BF090703.D
 Acq On : 21 Oct 2016 10:31
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Oct 21 23:11:34 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 21 17:36:31 2016
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	9.14	196	334021	42.68	ng	99
43) 2,4,5-Trichlorophenol	9.18	196	322338	40.59	ng	97
45) 1,1'-Biphenyl	9.32	154	1394716	38.00	ng	90
46) 2-Chloronaphthalene	9.34	162	1037856	38.05	ng	# 87
47) 2-Nitroaniline	9.45	65	412548	45.56	ng	# 74
48) Acenaphthylene	9.77	152	1698663	40.81	ng	96
49) Dimethylphthalate	9.63	163	1290116	43.93	ng	# 98
50) 2,6-Dinitrotoluene	9.69	165	265703	42.08	ng	# 50
51) Acenaphthene	9.94	154	1178629	42.66	ng	88
52) 3-Nitroaniline	9.86	138	299129	41.21	ng	# 58
53) 2,4-Dinitrophenol	9.96	184	126909	45.36	ng	# 43
54) Dibenzofuran	10.11	168	1432048	42.79	ng	# 87
55) 4-Nitrophenol	10.02	139	179504	43.20	ng	# 23
56) 2,4-Dinitrotoluene	10.09	165	308612	39.99	ng	# 47
57) Fluorene	10.45	166	1230020	44.34	ng	94
58) 2,3,4,6-Tetrachlorophenol	10.22	232	270594	43.42	ng	# 87
59) Diethylphthalate	10.33	149	1308276	42.96	ng	99
60) 4-Chlorophenyl-phenylether	10.44	204	569697	44.94	ng	# 88
61) 4-Nitroaniline	10.47	138	324819	44.18	ng	# 50
62) Azobenzene	10.60	77	1501041	42.09	ng	86
64) 4,6-Dinitro-2-methylphenol	10.50	198	181265	40.71	ng	89
65) n-Nitrosodiphenylamine	10.57	169	951673	41.86	ng	91
66) 4-Bromophenyl-phenylether	10.93	248	311867	45.33	ng	98
67) Hexachlorobenzene	11.00	284	352643	43.44	ng	# 80
68) Atrazine	11.09	200	303291	48.33	ng	84
69) Pentachlorophenol	11.19	266	177649	42.54	ng	99
70) Phenanthrene	11.41	178	1615406	44.66	ng	95
71) Anthracene	11.46	178	1580613	39.40	ng	96
72) Carbazole	11.62	167	1498514	44.68	ng	# 93
73) Di-n-butylphthalate	11.95	149	1927269	44.30	ng	# 96
74) Fluoranthene	12.59	202	1463431	41.82	ng	97
76) Benzidine	12.71	184	492049	44.36	ng	98
77) Pyrene	12.59	202	1463409	41.94	ng	95
79) Butylbenzylphthalate	13.45	149	770172	48.02	ng	93
80) Benzo(a)anthracene	14.01	228	1170231	42.80	ng	95
81) 3,3'-Dichlorobenzidine	13.97	252	399583	42.18	ng	99
82) Chrysene	14.05	228	1276071	47.82	ng	93
83) Bis(2-ethylhexyl)phthalate	14.01	149	1037747	45.22	ng	# 95
84) Di-n-octyl phthalate	14.61	149	1403277	40.29	ng	95
85) Indeno(1,2,3-cd)pyrene	16.86	276	799341	34.34	ng	# 93
87) Benzo(b)fluoranthene	15.05	252	879514	42.27	ng	# 88
88) Benzo(k)fluoranthene	15.05	252	879514	42.07	ng	# 89
89) Benzo(a)pyrene	15.40	252	785639	40.25	ng	# 86
90) Dibenzo(a,h)anthracene	16.88	278	703088	37.51	ng	# 88
91) Benzo(g,h,i)perylene	17.29	276	655694	36.07	ng	# 85


(#) = qualifier out of range (m) = manual integration (+) = signals summed

```
Data Path : Z:\HPCHEM1\BNA_F\DATA\BF102116\
Data File : BF090703.D
Acq On    : 21 Oct 2016   10:31
Operator  : UM/SJ
Sample    : SSTCCCC040
Misc      :
ALS Vial  : 5      Sample Multiplier: 1
```

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Quant Time: Oct 21 23:11:34 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 21 17:36:31 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 6.86 min Scan# 396

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:152 Resp: 219755

Ion Ratio Lower Upper

152 100

150 156.7 124.7 187.1




115 59.3 39.0 58.4#



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 36.75 ng

RT: 2.49 min Scan# 13

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31


Tgt Ion: 88 Resp: 289740

Ion Ratio Lower Upper

88 100

58 84.8 77.7 116.5



43 28.9 26.6 40.0




Abundance Ion 88.00 (87.70 to 88.70): BF0

Ion 58.00 (57.70 to 58.70): BF0

Ion 43.00 (42.70 to 43.70): BF0





#3

Pyridine

Concen: 39.82 ng

RT: 3.21 min Scan# 76

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 79 Resp: 735704

Ion Ratio Lower Upper

79 100

52 71.5 76.8 115.2#

51 34.2 32.2 48.4



#4

n-Nitrosodimethylamine

Concen: 41.71 ng

RT: 3.16 min Scan# 72

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31




Tgt Ion: 42 Resp: 263662


Ion Ratio Lower Upper

42 100

74 147.1 105.0 157.6

44 7.6 6.6 10.0





#5

2-Fluorophenol

Concen: 79.74 ng

RT: 5.46 min Scan# 273

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 112 Resp: 1086552

Ion Ratio Lower Upper

112 100

64 70.3 39.7 59.5#




63 35.1 17.4 26.0#



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#6

Aniline

Concen: 42.70 ng

RT: 6.52 min Scan# 366

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31


Tgt Ion: 93 Resp: 893511

Ion Ratio Lower Upper

93 100

66 52.5 27.2 40.8#



65 28.6 14.7 22.1#




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#7

Phenol-d6

Concen: 80.50 ng

RT: 6.50 min Scan# 364

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 99 Resp: 1485690

Ion Ratio Lower Upper

99 100

42 17.3 13.7 20.5

71 34.1 14.2 21.2#



#8

2-Chlorophenol

Concen: 42.09 ng

RT: 6.65 min Scan# 377

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31




Tgt Ion: 128 Resp: 688495


Ion Ratio Lower Upper

128 100

130 33.4 12.2 52.2

64 49.3 20.5 60.5





#9

Benzaldehyde

Concen: 40.08 ng

RT: 6.41 min Scan# 356

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 77 Resp: 404420

Ion Ratio Lower Upper

77 100

105 94.6 91.6 131.6

106 88.4 102.6 142.6#



#10

Phenol

Concen: 38.38 ng

RT: 6.51 min Scan# 365

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31




Tgt Ion: 94 Resp: 796219


Ion Ratio Lower Upper

94 100

65 36.0 0.7 40.7

66 50.2 0.8 40.8#





#11

bis(2-Chloroethyl)ether

Concen: 42.24 ng

RT: 6.60 min Scan# 373

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 93 Resp: 736352

Ion Ratio Lower Upper

93 100

63 76.9 65.6 105.6




95 33.0 13.6 53.6



Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 63.00 (62.70 to 63.70): BF0

Ion 95.00 (94.70 to 95.70): BF0



#12

1,3-Dichlorobenzene

Concen: 38.31 ng

RT: 6.79 min Scan# 390

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31


Tgt Ion: 146 Resp: 629389

Ion Ratio Lower Upper

146 100

148 62.7 51.0 76.4



75 43.4 15.0 22.6#




Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 75.00 (74.70 to 75.70): BF0









#13
 1,4-Dichlorobenzene
 Concen: 39.47 ng
 RT: 6.87 min Scan# 397
 Delta R.T. 0.00 min
 Lab File: BF090703.D
 Acq: 21 Oct 2016 10:31




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:146 Resp: 687935
 Ion Ratio Lower Upper
 146 100
 148 63.5 51.8 77.6
 111 47.6 24.9 37.3#



#14
 1,2-Dichlorobenzene
 Concen: 39.02 ng
 RT: 7.03 min Scan# 411
 Delta R.T. 0.00 min
 Lab File: BF090703.D
 Acq: 21 Oct 2016 10:31

Tgt Ion:146 Resp: 665692
 Ion Ratio Lower Upper
 146 100
 148 65.7 52.7 79.1
 111 41.9 28.8 43.2





#15

Benzyl Alcohol

Concen: 42.73 ng

RT: 7.00 min Scan# 408

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 79 Resp: 535669

Ion Ratio Lower Upper

79 100

108 68.8 88.6 132.8#

77 63.0 47.0 70.4



#16

2,2'-oxybis(1-chloropropane)

Concen: 39.43 ng

RT: 7.14 min Scan# 420

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31




Tgt Ion: 45 Resp: 1029379


Ion Ratio Lower Upper

45 100

77 16.7 0.0 28.1

79 13.3 0.0 26.0





#17

2-Methylphenol

Concen: 41.58 ng

RT: 7.11 min Scan# 418

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:107 Resp: 512411


Ion Ratio Lower Upper

107 100

108 114.7 96.6 145.0

77 53.2 23.3 34.9#

79 51.5 22.0 33.0#






Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BF0

Ion 79.00 (78.70 to 79.70): BF0



#18

Hexachloroethane

Concen: 38.46 ng

RT: 7.37 min Scan# 440

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31


Tgt Ion:117 Resp: 272124

Ion Ratio Lower Upper

117 100

119 96.1 83.8 125.6



201 60.6 55.1 82.7




Abundance Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E









#19
n-Nitroso-di-n-propylamine
Concen: 35.12 ng
RT: 7.27 min Scan# 432
Delta R.T. 0.00 min
Lab File: BF090703.D
Acq: 21 Oct 2016 10:31




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 70 Resp: 482702
Ion Ratio Lower Upper
70 100
42 53.0 63.8 95.6#
101 8.5 9.2 13.8#
130 18.0 27.3 40.9#



#20
3+4-Methylphenols
Concen: 41.88 ng
RT: 7.27 min Scan# 432
Delta R.T. 0.00 min
Lab File: BF090703.D
Acq: 21 Oct 2016 10:31

Tgt Ion: 107 Resp: 611853
Ion Ratio Lower Upper
107 100
108 91.2 74.6 114.6
77 119.0 0.7 40.7#
79 29.9 0.0 38.9





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 8.14 min Scan# 508

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:136 Resp: 923498


Ion Ratio Lower Upper

136 100

137 11.4 9.0 13.6

54 10.1 8.2 12.2

68 8.5 5.8 8.6




Abundance Ion 136.00 (135.70 to 136.70): E



Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0



Time-->



#22

Acetophenone

Concen: 41.34 ng

RT: 7.27 min Scan# 432

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Tgt Ion:105 Resp: 850817


Ion Ratio Lower Upper

105 100

71 8.7 4.7 7.1#

51 34.6 22.0 33.0#

120 21.5 30.1 45.1#




Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF0


Ion 51.00 (50.70 to 51.70): BF0

Ion 120.00 (119.70 to 120.70): E



Time-->





#23

Nitrobenzene-d5

Concen: 82.95 ng

RT: 7.42 min Scan# 445

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 82 Resp: 1396679

Ion Ratio Lower Upper




82 100

128 38.3 51.0 76.6#

54 49.6 47.5 71.3



Abundance Ion 82.00 (81.70 to 82.70): BF0
Ion 128.00 (127.70 to 128.70): B
Ion 54.00 (53.70 to 54.70): BF0



#24

Nitrobenzene

Concen: 43.66 ng

RT: 7.45 min Scan# 447

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31


Tgt Ion: 77 Resp: 796961

Ion Ratio Lower Upper



77 100


123 43.6 59.8 89.8#

65 13.9 10.2 15.4



Abundance Ion 77.00 (76.70 to 77.70): BF0
Ion 123.00 (122.70 to 123.70): B
Ion 65.00 (64.70 to 65.70): BF0





#25

Isophorone

Concen: 40.08 ng

RT: 7.69 min Scan# 468

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 82 Resp: 1277470

Ion Ratio Lower Upper

82 100

95 6.7 4.0 6.0#

138 16.0 18.3 27.5#



#26

2-Nitrophenol

Concen: 43.73 ng

RT: 7.77 min Scan# 475

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31




Tgt Ion: 139 Resp: 326092


Ion Ratio Lower Upper

139 100

109 24.8 11.0 16.4#

65 42.6 24.7 37.1#









#27
 2,4-Dimethylphenol
 Concen: 39.22 ng
 RT: 7.80 min Scan# 478
 Delta R.T. 0.00 min
 Lab File: BF090703.D
 Acq: 21 Oct 2016 10:31




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:122 Resp: 518596
 Ion Ratio Lower Upper
 122 100
 107 118.1 71.8 107.6#
 121 58.4 42.3 63.5



#28
 bis(2-Chloroethoxy)methane
 Concen: 40.98 ng
 RT: 7.89 min Scan# 486
 Delta R.T. 0.00 min
 Lab File: BF090703.D
 Acq: 21 Oct 2016 10:31

Tgt Ion: 93 Resp: 714626
 Ion Ratio Lower Upper
 93 100
 95 32.3 27.5 41.3
 123 10.1 13.7 20.5#









#29
 2,4-Dichlorophenol
 Concen: 44.72 ng
 RT: 8.01 min Scan# 496
 Delta R.T. 0.00 min
 Lab File: BF090703.D
 Acq: 21 Oct 2016 10:31




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:162 Resp: 485249
 Ion Ratio Lower Upper
 162 100
 164 64.3 44.9 84.9
 98 42.3 13.0 53.0



#30
 1,2,4-Trichlorobenzene
 Concen: 41.38 ng
 RT: 8.09 min Scan# 503
 Delta R.T. 0.00 min
 Lab File: BF090703.D
 Acq: 21 Oct 2016 10:31

Tgt Ion:180 Resp: 528411
 Ion Ratio Lower Upper
 180 100
 182 95.0 77.1 115.7
 145 35.0 23.3 34.9#





#31

Naphthalene

Concen: 39.12 ng

RT: 8.17 min Scan# 510

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:128 Resp: 1767196

Ion Ratio Lower Upper

128 100

129 11.9 8.4 12.6




127 13.9 9.9 14.9



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 40.98 ng

RT: 7.94 min Scan# 490

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31


Tgt Ion:122 Resp: 369764

Ion Ratio Lower Upper

122 100

105 125.8 76.1 116.1#



77 104.5 40.5 80.5#




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF0









#33
4-Chloroaniline
Concen: 43.38 ng
RT: 8.22 min Scan# 515
Delta R.T. 0.00 min
Lab File: BF090703.D
Acq: 21 Oct 2016 10:31




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:	127	Resp:	726769
Ion	Ratio	Lower	Upper
127	100		
129	32.8	25.8	38.6
65	28.1	17.9	26.9#
92	17.9	10.5	15.7#



#34
Hexachlorobutadiene
Concen: 41.39 ng
RT: 8.29 min Scan# 521
Delta R.T. 0.00 min
Lab File: BF090703.D
Acq: 21 Oct 2016 10:31

Tgt Ion:	225	Resp:	297182
Ion	Ratio	Lower	Upper
225	100		
223	63.5	50.2	75.2
227	66.1	52.2	78.2









#35
 Caprolactam
 Concen: 45.72 ng
 RT: 8.59 min Scan# 547
 Delta R.T. 0.00 min
 Lab File: BF090703.D
 Acq: 21 Oct 2016 10:31




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:113 Resp: 141703
 Ion Ratio Lower Upper
 113 100
 55 214.8 152.4 192.4#
 56 160.4 104.6 144.6#



#36
 4-Chloro-3-methylphenol
 Concen: 44.69 ng
 RT: 8.70 min Scan# 557
 Delta R.T. 0.00 min
 Lab File: BF090703.D
 Acq: 21 Oct 2016 10:31

Tgt Ion:107 Resp: 564102
 Ion Ratio Lower Upper
 107 100
 144 25.1 25.4 38.0#
 142 77.3 77.3 115.9









#37
 2-Methylnaphthalene
 Concen: 42.97 ng
 RT: 8.86 min Scan# 571
 Delta R.T. 0.00 min
 Lab File: BF090703.D
 Acq: 21 Oct 2016 10:31




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:142 Resp: 1134555
 Ion Ratio Lower Upper
 142 100
 141 88.4 67.5 101.3
 115 30.8 18.2 27.2#



#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 9.90 min Scan# 662
 Delta R.T. 0.00 min
 Lab File: BF090703.D
 Acq: 21 Oct 2016 10:31

Tgt Ion:164 Resp: 467316
 Ion Ratio Lower Upper
 164 100
 162 103.5 75.2 112.8
 160 47.2 31.5 47.3





#39

1,2,4,5-Tetrachlorobenzene

Concen: 39.52 ng

RT: 9.02 min Scan# 585

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:216 Resp: 498679


Ion Ratio Lower Upper

216 100

214 79.4 63.5 95.3

179 24.0 16.6 24.8

108 23.9 16.3 24.5






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E



#40

Hexachlorocyclopentadiene

Concen: 39.91 ng

RT: 9.01 min Scan# 584

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31


Tgt Ion:237 Resp: 236436

Ion Ratio Lower Upper

237 100

235 63.8 42.0 82.0



272 12.0 0.0 36.1




Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 97.02 ng
 RT: 10.69 min Scan# 731
 Delta R.T. 0.00 min
 Lab File: BF090703.D
 Acq: 21 Oct 2016 10:31




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:330 Resp: 358867
 Ion Ratio Lower Upper
 330 100
 332 95.1 76.6 114.8
 141 39.7 29.7 44.5



#42
 2,4,6-Trichlorophenol
 Concen: 42.68 ng
 RT: 9.14 min Scan# 595
 Delta R.T. 0.00 min
 Lab File: BF090703.D
 Acq: 21 Oct 2016 10:31

Tgt Ion:196 Resp: 334021
 Ion Ratio Lower Upper
 196 100
 198 94.0 75.7 113.5
 200 30.5 23.9 35.9









#43
 2,4,5-Trichlorophenol
 Concen: 40.59 ng
 RT: 9.18 min Scan# 599
 Delta R.T. 0.00 min
 Lab File: BF090703.D
 Acq: 21 Oct 2016 10:31

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:196 Resp: 322338
 Ion Ratio Lower Upper
 196 100
 198 96.0 75.4 113.0
 97 42.9 33.3 49.9
 132 26.8 24.6 37.0




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 76.49 ng
 RT: 9.23 min Scan# 603
 Delta R.T. 0.00 min
 Lab File: BF090703.D
 Acq: 21 Oct 2016 10:31

Tgt Ion:172 Resp: 2302519
 Ion Ratio Lower Upper
 172 100
 171 36.0 26.1 39.1
 170 24.8 17.4 26.2



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 38.00 ng
 RT: 9.32 min Scan# 611
 Delta R.T. 0.00 min
 Lab File: BF090703.D
 Acq: 21 Oct 2016 10:31

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:154 Resp: 1394716
 Ion Ratio Lower Upper
 154 100
 153 41.6 17.1 57.1
 76 18.4 0.0 31.6




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BF0






#46
 2-Chloronaphthalene
 Concen: 38.05 ng
 RT: 9.34 min Scan# 613
 Delta R.T. 0.00 min
 Lab File: BF090703.D
 Acq: 21 Oct 2016 10:31

Tgt Ion:162 Resp: 1037856
 Ion Ratio Lower Upper
 162 100
 127 47.5 27.6 41.4#
 164 33.3 25.4 38.2



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E





#47

2-Nitroaniline

Concen: 45.56 ng

RT: 9.45 min Scan# 622

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 65 Resp: 412548

Ion Ratio Lower Upper

65 100

92 66.5 55.4 83.0

138 99.1 115.5 173.3#



#48

Acenaphthylene

Concen: 40.81 ng

RT: 9.77 min Scan# 650

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31




Tgt Ion: 152 Resp: 1698663


Ion Ratio Lower Upper

152 100

151 20.8 14.6 22.0

153 13.4 10.3 15.5





#49

Dimethylphthalate

Concen: 43.93 ng

RT: 9.63 min Scan# 638

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:163 Resp: 1290116

Ion Ratio Lower Upper

163 100

194 3.5 4.4 6.6#




164 10.5 8.3 12.5



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 42.08 ng

RT: 9.69 min Scan# 643

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31


Tgt Ion:165 Resp: 265703

Ion Ratio Lower Upper

165 100

63 69.8 32.3 48.5#



89 66.9 28.6 43.0#




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF0

Ion 89.00 (88.70 to 89.70): BF0





#51

Acenaphthene

Concen: 42.66 ng

RT: 9.94 min Scan# 665

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:154 Resp: 1178629

Ion Ratio Lower Upper

154 100

153 114.8 82.1 123.1




152 55.6 37.9 56.9



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 41.21 ng

RT: 9.86 min Scan# 658

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31


Tgt Ion:138 Resp: 299129

Ion Ratio Lower Upper

138 100

108 10.3 5.7 8.5#



92 125.2 67.7 101.5#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0









#53
2,4-Dinitrophenol
Concen: 45.36 ng
RT: 9.96 min Scan# 667
Delta R.T. 0.00 min
Lab File: BF090703.D
Acq: 21 Oct 2016 10:31

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:184 Resp: 126909
Ion Ratio Lower Upper
184 100
63 81.9 35.4 53.2#
154 74.4 32.2 48.4#




Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 154.00 (153.70 to 154.70): E






#54
Dibenzofuran
Concen: 42.79 ng
RT: 10.11 min Scan# 680
Delta R.T. 0.00 min
Lab File: BF090703.D
Acq: 21 Oct 2016 10:31

Tgt Ion:168 Resp: 1432048
Ion Ratio Lower Upper
168 100
139 39.6 24.2 36.2#
169 13.9 10.6 15.8



Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E





#55

4-Nitrophenol

Concen: 43.20 ng

RT: 10.02 min Scan# 672

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:139 Resp: 179504

Ion Ratio Lower Upper

139 100

109 72.8 12.7 52.7#




65 129.7 45.9 85.9#



Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF0



#56

2,4-Dinitrotoluene

Concen: 39.99 ng

RT: 10.09 min Scan# 678

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Tgt Ion:165 Resp: 308612


Ion Ratio Lower Upper

165 100

63 65.3 29.8 44.6#

89 99.4 44.5 66.7#

182 1.9 3.0 4.4#






Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF0

Ion 89.00 (88.70 to 89.70): BF0

Ion 182.00 (181.70 to 182.70): E





#57

Fluorene

Concen: 44.34 ng

RT: 10.45 min Scan# 710

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:166 Resp: 1230020

Ion Ratio Lower Upper

166 100

165 97.7 72.6 109.0




167 13.3 10.6 16.0



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#58

2,3,4,6-Tetrachlorophenol

Concen: 43.42 ng

RT: 10.22 min Scan# 690

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Tgt Ion:232 Resp: 270594


Ion Ratio Lower Upper

232 100

131 58.8 38.5 57.7#

130 2.7 1.8 2.6#

166 35.9 25.0 37.6






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E





#59

Diethylphthalate

Concen: 42.96 ng

RT: 10.33 min Scan# 699

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:149 Resp: 1308276

Ion Ratio Lower Upper

149 100

177 22.3 17.5 26.3




150 12.7 9.4 14.2



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E



#60

4-Chlorophenyl-phenylether

Concen: 44.94 ng

RT: 10.44 min Scan# 709

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31


Tgt Ion:204 Resp: 569697

Ion Ratio Lower Upper

204 100

206 33.0 24.8 37.2



141 64.8 42.2 63.4#




Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E





#61

4-Nitroaniline

Concen: 44.18 ng

RT: 10.47 min Scan# 712

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:138 Resp: 324819

Ion Ratio Lower Upper

138 100

92 52.8 12.6 52.6#




108 68.8 12.4 52.4#



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BF0

Ion 108.00 (107.70 to 108.70): E



#62

Azobenzene

Concen: 42.09 ng

RT: 10.60 min Scan# 723

Delta R.T. 0.00 min

Lab File: BF090703.D

Acq: 21 Oct 2016 10:31

Tgt Ion: 77 Resp: 1501041


Ion Ratio Lower Upper

77 100

182 19.5 15.1 55.1

105 23.5 3.4 43.4

51 26.2 12.0 52.0






Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 182.00 (181.70 to 182.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 51.00 (50.70 to 51.70): BF0









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.39 min Scan# 792
Delta R.T. 0.00 min
Lab File: BF090703.D
Acq: 21 Oct 2016 10:31

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:188 Resp: 709691
Ion Ratio Lower Upper
188 100
94 8.7 11.1 16.7#
80 9.3 11.0 16.4#



#64
4,6-Dinitro-2-methylphenol
Concen: 40.71 ng
RT: 10.50 min Scan# 714
Delta R.T. 0.00 min
Lab File: BF090703.D
Acq: 21 Oct 2016 10:31

Tgt Ion:198 Resp: 181265
Ion Ratio Lower Upper
198 100
51 62.5 39.6 79.6
105 61.3 28.5 68.5

