


Data Path : Z:\HPCHEM1\BNA F\DATA\BF102116\
 Data File : BF090768.D
 Acq On : 22 Oct 2016 17:54
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102116.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Oct 21 17:36:31 2016



AutoFind: Scans 813, 814, 815; Background Corrected with Scan 810

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.1	14623	PASS
68	69	0.00	2	1.7	277	PASS
69	198	0.00	100	40.7	16047	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	54.1	21349	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	39444	PASS
199	198	5	9	6.9	2723	PASS
275	198	10	60	25.0	9857	PASS
365	198	1	100	2.8	1085	PASS
441	442	0.01	24	14.9	4379	PASS
442	198	50	100	74.3	29301	PASS
443	442	15	24	19.7	5768	PASS