

Data Path : Z:\HPCHEM1\BNA F\Data\BF102117\
 Data File : BF099721.D
 Acq On : 21 Oct 2017 12:08
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 22 00:59:05 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 20 13:11:43 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	60	0.00
2	1,4-Dioxane	0.600	0.601	-0.2	59	-0.02
3	Pyridine	1.624	1.486	8.5	56	-0.01
4	n-Nitrosodimethylamine	0.688	0.558	18.9	49#	0.00
5 S	2-Fluorophenol	1.256	1.349	-7.4	65	0.00
6	Aniline	2.092	2.078	0.7	60	0.00
7 S	Phenol-d6	1.550	1.626	-4.9	63	0.00
8	2-Chlorophenol	1.423	1.511	-6.2	65	0.00
9	Benzaldehyde	0.899	0.723	19.6	50#	0.00
10 C	Phenol	1.733	1.882	-8.6	66	0.00
11	bis(2-Chloroethyl)ether	1.348	1.363	-1.1	62	0.00
12	1,3-Dichlorobenzene	1.490	1.598	-7.2	66	0.00
13 C	1,4-Dichlorobenzene	1.516	1.612	-6.3	65	0.00
14	1,2-Dichlorobenzene	1.401	1.481	-5.7	64	0.00
15	Benzyl Alcohol	1.137	0.998	12.2	53	0.00
16	2,2'-oxybis(1-Chloropropane	2.099	1.692	19.4	50	0.00
17	2-Methylphenol	1.164	1.188	-2.1	62	0.00
18	Hexachloroethane	0.545	0.536	1.7	61	0.00
19 P	n-Nitroso-di-n-propylamine	0.990	0.861	13.0	55	0.00
20	3+4-Methylphenols	1.473	1.407	4.5	58	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	61	0.00
22	Acetophenone	0.485	0.454	6.4	59	0.00
23 S	Nitrobenzene-d5	0.312	0.309	1.0	60	0.00
24	Nitrobenzene	0.354	0.340	4.0	59	0.00
25	Isophorone	0.645	0.618	4.2	61	0.00
26 C	2-Nitrophenol	0.170	0.203	-19.4	71	0.00
27	2,4-Dimethylphenol	0.321	0.312	2.8	61	0.00
28	bis(2-Chloroethoxy)methane	0.402	0.401	0.2	63	0.00
29 C	2,4-Dichlorophenol	0.276	0.298	-8.0	67	0.00
30	1,2,4-Trichlorobenzene	0.276	0.295	-6.9	66	0.00
31	Naphthalene	0.939	0.979	-4.3	66	0.00
32	Benzoic acid	0.264	0.253	4.2	57	0.00
33	4-Chloroaniline	0.392	0.414	-5.6	65	0.00
34 C	Hexachlorobutadiene	0.142	0.148	-4.2	66	0.00
35	Caprolactam	0.088	0.095	-8.0	68	0.00
36 C	4-Chloro-3-methylphenol	0.310	0.316	-1.9	63	0.00
37	2-Methylnaphthalene	0.611	0.639	-4.6	66	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	63	0.00
39	1,2,4,5-Tetrachlorobenzene	0.596	0.642	-7.7	69	0.00
40 P	Hexachlorocyclopentadiene	0.273	0.256	6.2	57	0.00
41 S	2,4,6-Tribromophenol	0.164	0.178	-8.5	66	0.00
42 C	2,4,6-Trichlorophenol	0.423	0.427	-0.9	64	0.00
43	2,4,5-Trichlorophenol	0.422	0.438	-3.8	64	0.00
44 S	2-Fluorobiphenyl	1.198	1.283	-7.1	67	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF102117\
 Data File : BF099721.D
 Acq On : 21 Oct 2017 12:08
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 22 00:59:05 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 20 13:11:43 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.719	1.803	-4.9	67	0.00
46	2-Chloronaphthalene	1.291	1.375	-6.5	68	0.00
47	2-Nitroaniline	0.395	0.381	3.5	59	0.00
48	Acenaphthylene	1.976	2.027	-2.6	65	0.00
49	Dimethylphthalate	1.509	1.596	-5.8	68	0.00
50	2,6-Dinitrotoluene	0.329	0.365	-10.9	68	0.00
51 C	Acenaphthene	1.251	1.225	2.1	62	0.00
52	3-Nitroaniline	0.383	0.424	-10.7	67	0.00
53 P	2,4-Dinitrophenol	0.148	0.146	1.4	59	0.00
54	Dibenzofuran	1.666	1.649	1.0	63	0.00
55 P	4-Nitrophenol	0.324	0.320	1.2	60	0.00
56	2,4-Dinitrotoluene	0.426	0.429	-0.7	61	0.00
57	Fluorene	1.339	1.450	-8.3	70	0.00
58	2,3,4,6-Tetrachlorophenol	0.291	0.297	-2.1	64	0.00
59	Diethylphthalate	1.475	1.499	-1.6	65	0.00
60	4-Chlorophenyl-phenylether	0.602	0.651	-8.1	69	0.00
61	4-Nitroaniline	0.370	0.400	-8.1	66	0.00
62	Azobenzene	1.356	1.245	8.2	59	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	63	0.00
64	4,6-Dinitro-2-methylphenol	0.122	0.139	-13.9	71	0.00
65 c	n-Nitrosodiphenylamine	0.693	0.717	-3.5	68	0.00
66	4-Bromophenyl-phenylether	0.196	0.207	-5.6	69	0.00
67	Hexachlorobenzene	0.209	0.224	-7.2	70	0.00
68	Atrazine	0.208	0.201	3.4	62	0.00
69 C	Pentachlorophenol	0.130	0.108	16.9	51	0.00
70	Phenanthrene	1.071	1.114	-4.0	69	0.00
71	Anthracene	1.095	1.139	-4.0	68	0.00
72	Carbazole	1.073	1.093	-1.9	66	0.00
73	Di-n-butylphthalate	1.291	1.292	-0.1	64	0.00
74 C	Fluoranthene	1.084	1.119	-3.2	68	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	59	0.00
76	Benzidine	0.740	0.677	8.5	55	0.00
77	Pyrene	1.525	1.688	-10.7	67	0.00
78 S	Terphenyl-d14	0.879	0.981	-11.6	67	0.00
79	Butylbenzylphthalate	0.717	0.793	-10.6	65	0.00
80	Benzo(a)anthracene	1.211	1.296	-7.0	66	0.00
81	3,3'-Dichlorobenzidine	0.444	0.408	8.1	55	0.00
82	Chrysene	1.153	1.232	-6.9	65	0.00
83	Bis(2-ethylhexyl)phthalate	0.987	1.049	-6.3	63	0.00
84 c	Di-n-octyl phthalate	1.589	1.742	-9.6	67	0.00
85	Indeno(1,2,3-cd)pyrene	0.922	0.939	-1.8	67	0.00
86 I	Perylene-d12	1.000	1.000	0.0	56	0.00
87	Benzo(b)fluoranthene	1.230	1.406	-14.3	63	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF102117\
Data File : BF099721.D
Acq On : 21 Oct 2017 12:08
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 22 00:59:05 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 20 13:11:43 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.215	1.254	-3.2	59	0.00
89 C	Benzo(a)pyrene	1.129	1.199	-6.2	60	0.00
90	Dibenzo(a,h)anthracene	0.903	1.027	-13.7	66	0.00
91	Benzo(a,h,i)perylene	0.893	1.072	-20.0	69	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0