

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102122\
 Data File : BF130750.D
 Acq On : 21 Oct 2022 17:52
 Operator : CG\JU
 Sample : N5202-01 2X
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SU-04-101922

Quant Time: Oct 21 18:19:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 21 06:12:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.563	152	96425	20.000	ng	0.00
21) Naphthalene-d8	7.845	136	318142	20.000	ng	0.00
39) Acenaphthene-d10	9.592	164	152104	20.000	ng	0.00
64) Phenanthrene-d10	11.075	188	192959	20.000	ng	0.00
76) Chrysene-d12	13.710	240	152022	20.000	ng	0.00
86) Perylene-d12	15.086	264	121028	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.169	112	279014	49.897	ng	0.00
7) Phenol-d6	6.216	99	333401	49.279	ng	0.00
23) Nitrobenzene-d5	7.134	82	188831	31.486	ng	0.00
42) 2,4,6-Tribromophenol	10.381	330	56002	38.715	ng	0.00
45) 2-Fluorobiphenyl	8.916	172	355805	35.648	ng	-0.01
79) Terphenyl-d14	12.651	244	281802	31.238	ng	0.00
Target Compounds						Qvalue
19) n-Nitroso-di-n-propyla...	7.134	70	25924	5.671	ng	# 75
25) Isophorone	7.134	82	188829	19.110	ng	# 65
31) Naphthalene	7.869	128	824337	50.253	ng	99
32) Benzoic acid	7.681	122	320	6.126	ng	# 80
33) 4-Chloroaniline	7.863	127	107683	17.063	ng	# 33
37) 2-Methylnaphthalene	8.557	142	258711	24.071	ng	99
38) 1-Methylnaphthalene	8.651	142	209538	19.664	ng	97
46) 1,1'-Biphenyl	9.016	154	71801	6.768	ng	96
49) Acenaphthylene	9.451	152	72985	5.614	ng	# 95
52) Acenaphthene	9.622	154	202811	25.100	ng	99
54) 2,4-Dinitrophenol	9.657	184	53	8.574	ng	# 1
55) Dibenzofuran	9.798	168	348689	28.036	ng	96
56) 4-Nitrophenol	9.798	139	142131	125.680	ng	# 8
58) Fluorene	10.139	166	499078	48.250	ng	99
71) Phenanthrene	11.110	178	2380909	225.010	ng	98
72) Anthracene	11.151	178	581673	55.087	ng	99
73) Carbazole	11.316	167	323133	33.145	ng	99
75) Fluoranthene	12.292	202	2144168	200.441	ng	99
77) Benzidine	12.439	184	971	Below Cal		# 28
78) Pyrene	12.522	202	1726613	135.309	ng	98
81) Benzo(a)anthracene	13.698	228	795554	78.319	ng	98
83) Chrysene	13.739	228	650132	66.057	ng	96
87) Indeno(1,2,3-cd)pyrene	16.380	276	256968	29.604	ng	96
88) Benzo(b)fluoranthene	14.710	252	918384	113.966	ng	# 96
89) Benzo(k)fluoranthene	14.710	252	918384	113.436	ng	98
90) Benzo(a)pyrene	15.033	252	454752	70.262	ng	97
91) Dibenzo(a,h)anthracene	16.380	278	66950	9.334	ng	# 80
92) Benzo(g,h,i)perylene	16.768	276	197677	27.480	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102122\
Data File : BF130750.D
Acq On : 21 Oct 2022 17:52
Operator : CG\JU
Sample : N5202-01 2X
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SU-04-101922

Quant Time: Oct 21 18:19:30 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 21 06:12:36 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.563 min Scan# 795
Delta R.T. 0.000 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52




Instrument :
BNA_F
ClientSampleId :
SU-04-101922


Tgt Ion:152 Resp: 96425
Ion Ratio Lower Upper
152 100
150 157.0 127.0 190.6
115 62.1 49.9 74.9



#5
2-Fluorophenol
Concen: 49.897 ng
RT: 5.169 min Scan# 558
Delta R.T. 0.006 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52

Tgt Ion:112 Resp: 279014
Ion Ratio Lower Upper
112 100
64 54.6 46.2 69.4
63 28.4 25.1 37.7









#7
Phenol-d6
Concen: 49.279 ng
RT: 6.216 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52




Instrument :
BNA_F
ClientSampleId :
SU-04-101922


Tgt Ion: 99 Resp: 333401
Ion Ratio Lower Upper
99 100
42 17.3 13.8 20.8
71 35.6 28.3 42.5



#19
n-Nitroso-di-n-propylamine
Concen: 5.671 ng
RT: 7.134 min Scan# 892
Delta R.T. 0.141 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52

Tgt Ion: 70 Resp: 25924
Ion Ratio Lower Upper
70 100
42 43.1 44.1 66.1#
101 0.0 8.3 12.5#
130 1.6 17.8 26.6#









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 7.845 min Scan# 1013
Delta R.T. 0.000 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52




Instrument :
BNA_F
ClientSampleId :
SU-04-101922


Tgt Ion:136 Resp: 318142
Ion Ratio Lower Upper
136 100
137 10.8 9.1 13.7
54 6.7 5.8 8.8
68 4.8 4.2 6.2



#23
Nitrobenzene-d5
Concen: 31.486 ng
RT: 7.134 min Scan# 892
Delta R.T. -0.005 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52

Tgt Ion: 82 Resp: 188831
Ion Ratio Lower Upper
82 100
128 47.9 35.9 53.9
54 47.3 39.2 58.8









#25
Isophorone
Concen: 19.110 ng
RT: 7.134 min Scan# 892
Delta R.T. -0.265 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52




Instrument :
BNA_F
ClientSampleId :
SU-04-101922


Tgt Ion: 82 Resp: 188829
Ion Ratio Lower Upper
82 100
95 0.0 6.3 9.5#
138 0.0 14.6 21.8#



#31
Naphthalene
Concen: 50.253 ng
RT: 7.869 min Scan# 1017
Delta R.T. 0.000 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52

Tgt Ion: 128 Resp: 824337
Ion Ratio Lower Upper
128 100
129 10.9 8.9 13.3
127 13.0 10.6 16.0









#32
Benzoic acid
Concen: 6.126 ng
RT: 7.681 min Scan# 91
Delta R.T. -0.017 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52




Instrument :
BNA_F
ClientSampleId :
SU-04-101922


Tgt Ion:122 Resp: 320
Ion Ratio Lower Upper
122 100
105 103.1 109.2 149.2#
77 78.4 75.7 115.7



#33
4-Chloroaniline
Concen: 17.063 ng
RT: 7.863 min Scan# 1016
Delta R.T. -0.070 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52

Tgt Ion:127 Resp: 107683
Ion Ratio Lower Upper
127 100
129 85.8 26.2 39.4#
65 0.0 22.6 33.8#
92 0.0 15.0 22.4#





#37

2-Methylnaphthalene

Concen: 24.071 ng

RT: 8.557 min Scan# 1134

Delta R.T. 0.000 min

Lab File: BF130750.D

Acq: 21 Oct 2022 17:52

Instrument :

BNA_F

ClientSampleId :





SU-04-101922

Tgt Ion:142 Resp: 258711

Ion Ratio Lower Upper

142 100

141 89.7 71.0 106.4



#38

1-Methylnaphthalene

Concen: 19.664 ng

RT: 8.651 min Scan# 1150

Delta R.T. -0.006 min

Lab File: BF130750.D




Acq: 21 Oct 2022 17:52


Tgt Ion:142 Resp: 209538

Ion Ratio Lower Upper

142 100

141 88.5 73.0 109.4









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.592 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52




Instrument :
BNA_F
ClientSampleId :
SU-04-101922


Tgt Ion:164 Resp: 152104
Ion Ratio Lower Upper
164 100
162 100.6 81.4 122.0
160 43.9 35.0 52.4



#42
2,4,6-Tribromophenol
Concen: 38.715 ng
RT: 10.381 min Scan# 1444
Delta R.T. -0.005 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52

Tgt Ion:330 Resp: 56002
Ion Ratio Lower Upper
330 100
332 101.3 81.4 122.2
141 35.1 24.6 37.0





#45

2-Fluorobiphenyl

Concen: 35.648 ng

RT: 8.916 min Scan# 1195

Delta R.T. -0.012 min

Lab File: BF130750.D

Acq: 21 Oct 2022 17:52

Instrument :

BNA_F

ClientSampleId :

SU-04-101922





Tgt Ion:172 Resp: 355805

Ion Ratio Lower Upper

172 100

171 34.6 28.2 42.2

170 22.4 18.5 27.7



#46

1,1'-Biphenyl

Concen: 6.768 ng

RT: 9.016 min Scan# 1212

Delta R.T. -0.006 min

Lab File: BF130750.D

Acq: 21 Oct 2022 17:52




Tgt Ion:154 Resp: 71801


Ion Ratio Lower Upper

154 100

153 37.8 20.7 60.7

76 12.6 0.0 31.7





#49

Acenaphthylene

Concen: 5.614 ng

RT: 9.451 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF130750.D

Acq: 21 Oct 2022 17:52

Instrument :

BNA_F

ClientSampleId :

SU-04-101922





Tgt Ion:152 Resp: 72985

Ion Ratio Lower Upper

152 100

151 20.0 16.0 24.0

153 17.9 10.5 15.7#



#52

Acenaphthene

Concen: 25.100 ng

RT: 9.622 min Scan# 1315

Delta R.T. -0.006 min

Lab File: BF130750.D

Acq: 21 Oct 2022 17:52




Tgt Ion:154 Resp: 202811


Ion Ratio Lower Upper

154 100

153 109.9 88.4 132.6

152 53.2 42.0 63.0





#54

2,4-Dinitrophenol

Concen: 8.574 ng

RT: 9.657 min Scan# 1321

Delta R.T. -0.041 min

Lab File: BF130750.D

Acq: 21 Oct 2022 17:52

Instrument :

BNA_F

ClientSampleId :

SU-04-101922





Tgt Ion:184 Resp: 53

Ion Ratio Lower Upper

184 100

63 300.7 47.7 71.5#

154 564.0 55.0 82.4#



#55

Dibenzofuran

Concen: 28.036 ng

RT: 9.798 min Scan# 1345

Delta R.T. 0.000 min

Lab File: BF130750.D

Acq: 21 Oct 2022 17:52




Tgt Ion:168 Resp: 348689


Ion Ratio Lower Upper

168 100

139 39.7 34.3 51.5

169 14.4 11.0 16.6









#56
4-Nitrophenol
Concen: 125.680 ng
RT: 9.798 min Scan# 11
Delta R.T. 0.029 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52




Instrument :
BNA_F
ClientSampleId :
SU-04-101922


Tgt Ion:139 Resp: 142131
Ion Ratio Lower Upper
139 100
109 0.9 55.1 95.1#
65 1.3 72.2 112.2#



#58
Fluorene
Concen: 48.250 ng
RT: 10.139 min Scan# 1403
Delta R.T. -0.006 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52

Tgt Ion:166 Resp: 499078
Ion Ratio Lower Upper
166 100
165 95.8 77.6 116.4
167 13.6 10.3 15.5









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.075 min Scan# 1562
Delta R.T. 0.000 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52




Instrument :
BNA_F
ClientSampleId :
SU-04-101922


Tgt Ion:188 Resp: 192959
Ion Ratio Lower Upper
188 100
94 9.0 6.7 10.1
80 8.6 7.4 11.0



#71
Phenanthrene
Concen: 225.010 ng
RT: 11.110 min Scan# 1568
Delta R.T. 0.012 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52

Tgt Ion:178 Resp: 2380909
Ion Ratio Lower Upper
178 100
176 19.8 15.1 22.7
179 16.3 12.4 18.6





#72

Anthracene

Concen: 55.087 ng

RT: 11.151 min Scan# 1575

Delta R.T. 0.000 min

Lab File: BF130750.D

Acq: 21 Oct 2022 17:52

Instrument :

BNA_F

ClientSampleId :

SU-04-101922





Tgt Ion:178 Resp: 581673

Ion Ratio Lower Upper

178 100

176 18.5 14.6 21.8

179 15.9 12.4 18.6



#73

Carbazole

Concen: 33.145 ng

RT: 11.316 min Scan# 1603

Delta R.T. 0.000 min

Lab File: BF130750.D

Acq: 21 Oct 2022 17:52




Tgt Ion:167 Resp: 323133


Ion Ratio Lower Upper

167 100

166 20.9 17.3 25.9

139 13.0 10.7 16.1









#75
Fluoranthene
Concen: 200.441 ng
RT: 12.292 min Scan# 1
Delta R.T. 0.012 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52




Instrument :
BNA_F
ClientSampleId :
SU-04-101922


Tgt Ion:202 Resp: 2144168
Ion Ratio Lower Upper
202 100
101 11.5 0.0 31.0
203 18.2 0.0 37.6



#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.710 min Scan# 2010
Delta R.T. 0.006 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52

Tgt Ion:240 Resp: 152022
Ion Ratio Lower Upper
240 100
120 10.3 9.6 14.4
236 25.2 20.2 30.2





#78

Pyrene

Concen: 135.309 ng

RT: 12.522 min Scan# 1808

Delta R.T. 0.012 min

Lab File: BF130750.D

Acq: 21 Oct 2022 17:52

Instrument :

BNA_F

ClientSampleId :

SU-04-101922





Tgt Ion:202 Resp: 1726613

Ion Ratio Lower Upper

202 100

200 21.5 16.8 25.2

203 18.7 13.8 20.8



#79

Terphenyl-d14

Concen: 31.238 ng

RT: 12.651 min Scan# 1830

Delta R.T. -0.006 min

Lab File: BF130750.D

Acq: 21 Oct 2022 17:52




Tgt Ion:244 Resp: 281802


Ion Ratio Lower Upper

244 100

212 7.0 5.4 8.0

122 11.3 8.2 12.4









#81
Benzo(a)anthracene
Concen: 78.319 ng
RT: 13.698 min Scan# 20
Delta R.T. -0.000 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52




Instrument :
BNA_F
ClientSampleId :
SU-04-101922


Tgt Ion:228 Resp: 795554
Ion Ratio Lower Upper
228 100
226 27.9 21.8 32.8
229 21.1 15.4 23.2



#83
Chrysene
Concen: 66.057 ng
RT: 13.739 min Scan# 2015
Delta R.T. 0.006 min
Lab File: BF130750.D
Acq: 21 Oct 2022 17:52

Tgt Ion:228 Resp: 650132
Ion Ratio Lower Upper
228 100
226 30.1 23.6 35.4
229 23.2 15.8 23.6





#86

Perylene-d12

Concen: 20.000 ng

RT: 15.086 min Scan# 21

Delta R.T. 0.006 min

Lab File: BF130750.D

Acq: 21 Oct 2022 17:52

Instrument :

BNA_F

ClientSampleId :

SU-04-101922





Tgt Ion:264 Resp: 121028

Ion Ratio Lower Upper

264 100

260 23.8 18.8 28.2

265 23.9 16.2 24.4



#87

Indeno(1,2,3-cd)pyrene

Concen: 29.604 ng

RT: 16.380 min Scan# 2464

Delta R.T. 0.006 min

Lab File: BF130750.D

Acq: 21 Oct 2022 17:52




Tgt Ion:276 Resp: 256968


Ion Ratio Lower Upper

276 100

138 19.4 17.2 25.8

277 23.5 20.1 30.1





#88

Benzo(b)fluoranthene

Concen: 113.966 ng

RT: 14.710 min Scan# 2180

Delta R.T. 0.006 min

Lab File: BF130750.D

Acq: 21 Oct 2022 17:52

Instrument :

BNA_F

ClientSampleId :

SU-04-101922





Tgt Ion:252 Resp: 918384

Ion Ratio Lower Upper

252 100

253 22.8 17.4 26.0

125 10.6 6.4 9.6#



#89

Benzo(k)fluoranthene

Concen: 113.436 ng

RT: 14.710 min Scan# 2180

Delta R.T. -0.018 min

Lab File: BF130750.D

Acq: 21 Oct 2022 17:52




Tgt Ion:252 Resp: 918384


Ion Ratio Lower Upper

252 100

253 22.8 17.7 26.5

125 10.6 7.2 10.8





#90

Benzo(a)pyrene

Concen: 70.262 ng

RT: 15.033 min Scan# 21

Delta R.T. 0.012 min

Lab File: BF130750.D

Acq: 21 Oct 2022 17:52

Instrument :

BNA_F

ClientSampleId :

SU-04-101922





Tgt Ion:252 Resp: 454752

Ion Ratio Lower Upper

252 100

253 23.3 17.4 26.0

125 9.9 8.3 12.5



#91

Dibenzo(a,h)anthracene

Concen: 9.334 ng

RT: 16.380 min Scan# 2464

Delta R.T. 0.006 min

Lab File: BF130750.D

Acq: 21 Oct 2022 17:52




Tgt Ion:278 Resp: 66950


Ion Ratio Lower Upper

278 100

139 0.0 11.4 17.0#


279 28.1 18.4 27.6#





#92
 Benzo(g,h,i)perylene
 Concen: 27.480 ng
 RT: 16.768 min Scan# 21
 Delta R.T. 0.012 min
 Lab File: BF130750.D
 Acq: 21 Oct 2022 17:52

Instrument :
 BNA_F
 ClientSampleId :
 SU-04-101922



Tgt Ion:276 Resp: 197677
 Ion Ratio Lower Upper
 276 100
 277 25.6 19.1 28.7
 138 19.4 14.2 21.2

