

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102220\
 Data File : BF122080.D
 Acq On : 22 Oct 2020 13:26
 Operator : JU/CG
 Sample : L4467-03MSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 VNJ-240MSD

Manual Integrations
 APPROVED

mohammad
 10/23/2020 1:31:52 PM

Quant Time: Oct 22 13:54:19 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 21 12:04:42 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.728	152	104308	20.00	ng	0.00
21) Naphthalene-d8	8.010	136	415298	20.00	ng	0.00
39) Acenaphthene-d10	9.763	164	216234	20.00	ng	0.00
64) Phenanthrene-d10	11.251	188	408580	20.00	ng	0.00
76) Chrysene-d12	13.892	240	272910	20.00	ng	0.00
86) Perylene-d12	15.321	264	239792	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.357	112	611264	86.49	ng	0.01
7) Phenol-d6	6.386	99	755762	83.07	ng	0.00
23) Nitrobenzene-d5	7.298	82	536665	66.27	ng	0.00
42) 2,4,6-Tribromophenol	10.557	330	237989	88.97	ng	0.00
45) 2-Fluorobiphenyl	9.086	172	908816	61.84	ng	0.00
79) Terphenyl-d14	12.827	244	949808	66.33	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.457	88	99984	32.17	ng	98
3) Pyridine	3.181	79	278134	34.03	ng	97
4) n-Nitrosodimethylamine	3.140	42	141343	35.78	ng	98
6) Aniline	6.392	93	217908	19.45	ng	# 3
8) 2-Chlorophenol	6.522	128	266778	37.35	ng	97
9) Benzaldehyde	6.281	77	115997	19.28	ng	99
10) Phenol	6.398	94	346901	36.95	ng	100
11) bis(2-Chloroethyl)ether	6.469	93	265880	36.63	ng	96
12) 1,3-Dichlorobenzene	6.669	146	289531	36.03	ng	99
13) 1,4-Dichlorobenzene	6.745	146	294078	36.45	ng	99
14) 1,2-Dichlorobenzene	6.898	146	273410	37.07	ng	98
15) Benzyl Alcohol	6.886	79	237740	40.40	ng	97
16) 2,2'-oxybis(1-Chloropr...	7.004	45	406468	36.39	ng	# 23
17) 2-Methylphenol	7.004	107	213767	35.00	ng	# 88
18) Hexachloroethane	7.233	117	108528	37.24	ng	100
19) n-Nitroso-di-n-propyla...	7.151	70	198806	38.38	ng	99
20) 3+4-Methylphenols	7.157	107	287316	39.01	ng	# 65
22) Acetophenone	7.145	105	370037	34.63	ng	# 90
24) Nitrobenzene	7.322	77	297934	34.42	ng	95
25) Isophorone	7.557	82	535380	34.96	ng	100
26) 2-Nitrophenol	7.633	139	142125	35.66	ng	91
27) 2,4-Dimethylphenol	7.681	122	234244	34.47	ng	99
28) bis(2-Chloroethoxy)met...	7.769	93	337244	35.20	ng	100
29) 2,4-Dichlorophenol	7.886	162	226002	35.52	ng	98
30) 1,2,4-Trichlorobenzene	7.957	180	232837	34.41	ng	100
31) Naphthalene	8.033	128	764938	34.38	ng	100
32) Benzoic acid	7.839	122	126217	34.98	ng	96
33) 4-Chloroaniline	8.092	127	99339	10.48	ng	99
34) Hexachlorobutadiene	8.145	225	143369	35.74	ng	98
35) Caprolactam	8.469	113	71201m	35.23	ng	
36) 4-Chloro-3-methylphenol	8.592	107	247867	36.27	ng	96
37) 2-Methylnaphthalene	8.722	142	506748	35.17	ng	99
38) 1-Methylnaphthalene	8.822	142	460537	34.51	ng	100
40) 1,2,4,5-Tetrachloroben...	8.892	216	237492	34.03	ng	99
41) Hexachlorocyclopentadiene	8.875	237	94208	54.08	ng	99
43) 2,4,6-Trichlorophenol	9.010	196	153841	35.73	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102220\
 Data File : BF122080.D
 Acq On : 22 Oct 2020 13:26
 Operator : JU/CG
 Sample : L4467-03MSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 VNJ-240MSD


Manual Integrations
 APPROVED

mohammad
 10/23/2020 1:31:52 PM


Quant Time: Oct 22 13:54:19 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 21 12:04:42 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.063	196	161975	34.83	ng	100
46) 1,1'-Biphenyl	9.186	154	605537	34.44	ng	99
47) 2-Chloronaphthalene	9.210	162	472917	34.59	ng	99
48) 2-Nitroaniline	9.322	65	163996	36.36	ng	98
49) Acenaphthylene	9.627	152	715764	34.08	ng	100
50) Dimethylphthalate	9.492	163	681351	43.09	ng	99
51) 2,6-Dinitrotoluene	9.563	165	124328	35.85	ng	94
52) Acenaphthene	9.798	154	436822	34.97	ng	99
53) 3-Nitroaniline	9.733	138	71523	18.13	ng	99
54) 2,4-Dinitrophenol	9.857	184	107479	62.82	ng	90
55) Dibenzofuran	9.969	168	630254	34.50	ng	97
56) 4-Nitrophenol	9.922	139	115437	53.80	ng	84
57) 2,4-Dinitrotoluene	9.974	165	157569	36.90	ng	97
58) Fluorene	10.310	166	525002	35.67	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.098	232	136362	37.90	ng	96
60) Diethylphthalate	10.186	149	571251	37.47	ng	99
61) 4-Chlorophenyl-phenyle...	10.304	204	259382	36.75	ng	100
62) 4-Nitroaniline	10.351	138	118014	29.95	ng	95
63) Azobenzene	10.463	77	583898	36.21	ng	99
65) 4,6-Dinitro-2-methylph...	10.386	198	89694	34.17	ng	99
66) n-Nitrosodiphenylamine	10.427	169	488536	34.71	ng	100
67) 4-Bromophenyl-phenylether	10.792	248	167263	35.43	ng	97
68) Hexachlorobenzene	10.863	284	185756	34.77	ng	96
69) Atrazine	10.957	200	136981	39.80	ng	98
70) Pentachlorophenol	11.069	266	130084	63.72	ng	98
71) Phenanthrene	11.274	178	776026	34.64	ng	99
72) Anthracene	11.327	178	802084	34.52	ng	99
73) Carbazole	11.486	167	702727	34.01	ng	100
74) Di-n-butylphthalate	11.804	149	917064	38.58	ng	100
75) Fluoranthene	12.463	202	828753	34.50	ng	98
77) Benzidine	12.586	184	225945	26.85	ng	100
78) Pyrene	12.692	202	821223	39.49	ng	100
80) Butylbenzylphthalate	13.304	149	365714	44.34	ng	96
81) Benzo(a)anthracene	13.880	228	644711	36.05	ng	99
82) 3,3'-Dichlorobenzidine	13.839	252	141900	21.39	ng	99
83) Chrysene	13.915	228	624773	35.59	ng	100
84) Bis(2-ethylhexyl)phtha...	13.857	149	526292	47.00	ng	99
85) Di-n-octyl phthalate	14.468	149	786387	43.42	ng	100
87) Indeno(1,2,3-cd)pyrene	16.739	276	656372	42.16	ng	98
88) Benzo(b)fluoranthene	14.909	252	604527	37.30	ng	99
89) Benzo(k)fluoranthene	14.939	252	522692	33.95	ng	99
90) Benzo(a)pyrene	15.262	252	505545	36.11	ng	99
91) Dibenzo(a,h)anthracene	16.739	278	541161	42.21	ng	99
92) Benzo(g,h,i)perylene	17.156	276	561146	43.74	ng	98




(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.728 min Scan# 789
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26





Tgt Ion:152 Resp: 104308
Ion Ratio Lower Upper
152 100
150 156.0 125.3 187.9
115 62.4 51.9 77.9



#2
1,4-Dioxane
Concen: 32.17 ng
RT: 2.457 min Scan# 63
Delta R.T. 0.053 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26




Tgt Ion: 88 Resp: 99984
Ion Ratio Lower Upper
88 100
58 80.4 65.8 98.8
43 32.9 25.8 38.8



Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD

Manual Integrations
APPROVED

mohammad
10/23/2020 1:31:52 PM







#3
Pyridine
Concen: 34.03 ng
RT: 3.181 min Scan# 186
Delta R.T. 0.059 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion: 79 Resp: 278134
Ion Ratio Lower Upper
79 100
52 80.8 62.2 93.4
51 39.6 31.1 46.7

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#4
n-Nitrosodimethylamine
Concen: 35.78 ng
RT: 3.140 min Scan# 179
Delta R.T. 0.047 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion: 42 Resp: 141343
Ion Ratio Lower Upper
42 100
74 123.3 100.2 150.2
44 7.4 5.8 8.6









#5
2-Fluorophenol
Concen: 86.49 ng
RT: 5.357 min Scan# 556
Delta R.T. 0.012 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion: 112 Resp: 611264
Ion Ratio Lower Upper
112 100
64 64.8 50.4 75.6
63 33.7 25.8 38.6

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#6
Aniline
Concen: 19.45 ng
RT: 6.392 min Scan# 732
Delta R.T. -0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion: 93 Resp: 217908
Ion Ratio Lower Upper
93 100
66 102.1 32.7 49.1#
65 64.9 16.6 24.8#









#7
Phenol-d6
Concen: 83.07 ng
RT: 6.386 min Scan# 731
Delta R.T. 0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion: 99 Resp: 755762
Ion Ratio Lower Upper
99 100
42 20.4 15.9 23.9
71 35.3 27.1 40.7

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#8
2-Chlorophenol
Concen: 37.35 ng
RT: 6.522 min Scan# 754
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26


Tgt Ion: 128 Resp: 266778
Ion Ratio Lower Upper
128 100
130 32.9 12.4 52.4
64 51.6 34.8 74.8





#9
Benzaldehyde
Concen: 19.28 ng
RT: 6.281 min Scan# 713
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD






Tgt Ion: 77 Resp: 115997


Ion	Ratio	Lower	Upper
77	100		
105	96.4	76.7	116.7
106	89.8	71.3	111.3

Manual Integrations
APPROVED

mohammad
10/23/2020 1:31:52 PM






#10
Phenol
Concen: 36.95 ng
RT: 6.398 min Scan# 733
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26



Tgt Ion: 94 Resp: 346901

Ion	Ratio	Lower	Upper
94	100		
65	37.2	17.1	57.1
66	55.7	36.0	76.0









#11
bis(2-Chloroethyl)ether
Concen: 36.63 ng
RT: 6.469 min Scan# 745
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion: 93 Resp: 265880
Ion Ratio Lower Upper
93 100
63 78.3 62.4 102.4
95 32.0 12.6 52.6

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#12
1,3-Dichlorobenzene
Concen: 36.03 ng
RT: 6.669 min Scan# 779
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion: 146 Resp: 289531
Ion Ratio Lower Upper
146 100
148 63.9 51.0 76.4
75 32.4 27.6 41.4









#13
1,4-Dichlorobenzene
Concen: 36.45 ng
RT: 6.745 min Scan# 792
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:146 Resp: 294078
Ion Ratio Lower Upper
146 100
148 63.0 50.1 75.1
111 42.9 34.6 52.0

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#14
1,2-Dichlorobenzene
Concen: 37.07 ng
RT: 6.898 min Scan# 818
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:146 Resp: 273410
Ion Ratio Lower Upper
146 100
148 64.5 51.1 76.7
111 45.6 34.2 51.4









#15
Benzyl Alcohol
Concen: 40.40 ng
RT: 6.886 min Scan# 816
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion: 79 Resp: 237740
Ion Ratio Lower Upper
79 100
108 75.0 62.3 93.5
77 61.7 50.6 75.8

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#16
2,2'-oxybis(1-Chloropropane)
Concen: 36.39 ng
RT: 7.004 min Scan# 836
Delta R.T. -0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion: 45 Resp: 406468
Ion Ratio Lower Upper
45 100
77 43.5 0.0 33.3#
79 40.0 0.0 30.4#









#17
2-Methylphenol
Concen: 35.00 ng
RT: 7.004 min Scan# 836
Delta R.T. 0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:	107	Resp:	213767
Ion	Ratio	Lower	Upper
107	100		
108	112.5	90.9	136.3
77	66.4	37.5	56.3#
79	61.1	37.4	56.0#

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#18
Hexachloroethane
Concen: 37.24 ng
RT: 7.233 min Scan# 875
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:	117	Resp:	108528
Ion	Ratio	Lower	Upper
117	100		
119	95.3	76.3	114.5
201	89.5	71.8	107.6









#19

n-Nitroso-di-n-propylamine
Concen: 38.38 ng
RT: 7.151 min Scan# 861
Delta R.T. -0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26




Tgt Ion:	70	Resp:	198806
Ion	Ratio	Lower	Upper
70	100		
42	63.2	50.2	75.4
101	9.2	7.5	11.3
130	19.8	15.5	23.3



#20

3+4-Methylphenols
Concen: 39.01 ng
RT: 7.157 min Scan# 862
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26


Tgt Ion:	107	Resp:	287316
Ion	Ratio	Lower	Upper
107	100		
108	86.2	68.9	108.9
77	53.3	108.0	148.0#
79	27.8	10.5	50.5



Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD

Manual Integrations
APPROVED

mohammad
10/23/2020 1:31:52 PM







#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.010 min Scan# 1007
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:	136	Resp:	415298
Ion Ratio	Lower	Upper	
136	100		
137	10.8	8.6	13.0
54	9.6	7.1	10.7
68	6.6	5.0	7.4

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#22
Acetophenone
Concen: 34.63 ng
RT: 7.145 min Scan# 860
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:	105	Resp:	370037
Ion Ratio	Lower	Upper	
105	100		
71	6.5	5.3	7.9
51	31.2	33.2	49.8#
120	21.9	17.4	26.0





#23

Nitrobenzene-d5

Concen: 66.27 ng

RT: 7.298 min Scan# 886

Delta R.T. -0.006 min

Lab File: BF122080.D

Acq: 22 Oct 2020 13:26

Tgt Ion: 82 Resp: 536665

Ion	Ratio	Lower	Upper
82	100		
128	40.5	35.0	52.4
54	54.8	43.9	65.9

Instrument :

BNA_F

ClientSampleId :





VNJ-240MSD

Manual Integrations

APPROVED

mohammad

10/23/2020 1:31:52 PM



#24

Nitrobenzene

Concen: 34.42 ng

RT: 7.322 min Scan# 890




Delta R.T. -0.000 min


Lab File: BF122080.D

Acq: 22 Oct 2020 13:26

Tgt Ion: 77 Resp: 297934

Ion	Ratio	Lower	Upper
77	100		
123	48.8	35.6	53.4
65	13.6	11.2	16.8









#25
Isophorone
Concen: 34.96 ng
RT: 7.557 min Scan# 930
Delta R.T. -0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion: 82 Resp: 535380
Ion Ratio Lower Upper
82 100
95 6.9 5.4 8.0
138 16.8 13.4 20.0

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#26
2-Nitrophenol
Concen: 35.66 ng
RT: 7.633 min Scan# 943
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion: 139 Resp: 142125
Ion Ratio Lower Upper
139 100
109 31.6 23.0 34.4
65 57.1 39.9 59.9









#27
2,4-Dimethylphenol
Concen: 34.47 ng
RT: 7.681 min Scan# 951
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:	122	Resp:	234244
Ion	Ratio	Lower	Upper
122	100		
107	114.9	90.8	136.2
121	58.3	47.5	71.3

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#28
bis(2-Chloroethoxy)methane
Concen: 35.20 ng
RT: 7.769 min Scan# 966
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26


Tgt Ion:	93	Resp:	337244
Ion	Ratio	Lower	Upper
93	100		
95	32.5	26.2	39.4
123	11.0	8.8	13.2





#29
2,4-Dichlorophenol
Concen: 35.52 ng
RT: 7.886 min Scan# 986
Delta R.T. 0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26




Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Tgt Ion:162 Resp: 226002
Ion Ratio Lower Upper
162 100
164 64.0 42.5 82.5
98 38.1 20.1 60.1

Manual Integrations
APPROVED



mohammad
10/23/2020 1:31:52 PM




#30
1,2,4-Trichlorobenzene
Concen: 34.41 ng
RT: 7.957 min Scan# 998
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26



Tgt Ion:180 Resp: 232837
Ion Ratio Lower Upper
180 100
182 99.6 79.4 119.2
145 31.7 25.5 38.3









#31
Naphthalene
Concen: 34.38 ng
RT: 8.033 min Scan# 1011
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:128 Resp: 764938
Ion Ratio Lower Upper
128 100
129 11.3 8.7 13.1
127 13.5 10.8 16.2

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#32
Benzoic acid
Concen: 34.98 ng
RT: 7.839 min Scan# 978
Delta R.T. -0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:122 Resp: 126217
Ion Ratio Lower Upper
122 100
105 127.0 103.3 143.3
77 98.8 73.9 113.9









#33
4-Chloroaniline
Concen: 10.48 ng
RT: 8.092 min Scan# 1021
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:	127	Resp:	99339
Ion Ratio	Lower	Upper	
127	100		
129	33.0	26.3	39.5
65	34.0	26.2	39.2
92	20.5	15.9	23.9

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM




#34
Hexachlorobutadiene
Concen: 35.74 ng
RT: 8.145 min Scan# 1030
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:	225	Resp:	143369
Ion Ratio	Lower	Upper	
225	100		
223	62.6	51.1	76.7
227	62.3	51.0	76.6









#35
Caprolactam
Concen: 35.23 ng m
RT: 8.469 min Scan# 1085
Delta R.T. -0.012 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26





Tgt Ion:113 Resp: 71201
Ion Ratio Lower Upper
113 100
55 197.2 167.5 207.5
56 148.0 121.2 161.2



#36
4-Chloro-3-methylphenol
Concen: 36.27 ng
RT: 8.592 min Scan# 1106
Delta R.T. 0.012 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26




Tgt Ion:107 Resp: 247867
Ion Ratio Lower Upper
107 100
144 26.0 20.1 30.1
142 78.4 59.3 88.9



Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD

Manual Integrations
APPROVED

mohammad
10/23/2020 1:31:52 PM







#37
2-Methylnaphthalene
Concen: 35.17 ng
RT: 8.722 min Scan# 1128
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD

Tgt Ion:142 Resp: 506748
Ion Ratio Lower Upper
142 100
141 88.6 70.2 105.4
115 37.5 28.6 43.0




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#38
1-Methylnaphthalene
Concen: 34.51 ng
RT: 8.822 min Scan# 1145
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26


Tgt Ion:142 Resp: 460537
Ion Ratio Lower Upper
142 100
141 92.6 74.5 111.7
116 4.0 3.2 4.8





#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.763 min Scan# 1305
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26




Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Tgt Ion:164 Resp: 216234
Ion Ratio Lower Upper
164 100
162 101.5 83.2 124.8
160 46.1 37.5 56.3

Manual Integrations
APPROVED



mohammad
10/23/2020 1:31:52 PM




#40
1,2,4,5-Tetrachlorobenzene
Concen: 34.03 ng
RT: 8.892 min Scan# 1157
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26



Tgt Ion:216 Resp: 237492
Ion Ratio Lower Upper
216 100
214 79.8 63.4 95.2
179 21.0 17.0 25.6
108 19.5 15.5 23.3









#41
Hexachlorocyclopentadiene
Concen: 54.08 ng
RT: 8.875 min Scan# 1154
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:	237	Resp:	94208
Ion	Ratio	Lower	Upper
237	100		
235	62.0	41.8	81.8
272	14.5	0.0	33.5

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#42
2,4,6-Tribromophenol
Concen: 88.97 ng
RT: 10.557 min Scan# 1440
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:	330	Resp:	237989
Ion	Ratio	Lower	Upper
330	100		
332	96.5	77.3	115.9
141	28.4	22.3	33.5









#43
2,4,6-Trichlorophenol
Concen: 35.73 ng
RT: 9.010 min Scan# 1177
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:196 Resp: 153841
Ion Ratio Lower Upper
196 100
198 94.2 75.8 113.6
200 30.2 24.6 36.8

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#44
2,4,5-Trichlorophenol
Concen: 34.83 ng
RT: 9.063 min Scan# 1186
Delta R.T. 0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:196 Resp: 161975
Ion Ratio Lower Upper
196 100
198 94.8 76.1 114.1
97 50.9 41.0 61.6
132 27.7 22.4 33.6









#45
2-Fluorobiphenyl
Concen: 61.84 ng
RT: 9.086 min Scan# 1190
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:172 Resp: 908816
Ion Ratio Lower Upper
172 100
171 35.6 29.1 43.7
170 24.1 19.8 29.6

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#46
1,1'-Biphenyl
Concen: 34.44 ng
RT: 9.186 min Scan# 1207
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:154 Resp: 605537
Ion Ratio Lower Upper
154 100
153 40.5 21.3 61.3
76 13.8 0.0 34.7









#47
2-Chloronaphthalene
Concen: 34.59 ng
RT: 9.210 min Scan# 1211
Delta R.T. -0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:	162	Resp:	472917
Ion	Ratio	Lower	Upper
162	100		
127	42.6	34.4	51.6
164	32.4	26.3	39.5

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#48
2-Nitroaniline
Concen: 36.36 ng
RT: 9.322 min Scan# 1230
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:	65	Resp:	163996
Ion	Ratio	Lower	Upper
65	100		
92	68.6	55.1	82.7
138	103.1	80.5	120.7





#49

Acenaphthylene

Concen: 34.08 ng

RT: 9.627 min Scan# 1282

Delta R.T. -0.000 min

Lab File: BF122080.D

Acq: 22 Oct 2020 13:26

Tgt Ion:	152	Resp:	715764
Ion	Ratio	Lower	Upper
152	100		
151	20.4	16.4	24.6
153	12.9	10.6	15.8

Instrument :





BNA_F

ClientSampleId :

VNJ-240MSD

Manual Integrations
APPROVED

mohammad
10/23/2020 1:31:52 PM



#50

Dimethylphthalate

Concen: 43.09 ng




RT: 9.492 min Scan# 1259


Delta R.T. -0.000 min

Lab File: BF122080.D

Acq: 22 Oct 2020 13:26

Tgt Ion:	163	Resp:	681351
Ion	Ratio	Lower	Upper
163	100		
194	4.0	3.3	4.9
164	10.0	8.2	12.4









#51
2,6-Dinitrotoluene
Concen: 35.85 ng
RT: 9.563 min Scan# 1271
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:165 Resp: 124328
Ion Ratio Lower Upper
165 100
63 59.7 51.8 77.6
89 56.9 49.0 73.6

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#52
Acenaphthene
Concen: 34.97 ng
RT: 9.798 min Scan# 1311
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:154 Resp: 436822
Ion Ratio Lower Upper
154 100
153 111.0 89.8 134.8
152 53.4 43.2 64.8









#53
3-Nitroaniline
Concen: 18.13 ng
RT: 9.733 min Scan# 1300
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:138 Resp: 71523
Ion Ratio Lower Upper
138 100
108 11.8 8.6 13.0
92 119.2 96.6 144.8

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM




#54
2,4-Dinitrophenol
Concen: 62.82 ng
RT: 9.857 min Scan# 1321
Delta R.T. 0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:184 Resp: 107479
Ion Ratio Lower Upper
184 100
63 66.6 45.9 68.9
154 63.0 45.8 68.6









#55
Dibenzofuran
Concen: 34.50 ng
RT: 9.969 min Scan# 1340
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26





Tgt Ion:168 Resp: 630254
Ion Ratio Lower Upper
168 100
139 41.0 31.0 46.6
169 14.5 11.3 16.9



#56
4-Nitrophenol
Concen: 53.80 ng
RT: 9.922 min Scan# 1332
Delta R.T. 0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26




Tgt Ion:139 Resp: 115437
Ion Ratio Lower Upper
139 100
109 72.8 45.4 85.4
65 107.7 68.4 108.4



Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD

Manual Integrations
APPROVED

mohammad
10/23/2020 1:31:52 PM







#57
2,4-Dinitrotoluene
Concen: 36.90 ng
RT: 9.974 min Scan# 1341
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:	165	Resp:	157569
Ion	Ratio	Lower	Upper
165	100		
63	56.9	41.3	61.9
89	77.0	61.8	92.8
182	2.7	2.1	3.1

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM




#58
Fluorene
Concen: 35.67 ng
RT: 10.310 min Scan# 1398
Delta R.T. -0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:	166	Resp:	525002
Ion	Ratio	Lower	Upper
166	100		
165	98.9	79.1	118.7
167	13.5	11.0	16.6









#59
2,3,4,6-Tetrachlorophenol
Concen: 37.90 ng
RT: 10.098 min Scan# 1362
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26





Tgt Ion:232 Resp: 136362
Ion Ratio Lower Upper
232 100
131 41.5 35.9 53.9
130 2.3 1.8 2.6
166 27.6 23.5 35.3



#60
Diethylphthalate
Concen: 37.47 ng
RT: 10.186 min Scan# 1377
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26




Tgt Ion:149 Resp: 571251
Ion Ratio Lower Upper
149 100
177 21.8 18.1 27.1
150 12.8 10.3 15.5



Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD

Manual Integrations
APPROVED

mohammad
10/23/2020 1:31:52 PM







#61
4-Chlorophenyl-phenylether
Concen: 36.75 ng
RT: 10.304 min Scan# 1397
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:204 Resp: 259382
Ion Ratio Lower Upper
204 100
206 32.3 26.0 39.0
141 58.0 46.2 69.4

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM




#62
4-Nitroaniline
Concen: 29.95 ng
RT: 10.351 min Scan# 1405
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:138 Resp: 118014
Ion Ratio Lower Upper
138 100
92 52.3 31.0 71.0
108 77.7 51.5 91.5









#63
Azobenzene
Concen: 36.21 ng
RT: 10.463 min Scan# 1424
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26





Tgt Ion: 77 Resp: 583898
Ion Ratio Lower Upper
77 100
182 23.8 3.4 43.4
105 21.2 1.7 41.7
51 32.7 11.6 51.6



#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.251 min Scan# 1558
Delta R.T. 0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26




Tgt Ion: 188 Resp: 408580
Ion Ratio Lower Upper
188 100
94 9.2 8.3 12.5
80 9.7 9.1 13.7



Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD

Manual Integrations
APPROVED

mohammad
10/23/2020 1:31:52 PM







#65
4,6-Dinitro-2-methylphenol
Concen: 34.17 ng
RT: 10.386 min Scan# 1411
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:	198	Resp:	89694
Ion Ratio	Lower	Upper	
198	100		
51	48.4	29.6	69.6
105	44.1	24.4	64.4

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#66
n-Nitrosodiphenylamine
Concen: 34.71 ng
RT: 10.427 min Scan# 1418
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:	169	Resp:	488536
Ion Ratio	Lower	Upper	
169	100		
168	62.7	50.0	75.0
167	34.7	27.8	41.6









#67
4-Bromophenyl-phenylether
Concen: 35.43 ng
RT: 10.792 min Scan# 1480
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:248 Resp: 167263
Ion Ratio Lower Upper
248 100
250 97.9 77.0 115.6
141 63.3 53.7 80.5

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#68
Hexachlorobenzene
Concen: 34.77 ng
RT: 10.863 min Scan# 1492
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:284 Resp: 185756
Ion Ratio Lower Upper
284 100
142 30.6 27.4 41.2
249 29.9 24.2 36.4









#69
Atrazine
Concen: 39.80 ng
RT: 10.957 min Scan# 1508
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:200 Resp: 136981
Ion Ratio Lower Upper
200 100
173 25.6 7.5 47.5
215 48.0 27.1 67.1

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#70
Pentachlorophenol
Concen: 63.72 ng
RT: 11.069 min Scan# 1527
Delta R.T. 0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:266 Resp: 130084
Ion Ratio Lower Upper
266 100
268 68.4 52.4 78.6
264 63.3 49.9 74.9









#71
Phenanthrene
Concen: 34.64 ng
RT: 11.274 min Scan# 1562
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:178 Resp: 776026
Ion Ratio Lower Upper
178 100
176 19.6 15.9 23.9
179 15.5 12.2 18.4

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM




#72
Anthracene
Concen: 34.52 ng
RT: 11.327 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:178 Resp: 802084
Ion Ratio Lower Upper
178 100
176 19.1 15.7 23.5
179 15.4 12.7 19.1






#73
Carbazole
Concen: 34.01 ng
RT: 11.486 min Scan# 1598
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26




Tgt Ion:167 Resp: 702727
Ion Ratio Lower Upper
167 100
166 22.0 17.8 26.6
139 13.4 10.8 16.2




Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD

Manual Integrations
APPROVED



mohammad
10/23/2020 1:31:52 PM




#74
Di-n-butylphthalate
Concen: 38.58 ng
RT: 11.804 min Scan# 1652
Delta R.T. -0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26



Tgt Ion:149 Resp: 917064
Ion Ratio Lower Upper
149 100
150 9.5 7.7 11.5
104 5.4 4.2 6.2









#75
Fluoranthene
Concen: 34.50 ng
RT: 12.463 min Scan# 1764
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD

Tgt Ion:202 Resp: 828753
Ion Ratio Lower Upper
202 100
101 11.6 0.0 33.4
203 17.6 0.0 37.9




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#76
Chrysene-d12
Concen: 20.00 ng
RT: 13.892 min Scan# 2007
Delta R.T. 0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26


Tgt Ion:240 Resp: 272910
Ion Ratio Lower Upper
240 100
120 10.0 9.2 13.8
236 25.9 20.9 31.3





#77
Benzidine
Concen: 26.85 ng
RT: 12.586 min Scan# 1785
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26




Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD



Tgt Ion:184 Resp: 225945
Ion Ratio Lower Upper
184 100
185 14.6 11.6 17.4
183 11.5 9.4 14.0




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM




#78
Pyrene
Concen: 39.49 ng
RT: 12.692 min Scan# 1803
Delta R.T. 0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:202 Resp: 821223
Ion Ratio Lower Upper
202 100
200 21.7 17.4 26.2
203 17.8 14.6 21.8









#79
Terphenyl-d14
Concen: 66.33 ng
RT: 12.827 min Scan# 1826
Delta R.T. -0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26





Tgt Ion:244 Resp: 949808
Ion Ratio Lower Upper
244 100
212 7.1 5.4 8.2
122 11.4 8.1 12.1



#80
Butylbenzylphthalate
Concen: 44.34 ng
RT: 13.304 min Scan# 1907
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26




Tgt Ion:149 Resp: 365714
Ion Ratio Lower Upper
149 100
91 72.2 61.0 91.4
206 21.0 16.7 25.1



Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD

Manual Integrations
APPROVED

mohammad
10/23/2020 1:31:52 PM







#81
Benzo(a)anthracene
Concen: 36.05 ng
RT: 13.880 min Scan# 2005
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD

Tgt Ion:228 Resp: 644711
Ion Ratio Lower Upper
228 100
226 27.4 22.3 33.5
229 19.7 16.2 24.4




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM



#82
3,3'-Dichlorobenzidine
Concen: 21.39 ng
RT: 13.839 min Scan# 1998
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:252 Resp: 141900
Ion Ratio Lower Upper
252 100
254 64.5 51.8 77.6
126 12.1 8.8 13.2









#83
Chrysene
Concen: 35.59 ng
RT: 13.915 min Scan# 2011
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:228 Resp: 624773
Ion Ratio Lower Upper
228 100
226 30.7 24.6 37.0
229 19.7 15.9 23.9

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM




#84
Bis(2-ethylhexyl)phthalate
Concen: 47.00 ng
RT: 13.857 min Scan# 2001
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:149 Resp: 526292
Ion Ratio Lower Upper
149 100
167 26.4 20.9 31.3
279 3.9 3.2 4.8









#85
Di-n-octyl phthalate
Concen: 43.42 ng
RT: 14.468 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26





Tgt Ion:149 Resp: 786387
Ion Ratio Lower Upper
149 100
167 1.5 1.2 1.8
43 11.4 9.0 13.6



#86
Perylene-d12
Concen: 20.00 ng
RT: 15.321 min Scan# 2250
Delta R.T. 0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26




Tgt Ion:264 Resp: 239792
Ion Ratio Lower Upper
264 100
260 23.8 18.7 28.1
265 20.9 17.3 25.9



Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD

Manual Integrations
APPROVED

mohammad
10/23/2020 1:31:52 PM







#87
Indeno(1,2,3-cd)pyrene
Concen: 42.16 ng
RT: 16.739 min Scan# 2491
Delta R.T. 0.012 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:	276	Resp:	656372
Ion Ratio	Lower	Upper	
276	100		
138	23.5	19.9	29.9
277	25.3	20.4	30.6

Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD




Manual Integrations
APPROVED


mohammad
10/23/2020 1:31:52 PM




#88
Benzo(b)fluoranthene
Concen: 37.30 ng
RT: 14.909 min Scan# 2180
Delta R.T. -0.000 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26

Tgt Ion:	252	Resp:	604527
Ion Ratio	Lower	Upper	
252	100		
253	22.5	18.2	27.2
125	10.8	8.0	12.0









#89
Benzo(k)fluoranthene
Concen: 33.95 ng
RT: 14.939 min Scan# 2185
Delta R.T. 0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26





Tgt Ion:252 Resp: 522692
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.6
125 10.5 7.7 11.5



#90
Benzo(a)pyrene
Concen: 36.11 ng
RT: 15.262 min Scan# 2240
Delta R.T. 0.006 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26




Tgt Ion:252 Resp: 505545
Ion Ratio Lower Upper
252 100
253 22.0 17.3 25.9
125 11.1 8.6 13.0




Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD

Manual Integrations
APPROVED




mohammad
10/23/2020 1:31:52 PM




#91
Dibenzo(a,h)anthracene
Concen: 42.21 ng
RT: 16.739 min Scan# 2491
Delta R.T. 0.012 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26





Tgt Ion: 278 Resp: 541161
Ion Ratio Lower Upper
278 100
139 15.3 13.0 19.4
279 23.9 19.0 28.4



#92
Benzo(g,h,i)perylene
Concen: 43.74 ng
RT: 17.156 min Scan# 2562
Delta R.T. 0.017 min
Lab File: BF122080.D
Acq: 22 Oct 2020 13:26



Tgt Ion: 276 Resp: 561146
Ion Ratio Lower Upper
276 100
277 23.9 19.4 29.2
138 20.3 17.8 26.6



Instrument :
BNA_F
ClientSampleId :
VNJ-240MSD

Manual Integrations
APPROVED

mohammad
10/23/2020 1:31:52 PM