

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102224\
 Data File : BF139935.D
 Acq On : 22 Oct 2024 18:27
 Operator : RC/JU
 Sample : P4443-10
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 OG-315-HR-502-COMP-30

Quant Time: Oct 23 01:12:10 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.892	152	151455	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	577359	20.000	ng	0.00
39) Acenaphthene-d10	9.922	164	309203	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	475855	20.000	ng	0.00
76) Chrysene-d12	14.051	240	325699	20.000	ng	0.00
86) Perylene-d12	15.527	264	354964	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	1227656	126.895	ng	0.02
7) Phenol-d6	6.510	99	1434259	114.464	ng	0.00
23) Nitrobenzene-d5	7.451	82	1058903	101.649	ng	0.00
42) 2,4,6-Tribromophenol	10.716	330	402783	139.266	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1830776	97.855	ng	0.00
79) Terphenyl-d14	12.998	244	1619010	80.999	ng	0.00
Target Compounds						
77) Benzidine	12.745	184	495850	92.585	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102224\
Data File : BF139935.D
Acq On : 22 Oct 2024 18:27
Operator : RC/JU
Sample : P4443-10
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
OG-315-HR-502-COMP-30

Quant Time: Oct 23 01:12:10 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

