

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102224\
 Data File : BF139917.D
 Acq On : 22 Oct 2024 09:39
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 22 11:02:29 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	1,4-Dioxane	0.596	0.558	6.4	82	-0.02
3	Pyridine	1.476	1.403	4.9	81	-0.02
4	n-Nitrosodimethylamine	0.793	0.733	7.6	79	-0.02
5 S	2-Fluorophenol	1.278	1.261	1.3	85	0.00
6	Aniline	1.542	1.510	2.1	82	0.00
7 S	Phenol-d6	1.655	1.609	2.8	84	0.00
8	2-Chlorophenol	1.306	1.309	-0.2	86	0.00
9	Benzaldehyde	0.931	0.912	2.0	84	0.00
10 C	Phenol	1.706	1.663	2.5	84	0.00
11	bis(2-Chloroethyl)ether	1.319	1.282	2.8	85	0.00
12	1,3-Dichlorobenzene	1.499	1.497	0.1	86	0.00
13 C	1,4-Dichlorobenzene	1.498	1.509	-0.7	88	0.00
14	1,2-Dichlorobenzene	1.398	1.406	-0.6	88	0.00
15	Benzyl Alcohol	1.214	1.201	1.1	85	0.00
16	2,2'-oxybis(1-Chloropropane	2.248	2.110	6.1	81	0.00
17	2-Methylphenol	1.114	1.094	1.8	85	0.00
18	Hexachloroethane	0.534	0.540	-1.1	86	0.00
19 P	n-Nitroso-di-n-propylamine	1.004	0.945	5.9	84	0.00
20	3+4-Methylphenols	1.424	1.403	1.5	85	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
22	Acetophenone	0.490	0.478	2.4	86	0.00
23 S	Nitrobenzene-d5	0.361	0.380	-5.3	88	0.00
24	Nitrobenzene	0.396	0.402	-1.5	86	0.00
25	Isophorone	0.685	0.670	2.2	85	0.00
26 C	2-Nitrophenol	0.149	0.173	-16.1	95	0.00
27	2,4-Dimethylphenol	0.249	0.244	2.0	85	0.00
28	bis(2-Chloroethoxy)methane	0.415	0.406	2.2	85	0.00
29 C	2,4-Dichlorophenol	0.283	0.288	-1.8	88	0.00
30	1,2,4-Trichlorobenzene	0.314	0.320	-1.9	88	0.00
31	Naphthalene	1.031	1.032	-0.1	87	0.00
32	Benzoic acid	0.217	0.241	-11.1	95	0.00
33	4-Chloroaniline	0.352	0.351	0.3	86	0.00
34 C	Hexachlorobutadiene	0.197	0.204	-3.6	89	0.00
35	Caprolactam	0.090	0.089	1.1	84	0.00
36 C	4-Chloro-3-methylphenol	0.314	0.312	0.6	85	0.00
37	2-Methylnaphthalene	0.632	0.628	0.6	87	0.00
38	1-Methylnaphthalene	0.620	0.616	0.6	87	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
40	1,2,4,5-Tetrachlorobenzene	0.540	0.546	-1.1	88	0.00
41 P	Hexachlorocyclopentadiene	0.188	0.217	-15.4	95	0.00
42 S	2,4,6-Tribromophenol	0.187	0.203	-8.6	94	0.00
43 C	2,4,6-Trichlorophenol	0.382	0.408	-6.8	92	0.00
44	2,4,5-Trichlorophenol	0.394	0.398	-1.0	87	0.00
45 S	2-Fluorobiphenyl	1.210	1.182	2.3	88	0.00
46	1,1'-Biphenyl	1.392	1.379	0.9	87	0.00
47	2-Chloronaphthalene	1.121	1.123	-0.2	88	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102224\
 Data File : BF139917.D
 Acq On : 22 Oct 2024 09:39
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 22 11:02:29 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.343	0.368	-7.3	88	0.00
49	Acenaphthylene	1.621	1.606	0.9	87	0.00
50	Dimethylphthalate	1.246	1.243	0.2	87	0.00
51	2,6-Dinitrotoluene	0.270	0.288	-6.7	89	0.00
52 C	Acenaphthene	1.049	1.063	-1.3	89	0.00
53	3-Nitroaniline	0.278	0.290	-4.3	86	0.00
54 P	2,4-Dinitrophenol	0.093	0.135	-45.2#	116	0.00
55	Dibenzofuran	1.505	1.486	1.3	87	0.00
56 P	4-Nitrophenol	0.214	0.236	-10.3	89	0.00
57	2,4-Dinitrotoluene	0.332	0.383	-15.4	93	0.00
58	Fluorene	1.146	1.131	1.3	89	0.00
59	2,3,4,6-Tetrachlorophenol	0.309	0.325	-5.2	91	0.00
60	Diethylphthalate	1.223	1.207	1.3	86	0.00
61	4-Chlorophenyl-phenylether	0.579	0.588	-1.6	90	0.00
62	4-Nitroaniline	0.263	0.284	-8.0	90	0.00
63	Azobenzene	1.297	1.249	3.7	83	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
65	4,6-Dinitro-2-methylphenol	0.082	0.110	-34.1#	110	0.00
66 c	n-Nitrosodiphenylamine	0.599	0.590	1.5	86	0.00
67	4-Bromophenyl-phenylether	0.206	0.210	-1.9	90	0.00
68	Hexachlorobenzene	0.232	0.234	-0.9	88	0.00
69	Atrazine	0.162	0.167	-3.1	83	0.00
70 C	Pentachlorophenol	0.141	0.161	-14.2	93	0.00
71	Phenanthrene	0.945	0.922	2.4	86	0.00
72	Anthracene	0.922	0.911	1.2	87	0.00
73	Carbazole	0.857	0.835	2.6	86	0.00
74	Di-n-butylphthalate	0.990	0.962	2.8	83	0.00
75 C	Fluoranthene	0.955	0.943	1.3	87	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
77	Benzydine	0.329	0.416	-26.4#	92	0.00
78	Pyrene	1.751	1.947	-11.2	88	0.00
79 S	Terphenyl-d14	1.227	1.333	-8.6	87	0.00
80	Butylbenzylphthalate	0.525	0.554	-5.5	81	0.00
81	Benzo(a)anthracene	1.302	1.298	0.3	79	0.00
82	3,3'-Dichlorobenzidine	0.380	0.418	-10.0	88	0.00
83	Chrysene	1.194	1.200	-0.5	83	0.00
84	Bis(2-ethylhexyl)phthalate	0.589	0.653	-10.9	85	0.00
85 c	Di-n-octyl phthalate	1.071	1.174	-9.6	84	0.00
86 I	Perylene-d12	1.000	1.000	0.0	83	0.00
87	Indeno(1,2,3-cd)pyrene	1.287	1.475	-14.6	91	0.01
88	Benzo(b)fluoranthene	1.218	1.212	0.5	85	0.00
89	Benzo(k)fluoranthene	1.051	0.968	7.9	76	0.00
90 C	Benzo(a)pyrene	1.002	1.008	-0.6	82	0.00
91	Dibenzo(a,h)anthracene	1.073	1.220	-13.7	91	0.01
92	Benzo(g,h,i)perylene	1.072	1.265	-18.0	94	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102224\
Data File : BF139917.D
Acq On : 22 Oct 2024 09:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 22 11:02:29 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0