

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF102315\
 Data File : BF082484.D
 Acq On : 24 Oct 2015 00:14
 Operator : UM/IZ
 Sample : G4140-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB15-45-GW

Manual Integrations
 APPROVED

MmDadoda
 10/26/2015 7:22:23 PM

Quant Time: Oct 26 14:20:02 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF101015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 24 04:47:19 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.75	152	100583m	20.00	ng	0.00
21) Naphthalene-d8	8.05	136	405891	20.00	ng	0.00
38) Acenaphthene-d10	9.81	164	191950	20.00	ng	0.00
63) Phenanthrene-d10	11.29	188	373296	20.00	ng	0.00
75) Chrysene-d12	13.96	240	296514	20.00	ng	0.00
86) Perylene-d12	15.44	264	222010	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.35	112	68117	13.67	ng	0.00
7) Phenol-d6	6.40	99	67588	10.42	ng	-0.01
23) Nitrobenzene-d5	7.33	82	325132	53.79	ng	-0.01
41) 2,4,6-Tribromophenol	10.59	330	42943	23.86	ng	0.00
44) 2-Fluorobiphenyl	9.12	172	594457	51.36	ng	-0.01
78) Terphenyl-d14	12.88	244	584000	52.19	ng	0.00
Target Compounds						Qvalue
10) Phenol	6.42	94	16617	2.22	ng	96
31) Naphthalene	8.07	128	142752	8.11	ng	99
37) 2-Methylnaphthalene	8.75	142	29408	2.41	ng	# 91
49) Dimethylphthalate	9.52	163	170003	12.58	ng	100
59) Diethylphthalate	10.22	149	30865	2.45	ng	# 92
70) Phenanthrene	11.31	178	47522	2.60	ng	# 95
74) Fluoranthene	12.50	202	39757	2.02	ng	95
77) Pyrene	12.73	202	44732	2.54	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF102315\
Data File : BF082484.D
Acq On : 24 Oct 2015 00:14
Operator : UM/IZ
Sample : G4140-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB15-45-GW

Quant Time: Oct 26 14:20:02 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF101015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 24 04:47:19 2015
Response via : Initial Calibration

Manual Integrations
APPROVED

MmDadoda
10/26/2015 7:22:23 PM

