

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF102315\
 Data File : BF082508.D
 Acq On : 24 Oct 2015 12:31
 Operator : UM/IZ
 Sample : G4119-04RE
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB15-27-2.5-3RE

Manual Integrations
 APPROVED

MmDadoda
 10/26/2015 7:23:02 PM

Quant Time: Oct 26 15:27:30 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF101015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 23 15:17:13 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.75	152	100896	20.00	ng	0.00
21) Naphthalene-d8	8.05	136	377230	20.00	ng	0.00
38) Acenaphthene-d10	9.81	164	169151	20.00	ng	0.00
63) Phenanthrene-d10	11.29	188	246106	20.00	ng	0.00
75) Chrysene-d12	13.99	240	212291	20.00	ng	-0.05
86) Perylene-d12	15.51	264	176929	20.00	ng	-0.11
System Monitoring Compounds						
5) 2-Fluorophenol	5.38	112	122384	24.48	ng	0.03
7) Phenol-d6	6.42	99	623323	95.81	ng	0.00
23) Nitrobenzene-d5	7.33	82	382817	68.15	ng	-0.01
41) 2,4,6-Tribromophenol	10.61	330	1732	1.09	ng	0.00
44) 2-Fluorobiphenyl	9.13	172	842658	82.61	ng	0.00
78) Terphenyl-d14	12.89	244	565817	70.63	ng	-0.01
Target Compounds						Qvalue
10) Phenol	6.43	94	16492	2.20	ng	81
37) 2-Methylnaphthalene	8.77	142	47347	4.18	ng	100
49) Dimethylphthalate	9.53	163	236705	19.87	ng	98
70) Phenanthrene	11.31	178	109027	9.04	ng	98
71) Anthracene	11.37	178	25969m	2.09	ng	
74) Fluoranthene	12.51	202	134989	10.42	ng	97
77) Pyrene	12.74	202	135371	10.74	ng	98
79) Butylbenzylphthalate	13.37	149	38971	6.78	ng	94
80) Benzo(a)anthracene	13.98	228	49544m	4.49	ng	
82) Chrysene	14.01	228	48416	4.16	ng	96
83) Bis(2-ethylhexyl)phthalate	13.97	149	352713	43.00	ng	98
87) Benzo(b)fluoranthene	15.10	252	49726m	4.62	ng	
91) Benzo(g,h,i)perylene	17.33	276	14963	2.03	ng	# 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF102315\
Data File : BF082508.D
Acq On : 24 Oct 2015 12:31
Operator : UM/IZ
Sample : G4119-04RE
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB15-27-2.5-3RE

Manual Integrations
APPROVED

MmDadoda
10/26/2015 7:23:02 PM

Quant Time: Oct 26 15:27:30 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF101015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 23 15:17:13 2015
Response via : Initial Calibration

