

Data Path : Z:\HPCHEM1\BNA F\Data\BF102317\
 Data File : BF099761.D
 Acq On : 23 Oct 2017 16:06
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 23 17:54:05 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 23 15:08:33 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	93	0.00
2	1,4-Dioxane	40.000	40.443	-1.1	93	-0.01
3	Pyridine	40.000	43.359	-8.4	93	-0.01
4	n-Nitrosodimethylamine	40.000	42.549	-6.4	91	-0.01
5 S	2-Fluorophenol	80.000	81.041	-1.3	92	0.00
6	Aniline	40.000	39.910	0.2	92	0.00
7 S	Phenol-d6	80.000	79.835	0.2	92	0.00
8	2-Chlorophenol	40.000	40.286	-0.7	91	0.00
9	Benzaldehyde	40.000	41.109	-2.8	94	0.00
10 C	Phenol	40.000	40.762	-1.9	93	0.00
11	bis(2-Chloroethyl)ether	40.000	40.662	-1.7	91	0.00
12	1,3-Dichlorobenzene	40.000	39.566	1.1	91	0.00
13 C	1,4-Dichlorobenzene	40.000	39.815	0.5	91	0.00
14	1,2-Dichlorobenzene	40.000	40.320	-0.8	93	0.00
15	Benzyl Alcohol	40.000	40.249	-0.6	94	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	39.952	0.1	92	0.00
17	2-Methylphenol	40.000	39.719	0.7	91	0.00
18	Hexachloroethane	40.000	39.888	0.3	92	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.115	2.2	92	0.00
20	3+4-Methylphenols	40.000	40.020	-0.1	94	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	94	0.00
22	Acetophenone	40.000	38.473	3.8	92	0.00
23 S	Nitrobenzene-d5	80.000	78.416	2.0	92	0.00
24	Nitrobenzene	40.000	39.618	1.0	91	0.00
25	Isophorone	40.000	38.968	2.6	91	0.00
26 C	2-Nitrophenol	40.000	40.233	-0.6	89	0.00
27	2,4-Dimethylphenol	40.000	39.644	0.9	92	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.124	2.2	92	0.00
29 C	2,4-Dichlorophenol	40.000	40.139	-0.3	92	0.00
30	1,2,4-Trichlorobenzene	40.000	39.603	1.0	92	0.00
31	Naphthalene	40.000	39.378	1.6	92	0.00
32	Benzoic acid	40.000	41.012	-2.5	97	0.00
33	4-Chloroaniline	40.000	39.641	0.9	91	0.00
34 C	Hexachlorobutadiene	40.000	39.771	0.6	94	0.00
35	Caprolactam	40.000	39.863	0.3	91	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.742	3.1	90	0.00
37	2-Methylnaphthalene	40.000	39.052	2.4	93	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	94	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	39.314	1.7	92	0.00
40 P	Hexachlorocyclopentadiene	40.000	38.854	2.9	91	0.00
41 S	2,4,6-Tribromophenol	80.000	76.200	4.7	89	0.00
42 C	2,4,6-Trichlorophenol	40.000	39.486	1.3	93	0.00
43	2,4,5-Trichlorophenol	40.000	40.088	-0.2	90	0.00
44 S	2-Fluorobiphenyl	80.000	79.074	1.2	91	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF102317\
 Data File : BF099761.D
 Acq On : 23 Oct 2017 16:06
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 23 17:54:05 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 23 15:08:33 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	38.516	3.7	91	0.00
46	2-Chloronaphthalene	40.000	38.876	2.8	91	0.00
47	2-Nitroaniline	40.000	39.026	2.4	89	0.00
48	Acenaphthylene	40.000	38.973	2.6	92	0.00
49	Dimethylphthalate	40.000	38.603	3.5	90	0.00
50	2,6-Dinitrotoluene	40.000	39.247	1.9	90	0.00
51 C	Acenaphthene	40.000	38.361	4.1	91	0.00
52	3-Nitroaniline	40.000	39.571	1.1	91	0.00
53 P	2,4-Dinitrophenol	40.000	36.142	9.6	84	0.00
54	Dibenzofuran	40.000	38.261	4.3	91	0.00
55 P	4-Nitrophenol	40.000	38.747	3.1	85	0.00
56	2,4-Dinitrotoluene	40.000	38.835	2.9	90	0.00
57	Fluorene	40.000	38.431	3.9	91	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	39.498	1.3	90	0.00
59	Diethylphthalate	40.000	37.812	5.5	89	0.00
60	4-Chlorophenyl-phenylether	40.000	38.388	4.0	92	0.00
61	4-Nitroaniline	40.000	40.147	-0.4	91	0.00
62	Azobenzene	40.000	38.567	3.6	89	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	90	0.00
64	4,6-Dinitro-2-methylphenol	40.000	40.322	-0.8	88	0.00
65 c	n-Nitrosodiphenylamine	40.000	39.493	1.3	91	0.00
66	4-Bromophenyl-phenylether	40.000	39.821	0.4	91	0.00
67	Hexachlorobenzene	40.000	38.959	2.6	89	0.00
68	Atrazine	40.000	38.758	3.1	86	0.00
69 C	Pentachlorophenol	40.000	38.261	4.3	87	0.00
70	Phenanthrene	40.000	39.289	1.8	91	0.00
71	Anthracene	40.000	39.205	2.0	91	0.00
72	Carbazole	40.000	39.069	2.3	89	0.00
73	Di-n-butylphthalate	40.000	38.514	3.7	87	0.00
74 C	Fluoranthene	40.000	38.634	3.4	89	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	92	0.00
76	Benzidine	40.000	40.327	-0.8	94	0.00
77	Pyrene	40.000	39.557	1.1	90	0.00
78 S	Terphenyl-d14	80.000	76.934	3.8	90	0.00
79	Butylbenzylphthalate	40.000	39.049	2.4	86	0.00
80	Benzo(a)anthracene	40.000	39.780	0.5	90	0.00
81	3,3'-Dichlorobenzidine	40.000	40.477	-1.2	91	0.00
82	Chrysene	40.000	39.350	1.6	89	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	37.856	5.4	88	0.00
84 c	Di-n-octyl phthalate	40.000	38.840	2.9	87	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	39.003	2.5	93	0.00
86 I	Perylene-d12	20.000	20.000	0.0	96	0.00
87	Benzo(b)fluoranthene	40.000	37.326	6.7	94	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF102317\
Data File : BF099761.D
Acq On : 23 Oct 2017 16:06
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 23 17:54:05 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 23 15:08:33 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	40.598	-1.5	93	0.00
89 C	Benzo(a)pyrene	40.000	39.190	2.0	94	0.00
90	Dibenzo(a,h)anthracene	40.000	37.544	6.1	93	0.00
91	Benzo(a,h,i)perylene	40.000	37.196	7.0	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0