

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102320\
 Data File : BF122096.D
 Acq On : 23 Oct 2020 13:16
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 23 14:45:06 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 21 12:04:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
2	1,4-Dioxane	0.596	0.557	6.5	80	-0.02
3	Pyridine	1.567	1.560	0.4	83	-0.01
4	n-Nitrosodimethylamine	0.757	0.752	0.7	84	-0.02
5 S	2-Fluorophenol	1.355	1.348	0.5	86	0.00
6	Aniline	2.148	2.041	5.0	83	0.00
7 S	Phenol-d6	1.744	1.672	4.1	84	0.00
8	2-Chlorophenol	1.370	1.339	2.3	84	0.00
9	Benzaldehyde	1.006	0.966	4.0	85	0.00
10 C	Phenol	1.800	1.722	4.3	84	0.00
11	bis(2-Chloroethyl)ether	1.392	1.341	3.7	82	0.00
12	1,3-Dichlorobenzene	1.541	1.490	3.3	84	0.00
13 C	1,4-Dichlorobenzene	1.547	1.516	2.0	85	0.00
14	1,2-Dichlorobenzene	1.414	1.371	3.0	83	0.00
15	Benzyl Alcohol	1.128	1.148	-1.8	87	0.00
16	2,2'-oxybis(1-Chloropropane	2.142	2.027	5.4	81	0.00
17	2-Methylphenol	1.171	1.093	6.7	81	0.00
18	Hexachloroethane	0.559	0.551	1.4	84	0.00
19 P	n-Nitroso-di-n-propylamine	0.993	0.980	1.3	87	0.00
20	3+4-Methylphenols	1.412	1.427	-1.1	89	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
22	Acetophenone	0.515	0.509	1.2	88	0.00
23 S	Nitrobenzene-d5	0.390	0.386	1.0	86	0.00
24	Nitrobenzene	0.417	0.409	1.9	85	0.00
25	Isophorone	0.737	0.721	2.2	86	0.00
26 C	2-Nitrophenol	0.192	0.192	0.0	84	0.00
27	2,4-Dimethylphenol	0.327	0.314	4.0	83	0.00
28	bis(2-Chloroethoxy)methane	0.461	0.447	3.0	84	0.00
29 C	2,4-Dichlorophenol	0.306	0.303	1.0	85	0.00
30	1,2,4-Trichlorobenzene	0.326	0.317	2.8	85	0.00
31	Naphthalene	1.071	1.028	4.0	84	0.00
32	Benzoic acid	0.165	0.167	-1.2	84	0.00
33	4-Chloroaniline	0.456	0.452	0.9	86	0.00
34 C	Hexachlorobutadiene	0.193	0.192	0.5	85	0.00
35	Caprolactam	0.097	0.097	0.0	87	0.00
36 C	4-Chloro-3-methylphenol	0.329	0.331	-0.6	87	0.00
37	2-Methylnaphthalene	0.694	0.669	3.6	85	0.00
38	1-Methylnaphthalene	0.643	0.619	3.7	85	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	84	0.00
40	1,2,4,5-Tetrachlorobenzene	0.646	0.632	2.2	85	0.00
41 P	Hexachlorocyclopentadiene	0.127	0.116	8.7	76	0.00
42 S	2,4,6-Tribromophenol	0.247	0.250	-1.2	86	0.00
43 C	2,4,6-Trichlorophenol	0.398	0.396	0.5	85	0.00
44	2,4,5-Trichlorophenol	0.430	0.414	3.7	83	0.00
45 S	2-Fluorobiphenyl	1.359	1.290	5.1	86	0.00
46	1,1'-Biphenyl	1.626	1.583	2.6	86	0.00
47	2-Chloronaphthalene	1.265	1.243	1.7	85	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102320\
 Data File : BF122096.D
 Acq On : 23 Oct 2020 13:16
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 23 14:45:06 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 21 12:04:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.417	0.424	-1.7	87	0.00
49	Acenaphthylene	1.942	1.849	4.8	84	0.00
50	Dimethylphthalate	1.462	1.436	1.8	85	0.00
51	2,6-Dinitrotoluene	0.321	0.319	0.6	84	0.00
52 C	Acenaphthene	1.155	1.125	2.6	86	0.00
53	3-Nitroaniline	0.365	0.359	1.6	85	0.00
54 P	2,4-Dinitrophenol	0.129	0.127	1.6	81	0.00
55	Dibenzofuran	1.690	1.619	4.2	84	0.00
56 P	4-Nitrophenol	0.198	0.214	-8.1	91	0.00
57	2,4-Dinitrotoluene	0.395	0.400	-1.3	87	0.00
58	Fluorene	1.361	1.359	0.1	88	0.00
59	2,3,4,6-Tetrachlorophenol	0.333	0.344	-3.3	86	0.00
60	Diethylphthalate	1.410	1.423	-0.9	88	0.00
61	4-Chlorophenyl-phenylether	0.653	0.644	1.4	87	0.00
62	4-Nitroaniline	0.365	0.338	7.4	80	0.00
63	Azobenzene	1.491	1.497	-0.4	87	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
65	4,6-Dinitro-2-methylphenol	0.115	0.123	-7.0	86	0.00
66 c	n-Nitrosodiphenylamine	0.689	0.664	3.6	87	0.00
67	4-Bromophenyl-phenylether	0.231	0.227	1.7	87	0.00
68	Hexachlorobenzene	0.262	0.258	1.5	88	0.00
69	Atrazine	0.168	0.167	0.6	87	0.00
70 C	Pentachlorophenol	0.084	0.074	11.9	77	0.00
71	Phenanthrene	1.097	1.055	3.8	87	0.00
72	Anthracene	1.137	1.110	2.4	88	0.00
73	Carbazole	1.011	0.991	2.0	88	0.00
74	Di-n-butylphthalate	1.163	1.165	-0.2	88	0.00
75 C	Fluoranthene	1.176	1.140	3.1	87	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	85	0.00
77	Benzydine	0.617	0.534	13.5	77	0.00
78	Pyrene	1.524	1.534	-0.7	87	0.00
79 S	Terphenyl-d14	1.049	1.064	-1.4	90	0.00
80	Butylbenzylphthalate	0.604	0.637	-5.5	89	0.00
81	Benzo(a)anthracene	1.311	1.302	0.7	87	0.00
82	3,3'-Dichlorobenzidine	0.486	0.475	2.3	82	0.00
83	Chrysene	1.287	1.262	1.9	84	0.00
84	Bis(2-ethylhexyl)phthalate	0.821	0.863	-5.1	89	0.00
85 c	Di-n-octyl phthalate	1.327	1.374	-3.5	87	0.00
86 I	Perylene-d12	1.000	1.000	0.0	81	0.01
87	Indeno(1,2,3-cd)pyrene	1.299	1.317	-1.4	83	0.02
88	Benzo(b)fluoranthene	1.352	1.476	-9.2	85	0.01
89	Benzo(k)fluoranthene	1.284	1.307	-1.8	80	0.01
90 C	Benzo(a)pyrene	1.168	1.215	-4.0	81	0.01
91	Dibenzo(a,h)anthracene	1.069	1.080	-1.0	83	0.02
92	Benzo(g,h,i)perylene	1.070	1.094	-2.2	84	0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102320\
Data File : BF122096.D
Acq On : 23 Oct 2020 13:16
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 23 14:45:06 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 21 12:04:42 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0