

Data Path : Z:\HPCHEM1\BNA F\DATA\BF102617\
 Data File : BF099861.D
 Acq On : 26 Oct 2017 15:06
 Operator : SJ/JU
 Sample : I6068-02
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-05

Quant Time: Oct 26 17:00:22 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 26 16:31:52 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.79	152	61695	20.00	ng	0.00
21) Naphthalene-d8	8.08	136	242243	20.00	ng	0.00
38) Acenaphthene-d10	9.85	164	108247	20.00	ng	0.00
63) Phenanthrene-d10	11.35	188	187573	20.00	ng	0.01
75) Chrysene-d12	13.99	240	175985	20.00	ng	0.00
86) Perylene-d12	15.47	264	178269	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.41	112	451057	114.78	ng	0.00
7) Phenol-d6	6.44	99	531616	114.09	ng	0.00
23) Nitrobenzene-d5	7.36	82	290035	77.08	ng	0.00
41) 2,4,6-Tribromophenol	10.65	330	113022	114.57	ng	0.01
44) 2-Fluorobiphenyl	9.16	172	535228	76.29	ng	0.00
78) Terphenyl-d14	12.92	244	512875	63.69	ng	0.00
Target Compounds						
10) Phenol	6.45	94	29232	5.714	ng	95
37) 2-Methylnaphthalene	8.80	142	390845	49.877	ng	98
51) Acenaphthene	9.88	154	30376	4.539	ng	# 70
57) Fluorene	10.40	166	77781	9.944	ng	# 91
70) Phenanthrene	11.37	178	237571	22.443	ng	96
74) Fluoranthene	12.56	202	39862	3.624	ng	92
77) Pyrene	12.79	202	58709	4.387	ng	96
79) Butylbenzylphthalate	13.39	149	14750	2.238	ng	93
83) Bis(2-ethylhexyl)phthalate	13.94	149	138557	14.973	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF102617\
Data File : BF099861.D
Acq On : 26 Oct 2017 15:06
Operator : SJ/JU
Sample : I6068-02
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-05

Quant Time: Oct 26 17:00:22 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 26 16:31:52 2017
Response via : Initial Calibration

